



REQUEST FOR QUOTATION
NEGOTIATED PROCUREMENT
SECTION 53.1

Date: November 24, 2020
 PR No: GF-20-08-01254

Name of Company : _____
 Address : _____
 Contact No. : _____
 Project Title : **Procurement of VITAL SIGNS 8-IN-1 SIMULATOR**
 Approved Budget for the Contract : **Php 2,000,000.00**
 End-User / Implementing Office : **ROSARIO MACLANG BAUTISTA GENERAL HOSPITAL**

Please quote your best offer for the item/s described below, subject to the Terms and Conditions provided. Submit your quotation duly signed by you or your duly authorized representative not later than **November 27, 2020 10:00 AM** Philippine Standard Time, together with the following documents of your company:

- 1 PhilGEPS certificate (not expired on the time of opening of quotations);
- 2 Business Registration (DTI / SEC);
- 3 Mayor's/Business Permit (2020);
- 4 Tax Clearance;
- 5 Income/Business Tax Return (for FY 2018) (For ABC P500,000.00 above);
- 6 Omnibus Sworn Statement prescribed by the **QC BAC- Goods and Services**
- 7 If applicable, the JVA in case the joint venture is already in existence, or duly notarized statements from all the potential joint venture partners stating that they will enter into and abide by the provisions of the JVA in the instance that the bid is successful.


in a **SEALED LONG BROWN ENVELOPE** shall:

- 1 Contain the Project Name and PR Number of the contract to be bid in capital letters;
- 2 Bear the name and address of the Bidder in capital letters;
- 3 Be addressed to the Procuring Entity's BAC.

Project Title: **Procurement of VITAL SIGNS 8-IN-1 SIMULATOR**

Quezon City Local Government
BIDS AND AWARDS COMMITTEE
2/F Procurement Department, Finance Building
Quezon City Hall Compound

For any clarification you may contact us at 89884242 loc. 8505/8709.


ATTY. DOMINIC B. GARCIA
 Officer-In-Charge/Head, BAC Secretariat

TERMS AND CONDITIONS

1. Bidders shall **provide correct and accurate** information required in this form.
2. Price quotation/s must be valid for a period of thirty (30) calendar days from the date of submission.
3. Price quotation/s, to be denominated in Philippine Peso shall include all taxes, duties and/or levies payable.
4. Quotation **exceeding** the Approved Budget for the Contract (ABC) shall be **rejected**.
5. Award of contract shall be made to the lowest quotation (for goods) or the highest rated offer (for consulting services) which complies with the minimum technical specifications and other terms and conditions stated herein.
6. Any interlineations, erasures or overwriting shall be valid only if they are signed or initialed by you or any of your duly authorized representative/s.
7. The City General Services Department (CGSD) shall have the right to inspect the goods.
8. Non-submission of eligibility documents shall mean disqualification of Quotation.
9. Liquidated damages equivalent to one tenth (1/10) of one percent (1%) of the value of the goods not delivered within the prescribed delivery period shall be imposed per day of delay. CGSD shall rescind the contract once the cumulative amount of liquidated damages reaches ten percent (10%) of the amount of the contract, without prejudice to other courses of action and remedies open to it.
10. Failure to follow these instructions will disqualify your entire quotation.

After having carefully read and accepted the Terms and Conditions, I/We submit our quotation/s as follows:

ITEM NO.	ITEM & DESCRIPTION	UNIT OF ISSUE	QTY.	UNIT PRICE	ITEM TOTAL
1	<p>VITAL SIGNS 8-IN-1 SIMULATOR 8-in-1 multifunction simulator tests ECG (including fetal ECG and arrhythmias), respiration temperature, IBP, cardiac output/cardiac catheterization NIBP, SPO2, and Rainbow multi wavelength waveforms. With authorized calibration laboratory test equipment in the Philippines</p> <ul style="list-style-type: none"> • Stay-connected ECG posts for easy/secure ECG snap and lead connections. • Custom SPO2 r-curve for accurate testing of the latest and future oximetry technologies • Static pressure linearity testing • Repeatable NIBP simulation (+/-2 mmHg) for dynamic pressure repeatability testing • Physiologically synchronized pulses across all parameters • Barcode scanning and direct data capture and printing functionality • Onboard, customizable patient pre-sets and auto sequences for fast/easy testing • Multi-language user interface offers choices of language selection • Integrated, easily-replaceable long-life battery • Optional PC-interface software offers customizable procedures checklists to replace bulky, service manuals and automated data capture/storage. • Wireless communication for remote PC control of test device, as well as data transfer and automated regulatory reporting. • Temperature Operating: 10°C to 40°C (50°F to 104°F) Storage: -20°C to +60°C (-4°F to 140 F) • Humidity: 10% to 90% non-condensing • Altitude: 3,000 meters (9,843 ft) • Dimension: (LxWxH) 14.5cm x 30.2cm x 8.6cm (5.7in x 11.9in x 3.4in) • Display: LCD color display • Communication USB device upstream port: Mini-B connector for control by a computer USB host controller port: Type A, 5V output, 0.5A max load. Connector for keyboard barcode reader, and printer • Wireless: IEEE 82.15.4 for control by a computer • Power: Lithium-ion rechargeable battery 	unit	1		

<ul style="list-style-type: none"> • Battery charger: 100V to 240V input, 15V/2.0 A output. For best performance, the battery charger should be connected to a properly-grounded AC receptacle • Battery life: 9 hours (minimum), 100 NIBP cycles, typical • Weight: 1.87 kg (4.2lb) • Safety standards: IEC/EN 61010-1 3rd Edition; pollution degree 2 CAT None • Warranty: one (1) year parts and service with consumables 				
Total Quoted Amount				

Amount in Words: _____

Delivery Period : Ninety (90) Calendar Days
 Warranty : _____

Other requirement:
1. Copy of valid, current License to Operate from DOH Accreditation as Supplier, Distributor or Manufacturer for Medical or Hospital Equipment or Devices
2. Statement of Warranty - One (1) year parts and service with consumables

 Signature over printed name

 Office Telephone No./Fax/Mobile No.

 Date

 Email Address

GF-20-08-01254