



Republic of the Philippines
Quezon City
BIDS AND AWARDS COMMITTEE – GOODS
 Quezon City Hall, Elliptical Road, Diliman, Quezon Cit007



BID BULLETIN NO. 01-2019
Series of 2019

Notice is hereby given to all concerned bidders for the Procurement of
 Various projects scheduled for Public Bidding on 21 January 2019, viz:

Change of Specifications

- a) **Project Title: Procurement of Mobile Endoscopy System and Others**
Project No.: 1901-55083

- b) **Project Title: Mechanical Ventilators and Others**
Project No.: 1901-55121

| Item/Description | FROM | TO |
|-----------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| a) Mechanical Ventilator | | |
| Battery Life | 2.5 or more when fully charged with indicator | <i>2.5 hours or more when fully charged with indicator. May include an external battery that will be provided by the supplier to fulfil the minimum time of 2.5 hours or more when fully charged without additional cost to the end-user</i> |
| Specs | 1.Dimension: More or less at 20x45 inches (WxH) more or less with weight of 75kg 2. Equipped with 12" or better touch screen (colored) synchronized nebulizer | 1. Dimension: More or less at 20 x 45 inches (WxH) with 65kg to 75kg. overall weight. 2. Equipped with 12" or better touch screen (colored) with built-in synchronized nebulizer. |
| Accessories | Noted reusable circuit/expiratory sets tube holders O2 sensors ventilator cart | <i>Noted reusable circuit/expiratory sets, tube holders, O2 sensors that are not disposable</i> |
| b) Defibrillator | | |
| Monitor | Biphasic type of defibrillation AED modes with pads configurable to 2-200J Manual Mode with paddles or pads and energy selection between 2-200J LCD, large, colour approx.. more or less 10 inches | <i>Biphasic type of defibrillation AED modes with pads configurable to 50-360J Manual Mode with paddles or pads and energy selection between 50-360J</i> <i>LCD, large, colour minimum of 10 inches</i> |
| c) Volumetric Infusion Pump | | |
| | KVO flow rate: 1-10ml/h Volume limit Occlusion detection for either 2 way or 4 way steps | <i>KVO (Keep Vein Open) flow rate: 1-10ml/hour Volume limit Occlusion detection for presence of air or occlusion either 2 way or 4 way steps.</i> |

Additional Specs: Mobile Endoscopy System (Project No. 1901-55083)

| Item/Description | FROM | TO |
|-------------------------|---------------------------------------------------|------------------------------------------------------|
| Mobile Endoscopy System | Capable of recording to SD card and USB and DVDrw | <i>Capable of recording to SD card, USB or DVDrw</i> |

Please be guided accordingly

By: **QC-BIDS & AWARDS COMMITTEE**

FREDERIKA C. RENTOY
 Vice-Chairperson