



Republic of the Philippines
QUEZON CITY GOVERNMENT
BAC- Goods and Services
2nd Floor, Procurement Department,
Finance Building, Quezon City Hall Compound



REQUEST FOR QUOTATION
NEGOTIATED PROCUREMENT
(SECTION 53.1)

Date : **FEBRUARY 16, 2022**
Project No. : **RMBGH-21-EHSE-624**

Name of Company : _____
Address : _____
Contact No. : _____
Project Title : **SUPPLY AND INSTALLATION OF CEILING MOUNTED NEGATIVE PRESSURE AND OTHERS**
Approved budget of the Contract : **Php 8,163,000.00**
End-User /
Implementing Office : **ROSARIO MACLANG BAUTISTA GENERAL HOSPITAL**

Please quote your best offer for the item/s described below, subject to the Terms and Conditions provided. Submit your quotation duly signed by you or your duly authorized representative not later than 11:00 A.M. **FEBRUARY 21, 2022** Philippine Standard Time, together with the following documents of your company:

- 1 PhilGEPS certificate (not expired on the time of opening of quotations);
- 2 Business Registration (DTI/SEC)
- 3 Mayor's/Business Permit (2022);
- 4 Tax Clearance; and
- 5 Omnibus Sworn Statement prescribed by **QC BAC – Goods and Services**
- 6 Income/Business Tax Return (for FY 2020) (For ABCs above P500,000.00)
- 7 If applicable, the JVA in case the joint venture is already in existence, or duly notarized statements from all the potential joint venture partners stating that they will enter into and abide by the provisions of the JVA in the instance that the bid is successful.

in a **SEALED LONG BROWN ENVELOPE** shall:

- 1 Contain the Project Name and PR Number of the contract to be bid in capital letters;
- 2 Bear the name and address of the Bidder in capital letters;
- 3 Be addressed to the Procuring Entity's BAC.

Project Title : **SUPPLY AND INSTALLATION OF CEILING MOUNTED NEGATIVE PRESSURE AND OTHERS**
Quezon City Local Government
BIDS AND AWARDS COMMITTEE
2/F Procurement Department, Finance Building
Quezon City Hall Compound

For any clarification you may contact us at 89884242 loc. 8506/8709.


MA. MARGARITA T. SANTOS
Chairperson, BAC- Goods and Services

TERMS AND CONDITIONS

1. Bidders shall **provide correct and accurate** information required in this form.
2. Price quotation/s must be valid for a period of thirty (30) calendar days from the date of submission.
3. Price quotation/s, to be denominated in Philippine Peso shall include all taxes, duties and/or levies payable.
4. Quotation **exceeding** the Approved Budget for the Contract (ABC) shall be **rejected**.
5. Award of contract shall be made to the lowest quotation (for goods) or the highest rated offer (for consulting services) which complies with the minimum technical specifications and other terms and conditions stated herein.
6. Any interlineations, erasures or overwriting shall be valid only if they are signed or initiated by you or any of your duly authorized representative/s.
7. The City General Services Department (CGSD) shall have the right to inspect the goods.
8. Non-submission of eligibility documents shall mean disqualification of Quotation.
9. Liquidated damages equivalent to one tenth (1/10) of one percent (1%) of the value of the goods not delivered within the prescribed delivery period shall be imposed per day of delay. CGSD shall rescind the contract once the cumulative amount of liquidated damages reaches ten percent (10%) of the amount of the contract, without prejudice to other courses of action and remedies open to it.
10. Failure to follow these instructions will disqualify your entire quotation.

After having carefully read and accepted the Terms and Conditions, I/We submit our quotation/s as follows:

ITEM NO.	ITEM & DESCRIPTION	UNIT OF ISSUE	QTY.	UNIT PRICE	ITEM TOTAL
1	CEILING MOUNTED NEGATIVE PRESSURE , Isolation Unit, 1 piece for positive or negative room pressure Specification: * One Way airflow; Ceiling Mounted * with remote control key for hospital staff * With Degree of contamination indication at remote control With 4 Operational level control * Fan Stage can be adjusted to silent, basic, high, and intensive levels Clean air capacity of not less than 220-990 cubic meter per hour With ULPA 15 Quality filter cartridge capable of up to 99.9995% on 0.1 to 0.3 micrometer at 10 sq.m. With permanent pressure difference display, with integrated security alarm in case pressure drop in 60 seconds.* Complies with ISO 9001, ISO 14001, Safety Class 1, EN 1822 Dimension of not more than 54 x 54 x 27 centimeters * 220V, 60Hz, 8 to 175 watts; Complies with CE Standards Scope of work for Ceiling Mounted Negative Pressure System Supply and Installation of Unit Cutting and Fit out works for the Ceiling * Installation of support / hangers * Installation of ducting system for exhaust / supply of air * Installation of louver for air exhaust / intake * Restoration and repainting of affected areas * Sealing of pressure room to prevent too much air leak * Installation of electrical supply and control * Restoration of affected areas during electrical and control installation * Testing and Commissioning * Warranty : One year parts and service *	Unit ,	3 ,		
2	MEDICAL GRADE AIR PURIFIER WITH HEPA FILTER * H13 Grade HEPA Filter, includes Carbon Filter * Effectively filters 99.95% of PM2.5, dust, pollen and viruses. H13 HEPA filter effectively traps dust, allergens, PM2.5, bacteria and viruses * 1 x Blast (includes 1 x Blast HEPA Filter), 1 x Wire cord Dimensions L x W x H (cm) at least 57.5 x 33 x 123 * CADR (m3/h)*507 - 680 - 890 (low - mid - high) * Sound (dB)37.9 - 43.3 - 48 (low - mid - high) Power Consumption (W)64 - 88 (low - high) * Electricity Voltage (V)220 - 240 * Recommended Room SizeUp to 130m2 HEPA Lifespan 19-22 months * Warranty : One Year on parts and service *	Unit ,	12 ,		

✓A

3	PAPR - (POWERED AIR PURIFYING RESPIRATOR) Equipped with replaceable H13 HEPA filter, Retains >99.97% of 0.3 micron particles Extended life with large pleated surface area (5X that found in common respirators) Aircube filter unit dimensions L x W x H: 140mm x 90mm x 120mm Light-weight of 1.2kg including the filter unit, powerpack, and belt Includes high-quality nylon belt with easy to clip buckle Comes with external Li-ion powerpack with 30,000 mAh capacity Powerpack dimensions L x W x H: 85mm x 45mm x 170mm Up to 9 hours operating time from a single charge of the included powerpack External powerpack enables swapping of batteries for extended operating time comes with 18W Fast Charger that charges the powerpack in 6-7 hours. Flexible breathing hose length: at least 0.5m-1.5m Easy to connect and secure with durable quick-release mechanisms Fine adjustment 130L/min-400 L/min Comfortable breathing at low settings for half-masks	Unit	10		

	Total Quoted Amount				

Amount in Words: _____

Other Requirements:
1. Statement of Warranty (Parts and Services) – One (1) Year

Delivery Period : Thirty (30) Calendar Days
Warranty : _____

Signature over printed name

Office Telephone No./Fax/Mobile No.

Date

E-mail Address

