



REPUBLIC OF THE PHILIPPINES
QUEZON CITY GOVERNMENT
BIDS AND AWARDS COMMITTEE –
GOODS AND SERVICES



PHILIPPINE BIDDING DOCUMENTS

(As Harmonized with Development Partners)

PROCUREMENT OF ADVANCE CARDIAC LIFE SUPPORT AMBULANCE

PROJECT NO. QCDRRMO-22-VEHICLES-1372

Government of the Republic of the Philippines

**Sixth Edition
July 2020**

Preface

These Philippine Bidding Documents (PBDs) for the procurement of Goods through Competitive Bidding have been prepared by the Government of the Philippines for use by any branch, constitutional commission or office, agency, department, bureau, office, or instrumentality of the Government of the Philippines, National Government Agencies, including Government-Owned and/or Controlled Corporations, Government Financing Institutions, State Universities and Colleges, and Local Government Unit. The procedures and practices presented in this document have been developed through broad experience, and are for mandatory use in projects that are financed in whole or in part by the Government of the Philippines or any foreign government/foreign or international financing institution in accordance with the provisions of the 2016 revised Implementing Rules and Regulations of Republic Act No. 9184.

The Bidding Documents shall clearly and adequately define, among others: (i) the objectives, scope, and expected outputs and/or results of the proposed contract or Framework Agreement, as the case may be; (ii) the eligibility requirements of Bidders; (iii) the expected contract or Framework Agreement duration, the estimated quantity in the case of procurement of goods, delivery schedule and/or time frame; and (iv) the obligations, duties, and/or functions of the winning bidder.

Care should be taken to check the relevance of the provisions of the PBDs against the requirements of the specific Goods to be procured. If duplication of a subject is inevitable in other sections of the document prepared by the Procuring Entity, care must be exercised to avoid contradictions between clauses dealing with the same matter.

Moreover, each section is prepared with notes intended only as information for the Procuring Entity or the person drafting the Bidding Documents. They shall not be included in the final documents. The following general directions should be observed when using the documents:

- a. All the documents listed in the Table of Contents are normally required for the procurement of Goods. However, they should be adapted as necessary to the circumstances of the particular Procurement Project.
- b. Specific details, such as the “*name of the Procuring Entity*” and “*address for bid submission*,” should be furnished in the Instructions to Bidders, Bid Data Sheet, and Special Conditions of Contract. The final documents should contain neither blank spaces nor options.
- c. This Preface and the footnotes or notes in italics included in the Invitation to Bid, Bid Data Sheet, General Conditions of Contract, Special Conditions of Contract, Schedule of Requirements, and Specifications are not part of the text of the final document, although they contain instructions that the Procuring Entity should strictly follow.
- d. The cover should be modified as required to identify the Bidding Documents as to the Procurement Project, Project Identification Number, and Procuring Entity, in addition to the date of issue.

- e. Modifications for specific Procurement Project details should be provided in the Special Conditions of Contract as amendments to the Conditions of Contract. For easy completion, whenever reference has to be made to specific clauses in the Bid Data Sheet or Special Conditions of Contract, these terms shall be printed in bold typeface on Sections I (Instructions to Bidders) and III (General Conditions of Contract), respectively.
- f. For guidelines on the use of Bidding Forms and the procurement of Foreign-Assisted Projects, these will be covered by a separate issuance of the Government Procurement Policy Board.

Table of Contents

Glossary of Acronyms, Terms, and Abbreviations	4
Section I. Invitation to Bid	7
Section II. Instructions to Bidders	11
1. Scope of Bid	12
2. Funding Information	12
3. Bidding Requirements	12
4. Corrupt, Fraudulent, Collusive, and Coercive Practices	12
5. Eligible Bidders	13
6. Origin of Goods	13
7. Subcontracts	13
8. Pre-Bid Conference	14
9. Clarification and Amendment of Bidding Documents	14
10. Documents comprising the Bid: Eligibility and Technical Components	14
11. Documents comprising the Bid: Financial Component	14
12. Bid Prices	14
13. Bid and Payment Currencies	15
14. Bid Security	15
15. Sealing and Marking of Bids	16
16. Deadline for Submission of Bids	16
17. Opening and Preliminary Examination of Bids	16
18. Domestic Preference	16
19. Detailed Evaluation and Comparison of Bids	16
20. Post-Qualification	17
21. Signing of the Contract	17
Section III. Bid Data Sheet	18
Section IV. General Conditions of Contract	21
1. Scope of Contract	22
2. Advance Payment and Terms of Payment	22
3. Performance Security	22
4. Inspection and Tests	22
5. Warranty	22
6. Liability of the Supplier	23
Section V. Special Conditions of Contract	24
Section VI. Schedule of Requirements	28
Section VII. Technical Specifications	41
Section VIII. Checklist of Technical and Financial Documents	55

Glossary of Acronyms, Terms, and Abbreviations

ABC – Approved Budget for the Contract.

BAC – Bids and Awards Committee.

Bid – A signed offer or proposal to undertake a contract submitted by a bidder in response to and in consonance with the requirements of the bidding documents. Also referred to as *Proposal* and *Tender*. (2016 revised IRR, Section 5[c])

Bidder – Refers to a contractor, manufacturer, supplier, distributor and/or consultant who submits a bid in response to the requirements of the Bidding Documents. (2016 revised IRR, Section 5[d])

Bidding Documents – The documents issued by the Procuring Entity as the bases for bids, furnishing all information necessary for a prospective bidder to prepare a bid for the Goods, Infrastructure Projects, and/or Consulting Services required by the Procuring Entity. (2016 revised IRR, Section 5[e])

BIR – Bureau of Internal Revenue.

BSP – Bangko Sentral ng Pilipinas.

Consulting Services – Refer to services for Infrastructure Projects and other types of projects or activities of the GOP requiring adequate external technical and professional expertise that are beyond the capability and/or capacity of the GOP to undertake such as, but not limited to: (i) advisory and review services; (ii) pre-investment or feasibility studies; (iii) design; (iv) construction supervision; (v) management and related services; and (vi) other technical services or special studies. (2016 revised IRR, Section 5[i])

CDA - Cooperative Development Authority.

Contract – Refers to the agreement entered into between the Procuring Entity and the Supplier or Manufacturer or Distributor or Service Provider for procurement of Goods and Services; Contractor for Procurement of Infrastructure Projects; or Consultant or Consulting Firm for Procurement of Consulting Services; as the case may be, as recorded in the Contract Form signed by the parties, including all attachments and appendices thereto and all documents incorporated by reference therein.

CIF – Cost Insurance and Freight.

CIP – Carriage and Insurance Paid.

CPI – Consumer Price Index.

DDP – Refers to the quoted price of the Goods, which means “delivered duty paid.”

DTI – Department of Trade and Industry.

EXW – Ex works.

FCA – “Free Carrier” shipping point.

FOB – “Free on Board” shipping point.

Foreign-funded Procurement or Foreign-Assisted Project– Refers to procurement whose funding source is from a foreign government, foreign or international financing institution as specified in the Treaty or International or Executive Agreement. (2016 revised IRR, Section 5[b]).

Framework Agreement – Refers to a written agreement between a procuring entity and a supplier or service provider that identifies the terms and conditions, under which specific purchases, otherwise known as “Call-Offs,” are made for the duration of the agreement. It is in the nature of an option contract between the procuring entity and the bidder(s) granting the procuring entity the option to either place an order for any of the goods or services identified in the Framework Agreement List or not buy at all, within a minimum period of one (1) year to a maximum period of three (3) years. (GPPB Resolution No. 27-2019)

GFI – Government Financial Institution.

GOCC – Government-owned and/or –controlled corporation.

Goods – Refer to all items, supplies, materials and general support services, except Consulting Services and Infrastructure Projects, which may be needed in the transaction of public businesses or in the pursuit of any government undertaking, project or activity, whether in the nature of equipment, furniture, stationery, materials for construction, or personal property of any kind, including non-personal or contractual services such as the repair and maintenance of equipment and furniture, as well as trucking, hauling, janitorial, security, and related or analogous services, as well as procurement of materials and supplies provided by the Procuring Entity for such services. The term “related” or “analogous services” shall include, but is not limited to, lease or purchase of office space, media advertisements, health maintenance services, and other services essential to the operation of the Procuring Entity. (2016 revised IRR, Section 5[r])

GOP – Government of the Philippines.

GPPB – Government Procurement Policy Board.

INCOTERMS – International Commercial Terms.

Infrastructure Projects – Include the construction, improvement, rehabilitation, demolition, repair, restoration or maintenance of roads and bridges, railways, airports, seaports, communication facilities, civil works components of information technology projects, irrigation, flood control and drainage, water supply, sanitation, sewerage and solid waste management systems, shore protection, energy/power and electrification facilities, national buildings, school buildings, hospital buildings, and other related construction projects of the government. Also referred to as *civil works or works*. (2016 revised IRR, Section 5[u])

LGUs – Local Government Units.

NFCC – Net Financial Contracting Capacity.

NGA – National Government Agency.

PhilGEPS - Philippine Government Electronic Procurement System.

Procurement Project – refers to a specific or identified procurement covering goods, infrastructure project or consulting services. A Procurement Project shall be described, detailed, and scheduled in the Project Procurement Management Plan prepared by the agency

which shall be consolidated in the procuring entity's Annual Procurement Plan. (GPPB Circular No. 06-2019 dated 17 July 2019)

PSA – Philippine Statistics Authority.

SEC – Securities and Exchange Commission.

SLCC – Single Largest Completed Contract.

Supplier – refers to a citizen, or any corporate body or commercial company duly organized and registered under the laws where it is established, habitually established in business and engaged in the manufacture or sale of the merchandise or performance of the general services covered by his bid. (Item 3.8 of GPPB Resolution No. 13-2019, dated 23 May 2019). Supplier as used in these Bidding Documents may likewise refer to a distributor, manufacturer, contractor, or consultant.

UN – United Nations.

Section I. Invitation to Bid

Notes on the Invitation to Bid

The Invitation to Bid (IB) provides information that enables potential Bidders to decide whether to participate in the procurement at hand. The IB shall be posted in accordance with Section 21.2 of the 2016 revised IRR of RA No. 9184.

Apart from the essential items listed in the Bidding Documents, the IB should also indicate the following:

- a. The date of availability of the Bidding Documents, which shall be from the time the IB is first advertised/posted until the deadline for the submission and receipt of bids;
- b. The place where the Bidding Documents may be acquired or the website where it may be downloaded;
- c. The deadline for the submission and receipt of bids; and
- d. Any important bid evaluation criteria (*e.g.*, the application of a margin of preference in bid evaluation).

The IB should be incorporated in the Bidding Documents. The information contained in the IB must conform to the Bidding Documents and in particular to the relevant information in the Bid Data Sheet.



QUEZON CITY GOVERNMENT
BAC – GOODS AND SERVICES



INVITATION TO BID

December 21, 2022

	PROJECT NO.	OFFICE	PROJECT NAME	AMOUNT	SOURCE OF FUND	DELIVERY PERIOD
1	OCM-22-HME-346	OFFICE OF THE CITY MAYOR	SMART URINE ANALYZER AND OTHERS	P 2,706,705.00	GENERAL FUND	30 CD
2	OCM(POPS)-22-JE-835B	OFFICE OF THE CITY MAYOR – POPS PLAN (DPOS)	ROAD SWEEPING MACHINE	P 1,192,016.80	GENERAL FUND	30 CD
3	OCM(POPS)-22-VEHICLES-1046B	OFFICE OF THE CITY MAYOR – POPS PLAN (QCPD)	QCPD UTILITY VEHICLE	P 9,973,600.00	GENERAL FUND	90 CD
4	OCM(POPS)-22-VEHICLES-1041B	OFFICE OF THE CITY MAYOR- POPS PLAN (QCPD)	SWAT VAN	P 6,000,000.00	GENERAL FUND	90 CD
5	CONSO-22-HCS-1704B	BARANGAY AND COMMUNITY RELATIONS DEPARTMENT	LINE 2: TENT (8FT X 10FT & 8FT X 8FT)	P 992,200.00	GENERAL FUND	30 CD
6	QCRRMO-22-SOP-1742	QUEZON CITY DISASTER RISK REDUCTION AND MANAGEMENT OFFICE	ELSAROC BOX	P 16,030,552.50	GENERAL FUND	30 CD
7	QCRRMO-22-HCS-1747	QUEZON CITY DISASTER RISK REDUCTION AND MANAGEMENT OFFICE	PNEUMATIC TENT	P 9,605,928.00	GENERAL FUND	30 CD
8	QCRRMO-22-VEHICLES-1372	QUEZON CITY DISASTER RISK REDUCTION AND MANAGEMENT OFFICE	ADVANCE CARDIAC LIFE SUPPORT AMBULANCE	P 73,500,000.00	GENERAL FUND	120 CD

1. The **QUEZON CITY LOCAL GOVERNMENT**, through the *General Fund of various years* intends to apply the sums stated above being the ABC to payments under the contract for *the above stated projects of contract for each lot/item*. Bids received in excess of the ABC shall be automatically rejected at bid opening.
2. The **QUEZON CITY LOCAL GOVERNMENT** now invites bids for various **Projects**. Delivery of the Goods is required *as stated above*. Bidders should have completed, within **the last three (3) years** from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II. Instructions to Bidders.
3. Bidding will be conducted through open competitive bidding procedures using a non-discretionary “*pass/fail*” criterion as specified in the 2016 revised Implementing Rules and Regulations (IRR) of Republic Act (RA) No. 9184.
 - a. Bidding is restricted to Filipino citizens/sole proprietorships, partnerships, or organizations with at least sixty percent (60%) interest or outstanding capital stock belonging to citizens of the Philippines, and to citizens or organizations of a country the laws or regulations of which grant similar rights or privileges to Filipino citizens, pursuant to RA No. 5183.
4. Prospective Bidders may obtain further information from **QUEZON CITY GOVERNMENT Bids and Awards Committee (BAC) Secretariat** and inspect the Bidding Documents at the address given below during *weekdays from 8:00 a.m. – 5:00 p.m.*
5. A complete set of Bidding Documents may be acquired by interested Bidders on **Thursday, December 22, 2022** from the given address and website(s) below *and upon payment of the applicable fee for the Bidding Documents, pursuant to the latest Guidelines issued by the GPPB*. The Procuring Entity shall allow the bidder to present its proof of payment for the fees *in person*.

STANDARD RATES:

Approved Budget for the Contract	Maximum Cost of Bidding Documents (in Philippine Peso)
500,000 and below	500.00
More than 500,000 up to 1 Million	1,000.00
More than 1 Million up to 5 Million	5,000.00
More than 5 Million up to 10 Million	10,000.00
More than 10 Million up to 50 Million	25,000.00
More than 50 Million up to 500 Million	50,000.00
More than 500 Million	75,000.00

The following are the requirements for purchase of Bidding Documents;

1. PhilGEPS Registration Certificate (Platinum – 3 pages)
 2. Document Request List (DRL)
 3. Authorization to Purchase Bidding Documents
 - 3.1 Corporate Secretary Certificate for corporation (specific for the project)
 - 3.2 Special Power of Attorney for single proprietorship (specific for the project)
 4. Notarized Joint Venture Agreement (as applicable)
6. The *Quezon City Local Government* will hold a Pre-Bid Conference on 10:30 A.M. of **Tuesday, January 03, 2023** at **2nd Floor, Procurement Department-Bidding Room, Finance Building, Quezon City Hall Compound**, and/or through video conferencing *via Zoom* which shall be open to prospective bidders.

Topic: BAC-GOODS Pre-Bid Conference Meeting

Join Zoom Meeting

<https://us02web.zoom.us/j/84835002246?pwd=OVRuVE0weXZMNXYwZG5LaWdldXk1QT09>

Meeting ID: 848 3500 2246

Passcode: 154733

7. Bids must be duly received by the BAC Secretariat through manual submission at the 2nd Floor, Procurement Department, Finance Building, Quezon City Hall Compound on or before 11:00 A.M. of **Monday, January 16, 2023**. Late bids shall not be accepted.
8. All Bids must be accompanied by a bid security in any of the acceptable forms and in the amount stated in ITB Clause 14.
9. Bid opening shall be on 01:00 P.M. of **Monday, January 16, 2023** at the given address below and/or via Zoom. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.

Topic: BAC-GOODS & SERVICES BIDDING

Join Zoom Meeting

<https://us02web.zoom.us/j/85850855933?pwd=R2dZUUp4Z3lyU29iZGVlWmdKRjZCdz09>

Meeting ID: 858 5085 5933

Passcode: 118682
10. The *Quezon City Local Government* reserves the right to reject any and all bids, declare a failure of bidding, or not award the contract at any time prior to contract award in accordance with Sections 35.6 and 41 of the 2016 revised IRR of RA No. 9184, without thereby incurring any liability to the affected bidder or bidders.

11. For further information, please refer to:

ATTY. DOMINIC B. GARCIA

OIC, Procurement Department

2nd Floor, Procurement Department,

Finance Building, Quezon City Hall Compound

Elliptical Road, Barangay Central Diliman, Quezon City.

Email Add: bacgoods.procurement@quezoncity.gov.ph

Tel. No. (02)8988-4242 loc. 8506/8710

Website: www.quezoncity.gov.ph

12. You may visit the following websites:

For downloading of Bidding Documents: www.quezoncity.gov.ph

By:


MA. MARGARITA T. SANTOS
Chairperson, QC-BAC-Goods and Services

Section II. Instructions to Bidders

Notes on the Instructions to Bidders

This Section on the Instruction to Bidders (ITB) provides the information necessary for bidders to prepare responsive bids, in accordance with the requirements of the Procuring Entity. It also provides information on bid submission, eligibility check, opening and evaluation of bids, post-qualification, and on the award of contract.

1. Scope of Bid

The Procuring Entity, **Quezon City Local Government** wishes to receive Bids for the **PROCUREMENT OF ADVANCE CARDIAC LIFE SUPPORT AMBULANCE** with identification number **QCRRMO-22-VEHICLES-1372**.

[Note: The Project Identification Number is assigned by the Procuring Entity based on its own coding scheme and is not the same as the PhilGEPS reference number, which is generated after the posting of the bid opportunity on the PhilGEPS website.]

The Procurement Project (referred to herein as “Project”) is composed of **One (1) item**, the details of which are described in Section VII (Technical Specifications).

2. Funding Information

2.1. The GOP through the source of funding as indicated below for **2022** in the amount of **SEVENTY THREE MILLION FIVE HUNDRED THOUSAND PESOS AND 00/100 ONLY (Php73,500,000.00)**.

2.2. The source of funding is:

a. LGUs, the Annual or Supplemental Budget, as approved by the Sanggunian.

3. Bidding Requirements

The Bidding for the Project shall be governed by all the provisions of RA No. 9184 and its 2016 revised IRR, including its Generic Procurement Manuals and associated policies, rules and regulations as the primary source thereof, while the herein clauses shall serve as the secondary source thereof.

Any amendments made to the IRR and other GPPB issuances shall be applicable only to the ongoing posting, advertisement, or **IB** by the BAC through the issuance of a supplemental or bid bulletin.

The Bidder, by the act of submitting its Bid, shall be deemed to have verified and accepted the general requirements of this Project, including other factors that may affect the cost, duration and execution or implementation of the contract, project, or work and examine all instructions, forms, terms, and project requirements in the Bidding Documents.

4. Corrupt, Fraudulent, Collusive, and Coercive Practices

The Procuring Entity, as well as the Bidders and Suppliers, shall observe the highest standard of ethics during the procurement and execution of the contract. They or through an agent shall not engage in corrupt, fraudulent, collusive, coercive, and obstructive practices defined under Annex “I” of the 2016 revised IRR of RA No. 9184 or other integrity violations in competing for the Project.

5. Eligible Bidders

- 5.1. Only Bids of Bidders found to be legally, technically, and financially capable will be evaluated.
- 5.2. Foreign ownership exceeding those allowed under the rules may participate pursuant to:
 - i. When a Treaty or International or Executive Agreement as provided in Section 4 of the RA No. 9184 and its 2016 revised IRR allow foreign bidders to participate;
 - ii. Citizens, corporations, or associations of a country, included in the list issued by the GPPB, the laws or regulations of which grant reciprocal rights or privileges to citizens, corporations, or associations of the Philippines;
 - iii. When the Goods sought to be procured are not available from local suppliers; or
 - iv. When there is a need to prevent situations that defeat competition or restrain trade.
- 5.3. Pursuant to Section 23.4.1.3 of the 2016 revised IRR of RA No.9184, the Bidder shall have an SLCC that is at least one (1) contract similar to the Project the value of which, adjusted to current prices using the PSA's CPI, must be at least equivalent to:
 - a. For the procurement of Non-Expendable Supplies and Services: The Bidder must have completed a single contract that is similar to this Project, equivalent to at least fifty percent (50%) of the ABC.
- 5.4. The Bidders shall comply with the eligibility criteria under Section 23.4.1 of the 2016 IRR of RA No. 9184.

6. Origin of Goods

There is no restriction on the origin of goods other than those prohibited by a decision of the UN Security Council taken under Chapter VII of the Charter of the UN, subject to Domestic Preference requirements under **ITB** Clause 18.

7. Subcontracts

- 7.1. The Bidder may subcontract portions of the Project to the extent allowed by the Procuring Entity as stated herein, but in no case more than twenty percent (20%) of the Project.

The Procuring Entity has prescribed that: Subcontracting is not allowed.

8. Pre-Bid Conference

The Procuring Entity will hold a pre-bid conference for this Project on the specified date and time and either at its physical address and/or through videoconferencing as indicated in paragraph 6 of the **IB**.

9. Clarification and Amendment of Bidding Documents

Prospective bidders may request for clarification on and/or interpretation of any part of the Bidding Documents. Such requests must be in writing and received by the Procuring Entity, either at its given address or through electronic mail indicated in the **IB**, at least ten (10) calendar days before the deadline set for the submission and receipt of Bids.

10. Documents comprising the Bid: Eligibility and Technical Components

- 10.1. The first envelope shall contain the eligibility and technical documents of the Bid as specified in **Section VIII (Checklist of Technical and Financial Documents)**.
- 10.2. The Bidder's SLCC as indicated in **ITB** Clause 5.3 should have been completed within *the last three (3) years* prior to the deadline for the submission and receipt of bids.
- 10.3. If the eligibility requirements or statements, the bids, and all other documents for submission to the BAC are in foreign language other than English, it must be accompanied by a translation in English, which shall be authenticated by the appropriate Philippine foreign service establishment, post, or the equivalent office having jurisdiction over the foreign bidder's affairs in the Philippines. Similar to the required authentication above, for Contracting Parties to the Apostille Convention, only the translated documents shall be authenticated through an Apostille pursuant to GPPB Resolution No. 13-2019 dated 23 May 2019. The English translation shall govern, for purposes of interpretation of the bid.

11. Documents comprising the Bid: Financial Component

- 11.1. The second bid envelope shall contain the financial documents for the Bid as specified in **Section VIII (Checklist of Technical and Financial Documents)**.
- 11.2. If the Bidder claims preference as a Domestic Bidder or Domestic Entity, a certification issued by DTI shall be provided by the Bidder in accordance with Section 43.1.3 of the 2016 revised IRR of RA No. 9184.
- 11.3. Any bid exceeding the ABC indicated in paragraph 1 of the **IB** shall not be accepted.
- 11.4. For Foreign-funded Procurement, a ceiling may be applied to bid prices provided the conditions are met under Section 31.2 of the 2016 revised IRR of RA No. 9184.

12. Bid Prices

- 12.1. Prices indicated on the Price Schedule shall be entered separately in the following manner:

- a. For Goods offered from within the Procuring Entity's country:
 - i. The price of the Goods quoted EXW (ex-works, ex-factory, ex-warehouse, ex-showroom, or off-the-shelf, as applicable);
 - ii. The cost of all customs duties and sales and other taxes already paid or payable;
 - iii. The cost of transportation, insurance, and other costs incidental to delivery of the Goods to their final destination; and
 - iv. The price of other (incidental) services, if any, listed in e.
- b. For Goods offered from abroad:
 - i. Unless otherwise stated in the **BDS**, the price of the Goods shall be quoted delivered duty paid (DDP) with the place of destination in the Philippines as specified in the **BDS**. In quoting the price, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible source country.
 - ii. The price of other (incidental) services, if any, as listed in **Section VII (Technical Specifications)**.

13. Bid and Payment Currencies

- 13.1. For Goods that the Bidder will supply from outside the Philippines, the bid prices may be quoted in the local currency or tradeable currency accepted by the BSP at the discretion of the Bidder. However, for purposes of bid evaluation, Bids denominated in foreign currencies, shall be converted to Philippine currency based on the exchange rate as published in the BSP reference rate bulletin on the day of the bid opening.
- 13.2. Payment of the contract price shall be made in:
 - a. Philippine Pesos.

14. Bid Security

- 14.1. The Bidder shall submit a Bid Securing Declaration¹ or any form of Bid Security in the amount indicated in the **BDS**, which shall be not less than the percentage of the ABC in accordance with the schedule in the **BDS**.
- 14.2. The Bid and bid security *in no case shall exceed One Hundred Twenty (120) calendar days from the date of opening of bids, unless duly extended by the bidder upon the request of the Head of the Procuring Entity (HoPE) of the Quezon City Local Government*. Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.

¹ In the case of Framework Agreement, the undertaking shall refer to entering into contract with the Procuring Entity and furnishing of the performance security or the performance securing declaration within ten (10) calendar days from receipt of Notice to Execute Framework Agreement.

15. Sealing and Marking of Bids

Each Bidder shall submit one copy of the first and second components of its Bid.

The Procuring Entity may request additional hard copies and/or electronic copies of the Bid. However, failure of the Bidders to comply with the said request shall not be a ground for disqualification.

If the Procuring Entity allows the submission of bids through online submission or any other electronic means, the Bidder shall submit an electronic copy of its Bid, which must be digitally signed. An electronic copy that cannot be opened or is corrupted shall be considered non-responsive and, thus, automatically disqualified.

16. Deadline for Submission of Bids

16.1. The Bidders shall submit on the specified date and time through manual submission as indicated in paragraph 7 of the **IB**.

17. Opening and Preliminary Examination of Bids

17.1. The BAC shall open the Bids in public at the time, on the date, and at the place specified in paragraph 9 of the **IB**. The Bidders' representatives who are present shall sign a register evidencing their attendance. In case videoconferencing, webcasting or other similar technologies will be used, attendance of participants shall likewise be recorded by the BAC Secretariat.

In case the Bids cannot be opened as scheduled due to justifiable reasons, the rescheduling requirements under Section 29 of the 2016 revised IRR of RA No. 9184 shall prevail.

17.2. The preliminary examination of bids shall be governed by Section 30 of the 2016 revised IRR of RA No. 9184.

18. Domestic Preference

18.1. The Procuring Entity will grant a margin of preference for the purpose of comparison of Bids in accordance with Section 43.1.2 of the 2016 revised IRR of RA No. 9184.

19. Detailed Evaluation and Comparison of Bids

19.1. The Procuring BAC shall immediately conduct a detailed evaluation of all Bids rated "*passed*," using non-discretionary pass/fail criteria. The BAC shall consider the conditions in the evaluation of Bids under Section 32.2 of the 2016 revised IRR of RA No. 9184.

19.2. If the Project allows partial bids, bidders may submit a proposal on any of the lots or items, and evaluation will be undertaken on a per lot or item basis, as the case maybe. In this case, the Bid Security as required by **ITB** Clause 15 shall be submitted for each lot or item separately.

19.3. The descriptions of the lots or items shall be indicated in **Section VII (Technical Specifications)**, although the ABCs of these lots or items are indicated in the **BDS** for purposes of the NFCC computation pursuant to Section 23.4.2.6 of the 2016 revised IRR of RA No. 9184. The NFCC must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder.

19.4. The Project shall be awarded as follows:

One Project having several items that shall be awarded as one contract.

19.5. Except for bidders submitting a committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation, all Bids must include the NFCC computation pursuant to Section 23.4.1.4 of the 2016 revised IRR of RA No. 9184, which must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder. For bidders submitting the committed Line of Credit, it must be at least equal to ten percent (10%) of the ABCs for all the lots or items participated in by the prospective Bidder.

20. Post-Qualification

20.1. Within a non-extendible period of five (5) calendar days from receipt by the Bidder of the notice from the BAC that it submitted the Lowest Calculated Bid, the Bidder shall submit its latest income and business tax returns filed and paid through the BIR Electronic Filing and Payment System (eFPS) and other appropriate licenses and permits required by law and stated in the **BDS**.

21. Signing of the Contract

21.1. The documents required in Section 37.2 of the 2016 revised IRR of RA No. 9184 shall form part of the Contract. Additional Contract documents are indicated in the **BDS**.

Section III. Bid Data Sheet

Notes on the Bid Data Sheet

The Bid Data Sheet (BDS) consists of provisions that supplement, amend, or specify in detail, information, or requirements included in the ITB found in Section II, which are specific to each procurement.

This Section is intended to assist the Procuring Entity in providing the specific information in relation to corresponding clauses in the ITB and has to be prepared for each specific procurement.

The Procuring Entity should specify in the BDS information and requirements specific to the circumstances of the Procuring Entity, the processing of the procurement, and the bid evaluation criteria that will apply to the Bids. In preparing the BDS, the following aspects should be checked:

- a. Information that specifies and complements provisions of the ITB must be incorporated.
- b. Amendments and/or supplements, if any, to provisions of the ITB as necessitated by the circumstances of the specific procurement, must also be incorporated.

Bid Data Sheet

ITB Clause	
5.3	<p>For this purpose, contracts similar to the Project shall be:</p> <ol style="list-style-type: none"> a. <i>A single contract similar to the item/s to be bid and must be at least Fifty Percent (50%) of the ABC.</i> b. Completed within the last three (3) years prior to the deadline for the submission and receipt of bids substantially in a FORM prescribed by the QC-BAC-GOODS AND SERVICES, must be accompanied by a copy of Certificate of Acceptance by the end-user or Official Receipt (O.R) or Sales Invoice (S.I.) issued for the Contract.
7.1	Subcontracting is not allowed.
12	The price of the Goods shall be quoted DDP <i>within Quezon City</i> or the applicable International Commercial Terms (INCOTERMS) for this Project.
14.1	<p>The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:</p> <ol style="list-style-type: none"> a. The amount of not less than Php 1,470,000.00 or equivalent to two percent (2%) of ABC if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or b. The amount of not less than Php 3,675,000.00 or equivalent to five percent (5%) of ABC if bid security is in Surety Bond.
19.3	<p><i>[In case the Project will be awarded by lot, list the grouping of lots by specifying the group title, items, and the quantity for every identified lot, and the corresponding ABC for each lot.]</i></p> <p><i>[In case the project will be awarded by item, list each item indicating its quantity and ABC.]</i></p>
20.2	<p>List of required licenses and permits relevant to the Project and the corresponding law requiring it.</p> <ul style="list-style-type: none"> • No additional requirements
21.2	<p>Additional required documents relevant to the Project that are required by existing laws and/or the Procuring Entity.</p> <ul style="list-style-type: none"> • Copy of Authority to Sell from the Manufacturer/Distributor/ Dealer of the Vehicle being offered which includes a guarantee on after-sales services and availability of parts • List of after-sales/maintenance services facilities located within Metro Manila • Statement of Warranty <ul style="list-style-type: none"> - One (1) Year on parts and service - 100,000 mileage or three (3) years whichever comes first • Copy of valid, current License to Operate from DOH Accreditation as Supplier, Distributor or Manufacturer for Medical or Hospital Equipment or Devices

	<ul style="list-style-type: none">• Copy of valid, current License to Operate from FDA and/or DOH Accreditation as Supplier, Distributor or Manufacturer for Drugs and Medicines.• Copy of current and valid Certificate of Product Registration (CPR) for the drugs and medicines.• Statement of the bidder specifying that the expiration dates of the drugs and medicines from the date of delivery shall be two (2) to three (3) years for oral respiratory drugs, intravenous and peritoneal solutions, topical, ophthalmologic and ENT preparations; one (1) to two (2) years for injectables; one (1) year for anesthetics.
--	--

Section IV. General Conditions of Contract

Notes on the General Conditions of Contract

The General Conditions of Contract (GCC) in this Section, read in conjunction with the Special Conditions of Contract in Section V and other documents listed therein, should be a complete document expressing all the rights and obligations of the parties.

Matters governing performance of the Supplier, payments under the contract, or matters affecting the risks, rights, and obligations of the parties under the contract are included in the GCC and Special Conditions of Contract.

Any complementary information, which may be needed, shall be introduced only through the Special Conditions of Contract.

1. Scope of Contract

This Contract shall include all such items, although not specifically mentioned, that can be reasonably inferred as being required for its completion as if such items were expressly mentioned herein. All the provisions of RA No. 9184 and its 2016 revised IRR, including the Generic Procurement Manual, and associated issuances, constitute the primary source for the terms and conditions of the Contract, and thus, applicable in contract implementation. Herein clauses shall serve as the secondary source for the terms and conditions of the Contract.

This is without prejudice to Sections 74.1 and 74.2 of the 2016 revised IRR of RA No. 9184 allowing the GPPB to amend the IRR, which shall be applied to all procurement activities, the advertisement, posting, or invitation of which were issued after the effectivity of the said amendment.

Additional requirements for the completion of this Contract shall be provided in the **Special Conditions of Contract (SCC)**.

2. Advance Payment and Terms of Payment

2.1. Advance payment of the contract amount is provided under Annex “D” of the revised 2016 IRR of RA No. 9184.

2.2. The Procuring Entity is allowed to determine the terms of payment on the partial or staggered delivery of the Goods procured, provided such partial payment shall correspond to the value of the goods delivered and accepted in accordance with prevailing accounting and auditing rules and regulations. The terms of payment are indicated in the **SCC**.

3. Performance Security

Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than prior to the signing of the Contract by both parties, the successful Bidder shall furnish the performance security in any of the forms prescribed in Section 39 of the 2016 revised IRR of RA No. 9184.

4. Inspection and Tests

The Procuring Entity or its representative shall have the right to inspect and/or to test the Goods to confirm their conformity to the Project specifications at no extra cost to the Procuring Entity in accordance with the Generic Procurement Manual. In addition to tests in the **SCC, Section IV (Technical Specifications)** shall specify what inspections and/or tests the Procuring Entity requires, and where they are to be conducted. The Procuring Entity shall notify the Supplier in writing, in a timely manner, of the identity of any representatives retained for these purposes.

All reasonable facilities and assistance for the inspection and testing of Goods, including access to drawings and production data, shall be provided by the Supplier to the authorized inspectors at no charge to the Procuring Entity.

5. Warranty

6.1. In order to assure that manufacturing defects shall be corrected by the Supplier, a warranty shall be required from the Supplier as provided under Section 62.1 of the 2016 revised IRR of RA No. 9184.

- 6.2. The Procuring Entity shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall, repair or replace the defective Goods or parts thereof without cost to the Procuring Entity, pursuant to the Generic Procurement Manual.

6. Liability of the Supplier

The Supplier's liability under this Contract shall be as provided by the laws of the Republic of the Philippines.

If the Supplier is a joint venture, all partners to the joint venture shall be jointly and severally liable to the Procuring Entity.

Section V. Special Conditions of Contract

Notes on the Special Conditions of Contract

Similar to the BDS, the clauses in this Section are intended to assist the Procuring Entity in providing contract-specific information in relation to corresponding clauses in the GCC found in Section IV.

The Special Conditions of Contract (SCC) complement the GCC, specifying contractual requirements linked to the special circumstances of the Procuring Entity, the Procuring Entity's country, the sector, and the Goods purchased. In preparing this Section, the following aspects should be checked:

- a. Information that complements provisions of the GCC must be incorporated.
- b. Amendments and/or supplements to provisions of the GCC as necessitated by the circumstances of the specific purchase, must also be incorporated.

However, no special condition which defeats or negates the general intent and purpose of the provisions of the GCC should be incorporated herein.

Special Conditions of Contract

GCC Clause	
1	<p><i>[List here any additional requirements for the completion of this Contract. The following requirements and the corresponding provisions may be deleted, amended, or retained depending on its applicability to this Contract:]</i></p> <p>Delivery and Documents –</p> <p>For purposes of the Contract, “EXW,” “FOB,” “FCA,” “CIF,” “CIP,” “DDP” and other trade terms used to describe the obligations of the parties shall have the meanings assigned to them by the current edition of INCOTERMS published by the International Chamber of Commerce, Paris. The Delivery terms of this Contract shall be as follows:</p> <p><i>[For Goods supplied from abroad, state:]</i> “The delivery terms applicable to the Contract are DDP delivered <i>[indicate place of destination]</i>. In accordance with INCOTERMS.”</p> <p><i>[For Goods supplied from within the Philippines, state:]</i> “The delivery terms applicable to this Contract are delivered <i>[indicate place of destination]</i>. Risk and title will pass from the Supplier to the Procuring Entity upon receipt and final acceptance of the Goods at their final destination.”</p> <p>Delivery of the Goods shall be made by the Supplier in accordance with the terms specified in Section VI (Schedule of Requirements).</p> <p>For purposes of this Clause the Procuring Entity’s Representative at the Project Site is <i>[indicate name(s)]</i>.</p> <p>Incidental Services –</p> <p>The Supplier is required to provide all of the following services, including additional services, if any, specified in Section VI. Schedule of Requirements: <i>Select appropriate requirements and delete the rest.</i></p> <ol style="list-style-type: none"> a. performance or supervision of on-site assembly and/or start-up of the supplied Goods; b. furnishing of tools required for assembly and/or maintenance of the supplied Goods; c. furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied Goods; d. performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and
	<ol style="list-style-type: none"> e. training of the Procuring Entity’s personnel, at the Supplier’s plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied Goods. f. <i>[Specify additional incidental service requirements, as needed.]</i> <p>The Contract price for the Goods shall include the prices charged by the Supplier for incidental services and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.</p>

Spare Parts –

The Supplier is required to provide all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the Supplier:

Select appropriate requirements and delete the rest.

- a. such spare parts as the Procuring Entity may elect to purchase from the Supplier, provided that this election shall not relieve the Supplier of any warranty obligations under this Contract; and
- b. in the event of termination of production of the spare parts:
 - i. advance notification to the Procuring Entity of the pending termination, in sufficient time to permit the Procuring Entity to procure needed requirements; and
 - ii. following such termination, furnishing at no cost to the Procuring Entity, the blueprints, drawings, and specifications of the spare parts, if requested.

The spare parts and other components required are listed in **Section VI (Schedule of Requirements)** and the cost thereof are included in the contract price.

The Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spare parts or components for the Goods for a period of *[indicate here the time period specified. If not used indicate a time period of three times the warranty period]*.

Spare parts or components shall be supplied as promptly as possible, but in any case, within *[insert appropriate time period]* months of placing the order.

Packaging –

The Supplier shall provide such packaging of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in this Contract. The packaging shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit, and open storage. Packaging case size and weights shall take into consideration, where appropriate, the remoteness of the Goods' final destination and the absence of heavy handling facilities at all points in transit.

The packaging, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified below, and in any subsequent instructions ordered by the Procuring Entity.

The outer packaging must be clearly marked on at least four (4) sides as follows:

Name of the Procuring Entity
Name of the Supplier

	<p>Contract Description Final Destination Gross weight Any special lifting instructions Any special handling instructions Any relevant HAZCHEM classifications</p>
	<p>A packaging list identifying the contents and quantities of the package is to be placed on an accessible point of the outer packaging if practical. If not practical the packaging list is to be placed inside the outer packaging but outside the secondary packaging.</p> <p>Transportation –</p> <p>Where the Supplier is required under Contract to deliver the Goods CIF, CIP, or DDP, transport of the Goods to the port of destination or such other named place of destination in the Philippines, as shall be specified in this Contract, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract Price.</p> <p>Where the Supplier is required under this Contract to transport the Goods to a specified place of destination within the Philippines, defined as the Project Site, transport to such place of destination in the Philippines, including insurance and storage, as shall be specified in this Contract, shall be arranged by the Supplier, and related costs shall be included in the contract price.</p>
	<p>Where the Supplier is required under Contract to deliver the Goods CIF, CIP or DDP, Goods are to be transported on carriers of Philippine registry. In the event that no carrier of Philippine registry is available, Goods may be shipped by a carrier which is not of Philippine registry provided that the Supplier obtains and presents to the Procuring Entity certification to this effect from the nearest Philippine consulate to the port of dispatch. In the event that carriers of Philippine registry are available but their schedule delays the Supplier in its performance of this Contract the period from when the Goods were first ready for shipment and the actual date of shipment the period of delay will be considered force majeure.</p> <p>The Procuring Entity accepts no liability for the damage of Goods during transit other than those prescribed by INCOTERMS for DDP deliveries. In the case of Goods supplied from within the Philippines or supplied by domestic Suppliers risk and title will not be deemed to have passed to the Procuring Entity until their receipt and final acceptance at the final destination.</p> <p>Intellectual Property Rights –</p> <p>The Supplier shall indemnify the Procuring Entity against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the Goods or any part thereof.</p>
2.2	<p><i>[If partial payment is allowed, state]</i> “The terms of payment shall be as follows: _____.”</p>
4	<p>The inspections and tests that will be conducted are: <i>Product Presentation/Demonstration/Site Inspection, if applicable.</i></p>

Section VI. Schedule of Requirements

**PROJECT NAME: PROCUREMENT OF ADVANCE CARDIAC LIFE SUPPORT
AMBULANCE**

PROJECT NO. QCDRRMO-22-VEHICLES-1372

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item Number	Description	Unit of Issue	Quantity	Delivered, Weeks / Months
1	<p>Minimum Technical Specifications for the Following:</p> <p>ADVANCE CARDIAC LIFE SUPPORT AMBULANCE - Variant: 2.8 Diesel M/T</p> <p>OVERALL DIMENSIONS Length x Width x Height: 5,265 mm x 1,950 mm x 1,990 mm Wheelbase: 3,210mm Ground Clearance: 175mm Minimum Tire Turning Radius: 5.5mm Fuel Tank Capacity: 70L Seating Capacity: 15 pax</p> <p>ENGINE & TRANSMISSION Engine Technology: 4 Cylinders, In-Line Type Type: 16-Valve DOHC Displacement: 2,755 cc Maximum Output Ps/Rpm: 176 / 3,400 Maximum Torque Nm/Rpm: 420 / 1,400 - 2,600 Transmission: 6-Speed Manual</p> <p>CHASIS Suspension Front/ Rear: McPherson Strut / Rigid Axle, Leaf Spring Brakes Front/ Rear: Ventilated Discs / Drum Steering System: Variable Power Steering Tires/ Wheels: 215 / 70R16C Steel with Full Cap</p> <p>EXTERIOR Headlamp Bi-beam Multi-Reflector Halogen with integrated Turn Lamp Wiper Front/ Rear Intermittent with Mist Rear Window Defogger with Timer</p> <p>INTERIOR Instrument Panel Meter Gauge: Analog with Digital Trip Meter Multi-Info Display: With Accessory Connector: Front 12V x 2 Glove Box: With Steering Column: Manual Tilt and Telescopic Adjustable Shift Lever Knob: Urethane Dash- Mounted Front Personal Lamp: With (Driver and Passenger) Room Lamp: With Sun visor: Driver + Passenger Rear View Mirror: Day and Night Assist Grip: Front- With, Seat- With Seat: Driver- Slide, Recline, and Vertical Adjust</p>	Lot	3	<p>Within One Hundred Twenty (120) Calendar Days upon issuance of the Notice to Proceed</p>

<p>Front Passenger: With Recline Adjust Front Center Passenger: With Recline Adjust Headrest: Driver + Front Passenger (Adjustable) Rr. No.1: Bench Type/ 3+1 Rr. No. 2: Bench Type/ 2+1 Rr. No. 3: 2+2 (50:50 Space-Up) Material: Fabric Seat Under Compartment: With Carpet: Vinyl Cup / Bottle Holder: Front- Instrument Panel (2pcs.), Center Seat Back (2pcs.) Door Trim (2pcs.) Audio System: 2-DIN CD/ Tuner/ MP3/ AUX/ USB/ BT and 4 Speakers Air-conditioner: Type- Dual, Front Manual Control</p> <p>FUNCTION Power: Window- With (Driver/Front Passenger Auto Up-Down) Door Lock: With Outside Rearview Mirror: Adjust + Retract Illuminated Entry: Ignition Key + Lamp Light Control System: Sensor Only Door Ajar Warning: With Key Remind Warning: With (Buzzer) Light Remind Warning: With</p> <p>SAFETY ABS (Anti- Lock Brake System): With SRS Airbag: Driver + Front Passengers Seat Belt: Front- 3 pt. ELR (2 pcs.) + 2 pt. NR (1pc.) Rear- 2 pt. NR (12 pcs.) High Mount Stop Lamp: With Alarm with Immobilizer</p> <p>Ambulance Conversion:</p> <p>INSTALLATION AND FABRICATION</p> <ul style="list-style-type: none"> - Installation of captain seats and attendants with seatbelts. - Install 1/2 thick plywood and cover floor with rubber vinyl matting. -Ambulance stretcher (4 wheels), collapsible type with wheels, S/S side rails, foam cushion 2 inches thick -Install Ambulance stretcher looking assembly, oxygen tank holder stainless steel locally assembled. -Portable flashlight, 2 batteries, dry cell size D. -One portable medical tank with content and holder 201bs. capacity. -Window curtain, cotton fabric. -Fabricated laminated medicine cabinet, with front cover, (flexi glass). -Lettering of word "AMBULANCE" on four sides portion of the vehicle, 1 inch - 2 inches below the window and the letters spaced 1 inch apart using orange reflectorized plastic sticker with the following dimensions: 6 inches height, 4 inches width and 1 inch thickness. -Stickers includes company logo as specified by the client (maximum of 2 sets logo). -LED dome light, power inverter, electrical outlets. 			
--	--	--	--

<p>-Non-breakable Window Panel at the Back of the Driver Flexi glass at the Sliding Window -Flexi glass at the sliding window</p> <p>Other Equipment and Supplies</p> <ul style="list-style-type: none"> • FIRE EXTINGUISHER Valve Type: Chrome Plated Brass UL&ULC Rating: 2A:10B:C Capacity (lbs): 5 Shipping Weight (lbs): 10.5 Height: 15.5 inches Width: 8 inches (entire width of unit, including hose and handle) Depth: 4.25 inches (cylinder diameter) Range (Initial): 12-18 ft. Discharge time (seconds): 14 Stored Pressure Design Dependable Drawn Steel Cylinders Durable High Gloss Polyester Powder Paint All Metal Valve Construction Temperature Range -65°F to 120°F UL Rated: 2A:10B:C Easy and More Economical to Maintain and Service Large Loop Pull Pin Bar Coded and Bi-lingual Labels User Friendly • Digital Clock • LED Room lights • TV hook Holder Stainless • UV Light • Digital Mobile Radio Specification: General -Double time slot: Perfectly compatible with motorola/motrbo – Frequency Range: Dual Band VHF 136-174MHz &UHF 400-480MHz -Storage channel: 64 groups, 1024 -Stepped Frequency: 2.5/5/6.25/10/12.5/20/25KHz - Operating Voltage: DC 7.4V Frequency Stability: ±1.5ppm -Operating Temperature: -20°C ~ 50°C -Operating Mode: Same- frequency simple or different-frequency simplex Main features Support UV double segment, frequency range UHF400,000-480,000MHz/VHF 136,000- 174,000MHz DMO True2-Slot time slot Direct mode dual-slot communication achieves two groups of calls at a frequency point at the same time, so that the communication efficiency is doubled and frequency resources are saved Digital and Analog Compatible, smooth over Support DMR digital and analog two communication modes to ensure that the original analog products smooth transition to digital products to meet different communication needs Digital Signaling function Rich calling modes that support DMR protocol, including single call, group call and all call; supports remote inhibit function and other applications SMS function Supports SMS, text editing and drat in 64 characters in Chinese and English, so that text messages calls can be sent when voice call is inconvenient High Sound Quality This digital two-way radio uses advanced AMBE +2TM voice processing technology to achieve high quality Manual Writing-Frequency Support 			
---	--	--	--

	<p>channel / frequency mode, allows manual writing-frequency in frequency mode, and writing-frequency software can be used to set the required frequency without connecting the computer. High/low power switch Meets the power requirements of different call distances to save more energy Alarm Function Supports general alarm and silent alarm in emergency situations. Digital mode supports the use of relays Side buttons can be customized without changing the user's operating habits.</p> <ul style="list-style-type: none"> • One (1) set PA System with Siren (80 watts / heavy duty) • One (1) set 50 pounds Oxygen Tank (with gauge, regulator and holder) • One (1) set 5 pounds Portable Oxygen Tank • Two (2) sets Overhead Cabinets (Duco Finish) • One (1) pc IV Fluid Bottle Hanger • One (1) set Modified Seat on the right • One (1) set Modified Seat on the front • One (1) set 3M Reflectorized Tint • One (1) lot Aluminum Checkered Plate Flooring with ACL matting • One (1) set Power Inverter (2,500 watts) • Four (4) pcs Rear Blinkers (6 inches x 6 inches) • Two (2) pcs Front Grill Blinkers (2 inches x 4 inches) • Six (6) pcs Upper Right and Left Side (Slim type) • One (1) set Stretcher Ramp • One (1) set Negative Pressure System • Three (3) pcs Safety Handles • One (1) set Power Inverter (2,500 watts) • One (1) set Non-Porous Division with Acrylic Window (Duco Finish) • One (1) set Stainless overhead grab rail on ceiling on the top of patient/stretcher • One (1) unit Dash Cam • One (1) Rear Bumper Metal Protection • One (1) pair Emergency Hammer • Two (2) pcs Reflectorized Vest • Two (2) pcs Early Warning Device • One (1) set OVM Tools • One (1) pc Hydraulic Jack • One (1) pc Tire Wrench <p>Medical Equipment/Device:</p> <ul style="list-style-type: none"> • One (1) set ambulance/removable stretcher, floor mounted retainer on the left side (wheeled cot with mounted cot fastening system) <p>SPECIFICATIONS: High Position (LxWxH): 195cm x 55cm x 85cm, Low Position (LxWxH): 195cm x 55cm x 25cm, Self-Weight: 38kg, Load: 180kg. The height of loading wheels is adjustable to fit different height of ambulance board. The bed surface is made of foamed cushion, the back can be adjustable by gas spring, Which makes the wounded feel comfortable. It is made of high-quality aluminum alloy with features of light-weighted, use safely and easy for sterilization and clean. The joint is cast by aluminum alloy. The mechanical folding legs are controlled by the handles on both</p>			
--	---	--	--	--

	<p>sides. The stretcher with safety lock, it must be locked up when transport the patient and opened out when get on the ambulance. There are fixed bracket which can fix the stretcher on the ambulance board. Only one person can send the patient to the ambulance. All-Metal Frame Construction- provides strength, long service life, and exceptional performance with heavy patients.</p> <p>Rigid Cervical Collars SPECIFICATIONS: Small: 34.3~40.0 cm Medium: 40.0~46.4 cm Large: 46.4~52.7 cm Indication: Cervical spondylosis HIVD (Herniated Intervertebral Disc) Cervical fracture Post-surgical support Completely height-adjustable allowing desired flexion or extension positioning. Cradles point of chin to reduce rotation. Vinyl-covered foam padding for patient comfort and easy cleaning. Ventilated to allow air circulation.</p> <p>Thermometer Infrared SPECIFICATIONS: Product Size: 40mm*44mm*164mm (L x W x H) Product Weight: 61 g Memory: 32 groups Operating temp: 5°C~40°C Measurement range: 32.0°C~42.2°C (95° ~ 107.6°F) Accuracy: ±0.2°C (35.0°C ~ 42.0°C) ; ±0.4°F (95° ~ 107.6°F) Voltage source: 2 x 1.5 V Batteries: size AAA (Not included) Display: LCD/LED digital display Automatic power off: 60s</p> <p>One (1) set Lower Extremity Traction Device (Supporting, Slings, Traction Strap) Technical features: *Relieve the tissue pressure on fractured leg by means of traction applied thus avoids further hemorrhages, shock and secondary nervous, vascular, bone and muscular damages. *Length adjusted aluminum body durable nylon leg straps automatic traction mechanism *Set of adult and pediatric sizes together, with the carry bag. Model Stretch Size(LxWxH) Normal Size(LxWxH) Self-weight Packing Quantity Size(LxWxH) G.W For adult 136cmx22cmx18cm 90cmx22cmx18cm 2kg 498cmx23cmx20cm 10kg For child 118cmx18cmx17cm 80cmx18cmx17cm 1.8kg 491cm x23cmx20cm 9kg *One (1) pc Joint Above and Joint Below Fracture *One (1) pc Rigid Support appropriate material (cardboard, metal, pneumatic, vacuum, wood or plastic) - various sizes *Three (3) Resistant straps or cravats</p>			
--	---	--	--	--

<p>Orthopedic Stretcher Dimensions: 169cmx45cmx8cm Casualty handling for patients who cannot be moved or rocked No rolling or lifting of the patient is required Stretcher Orthopedic - Two Piece splits into two halves and slides under the casualty Lightweight aluminum construction with three retaining straps extendable length.</p> <p>Spine Board with Strap SPECIFICATIONS: Made of high strength patented plastic which can load up to 159kg. Spine Board Stretchers commonly used for safely transporting injured or immobile patients. Fully compatible with X-rays, CT Scan, MRI Can float on water A must have in first-aid equipment in medical centers, construction areas</p> <p>Ambu bag with Manual Resuscitator Adults/Child/Baby Manual Resuscitator PVC AmbuBag Oxygen Tube First Aid Kit Simple Breathing Apparatus Oxygen Machines Made of high quality ABS material Easy to install and operate Anti-skid, durable and light weight in use 100% brand new and high quality Features: Suitable for first aid training or first aid at ambulance. Silica gel transparent mask can perfectly fit for patient's face. Easy to use, just insert into oropharyngeal airway to keep the patient's s airway open. No need electric device, it can be used whether it has compressed oxygen source or not. Simple design, easy to install and use, suitable for both home use and professional use. Manual Resuscitator Set only, other accessories demo in the picture is not included. Specification: Ball capacity: 1650ml Airbag capacity: 2000ml (2500ml optional) Oxygen tube length: app.210cm / 82.68inches Connector diameter: 1.5/2.2cm (0.59/0.87inches) Pressure limiting safety valve: 60cmH2O Material: PVC Color: White Optional style: Adult, Child, Baby Quantity: 1 Set</p> <p>Foldable Wheelchair SPECIFICATIONS: 11 kg with portable/foldable Thickened high carbon steel frame Thick 3D bee mesh back cushion/durable/good breathability Front and rear hand brakes, safe and reliable Good grip/wear resistance Backpack size: 24cm * 28cm Inflatable flywheel, 360 degree rotation, free implementation Strengthen help, overcome obstacles and drive smoothly Foldable footrest saves more space when</p>			
---	--	--	--

<p>folded. The whole car weighs only 11 kg with portable/foldable</p> <p>Automated External Defibrillator (AED) SPECIFICATIONS: Self-test: Configurable automatic self-test from 1 to 7 days. Default = every 7 days. Monthly full-energy test (200J). Automatic Self-test Checks: Battery capacity, electrode connection, electrocardiogram and charge/discharge circuits, microprocessor hardware and software, CPR circuitry and CPR-D sensor, and audio circuitry CPR: Metronome Rate: Variable 60 to 100 CPM Depth: 3/4inch to 3.5inches; 1.9cm to 8.9 cm Defibrillation Advisory: Evaluates electrode connection and patient ECG to determine if defibrillation is required Shockable Rhythms: Ventricular fibrillation with average amplitude >100 microvolts and wide complex ventricular tachycardia with rates greater than 150 BPM for adults, 200 BPM for pediatrics. For ECG Analysis Algorithm sensitivity and specificity, refer to AED Plus Administrator's Guide. Patient Impedance Measurement Range: 0 to 300 ohms Defibrillator: Protected ECG circuitry Display Format: Optional LCD with moving bar Size: 2.6inches x 1.3 inches; 6.6 cm x 3.3 cm Optional ECG Viewing Window: 2.6 seconds Optional ECG Display Sweep Speed: 25 mm/sec; 1 inch/sec Battery Capacity: Typical new (20°C) = 5 years (225 shocks) or 13 hours continuous monitoring. End of life designated by Red X (typical remaining shocks = 9). Data Recording and Storage: 7 hours of ECG and CPR data. If audio recording option is ordered and enabled, 20 minutes of audio recording, ECG, and CPR data. PC Minimum Requirements for Configuration and Patient Data Recovery (not compatible with Mac OS): Windows XP, Windows 7, Windows 8, Windows 10, IBM compatible PII with 16550 UART (or higher) computer. 64MB RAM. VGA monitor or better. CD-ROM drive. IrDA port. 2MB disk space. Device Size: (H x W x D) 5.25 inches x 9.50 inches x 11.50 inches; 13.3 cm x 24.1 cm x 29.2 cm Weight: 6.7 lbs;3.1 kg Power: User-replaceable batteries. 10 Type 123A Photo Flash lithium manganese dioxide batteries Device Classification: Class III and internally powered per EN60601-1 Design Standards: Meets applicable requirements of UL 2601, AAMI DF-39, IEC 601-2-4, EN60601-1, IEC60601-1-2 Environmental Operating Temperature: 32° to 122°F; 0° to 50°C Storage Temperature: -22° to 140°F; -30° to 60°C Humidity: 10 to 95% relative humidity, non- condensing Vibration: MIL Std. 810F, Min. Helicopter Test Shock: IEC 68-2-27; 100G Altitude: -300 to 15,000 ft.; -91 m to 4,573 m Particle and Water Ingress: IP-55 CPR-D-padz Shelf Life: 5 years Conductive Gel: Polymer Hydrogel Conductive Element: Tin Packaging: Multilayer foil laminate pouch Impedance</p>			
---	--	--	--

<p>Class: Low Cable Length: 48 inches (1.2 m) Sternum: Length: 6.1 inches (15.5 cm); Width: 5.0 inches (12.7 cm); Length, conductive gel: 3.5 inches (8.9 cm); Width, conductive gel: 3.5 inches (8.9cm); Area, conductive gel: 12.3 sq inches (79.0 sq cm) Apex: Length: 6.1 inches (15.5 cm); Width: 5.6 inches (14.1cm); Length, conductive gel: 3.5 inches (8.9cm); Width, conductive gel: 3.5 inches (8.9 cm); Area, conductive gel: 12.3 sq inches (79.0 sq cm) Complete Assembly: Folded length: 7.6 inches (19.4 cm); Folded width: 7.0 inches (17.8 cm); Folded height: 1.5 inches (3.8 cm) Design Standards: Meets applicable requirements of ANSI/AAMI/ISO DF-39-1993 A</p> <p>Portable Suction Machine SPECIFICATIONS: Suction Machine Power voltage: 220v/60Hz Lightweight and Portable Oil-free Pump Heavy Duty</p> <p>Nebulizer Machine with kit SPECIFICATIONS: Compressor, disposable kit, REUSABLE KIT, filters CARRY BAG and five-year warranty Particle size: 0.5µm to 5µm Compact Design Liter low: 8 LPM</p> <p>Pulse Oximeter SPECIFICATIONS: Dual color LED display Low perfusion: 0.2% Anti-shaking Alarm value adjusted in friendly menu 6 display mode Big font display</p> <p>Laryngoscope Set (Adult and Pedia) Airway Intubation Kit - 4 Curved Blades 1 Handle Conventional style Easy to Use/Clean Comfortable/Smooth 4 Curved Blades + 1 Handle Included Lightweight Ergonomic Design + High Quality</p> <p>Guedel Oral Airway PVC (Adult and Pedia) Color-coded for easy size identification. These airways serve as a bite block. Keep the patient's airway from collapsing. Made of polyethylene plastic. Guedel oral airway kit 40mm-110mm</p> <p>Endotracheal Tubes (Adult and Pedia) Endotracheal Tubes are manufactured with strong tube walls to ensure the proper ventilation for all patients. Each tube is designed to be easily flexible to conform to patient positions. The graduation marks are drawn for reference along the side of the tube to indicate centimeters. Specifications Cuffed: Two Options (Cuffed - Barrel Type, Uncuffed) Volume High/Pressure Low Eye Style Murphy</p>			
--	--	--	--

<p>Several sizes of Inner Diameter Offered Latex Free</p> <p>Wrap/Blanket for newborn Material: Cotton Package included: 1 x Sleeping Bag + 1 x Hat</p> <p>Head Immobilization Device (Firm Padding) With waterproof, easy to cleaning and other features, reliable and convenient. Can be carried by many people at the same time, X-ray, MRI and CT can be performed with excellent penetrating effect to facilitate the examination of the injured and minimize the pain caused to the patients in the process of handling. It can be used for spine plate and shovel stretcher, durable and not easy to be aged. X-ray, CT scan, MRI and imaging examination can be performed without removing the head immobilizer. Made of high-density plastic material and has two round ear holes on either side for observation of bleeding or drainage.</p> <p>Portable Ventilator SPECIFICATIONS: 1 x Ventilator 1 x Battery 1 x Travelling Case 2 x Oxygen Input Quick Connect Inserts 1 x Whisper Swivel II 1 x Reusable Flexible tubing 6 ft. (1.83m) Grey 1 x PAP Tube Port Adapter 1 x Proximal Flow Sensor 1 x RP Pollen Filters 1 x Kit Active Exhalation Valve UPB 1 x Power Cord 5ft.</p> <p>Sphygmomanometer (Non-Mercurial / Adult and Pedia Cuff) 300 mmHg swivel-type (110° arc) clinical sphygmomanometer for wall mounting. Available with Latex or Non-Latex components. Neutral White ABS plastic manometer with chrome plated steel bracket. Unique high contrast scale and red pointer increase visibility and reduce parallax error. 2 convenient cuff storage options available: large basket or small basket. Instrument Finish: Neutral White ABS Plastic. Instrument Dimensions: 61/4 inches L x 61/4 inches W x 4 inches H Packaged Dimensions: 105/8 inches L x 71/4 inches W x 71/2 inches H Basket Color: Neutral White (standard) Instrument Weight: 1.10 lbs. Packaged Weight: 3.25 lbs. Cuff Size: Adult and Pedia Wall Basket (Nylon Coated Steel)</p> <p>Stethoscope (Pedia and Adult) Description: Dual-Head Stethoscope with Raised stem Soft ear tips Chest piece rim Chill-free</p>			
---	--	--	--

<p>Cardiac Defibrillator SPECIFICATIONS: CPR-D Pads with CPR feedback 1 Rescue Kit- CPR Barrier, Gloves, Towel, Scissors, Razor, Antiseptic Wipe 123 Lithium Batteries 1 Administrator Guide CD 1 Quick Setup CD</p> <p>Cardiac Monitoring Device SPECIFICATIONS: High-resolution display 7 inches Screen 3/5 Lead ECG Monitoring Biphasic waveform Manual and Automated Defibrillation 2 - 360 Joule energy range SpO2 function Rechargeable battery providing 3 hours of ECG monitoring or 100 shocks at 360j 50mm Thermal Printer Easy access user-friendly front panel Recessed menu buttons</p> <p>Trauma Kit with Bag SPECIFICATIONS: Material: 600D PU Oxford cloth size: 45cmx30cmx19cm 1 unit BP Apparatus 1 unit Alpk2 Sphygmomanometer 1 unit Nebulizer 1 unit Aluminum Medical Oxygen 1 bottle Hydrogen Peroxide 500ml 1 bottle Povidone Iodine 1 unit Alcohol 480ml 1 roll Ice Bag 1 roll Elastic Bandage 1 pc Gauze Roll 1 pc Infrared Thermometer 10 pcs 4x4 Sterile Gauze Pad 50 pcs Band Aid Strips 1 pc Arm Sling 1 set Sterile Gloves 1 set Minor Surgical Set 10 pcs Facial Mask 2 pcs Adhesive Plaster 1 pack Cotton Applicator (100's) 1 pc Hot Water Bag 1 pack Cotton Absorbent (50 balls) 10 pcs Tongue Depressor Sterile 1 box Mefenamic Acid 250mg (100's/box) 1 box Amoxicillin 250mg. (100's /box) 1 box Loperamide Capsule, 2mg (100's/box) 1 box Paracetamol 500mg (100's/box)</p> <p>Sterile Delivery Set SPECIFICATIONS: 1 BOITE CLOCHE REF. 189-22, EPAISSEUR 8 CM FOND+COUVERCLE 1 DRAPE, SURGICAL, 75cmx90 cm, without slit, green 1 KIDNEY DISH, medium, 250mmx140mmx40mm, stainless steel 1 SCISSORS, MAYO, 15.5 cm, curved 2 FORCEPS, HEMOSTATIC, ROCHESTER- PEAN, 16 cm, straight 1 FORCEPS, SPONGE, FOERSTER, 24cm, serrated jaws, straight 1 BOWL, ROUND, 100 ml, 80 x 35 mm, stainless steel</p>			
---	--	--	--

<p>1 SCISSORS, EPISIOTOMY, BRAUN-STADLER, 21 cm, curved</p> <p>Minor Surgery Set SPECIFICATIONS: 1 x Kelley Forcep Straight 1 x Kelley Forcep Curve 1 x Needle Holder 1 x Mosquito Forcep Str 1 x Mosquito Forcep Cur 1 x Iris Scissor Str 1 x Blade Handle #4 1 x Blade Handle #3 1 x Thumbs Forceps 1 x Tissue Forceps 1 x Stitch Scissors 1 x Surgical Scissors Str 1 x Bandage Scissors 1 x Probe with Eye 1 x Grove Director 1 x Probe</p> <p>Portable Oxygen Concentrator ADJUSTABLE OXYGEN CONCENTRATION: Supply continuous flow 1-7L/min adjustable, provides a stable 95%±3% (1L) pure oxygen flow. The outflow and its purity is as followings: 1L 95%; 2L 90%; 3L 65%; 4L 50%;5L 40%; 6L 35%;7L 30%.two people can simultaneously absorb oxygen. ANION FUNCTION: To purify the air around for clean oxygen inhalation. Silent and negative ion function: 40 dB mute, increase the negative ions in the air, relax your mind and protect your delicate sleep. PORTABLE & SILENT Dimension: 10.47 inches x 9.05 inches x 11.6 inches Net weight: 13.221b</p> <p>Bag Valve Resuscitator with rebreather bag for Adult, Pedia and Infant. Sanitary resuscitator designed to revive adults and pediatric patients Disposable valve with 3M Filtrate filters helps keep practitioner safe Designed to prevent cross-contamination</p> <p>Portable Suction Machine SPECIFICATIONS: Portable Phlegm Suction Unit Voltage: AC220V 50Hz Max negative pressure value: >0.075MPa Pumping rate: >18L/min Reservoir volume: >1000mL Power: 90VA</p> <p>Cot Patient Shifting Rolls SPECIFICATIONS: Material: Nylon, EPE cotton Bearing Weight: 150KG Size: Approx. 140cm x 70cm/55.1 inches x 27.6 inches</p> <p>Syringe Pump SPECIFICATIONS: Available Syringe size :10, 20, 30, 50ml sizes Delivery Rate: 0.1~1200ml/h (depends on syringe size) Delivery Limit: 0.1-1000ml Accuracy: Mechanical (±1%), With syringe (±3%) K.V.O. Rate: 1~10ml/h (Adjustable) Purge Rate: 0.1-1200ml/h(Adjustable)</p>			
--	--	--	--

<p>Bolus: 0.1~1200ml/h, 0.1-1000ml (Adjustable) Anti Bolus: ON/OFF Dimension: 220 X 130 X 108 (mm) / Weight: 3KG Power: AC 110-220V, 50/60Hz or DC9.6V, DC 24V(1 A) Battery type: Ni-MH/ Operating time: 5hours (at 5ml/h)</p> <p>IV Infusion Pump SPECIFICATIONS: 1. Delivery Rate: 1~1000ml/h, 0.1~999.9ml/h 2. Delivery Limit Range: 0.1-9999ml 3. Volume To Be Infused: 0.1~9999ml 4. Pumping Mechanism: Linear Peristaltic Finger 5. Accuracy: ±5 (ml/h control mode/ with approved IV set) 6. K.V.O. Rate: 1ml/h/10~ml/h (Adjustable) 7. Purge Rate: 1~1000ml/h (Adjustable) 8. Bolus: 1-1000ml/h, 1~9999ml (Adjustable) 9. Dimension: 100 X 190 X 250(mm) / Weight: 3.5KG 10. Power : AC 110-230V, 50/60Hz or DC9.6V, DC 24V(1A) 11. Battery, type: Ni-cd (standard) /Ni MH(option) operating time : 4hours (at 125ml/h)</p> <p>Oxygen Mask (Adult, Infant and Newborn) Material: PVC Sizes: Newborn, Infant and Newborn Soft and comfortable</p> <p>ECG MACHINE SPECIFICATIONS: Ultra-compact and lightweight design High resolution color touch screen Comprehensive filters & anti-noise technology Enhanced powerful technical specifications provides reliable signal quality 12-lead simultaneous acquisition and display Support linear/2D barcode scanner Signal quality identification Long time sampling Large storage capacity Support USB flash disk and micro SD card to extend memory Built-in rechargeable Li-ion battery Bi-directional communication with SE-1515 Data Management System LAN/build-in Wi-Fi (optional) connections Multi-format reports export (optional) Support external printer via USB port.</p> <p>MEDICINES AND FLUIDS: -Activated Charcoal -Salbutamol Nebules -Sterile water for irrigation 1 Liter -Sterile water for injection 10ml -Intravenous fluids -D5 LRS 1 Liter -D5 NSS 1 Liter -D5 Water 1 Liter -D5 0.3 NaCL 500ml -Plain LRS -Plain NSS -Normal saline water (injectable) 10 ml</p>			
--	--	--	--

<p>-Dextrose 50%/50ml vial -Plasma Expander</p> <p>CONTROLLED MEDICATIONS:</p> <p>-Atropine Sulfate 1mg/ml ampule -Epinephrine 1mg/1ml tubaxes (IM, Intracardial, TV) ampule -Diazepam 10mg ampule/vial or Diazepam 5mg/ml 2ml ampule -Dobutamine 250mg ampule or Dobutamine 50 mg/mL, 5 mL ampule -Lidocaine 1gm/25ml vial or Lidocaine 2%, 20 mL ampule -Adenosine 6mg/2ml ampule or Adenosine 3 mg/mL, 2 mL Vial -Human Regular Insulin 100mg/ml vial or Insulin, Regular (Recombinant DNA, human) 100IU/ml, 10 mL -Calcium Gluconate 10% 1mg/10ml ampule/vial -Potassium Chloride 20mg/10ml vial or Potassium Chloride 2mEq/mL, 20 mL Vial -Furosemide 100mg/10ml vial and 20mg/2ml ampule or four pcs of Furosemide 10 mg/mL, 2 mL Ampule -Magnesium Sulfate 50% 1gm/2ml ampule or Magnesium Sulfate 500mg/mL, 2mL Ampule -Dopamine 400mg/5ml vial or Dopamine 40 mg/mL, 5 mL Vial -Diphenhydramine 50mg/ml ampule or Diphenhydramine (as Hydrochloride) 50 mg/mL, 1 mL Ampule -Sodium bicarbonate 10ml ampule or Sodium Bicarbonate 1mEq/ mL, 20mL Ampule -Dogoxin 0.1mg/ml ampule and 0.5mg/2ml ampule or pcs of digoxin 250 mcg/ml, 2mL Ampule -Nitroglycerine spray -Nitroglycerine sublingual -Nitroglycerine patch -Verapamil 5mg/2ml Verapamil 2.5mg/ml, 2ml Ampule</p> <p>Warranty: One (1) Year on parts and service Inclusive of Three (3) Years LTO Registration, Insurance, Warranty: 100,000 mileage or Three (3) Years whichever comes first</p>			

I hereby certify to comply and deliver all the above requirements.

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____

Section VII. Technical Specifications

Notes for Preparing the Technical Specifications

A set of precise and clear specifications is a prerequisite for Bidders to respond realistically and competitively to the requirements of the Procuring Entity without qualifying their Bids. In the context of Competitive Bidding, the specifications (*e.g.* production/delivery schedule, manpower requirements, and after-sales service/parts, descriptions of the lots or items) must be prepared to permit the widest possible competition and, at the same time, present a clear statement of the required standards of workmanship, materials, and performance of the goods and services to be procured. Only if this is done will the objectives of transparency, equity, efficiency, fairness, and economy in procurement be realized, responsiveness of bids be ensured, and the subsequent task of bid evaluation and post-qualification facilitated. The specifications should require that all items, materials and accessories to be included or incorporated in the goods be new, unused, and of the most recent or current models, and that they include or incorporate all recent improvements in design and materials unless otherwise provided in the Contract.

Samples of specifications from previous similar procurements are useful in this respect. The use of metric units is encouraged. Depending on the complexity of the goods and the repetitiveness of the type of procurement, it may be advantageous to standardize the General Technical Specifications and incorporate them in a separate subsection. The General Technical Specifications should cover all classes of workmanship, materials, and equipment commonly involved in manufacturing similar goods. Deletions or addenda should then adapt the General Technical Specifications to the particular procurement.

Care must be taken in drafting specifications to ensure that they are not restrictive. In the specification of standards for equipment, materials, and workmanship, recognized Philippine and international standards should be used as much as possible. Where other particular standards are used, whether national standards or other standards, the specifications should state that equipment, materials, and workmanship that meet other authoritative standards, and which ensure at least a substantially equal quality than the standards mentioned, will also be acceptable. The following clause may be inserted in the Special Conditions of Contract or the Technical Specifications.

Sample Clause: Equivalency of Standards and Codes

Wherever reference is made in the Technical Specifications to specific standards and codes to be met by the goods and materials to be furnished or tested, the provisions of the latest edition or revision of the relevant standards and codes shall apply, unless otherwise expressly stated in the Contract. Where such standards and codes are national or relate to a particular country or region, other authoritative standards that ensure substantial equivalence to the standards and codes specified will be acceptable.

Reference to brand name and catalogue number should be avoided as far as possible; where unavoidable they should always be followed by the words “*or at least equivalent.*” References to brand names cannot be used when the funding source is the GOP.

Where appropriate, drawings, including site plans as required, may be furnished by the Procuring Entity with the Bidding Documents. Similarly, the Supplier may be requested to provide drawings or samples either with its Bid or for prior review by the Procuring Entity during contract execution.

Bidders are also required, as part of the technical specifications, to complete their statement of compliance demonstrating how the items comply with the specification.

Technical Specifications

PROJECT NAME: PROCUREMENT OF ADVANCE CARDIAC LIFE SUPPORT
 AMBULANCE
 PROJECT NO. QCDRRMO-22-VEHICLES-1372

Item	Specification	Statement of Compliance
		<p><i>[Bidders must state here either “Comply” or “Not Comply” against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of “Comply” or “Not Comply” must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer’s un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder’s statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]</i></p>
<p>A.1</p>	<p>Minimum Technical Specifications for the Following:</p> <p>ADVANCE CARDIAC LIFE SUPPORT AMBULANCE - Variant: 2.8 Diesel M/T</p> <p>OVERALL DIMENSIONS Length x Width x Height: 5,265 mm x 1,950 mm x 1,990 mm Wheelbase: 3,210mm Ground Clearance: 175mm Minimum Tire Turning Radius: 5.5mm Fuel Tank Capacity: 70L Seating Capacity: 15 pax</p> <p>ENGINE & TRANSMISSION Engine Technology: 4 Cylinders, In-Line Type Type: 16-Valve DOHC Displacement: 2,755 cc Maximum Output Ps/Rpm: 176 / 3,400 Maximum Torque Nm/Rpm: 420 / 1,400 - 2,600 Transmission: 6-Speed Manual</p> <p>CHASIS Suspension Front/ Rear: McPherson Strut / Rigid Axle, Leaf Spring Brakes Front/ Rear: Ventilated Discs / Drum Steering System: Variable Power Steering Tires/ Wheels: 215 / 70R16C Steel with Full Cap</p> <p>EXTERIOR Headlamp Bi-beam Multi-Reflector Halogen with integrated Turn Lamp Wiper Front/ Rear Intermittent with Mist Rear Window Defogger with Timer</p> <p>INTERIOR Instrument Panel Meter Gauge: Analog with Digital Trip Meter Multi-Info Display: With</p>	

<p>Accessory Connector: Front 12V x 2 Glove Box: With Steering Column: Manual Tilt and Telescopic Adjustable Shift Lever Knob: Urethane Dash- Mounted Front Personal Lamp: With (Driver and Passenger) Room Lamp: With Sun visor: Driver + Passenger Rear View Mirror: Day and Night Assist Grip: Front- With, Seat- With Seat: Driver- Slide, Recline, and Vertical Adjust Front Passenger: With Recline Adjust Front Center Passenger: With Recline Adjust Headrest: Driver + Front Passenger (Adjustable) Rr. No.1: Bench Type/ 3+1 Rr. No. 2: Bench Type/ 2+1 Rr. No. 3: 2+2 (50:50 Space-Up) Material: Fabric Seat Under Compartment: With Carpet: Vinyl Cup / Bottle Holder: Front- Instrument Panel (2pcs.), Center Seat Back (2pcs.) Door Trim (2pcs.) Audio System: 2-DIN CD/ Tuner/ MP3/ AUX/ USB/ BT and 4 Speakers Air-conditioner: Type- Dual, Front Manual Control</p> <p>FUNCTION Power: Window- With (Driver/Front Passenger Auto Up-Down) Door Lock: With Outside Rearview Mirror: Adjust + Retract Illuminated Entry: Ignition Key + Lamp Light Control System: Sensor Only Door Ajar Warning: With Key Remind Warning: With (Buzzer) Light Remind Warning: With</p> <p>SAFETY ABS (Anti- Lock Brake System): With SRS Airbag: Driver + Front Passengers Seat Belt: Front- 3 pt. ELR (2 pcs.) + 2 pt. NR (1pc.) Rear- 2 pt. NR (12 pcs.) High Mount Stop Lamp: With Alarm with Immobilizer</p> <p>Ambulance Conversion:</p> <p>INSTALLATION AND FABRICATION</p> <ul style="list-style-type: none"> - Installation of captain seats and attendants with seatbelts. - Install 1/2 thick plywood and cover floor with rubber vinyl matting. -Ambulance stretcher (4 wheels), collapsible type with wheels, S/S side rails, foam cushion 2 inches thick -Install Ambulance stretcher looking assembly, oxygen tank holder stainless steel locally assembled. -Portable flashlight, 2 batteries, dry cell size D. -One portable medical tank with content and holder 20lbs. capacity. -Window curtain, cotton fabric. -Fabricated laminated medicine cabinet, with front cover, (flexi glass). -Lettering of word "AMBULANCE" on four sides portion of the vehicle, 1 inch - 2 inches below the window and the letters spaced 1 inch apart using orange reflectorized plastic sticker with the 	
---	--

following dimensions: 6 inches height, 4 inches width and 1 inch thickness.
 -Stickers includes company logo as specified by the client (maximum of 2 sets logo).
 -LED dome light, power inverter, electrical outlets.
 -Non-breakable Window Panel at the Back of the Driver
 Flexi glass at the Sliding Window
 -Flexi glass at the sliding window

Other Equipment and Supplies

- FIRE EXTINGUISHER
 Valve Type: Chrome Plated Brass
 UL&ULC Rating: 2A:10B:C
 Capacity (lbs): 5
 Shipping Weight (lbs): 10.5
 Height: 15.5 inches
 Width: 8 inches (entire width of unit, including hose and handle)
 Depth: 4.25 inches (cylinder diameter)
 Range (Initial): 12-18 ft.
 Discharge time (seconds): 14
 Stored Pressure Design
 Dependable Drawn Steel Cylinders
 Durable High Gloss Polyester Powder Paint
 All Metal Valve Construction
 Temperature Range -65°F to 120°F
 UL Rated: 2A:10B:C
 Easy and More Economical to Maintain and Service
 Large Loop Pull Pin
 Bar Coded and Bi-lingual Labels
 User Friendly

 - Digital Clock
 - LED Room lights
 - TV hook Holder Stainless
 - UV Light
 - Digital Mobile Radio
- Specification: General -Double time slot:
 Perfectly compatible with motorola/mototrbo –
 Frequency Range: Dual Band VHF 136-174MHz &UHF 400-480MHz -Storage channel: 64 groups, 1024 -Stepped Frequency: 2.5/5/6.25/10/12.5/20/25KHz -
 Operating Voltage: DC 7.4V
 Frequency Stability: ±1.5ppm -Operating Temperature: -20°C ~ 50°C -Operating Mode: Same- frequency simple or different-frequency simplex Main features
 Support UV double segment, frequency range UHF400,000-480,000MHz/VHF 136,000- 174,000MHz DMO True2-Slot time slot Direct mode dual-slot communication achieves two groups of calls at a frequency point at the same time, so that the communication efficiency is doubled and frequency resources are saved Digital and Analog Compatible, smooth over Support DMR digital and analog two communication modes to ensure that the original analog products smooth transition to digital products to meet different communication needs Digital Signaling function Rich calling modes that support DMR protocol, including single call, group call and all call; supports remote inhibit function and other applications SMS function Supports SMS, text editing and drat in 64 characters in Chinese and English, so that text messages calls can be sent when voice call is inconvenient High Sound Quality This digital two-way radio uses advanced AMBE +2TM voice processing technology to achieve high quality Manual Writing-Frequency

Support channel / frequency mode, allows manual writing-frequency in frequency mode, and writing-frequency software can be used to set the required frequency without connecting the computer. High/low power switch Meets the power requirements of different call distances to save more energy Alarm Function Supports general alarm and silent alarm in emergency situations. Digital mode supports the use of relays Side buttons can be customized without changing the user's operating habits.

- One (1) set PA System with Siren (80 watts / heavy duty)
- One (1) set 50 pounds Oxygen Tank (with gauge, regulator and holder)
- One (1) set 5 pounds Portable Oxygen Tank
- Two (2) sets Overhead Cabinets (Duco Finish)
- One (1) pc IV Fluid Bottle Hanger
- One (1) set Modified Seat on the right
- One (1) set Modified Seat on the front
- One (1) set 3M Reflectorized Tint
- One (1) lot Aluminum Checkered Plate Flooring with ACL matting
- One (1) set Power Inverter (2,500 watts)
- Four (4) pcs Rear Blinkers (6 inches x 6 inches)
- Two (2) pcs Front Grill Blinkers (2 inches x 4 inches)
- Six (6) pcs Upper Right and Left Side (Slim type)
- One (1) set Stretcher Ramp
- One (1) set Negative Pressure System
- Three (3) pcs Safety Handles
- One (1) set Power Inverter (2,500 watts)
- One (1) set Non-Porous Division with Acrylic Window (Duco Finish)
- One (1) set Stainless overhead grab rail on ceiling on the top of patient/stretcher
- One (1) unit Dash Cam
- One (1) Rear Bumper Metal Protection
- One (1) pair Emergency Hammer
- Two (2) pcs Reflectorized Vest
- Two (2) pcs Early Warning Device
- One (1) set OVM Tools
- One (1) pc Hydraulic Jack
- One (1) pc Tire Wrench

Medical Equipment/Device:

- **One (1) set ambulance/removable stretcher, floor mounted retainer on the left side (wheeled cot with mounted cot fastening system)**

SPECIFICATIONS:

High Position (LxWxH): 195cm x 55cm x 85cm, Low Position (LxWxH): 195cm x 55cm x 25cm, Self-Weight: 38kg, Load: 180kg. The height of loading wheels is adjustable to fit different height of ambulance board. The bed surface is made of foamed cushion, the back can be adjustable by gas spring, Which makes the wounded feel comfortable. It is made of high- quality aluminum alloy with features of light-weighted, use safely and easy for sterilization and clean. The joint is cast by aluminum alloy. The mechanical folding legs are controlled by the handles on both sides. The stretcher with safety lock, it must be locked up when transport the patient and opened out when get on the ambulance. There are fixed bracket which can fix the stretcher on the ambulance board. Only one person can send the patient to the ambulance. All-Metal Frame Construction- provides strength, long service life, and exceptional performance with heavy patients.

Rigid Cervical Collars

SPECIFICATIONS:

Small: 34.3~40.0 cm

Medium: 40.0~46.4 cm

Large: 46.4~52.7 cm

Indication:

Cervical spondylosis

HIVD (Herniated Intervertebral Disc)

Cervical fracture

Post-surgical support

Completely height-adjustable allowing desired flexion or extension positioning. Cradles point of chin to reduce rotation. Vinyl-covered foam padding for patient comfort and easy cleaning. Ventilated to allow air circulation.

Thermometer Infrared

SPECIFICATIONS:

Product Size: 40mm*44mm*164mm (L x W x H)

Product Weight: 61 g

Memory: 32 groups

Operating temp: 5°C~40°C

Measurement range: 32.0°C~42.2°C (95° ~ 107.6°F)

Accuracy: ±0.2°C (35.0°C ~ 42.0°C) ; ±0.4°F (95° ~ 107.6°F)

Voltage source: 2 x 1.5 V Batteries: size AAA (Not included)

Display: LCD/LED digital display

Automatic power off: 60s

One (1) set Lower Extremity Traction Device (Supporting, Slings, Traction Strap)

Technical features:

*Relieve the tissue pressure on fractured leg by means of traction applied thus avoids further hemorrhages, shock and secondary nervous, vascular, bone and muscular damages.

*Length adjusted aluminum body durable nylon leg straps automatic traction mechanism

*Set of adult and pediatric sizes together, with the carry bag.

Model Stretch Size(LxWxH) Normal Size(LxWxH)

Self-weight Packing

Quantity Size(LxWxH) G.W

For adult 136cmx22cmx18cm 90cmx22cmx18cm
2kg 498cmx23cmx20cm 10kg

For child 118cm x 18cm x 17cm 80cm x 18cm x
17cm 1.8kg 491cm x 23cm x 20cm 9kg

*One (1) pc Joint Above and Joint Below
Fracture

*One (1) pc Rigid Support appropriate material
(cardboard, metal, pneumatic, vacuum, wood
or plastic) - various sizes

*Three (3) Resistant straps or cravats

Orthopedic Stretcher

Dimensions: 169cmx45cmx8cm

Casualty handling for patients who cannot be moved or rocked No rolling or lifting of the patient is required Stretcher Orthopedic - Two Piece splits into two halves and slides under the casualty Lightweight aluminum construction with three retaining straps extendable length.

Spine Board with Strap

SPECIFICATIONS:

Made of high strength patented plastic which can load up to 159kg.

Spine Board Stretchers commonly used for safely transporting injured or immobile patients.

<p>Fully compatible with X-rays, CT Scan, MRI Can float on water A must have in first-aid equipment in medical centers, construction areas</p> <p>Ambu bag with Manual Resuscitator Adults/Child/Baby Manual Resuscitator PVC AmbuBag Oxygen Tube First Aid Kit Simple Breathing Apparatus Oxygen Machines Made of high quality ABS material Easy to install and operate Anti-skid, durable and light weight in use 100% brand new and high quality Features: Suitable for first aid training or first aid at ambulance. Silica gel transparent mask can perfectly fit for patient's face. Easy to use, just insert into oropharyngeal airway to keep the patient's s airway open. No need electric device, it can be used whether it has compressed oxygen source or not. Simple design, easy to install and use, suitable for both home use and professional use. Manual Resuscitator Set only, other accessories demo in the picture is not included. Specification: Ball capacity: 1650ml Airbag capacity: 2000ml (2500ml optional) Oxygen tube length: app.210cm / 82.68inches Connector diameter: 1.5/2.2cm (0.59/0.87inches) Pressure limiting safety valve: 60cmH2O Material: PVC Color: White Optional style: Adult, Child, Baby Quantity: 1 Set</p> <p>Foldable Wheelchair SPECIFICATIONS: 11 kg with portable/foldable Thickened high carbon steel frame Thick 3D bee mesh back cushion/durable/good breathability Front and rear hand brakes, safe and reliable good grip/wear resistance Backpack size: 24cm * 28cm Inflatable flywheel, 360-degree rotation, free implementation Strengthen help, overcome obstacles and drive smoothly Foldable footrest saves more space when folded. The whole car weighs only 11 kg with portable/foldable</p> <p>Automated External Defibrillator (AED) SPECIFICATIONS: Self-test: Configurable automatic self-test from 1 to 7 days. Default = every 7 days. Monthly full-energy test (200J). Automatic Self-test Checks: Battery capacity, electrode connection, electrocardiogram and charge/discharge circuits, microprocessor hardware and software, CPR circuitry and CPR-D sensor, and audio circuitry CPR: Metronome Rate: Variable 60 to 100 CPM Depth: 3/4inch to 3.5inches; 1.9cm to 8.9 cm Defibrillation Advisory: Evaluates electrode connection and patient ECG to determine if defibrillation is required Shockable Rhythms: Ventricular fibrillation with average amplitude >100 microvolts and wide complex ventricular tachycardia with rates greater than 150 BPM for adults, 200 BPM for pediatrics. For ECG Analysis Algorithm sensitivity and specificity, refer to AED</p>	
---	--

<p>Plus Administrator's Guide. Patient Impedance Measurement Range: 0 to 300 ohms Defibrillator: Protected ECG circuitry Display Format: Optional LCD with moving bar Size: 2.6inches x 1.3 inches; 6.6 cm x 3.3 cm Optional ECG Viewing Window: 2.6 seconds Optional ECG Display Sweep Speed: 25 mm/sec; 1 inch/sec Battery Capacity: Typical new (20°C) = 5 years (225 shocks) or 13 hours continuous monitoring. End of life designated by Red X (typical remaining shocks = 9). Data Recording and Storage: 7 hours of ECG and CPR data. If audio recording option is ordered and enabled, 20 minutes of audio recording, ECG, and CPR data. PC Minimum Requirements for Configuration and Patient Data Recovery (not compatible with Mac OS): Windows XP, Windows 7, Windows 8, Windows 10, IBM compatible PII with 16550 UART (or higher) computer. 64MB RAM. VGA monitor or better. CD-ROM drive. IrDA port. 2MB disk space. Device Size: (H x W x D) 5.25 inches x 9.50 inches x 11.50 inches; 13.3 cm x 24.1 cm x 29.2 cm Weight: 6.7 lbs;3.1 kg Power: User-replaceable batteries. 10 Type 123A Photo Flash lithium manganese dioxide batteries Device Classification: Class III and internally powered per EN60601-1 Design Standards: Meets applicable requirements of UL 2601, AAMI DF-39, IEC 601-2-4, EN60601-1, IEC60601-1-2 Environmental Operating Temperature: 32° to 122°F; 0° to 50°C Storage Temperature: -22° to 140°F; -30° to 60°C Humidity: 10 to 95% relative humidity, non- condensing Vibration: MIL Std. 810F, Min. Helicopter Test Shock: IEC 68-2-27; 100G Altitude: -300 to 15,000 ft.; -91 m to 4,573 m Particle and Water Ingress: IP-55 CPR-D-padz Shelf Life: 5 years Conductive Gel: Polymer Hydrogel Conductive Element: Tin Packaging: Multilayer foil laminate pouch Impedance Class: Low Cable Length: 48 inches (1.2 m) Sternum: Length: 6.1 inches (15.5 cm); Width: 5.0 inches (12.7 cm); Length, conductive gel: 3.5 inches (8.9 cm); Width, conductive gel: 3.5 inches (8.9cm); Area, conductive gel: 12.3 sq inches (79.0 sq cm) Apex: Length: 6.1 inches (15.5 cm); Width: 5.6 inches (14.1cm); Length, conductive gel: 3.5 inches (8.9cm); Width, conductive gel: 3.5 inches (8.9 cm); Area, conductive gel: 12.3 sq inches (79.0 sq cm) Complete Assembly: Folded length: 7.6 inches (19.4 cm); Folded width: 7.0 inches (17.8 cm); Folded height: 1.5 inches (3.8 cm) Design Standards: Meets applicable requirements of ANSI/AAMI/ISO DF-39-1993 A</p> <p>Portable Suction Machine SPECIFICATIONS: Suction Machine Power voltage: 220v/60Hz Lightweight and Portable Oil-free Pump Heavy Duty</p> <p>Nebulizer Machine with kit SPECIFICATIONS: Compressor, disposable kit, REUSABLE KIT, filters CARRY BAG and five-year warranty Particle size: 0.5µm to 5µm Compact Design</p>	
---	--

<p>Liter low: 8 LPM</p> <p>Pulse Oximeter SPECIFICATIONS: Dual color LED display Low perfusion: 0.2% Anti-shaking Alarm value adjusted in friendly menu 6 display mode Big font display</p> <p>Laryngoscope Set (Adult and Pedia) Airway Intubation Kit - 4 Curved Blades 1 Handle Conventional style Easy to Use/Clean Comfortable/Smooth 4 Curved Blades + 1 Handle Included Lightweight Ergonomic Design + High Quality</p> <p>Guedel Oral Airway PVC (Adult and Pedia) Color-coded for easy size identification. These airways serve as a bite block. Keep the patient's airway from collapsing. Made of polyethylene plastic. Guedel oral airway kit 40mm-110mm</p> <p>Endotracheal Tubes (Adult and Pedia) Endotracheal Tubes are manufactured with strong tube walls to ensure the proper ventilation for all patients. Each tube is designed to be easily flexible to conform to patient positions. The graduation marks are drawn for reference along the side of the tube to indicate centimeters. Specifications Cuffed: Two Options (Cuffed - Barrel Type, Uncuffed) Volume High/Pressure Low Eye Style Murphy Several sizes of Inner Diameter Offered Latex Free</p> <p>Wrap/Blanket for newborn Material: Cotton Package included: 1 x Sleeping Bag + 1 x Hat</p> <p>Head Immobilization Device (Firm Padding) With waterproof, easy to cleaning and other features, reliable and convenient. Can be carried by many people at the same time, X-ray, MRI and CT can be performed with excellent penetrating effect to facilitate the examination of the injured and minimize the pain caused to the patients in the process of handling. It can be used for spine plate and shovel stretcher, durable and not easy to be aged. X-ray, CT scan, MRI and imaging examination can be performed without removing the head immobilizer. Made of high-density plastic material and has two round ear holes on either side for observation of bleeding or drainage.</p> <p>Portable Ventilator SPECIFICATIONS: 1 x Ventilator 1 x Battery 1 x Travelling Case 2 x Oxygen Input Quick Connect Inserts 1 x Whisper Swivel II 1 x Reusable Flexible tubing 6 ft. (1.83m) Grey 1 x PAP Tube Port Adapter 1 x Proximal Flow Sensor 1 x RP Pollen Filters</p>	
---	--

1 x Kit Active Exhalation Valve UPB
1 x Power Cord 5ft.

Sphygmomanometer (Non-Mercurial / Adult and Pedia Cuff)

300 mmHg swivel-type (110° arc) clinical sphygmomanometer for wall mounting. Available with Latex or Non-Latex components. Neutral White ABS plastic manometer with chrome plated steel bracket. Unique high contrast scale and red pointer increase visibility and reduce parallax error. 2 convenient cuff storage options available: large basket or small basket. Instrument Finish: Neutral White ABS Plastic. Instrument Dimensions: 61/4 inches L x 61/4 inches W x 4 inches H
Packaged Dimensions: 105/8 inches L x 71/4 inches W x 71/2 inches H
Basket Color: Neutral White (standard)
Instrument Weight: 1.10 lbs.
Packaged Weight: 3.25 lbs.
Cuff Size: Adult and Pedia
Wall Basket (Nylon Coated Steel)

Stethoscope (Pedia and Adult)

Description:
Dual-Head Stethoscope with Raised stem
Soft ear tips
Chest piece rim
Chill-free

Cardiac Defibrillator

SPECIFICATIONS:
CPR-D Pads with CPR feedback
1 Rescue Kit- CPR Barrier, Gloves, Towel, Scissors, Razor, Antiseptic Wipe
123 Lithium Batteries
1 Administrator Guide CD
1 Quick Setup CD

Cardiac Monitoring Device

SPECIFICATIONS:
High-resolution display 7 inches Screen
3/5 Lead ECG Monitoring
Biphasic waveform
Manual and Automated Defibrillation
2 - 360 Joule energy range
SpO2 function
Rechargeable battery providing 3 hours of ECG monitoring or
100 shocks at 360j
50mm Thermal Printer
Easy access user-friendly front panel
Recessed menu buttons

Trauma Kit with Bag

SPECIFICATIONS:
Material: 600D PU Oxford cloth
size: 45cmx30cmx19cm
1 unit BP Apparatus
1 unit Alpk2 Sphygmomanometer
1 unit Nebulizer
1 unit Aluminum Medical Oxygen
1 bottle Hydrogen Peroxide 500ml
1 bottle Povidone Iodine
1 unit Alcohol 480ml
1 roll Ice Bag
1 roll Elastic Bandage
1 pc Gauze Roll
1 pc Infrared Thermometer
10 pcs 4x4 Sterile Gauze Pad
50 pcs Band Aid Strips

<p> 1 pc Arm Sling 1 set Sterile Gloves 1 set Minor Surgical Set 10 pcs Facial Mask 2 pcs Adhesive Plaster 1 pack Cotton Applicator (100's) 1 pc Hot Water Bag 1 pack Cotton Absorbent (50 balls) 10 pcs Tongue Depressor Sterile 1 box Mefenamic Acid 250mg (100's/box) 1 box Amoxicillin 250mg. (100's /box) 1 box Loperamide Capsule, 2mg (100's/box) 1 box Paracetamol 500mg (100's/box) </p> <p> Sterile Delivery Set SPECIFICATIONS: 1 BOITE CLOCHE REF. 189-22, EPAISSEUR 8 CM FOND+COUVERCLE 1 DRAPE, SURGICAL, 75cmx90 cm, without slit, green 1 KIDNEY DISH, medium, 250mmx140mmx40mm, stainless steel 1 SCISSORS, MAYO, 15.5 cm, curved 2 FORCEPS, HEMOSTATIC, ROCHESTER-PEAN, 16 cm, straight 1 FORCEPS, SPONGE, FOERSTER, 24cm, serrated jaws, straight 1 BOWL, ROUND, 100 ml, 80 x 35 mm, stainless steel 1 SCISSORS, EPISIOTOMY, BRAUN-STADLER, 21 cm, curved </p> <p> Minor Surgery Set SPECIFICATIONS: 1 x Kelley Forcep Straight 1 x Kelley Forcep Curve 1 x Needle Holder 1 x Mosquito Forcep Str 1 x Mosquito Forcep Cur 1 x Iris Scissor Str 1 x Blade Handle #4 1 x Blade Handle #3 1 x Thumbs Forceps 1 x Tissue Forceps 1 x Stitch Scissors 1 x Surgical Scissors Str 1 x Bandage Scissors 1 x Probe with Eye 1 x Grove Director 1 x Probe </p> <p> Portable Oxygen Concentrator ADJUSTABLE OXYGEN CONCENTRATION: Supply continuous flow 1-7L/min adjustable, provides a stable 95%±3% (1L) pure oxygen flow. The outflow and its purity is as followings: 1L 95%; 2L 90%; 3L 65%; 4L 50%;5L 40%; 6L 35%;7L 30%.two people can simultaneously absorb oxygen. ANION FUNCTION: To purify the air around for clean oxygen inhalation. Silent and negative ion function: 40 dB mute, increase the negative ions in the air, relax your mind and protect your delicate sleep. PORTABLE & SILENT Dimension: 10.47 inches x 9.05 inches x 11.6 inches Net weight: 13.221b </p> <p> Bag Valve Resuscitator with rebreather bag for Adult, Pedia and Infant. Sanitary resuscitator designed to revive adults and pediatric patients Disposable valve with 3M Filtrate filters helps keep practitioner safe </p>	
--	--

<p>Designed to prevent cross-contamination</p> <p>Portable Suction Machine SPECIFICATIONS: Portable Phlegm Suction Unit Voltage: AC220V 50Hz Max negative pressure value: >0.075MPa Pumping rate: >18L/min Reservoir volume: >1000mL Power: 90VA</p> <p>Cot Patient Shifting Rolls SPECIFICATIONS: Material: Nylon, EPE cotton Bearing Weight: 150KG Size: Approx. 140cm x 70cm/55.1 inches x 27.6 inches</p> <p>Syringe Pump SPECIFICATIONS: Available Syringe size :10, 20, 30, 50ml sizes Delivery Rate: 0.1~1200ml/h (depends on syringe size) Delivery Limit: 0.1-1000ml Accuracy: Mechanical (±1%), With syringe (±3%) K.V.O. Rate: 1~10ml/h (Adjustable) Purge Rate: 0.1-1200ml/h(Adjustable) Bolus: 0.1~1200ml/h, 0.1-1000ml (Adjustable) Anti Bolus: ON/OFF Dimension: 220 X 130 X 108 (mm) / Weight: 3KG Power: AC 110-220V, 50/60Hz or DC9.6V, DC 24V(1 A) Battery type: Ni-MH/ Operating time: 5hours (at 5ml/h)</p> <p>IV Infusion Pump SPECIFICATIONS: 1. Delivery Rate: 1~1000ml/h, 0.1~999.9ml/h 2. Delivery Limit Range: 0.1-9999ml 3. Volume To Be Infused: 0.1~9999ml 4. Pumping Mechanism: Linear Peristaltic Finger 5. Accuracy: ±5 (ml/h control mode/ with approved IV set) 6. K.V.O. Rate: 1ml/h/10~ml/h (Adjustable) 7. Purge Rate: 1~1000ml/h (Adjustable) 8. Bolus: 1-1000ml/h, 1-9999ml (Adjustable) 9. Dimension: 100 X 190 X 250(mm) / Weight: 3.5KG 10. Power : AC 110-230V, 50/60Hz or DC9.6V, DC 24V(1A) 11. Battery, type: Ni-cd (standard) /Ni MH(option) operating time : 4hours (at 125ml/h)</p> <p>Oxygen Mask (Adult, Infant and Newborn) Material: PVC Sizes: Newborn, Infant and Newborn Soft and comfortable</p> <p>ECG MACHINE SPECIFICATIONS: Ultra-compact and lightweight design High resolution color touch screen Comprehensive filters & anti-noise technology Enhanced powerful technical specifications provides reliable signal quality 12-lead simultaneous acquisition and display Support linear/2D barcode scanner Signal quality identification Long time sampling Large storage capacity</p>	
--	--

<p>Support USB flash disk and micro SD card to extend memory Built-in rechargeable Li-ion battery Bi-directional communication with SE-1515 Data Management System LAN/build-in Wi-Fi (optional) connections Multi-format reports export (optional) Support external printer via USB port.</p> <p>MEDICINES AND FLUIDS:</p> <ul style="list-style-type: none"> -Activated Charcoal -Salbutamol Nebules -Sterile water for irrigation 1 Liter -Sterile water for injection 10ml -Intravenous fluids -D5 LRS 1 Liter -D5 NSS 1 Liter -D5 Water 1 Liter -D5 0.3 NaCL 500ml -Plain LRS -Plain NSS -Normal saline water (injectable) 10 ml -Dextrose 50%/50ml vial -Plasma Expander <p>CONTROLLED MEDICATIONS:</p> <ul style="list-style-type: none"> -Atropine Sulfate 1mg/ml ampule -Epinephrine 1mg/1ml tubaxes (IM, Intracardial, TV) ampule -Diazepam 10mg ampule/vial or Diazepam 5mg/ml 2ml ampule -Dobutamine 250mg ampule or Dobutamine 50 mg/mL, 5 mL ampule -Lidocaine 1gm/25ml vial or Lidocaine 2%, 20 mL ampule -Adenosine 6mg/2ml ampule or Adenosine 3 mg/mL, 2 mL Vial -Human Regular Insulin 100mg/ml vial or Insulin, Regular (Recombinant DNA, human) 100IU/ml, 10 mL -Calcium Gluconate 10% 1mg/10ml ampule/vial -Potassium Chloride 20mg/10ml vial or Potassium Chloride 2mEq/mL, 20 mL Vial -Furosemide 100mg/10ml vial and 20mg/2ml ampule or four pcs of Furosemide 10 mg/mL, 2 mL Ampule -Magnesium Sulfate 50% 1gm/2ml ampule or Magnesium Sulfate 500mg/mL, 2mL Ampule -Dopamine 400mg/5ml vial or Dopamine 40 mg/mL, 5 mL Vial -Diphenhydramine 50mg/ml ampule or Diphenhydramine (as Hydrochloride) 50 mg/mL, 1 mL Ampule -Sodium bicarbonate 10ml ampule or Sodium Bicarbonate 1mEq/ mL, 20mL Ampule -Dogoxin 0.1mg/ml ampule and 0.5mg/2ml ampule or pcs of digoxin 250 mcg/ml, 2mL Ampule -Nitroglycerine spray -Nitroglycerine sublingual -Nitroglycerine patch -Verapamil 5mg/2ml Verapamil 2.5mg/ml, 2ml Ampule <p>Warranty: One (1) Year on parts and service Inclusive of Three (3) Years LTO Registration, Insurance, Warranty: 100,000 mileage or Three (3) Years whichever comes first</p>	
---	--

B.	Compliance to the Schedule of Requirements (Section VI)	
-----------	--	--

I hereby certify to comply and deliver all the above requirements.

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____

Cost Derivation

**PROJECT NAME: PROCUREMENT OF ADVANCE CARDIAC LIFE SUPPORT
AMBULANCE
PROJECT NO. QCRRMO-22-VEHICLES-1372**

	PARTICULAR	UNIT OF ISSUE	QTY	UNIT COST	TOTAL
1	ADVANCE CARDIAC LIFE SUPPORT AMBULANCE	UNIT	1		
2	INSTALLATION AND FABRICATION	LOT	1		
3	FIRE EXTINGUISHER	PC	1		
4	Digital Clock	PC	1		
5	LED Room lights	PC	1		
6	TV hook Holder Stainless	PC	1		
7	UV Light	PC	1		
8	Digital Mobile Radio	PC	1		
9	PA System with Siren	SET	1		
10	50 pounds Oxygen Tank	SET	1		
11	5 pounds Portable Oxygen Tank	SET	1		
12	Overhead Cabinets	SET	2		
13	IV Fluid Bottle Hanger	PC	1		
14	Modified Seat on the right	SET	1		
15	Modified Seat on the front	SET	1		
16	3M Reflectorized Tint	SET	1		
17	Aluminum Checkered Plate Flooring with ACL matting	LOT	1		
18	Power Inverter	SET	1		
19	Rear Blinkers	PC	4		
20	Front Grill Blinkers	PC	2		
21	Upper Right and Left Side	PC	6		
22	Stretcher Ramp	SET	1		
23	Negative Pressure System	SET	1		
24	Safety Handles	PC	3		
25	Power Inverter (2,500 watts)	SET	1		
26	Non-Porous Division with Acrylic Window	SET	1		
27	Stainless overhead grab rail on ceiling on the top of patient/stretcher	SET	1		
28	Dash Cam	UNIT	1		
29	Rear Bumper Metal Protection	PC	1		
30	Emergency Hammer	PAIR	1		
31	Reflectorized Vest	PC	2		

32	Early Warning Device	PC	2		
33	OVM Tools	SET	1		
34	Hydraulic Jack	PC	1		
35	Tire Wrench	PC	1		
36	ambulance/removable stretcher, floor mounted retainer on the left side (wheeled cot with mounted cot fastening system)	SET	1		
37	Rigid Cervical Collars	PC	1		
38	Thermometer Infrared	PC	1		
39	Lower Extremity Traction Device (Supporting, Slings, Traction Strap)	SET	1		
40	Orthopedic Stretcher	PC	1		
41	Spine Board with Strap	SET	1		
42	Ambu bag with Manual Resuscitator	SET	1		
43	Foldable Wheelchair	PC	1		
44	Automated External Defibrillator (AED)	UNIT	1		
45	Portable Suction Machine	UNIT	1		
46	Nebulizer Machine with kit	KIT	1		
47	Pulse Oximeter	PC	1		
48	Laryngoscope Set (Adult and Pedia)	SET	1		
49	Guedel Oral Airway PVC (Adult and Pedia)	SET	1		
50	Endotracheal Tubes (Adult and Pedia)	SET	1		
51	Wrap/Blanket for newborn	PC	1		
52	Head Immobilization Device (Firm Padding)	PC	1		
53	Portable Ventilator	UNIT	1		
54	Sphygmomanometer (Non-Mercurial / Adult and Pedia Cuff)	SET	1		
55	Stethoscope (Pedia and Adult)	SET	1		
56	Cardiac Defibrillator	UNIT	1		
57	Cardiac Monitoring Device	UNIT	1		
58	Trauma Kit with Bag	KIT	1		
59	Sterile Delivery Set	SET	1		
60	Minor Surgery Set	SET	1		
61	Portable Oxygen Concentrator	UNIT	1		
62	Bag Valve Resuscitator with rebreather bag for Adult, Pedia and Infant.	SET	1		
63	Portable Suction Machine	UNIT	1		
64	Cot Patient Shifting Rolls	PC	1		
65	Syringe Pump	UNIT	1		
66	IV Infusion Pump	UNIT	1		

67	Oxygen Mask (Adult, Infant and Newborn)	SET	1		
68	ECG MACHINE	UNIT	1		
69	MEDICINES AND FLUIDS	LOT	1		
70	CONTROLLED MEDICATIONS:	LOT	1		
TOTAL AMOUNT PER UNIT/AMBULANCE					
MULTIPLY BY TOTAL QUANTITY					3
TOTAL AMOUNT BID AMOUNT					

I hereby certify to comply and deliver all the above requirements.

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____

Section VIII. Checklist of Technical and Financial Documents

Notes on the Checklist of Technical and Financial Documents

The prescribed documents in the checklist are mandatory to be submitted in the Bid, but shall be subject to the following:

- a. GPPB Resolution No. 09-2020 on the efficient procurement measures during a State of Calamity or other similar issuances that shall allow the use of alternate documents in lieu of the mandated requirements; or
- b. Any subsequent GPPB issuances adjusting the documentary requirements after the effectivity of the adoption of the PBDs.

The BAC shall be checking the submitted documents of each Bidder against this checklist to ascertain if they are all present, using a non-discretionary “pass/fail” criterion pursuant to Section 30 of the 2016 revised IRR of RA No. 9184.

Checklist of Technical and Financial Documents

I. TECHNICAL COMPONENT ENVELOPE

Class “A” Documents

Legal Documents

- (a) Valid PhilGEPS Registration Certificate (Platinum Membership) (all pages) **in accordance with Section 8.5.2 of the IRR;**

Technical Documents

- (b) Statement of the prospective bidder of all its ongoing government and private contracts, including contracts awarded but not yet started, if any, whether similar or not similar in nature and complexity to the contract to be bid (in a **FORM prescribed by the QC-BAC-GOODS AND SERVICES); and**
- (c) Statement of the bidder’s Single Largest Completed Contract (SLCC) similar to the contract to be bid, except under conditions provided for in Sections 23.4.1.3 and 23.4.2.4 of the 2016 revised IRR of RA No. 9184, within the relevant period as provided in the Bidding Documents (in a **FORM prescribed by the QC-BAC-GOODS AND SERVICES); and**
- (d) Original copy of Bid Security. If in the form of a Surety Bond, submit also a certification issued by the Insurance Commission;
or
Original copy of Notarized Bid Securing Declaration; **and**
- (e) Conformity with Section VI. (Schedule of Requirements) and Section VII. (Technical Specifications), which may include production/delivery schedule, manpower requirements, and/or after-sales/parts, if applicable; **and**
- (f) Original duly signed Omnibus Sworn Statement (OSS); **and** if applicable, Original Notarized Secretary’s Certificate in case of a corporation, partnership, or cooperative; or Original Special Power of Attorney of all members of the joint venture giving full power and authority to its officer to sign the OSS and do acts to represent the Bidder.

Financial Documents

- (g) The prospective bidder’s computation of Net Financial Contracting Capacity (NFCC) (in a **FORM prescribed by the QC-BAC-GOODS AND SERVICES);**
or
A committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation.

Class “B” Documents

- (h) If applicable, a duly signed joint venture agreement (JVA) in case the joint venture is already in existence;
or
duly notarized statements from all the potential joint venture partners stating that they will enter into and abide by the provisions of the JVA in the instance that the bid is successful.

Other documentary requirements under RA No. 9184 (as applicable)

- (i) *[For foreign bidders claiming by reason of their country’s extension of reciprocal rights to Filipinos]* Certification from the relevant government office of their country stating that Filipinos are allowed to participate in government procurement activities for the same item or product.
- (j) Certification from the DTI if the Bidder claims preference as a Domestic Bidder or Domestic Entity.

II. FINANCIAL COMPONENT ENVELOPE

- (a) Original of duly signed and accomplished Financial Bid Form;
- (b) Original of duly signed and accomplished Price Schedule(s) **and**
- (c) Cost Derivation.

III. REQUIRED DOCUMENTS in BDS SECTION 20.2 and 21.2

- Copy of Authority to Sell from the Manufacturer/Distributor/ Dealer of the Vehicle being offered which includes a guarantee on after-sales services and availability of parts
- List of after-sales/maintenance services facilities located within Metro Manila
- Statement of Warranty
 - One (1) Year on parts and service
 - 100,000 mileage or three (3) years whichever comes first
- Copy of valid, current License to Operate from DOH Accreditation as Supplier, Distributor or Manufacturer for Medical or Hospital Equipment or Devices
- Copy of valid, current License to Operate from FDA and/or DOH Accreditation as Supplier, Distributor or Manufacturer for Drugs and Medicines.
- Copy of current and valid Certificate of Product Registration (CPR) for the drugs and medicines.
- Statement of the bidder specifying that the expiration dates of the drugs and medicines from the date of delivery shall be two (2) to three (3) years for oral respiratory drugs, intravenous and peritoneal solutions, topical, ophthalmologic and ENT preparations; one (1) to two (2) years for injectables; one (1) year for anesthetics.

Note:

1. Please refer to [\[https://drive.google.com/file/d/1uiYurh5WrpBL5B_pqpzAb62yucAbIR1p/view?usp=sharing\]](https://drive.google.com/file/d/1uiYurh5WrpBL5B_pqpzAb62yucAbIR1p/view?usp=sharing) for the following requirements:
 - a. Computation of NFCC;
 - b. List of All Ongoing Contracts/List of Contracts already awarded but not yet started;
 - c. Statement of Single Largest Completed Contract
2. Please refer to GPPB Resolution No. 16-2020 for the following requirements:
 - a. Bid Form;
 - b. Price Schedule (for Goods Offered from Abroad/ Within the Philippines)
 - c. Bid Securing Declaration; and
 - d. Omnibus Sworn Statement

