

REPUBLIC OF THE PHILIPPINES QUEZON CITY GOVERNMENT BIDS AND AWARDS COMMITTEE – GOODS AND SERVICES



PHILIPPINE BIDDING DOCUMENTS

(As Harmonized with Development Partners)

SUPPLY, INSTALLATION, TESTING AND COMMISSIONING OF EMERGENCY OUTDOOR ALERTING SYSTEM FOR DISASTER RISK REDUCTION AND MANAGEMENT

PROJECT NO. QCDRRMO-23-IT-1078

Government of the Republic of the Philippines

Sixth Edition July 2020

Preface

These Philippine Bidding Documents (PBDs) for the procurement of Goods through Competitive Bidding have been prepared by the Government of the Philippines for use by any branch, constitutional commission or office, agency, department, bureau, office, or instrumentality of the Government of the Philippines, National Government Agencies, including Government-Owned and/or Controlled Corporations, Government Financing Institutions, State Universities and Colleges, and Local Government Unit. The procedures and practices presented in this document have been developed through broad experience, and are for mandatory use in projects that are financed in whole or in part by the Government of the Philippines or any foreign government/foreign or international financing institution in accordance with the provisions of the 2016 revised Implementing Rules and Regulations of Republic Act No. 9184.

The Bidding Documents shall clearly and adequately define, among others: (i) the objectives, scope, and expected outputs and/or results of the proposed contract or Framework Agreement, as the case may be; (ii) the eligibility requirements of Bidders; (iii) the expected contract or Framework Agreement duration, the estimated quantity in the case of procurement of goods, delivery schedule and/or time frame; and (iv) the obligations, duties, and/or functions of the winning bidder.

Care should be taken to check the relevance of the provisions of the PBDs against the requirements of the specific Goods to be procured. If duplication of a subject is inevitable in other sections of the document prepared by the Procuring Entity, care must be exercised to avoid contradictions between clauses dealing with the same matter.

Moreover, each section is prepared with notes intended only as information for the Procuring Entity or the person drafting the Bidding Documents. They shall not be included in the final documents. The following general directions should be observed when using the documents:

- a. All the documents listed in the Table of Contents are normally required for the procurement of Goods. However, they should be adapted as necessary to the circumstances of the particular Procurement Project.
- b. Specific details, such as the "*name of the Procuring Entity*" and "*address for bid submission*," should be furnished in the Instructions to Bidders, Bid Data Sheet, and Special Conditions of Contract. The final documents should contain neither blank spaces nor options.
- c. This Preface and the footnotes or notes in italics included in the Invitation to Bid, Bid Data Sheet, General Conditions of Contract, Special Conditions of Contract, Schedule of Requirements, and Specifications are not part of the text of the final document, although they contain instructions that the Procuring Entity should strictly follow.
- d. The cover should be modified as required to identify the Bidding Documents as to the Procurement Project, Project Identification Number, and Procuring Entity, in addition to the date of issue.

- e. Modifications for specific Procurement Project details should be provided in the Special Conditions of Contract as amendments to the Conditions of Contract. For easy completion, whenever reference has to be made to specific clauses in the Bid Data Sheet or Special Conditions of Contract, these terms shall be printed in bold typeface on Sections I (Instructions to Bidders) and III (General Conditions of Contract), respectively.
- f. For guidelines on the use of Bidding Forms and the procurement of Foreign-Assisted Projects, these will be covered by a separate issuance of the Government Procurement Policy Board.

Table of Contents

Glossa	ry of Acronyms, Terms, and Abbreviations	4
Section	n I. Invitation to Bid	7
Section	n II. Instructions to Bidders	.11
1.	Scope of Bid	.12
2.	Funding Information	.12
3.	Bidding Requirements	12
4.	Corrupt, Fraudulent, Collusive, and Coercive Practices	12
5.	Eligible Bidders	12
6.	Origin of Goods	13
7.	Subcontracts	13
8.	Pre-Bid Conference	13
9.	Clarification and Amendment of Bidding Documents	13
10.	Documents comprising the Bid: Eligibility and Technical Components	• • • • •
11.	Documents comprising the Bid: Financial Component	. 14
12.	Bid Prices	.14
13.	Bid and Payment Currencies	
14.	Bid Security	15
15.	Sealing and Marking of Bids	
16.	Deadline for Submission of Bids	.15
17.	Opening and Preliminary Examination of Bids	
18.	Domestic Preference	
19.	Detailed Evaluation and Comparison of Bids	
20.	Post-Qualification	
21.	Signing of the Contract	
Section	n III. Bid Data Sheet	.18
Section	n IV. General Conditions of Contract	20
1.	Scope of Contract	21
2.	Advance Payment and Terms of Payment	21
3.	Performance Security	21
4.	Inspection and Tests	21
5.	Warranty	21
6.	Liability of the Supplier	22
Section	n V. Special Conditions of Contract	.23
Section	n VI. Schedule of Requirements	.27
	n VII. Technical Specifications	
	n VIII. Checklist of Technical and Financial Documents	

Glossary of Acronyms, Terms, and Abbreviations

ABC – Approved Budget for the Contract.

BAC – Bids and Awards Committee.

Bid – A signed offer or proposal to undertake a contract submitted by a bidder in response to and in consonance with the requirements of the bidding documents. Also referred to as *Proposal* and *Tender*. (2016 revised IRR, Section 5[c])

Bidder – Refers to a contractor, manufacturer, supplier, distributor and/or consultant who submits a bid in response to the requirements of the Bidding Documents. (2016 revised IRR, Section 5[d])

Bidding Documents – The documents issued by the Procuring Entity as the bases for bids, furnishing all information necessary for a prospective bidder to prepare a bid for the Goods, Infrastructure Projects, and/or Consulting Services required by the Procuring Entity. (2016 revised IRR, Section 5[e])

BIR – Bureau of Internal Revenue.

BSP – Bangko Sentral ng Pilipinas.

Consulting Services – Refer to services for Infrastructure Projects and other types of projects or activities of the GOP requiring adequate external technical and professional expertise that are beyond the capability and/or capacity of the GOP to undertake such as, but not limited to: (i) advisory and review services; (ii) pre-investment or feasibility studies; (iii) design; (iv) construction supervision; (v) management and related services; and (vi) other technical services or special studies. (2016 revised IRR, Section 5[i])

CDA - Cooperative Development Authority.

Contract – Refers to the agreement entered into between the Procuring Entity and the Supplier or Manufacturer or Distributor or Service Provider for procurement of Goods and Services; Contractor for Procurement of Infrastructure Projects; or Consultant or Consulting Firm for Procurement of Consulting Services; as the case may be, as recorded in the Contract Form signed by the parties, including all attachments and appendices thereto and all documents incorporated by reference therein.

CIF – Cost Insurance and Freight.

CIP - Carriage and Insurance Paid.

CPI – Consumer Price Index.

DDP – Refers to the quoted price of the Goods, which means "delivered duty paid."

DTI – Department of Trade and Industry.

EXW – Ex works.

FCA – "Free Carrier" shipping point.

FOB – "Free on Board" shipping point.

Foreign-funded Procurement or Foreign-Assisted Project– Refers to procurement whose funding source is from a foreign government, foreign or international financing institution as specified in the Treaty or International or Executive Agreement. (2016 revised IRR, Section 5[b]).

Framework Agreement – Refers to a written agreement between a procuring entity and a supplier or service provider that identifies the terms and conditions, under which specific purchases, otherwise known as "Call-Offs," are made for the duration of the agreement. It is in the nature of an option contract between the procuring entity and the bidder(s) granting the procuring entity the option to either place an order for any of the goods or services identified in the Framework Agreement List or not buy at all, within a minimum period of one (1) year to a maximum period of three (3) years. (GPPB Resolution No. 27-2019)

GFI – Government Financial Institution.

GOCC – Government-owned and/or –controlled corporation.

Goods – Refer to all items, supplies, materials and general support services, except Consulting Services and Infrastructure Projects, which may be needed in the transaction of public businesses or in the pursuit of any government undertaking, project or activity, whether in the nature of equipment, furniture, stationery, materials for construction, or personal property of any kind, including non-personal or contractual services such as the repair and maintenance of equipment and furniture, as well as trucking, hauling, janitorial, security, and related or analogous services, as well as procurement of materials and supplies provided by the Procuring Entity for such services. The term "related" or "analogous services" shall include, but is not limited to, lease or purchase of office space, media advertisements, health maintenance services, and other services essential to the operation of the Procuring Entity. (2016 revised IRR, Section 5[r])

GOP – Government of the Philippines.

GPPB – Government Procurement Policy Board.

INCOTERMS – International Commercial Terms.

Infrastructure Projects – Include the construction, improvement, rehabilitation, demolition, repair, restoration or maintenance of roads and bridges, railways, airports, seaports, communication facilities, civil works components of information technology projects, irrigation, flood control and drainage, water supply, sanitation, sewerage and solid waste management systems, shore protection, energy/power and electrification facilities, national buildings, school buildings, hospital buildings, and other related construction projects of the government. Also referred to as *civil works or works*. (2016 revised IRR, Section 5[u])

LGUs – Local Government Units.

NFCC – Net Financial Contracting Capacity.

NGA – National Government Agency.

PhilGEPS - Philippine Government Electronic Procurement System.

Procurement Project – refers to a specific or identified procurement covering goods, infrastructure project or consulting services. A Procurement Project shall be described, detailed, and scheduled in the Project Procurement Management Plan prepared by the agency which shall be consolidated in the procuring entity's Annual Procurement Plan. (GPPB Circular No. 06-2019 dated 17 July 2019)

PSA – Philippine Statistics Authority.

SEC – Securities and Exchange Commission.

SLCC – Single Largest Completed Contract.

Supplier – refers to a citizen, or any corporate body or commercial company duly organized and registered under the laws where it is established, habitually established in business and engaged in the manufacture or sale of the merchandise or performance of the general services covered by his bid. (Item 3.8 of GPPB Resolution No. 13-2019, dated 23 May 2019). Supplier as used in these Bidding Documents may likewise refer to a distributor, manufacturer, contractor, or consultant.

UN – United Nations.

Notes on the Invitation to Bid

The Invitation to Bid (IB) provides information that enables potential Bidders to decide whether to participate in the procurement at hand. The IB shall be posted in accordance with Section 21.2 of the 2016 revised IRR of RA No. 9184.

Apart from the essential items listed in the Bidding Documents, the IB should also indicate the following:

- a. The date of availability of the Bidding Documents, which shall be from the time the IB is first advertised/posted until the deadline for the submission and receipt of bids;
- b. The place where the Bidding Documents may be acquired or the website where it may be downloaded;
- c. The deadline for the submission and receipt of bids; and
- d. Any important bid evaluation criteria (*e.g.*, the application of a margin of preference in bid evaluation).

The IB should be incorporated in the Bidding Documents. The information contained in the IB must conform to the Bidding Documents and in particular to the relevant information in the Bid Data Sheet.



QUEZON CITY GOVERNMENT BAC – GOODS AND SERVICES INVITATION TO BID



July 10, 2023 PROJECT NO. OFFICE PROJECT NAME AMOUNT SOURCE OF DELIVERY FUND PERIOD CITY GENERAL SERVICES GENERAL 1. CGSD-23-ES-0433 CABLE TIE, BATTERY AND OTHERS P 2,089,635.60 30 CD DEPARTMENT FUND CITY GENERAL SERVICES GENERAL 2. CGSD-23-PS2-0420 PRINTING OF STICKER AND OTHERS P 1.133.279.50 30 CD DEPARTMENT FUND OFFICE FOR THE SENIOR GENERAL 3. CONSO-23-CS1-1127 FOOD AND DRINKS AND OTHERS P 4,046,500.00 4 MONTHS CITIZENS AFFAIRS FUND FIRE EXTINGUISHER AND CONSO-23-FFRSE-CITY GENERAL SERVICES REFILLING/ PREVENTIVE GENERAL 4. P 1,042,500.00 30 CD 1154 DEPARTMENT MAINTENANCE OF FIRE FUND EXTINGUISHER (HCFC) INSULATION MULTI-METER AND ENGINEERING-23-DEPARTMENT OF GENERAL 5. P 1.643.393.34 30 CD CE2-0844 ENGINEERING OTHERS FUND HOUSING, COMMUNITY DEVELOPMENT AND SURVEY WORKS AND TITLING GENERAL HCDRD-23-SS2-0888 6. P 7,257,500.00 4 MONTHS RESETTLEMENT DOCUMENTATION FUND DEPARTMENT HEALTH-23-AAS2-OUEZON CITY HEALTH SUPPLY AND INSTALLATION OF GENERAL 7. P 1.207.201.50 30 CD 0976 DEPARTMENT AIRCONDITIONING UNITS FUND HEALTH-23-CCP-QUEZON CITY HEALTH GENERAL 8. DISINFECTANT AND OTHERS P 1.540,200.00 30 CD DEPARTMENT 0748 FUND HEALTH-23-DM-OUEZON CITY HEALTH LINE 7: VARIOUS DRUGS AND GENERAL 9. P 11,297,541.15 4 MONTHS 0183B DEPARTMENT MEDICINES FOR QC EMPLOYEES FUND HEALTH-23-PS2-QUEZON CITY HEALTH PRINTING OF PHILHEALTH 10. P1.504.100.00 TRUST FUND 30 CD 0974 DEPARTMENT SIGNAGES OCM(POPS)-23-OFFICE OF THE CITY HOTEL ACCOMMODATION AND **GENERAL** 11. P 5,297,100.00 4 MONTHS HLMF-1027 MAYOR - POPS PLAN OTHERS FUND OFFICE OF THE VICE GENERAL 12. OVM-23-EMS-0299 BACKPACK WITH SCHOOL SUPPLIES P 29,998,800.00 30 CD MAYOR FUND PROCUREMENT 13. PD-23-OE-1089 MULTI FUNCTION PRINTER P 1,953,000.00 TRUST FUND 60 CD DEPARTMENT SUPPLY, INSTALLATION, TESTING OUEZON CITY DISASTER AND COMMISSIONING OF QCDRRMO-23-IT-GENERAL 14. RISK REDUCTION AND EMERGENCY OUTDOOR ALERTING P 32,000,000.00 60 CD 1078 FUND MANAGEMENT OFFICE SYSTEM FOR DISASTER RISK REDUCTION AND MANAGEMENT QCPL-23-BMOP-PROCUREMENT OF LIBRARY BOOKS QUEZON CITY PUBLIC GENERAL 15. 0700B (A HANDBOOK OF MANAGEMENT 30 CD LIBRARY P 1,703,205.30 FUND TECHNIQUES AND OTHERS) QUEZON CITY SMALL BUSINESS AND QCSBCDPO-23-CS1-GENERAL 16. COOPERATIVES FOOD AND DRINKS P 1.500.000.00 4 MONTHS 1046 FUND DEVELOPMENT AND PROMOTIONS OFFICE SPECIAL SCHOOLS DIVISION VARIOUS READING BOOKS FOR SDO-23-BMOP-0951 17. P 14,699,200.00 EDUCATION 30 CD OFFICE GRADE 1 TO 3 FUND

IB FOR JULY 31, 2023 - ABOVE

Page 1 of 3



18.	SDO-23-BMOP-0952	SCHOOLS DIVISION OFFICE	VARIOUS BOOKS (ECCD LEARNING PACKAGE)	P 5,529,304.00	SPECIAL EDUCATION FUND	30 CD
19.	SDO-23-CS1-0815B	SCHOOLS DIVISION OFFICE	FOOD AND DRINKS	P 4,109,900.00	GENERAL FUND	4 MONTHS
20.	SDO-23-HME-0968	SCHOOLS DIVISION OFFICE	HEIGHT AND WEIGHING SCALE	P 1,132,000.00	SPECIAL EDUCATION FUND	30 CD
21.	SDO-23-PP-0955	SCHOOLS DIVISION OFFICE	TABLET ARMCHAIRS AND KIDDIE TABLE AND CHAIR SET	P 96,350,000.00	SPECIAL EDUCATION FUND	45 CD
22.	SSDD-23-GM-0911	SOCIAL SERVICES DEVELOPMENT DEPARTMENT	TOOTHPASTE AND OTHERS	P 1,411,266.80	GENERAL FUND	15 CD
23.	SSDD-23-PS2-0928	SOCIAL SERVICES DEVELOPMENT DEPARTMENT	PRINTING OF ECCD CHECKLIST AND MODULE GUIDE	P 2,034,330.00	GENERAL FUND	4 MONTHS

- 1. The QUEZON CITY LOCAL GOVERNMENT, through the General Fund, Special Education Fund and Trust Fund of various years intends to apply the sums stated above being the ABC to payments under the contract for the above stated projects of contract for each lot/item. Bids received in excess of the ABC shall be automatically rejected at bid opening.
- 2. The QUEZON CITY LOCAL GOVERNMENT now invites bids for various Projects. Delivery of the Goods is required as stated above. Bidders should have completed, within the last three (3) years from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II. Instructions to Bidders.
- Bidding will be conducted through open competitive bidding procedures using a non-discretionary "pass/fail" criterion as specified in the 2016 revised Implementing Rules and Regulations (IRR) of Republic Act (RA) No. 9184.
 - a. Bidding is restricted to Filipino citizens/sole proprietorships, partnerships, or organizations with at least sixty percent (60%) interest or outstanding capital stock belonging to citizens of the Philippines, and to citizens or organizations of a country the laws or regulations of which grant similar rights or privileges to Filipino citizens, pursuant to RA No. 5183.
- 4. Prospective Bidders may obtain further information from *QUEZON CITY GOVERNMENT Bids and Awards Committee (BAC) Secretariat* and inspect the Bidding Documents at the address given below during weekdays from 8:00 a.m. - 5:00 p.m.
- 5. A complete set of Bidding Documents may be acquired by interested Bidders on <u>Tuesday, July 11, 2023</u> from the given address and website(s) below *and upon payment of the applicable fee for the Bidding Documents, pursuant to the latest Guidelines issued by the GPPB.* The Procuring Entity shall allow the bidder to present its proof of payment for the fees *in person.*

Approved Budget for the Contract	Maximum Cost of Bidding Documents (in Philippine Peso)
500,000 and below	500.00
More than 500,000 up to 1 Million	1,000.00
More than 1 Million up to 5 Million	5,000.00
More than 5 Million up to 10 Million	10,000.00
More than 10 Million up to 50 Million	25,000.00
More than 50 Million up to 500 Million	50,000.00
More than 500 Million	75,000.00

Page 2 of 3

x."

STANDARD RATES:

The following are the requirements for purchase of Bidding Documents;

- 1. PhilGEPS Registration Certificate (Platinum 3 pages)
- 2. Document Request List (DRL)
- 3. Authorization to Purchase Bidding Documents
 - 3.1 Corporate Secretary Certificate for corporation (specific for the project)
 - 3.2 Special Power of Attorney for single proprietorship (specific for the project)
- 4. Notarized Joint Venture Agreement (as applicable)

IB FOR JULY 31, 2023 - ABOVE

 The Quezon City Local Government will hold a Pre-Bid Conference on 10:30 A.M. of <u>Tuesday, July</u> <u>18, 2023</u> at 2nd Floor, Procurement Department-Bidding Room, Finance Building, Quezon City Hall Compound, and/or through video conferencing via Zoom which shall be open to prospective bidders.

Topic: BAC-GOODS Pre-Bid Conference Meeting Join Zoom Meeting https://us02web.zoom.us/j/84835002246?pwd=OVRuVE0weXZMNXYwZG5LaWd1dXk1QT09

Meeting ID: 848 3500 2246 Passcode: 154733

- Bids must be duly received by the BAC Secretariat through manual submission at the 2nd Floor, Procurement Department, Finance Building, Quezon City Hall Compound on or before <u>10:00 A.M.</u> of <u>Monday, July 31, 2023</u>. Late bids shall not be accepted.
- 8. All Bids must be accompanied by a bid security in any of the acceptable forms and in the amount stated in **ITB** Clause 14.
- 9. Bid opening shall be on <u>11:00 A.M. of Monday, July 31, 2023</u> at the given address below and/or via Zoom. Bids will be opened in the presence of the bidders' representatives who choose to attend the activity.

Topic: BAC-GOODS & SERVICES BIDDING Join Zoom Meeting https://us02web.zoom.us/j/85850855933?pwd=R2dZUUp4Z3lyU29iZGV1WmdKRjZCdz09

Meeting ID: 858 5085 5933 Passcode: 118682

- 10. The *Quezon City Local Government* reserves the right to reject any and all bids, declare a failure of bidding, or not award the contract at any time prior to contract award in accordance with Sections 35.6 and 41 of the 2016 revised IRR of RA No. 9184, without thereby incurring any liability to the affected bidder or bidders.
- 11. For further information, please refer to:

ATTY. DOMINIC B. GARCIA OIC, Procurement Department 2nd Floor, Procurement Department, Finance Building, Quezon City Hall Compound Elliptical Road, Barangay Central Diliman, Quezon City. Email Add: <u>bacgoods.procurement@quezoncity.gov.ph</u> Tel. No. (02)8988-4242 loc. 8506/8710 Website: <u>www.quezoncity.gov.ph</u>

12. You may visit the following websites:

For downloading of Bidding Documents: www.quezoncity.gov.ph

By: MS. MA. MAKCARDA T. SANTOS, PhD Chairperson, QC-BAC-Goods and Services

Notes on the Instructions to Bidders

This Section on the Instruction to Bidders (ITB) provides the information necessary for bidders to prepare responsive bids, in accordance with the requirements of the Procuring Entity. It also provides information on bid submission, eligibility check, opening and evaluation of bids, post-qualification, and on the award of contract.

1. Scope of Bid

The Procuring Entity, Quezon City Local Government *wishes* to receive Bids for the SUPPLY, INSTALLATION, TESTING AND COMMISSIONING OF EMERGENCY OUTDOOR ALERTING SYSTEM FOR DISASTER RISK REDUCTION AND MANAGEMENT with identification number QCDRRMO-23-IT-1078.

[Note: The Project Identification Number is assigned by the Procuring Entity based on its own coding scheme and is not the same as the PhilGEPS reference number, which is generated after the posting of the bid opportunity on the PhilGEPS website.]

The Procurement Project (referred to herein as "Project") is composed of *ONE* (1) *item*, the details of which are described in Section VII (Technical Specifications).

2. Funding Information

- 2.1. The GOP through the source of funding as indicated below for 2023 in the amount of THIRTY TWO MILLION PESOS AND 00/100 ONLY (Php32,000,000.00).
- 2.2. The source of funding is:
 - *a.* LGUs, the Annual or Supplemental Budget, as approved by the Sanggunian.

3. Bidding Requirements

The Bidding for the Project shall be governed by all the provisions of RA No. 9184 and its 2016 revised IRR, including its Generic Procurement Manuals and associated policies, rules and regulations as the primary source thereof, while the herein clauses shall serve as the secondary source thereof.

Any amendments made to the IRR and other GPPB issuances shall be applicable only to the ongoing posting, advertisement, or **IB** by the BAC through the issuance of a supplemental or bid bulletin.

The Bidder, by the act of submitting its Bid, shall be deemed to have verified and accepted the general requirements of this Project, including other factors that may affect the cost, duration and execution or implementation of the contract, project, or work and examine all instructions, forms, terms, and project requirements in the Bidding Documents.

4. Corrupt, Fraudulent, Collusive, and Coercive Practices

The Procuring Entity, as well as the Bidders and Suppliers, shall observe the highest standard of ethics during the procurement and execution of the contract. They or through an agent shall not engage in corrupt, fraudulent, collusive, coercive, and obstructive practices defined under Annex "I" of the 2016 revised IRR of RA No. 9184 or other integrity violations in competing for the Project.

5. Eligible Bidders

- 5.1. Only Bids of Bidders found to be legally, technically, and financially capable will be evaluated.
- 5.2. Foreign ownership exceeding those allowed under the rules may participate pursuant to:

- i. When a Treaty or International or Executive Agreement as provided in Section 4 of the RA No. 9184 and its 2016 revised IRR allow foreign bidders to participate;
- ii. Citizens, corporations, or associations of a country, included in the list issued by the GPPB, the laws or regulations of which grant reciprocal rights or privileges to citizens, corporations, or associations of the Philippines;
- iii. When the Goods sought to be procured are not available from local suppliers; or
- iv. When there is a need to prevent situations that defeat competition or restrain trade.
- 5.3. Pursuant to Section 23.4.1.3 of the 2016 revised IRR of RA No.9184, the Bidder shall have an SLCC that is at least one (1) contract similar to the Project the value of which, adjusted to current prices using the PSA's CPI, must be at least equivalent to:
 - a. For the procurement of **Non-Expendable Supplies and Services**: The Bidder must have completed a single contract that is similar to this Project, equivalent to at least **fifty percent (50%)** of the ABC.
- 5.4. The Bidders shall comply with the eligibility criteria under Section 23.4.1 of the 2016 IRR of RA No. 9184.

6. Origin of Goods

There is no restriction on the origin of goods other than those prohibited by a decision of the UN Security Council taken under Chapter VII of the Charter of the UN, subject to Domestic Preference requirements under **ITB** Clause 18.

7. Subcontracts

7.1. The Bidder may subcontract portions of the Project to the extent allowed by the Procuring Entity as stated herein, but in no case more than twenty percent (20%) of the Project.

The Procuring Entity has prescribed that: Subcontracting is not allowed.

8. Pre-Bid Conference

The Procuring Entity will hold a pre-bid conference for this Project on the specified date and time and either at its physical address and/or through videoconferencing as indicated in paragraph 6 of the **IB**.

9. Clarification and Amendment of Bidding Documents

Prospective bidders may request for clarification on and/or interpretation of any part of the Bidding Documents. Such requests must be in writing and received by the Procuring Entity, either at its given address or through electronic mail indicated in the **IB**, at least ten (10) calendar days before the deadline set for the submission and receipt of Bids.

10. Documents comprising the Bid: Eligibility and Technical Components

- 10.1. The first envelope shall contain the eligibility and technical documents of the Bid as specified in Section VIII (Checklist of Technical and Financial Documents).
- 10.2. The Bidder's SLCC as indicated in **ITB** Clause 5.3 should have been completed within *the last three (3) years* prior to the deadline for the submission and receipt of bids.
- 10.3. If the eligibility requirements or statements, the bids, and all other documents for submission to the BAC are in foreign language other than English, it must be accompanied by a translation in English, which shall be authenticated by the appropriate Philippine foreign service establishment, post, or the equivalent office having jurisdiction over the foreign bidder's affairs in the Philippines. Similar to the required authentication above, for Contracting Parties to the Apostille Convention, only the translated documents shall be authenticated through an Apostille pursuant to GPPB Resolution No. 13-2019 dated 23 May 2019. The English translation shall govern, for purposes of interpretation of the bid.

11. Documents comprising the Bid: Financial Component

- 11.1. The second bid envelope shall contain the financial documents for the Bid as specified in **Section VIII (Checklist of Technical and Financial Documents)**.
- 11.2. If the Bidder claims preference as a Domestic Bidder or Domestic Entity, a certification issued by DTI shall be provided by the Bidder in accordance with Section 43.1.3 of the 2016 revised IRR of RA No. 9184.
- 11.3. Any bid exceeding the ABC indicated in paragraph 1 of the **IB** shall not be accepted.
- 11.4. For Foreign-funded Procurement, a ceiling may be applied to bid prices provided the conditions are met under Section 31.2 of the 2016 revised IRR of RA No. 9184.

12. Bid Prices

- 12.1. Prices indicated on the Price Schedule shall be entered separately in the following manner:
 - a. For Goods offered from within the Procuring Entity's country:
 - i. The price of the Goods quoted EXW (ex-works, ex-factory, exwarehouse, ex-showroom, or off-the-shelf, as applicable);
 - ii. The cost of all customs duties and sales and other taxes already paid or payable;
 - iii. The cost of transportation, insurance, and other costs incidental to delivery of the Goods to their final destination; and
 - iv. The price of other (incidental) services, if any, listed in e.
 - b. For Goods offered from abroad:
 - i. Unless otherwise stated in the **BDS**, the price of the Goods shall be quoted delivered duty paid (DDP) with the place of destination

in the Philippines as specified in the **BDS**. In quoting the price, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible source country.

ii. The price of other (incidental) services, if any, as listed in **Section VII (Technical Specifications).**

13. Bid and Payment Currencies

- 13.1. For Goods that the Bidder will supply from outside the Philippines, the bid prices may be quoted in the local currency or tradeable currency accepted by the BSP at the discretion of the Bidder. However, for purposes of bid evaluation, Bids denominated in foreign currencies, shall be converted to Philippine currency based on the exchange rate as published in the BSP reference rate bulletin on the day of the bid opening.
- 13.2. Payment of the contract price shall be made in:
 - a. Philippine Pesos.

14. Bid Security

- 14.1. The Bidder shall submit a Bid Securing Declaration¹ or any form of Bid Security in the amount indicated in the **BDS**, which shall be not less than the percentage of the ABC in accordance with the schedule in the **BDS**.
- 14.2. The Bid and bid security *in no case shall exceed One Hundred Twenty (120) calendar days from the date of opening of bids, unless duly extended by the bidder upon the request of the Head of the Procuring Entity (HoPE) of the Quezon City Local Government.* Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.

15. Sealing and Marking of Bids

Each Bidder shall submit one copy of the first and second components of its Bid.

The Procuring Entity may request additional hard copies and/or electronic copies of the Bid. However, failure of the Bidders to comply with the said request shall not be a ground for disqualification.

If the Procuring Entity allows the submission of bids through online submission or any other electronic means, the Bidder shall submit an electronic copy of its Bid, which must be digitally signed. An electronic copy that cannot be opened or is corrupted shall be considered non-responsive and, thus, automatically disqualified.

16. Deadline for Submission of Bids

16.1. The Bidders shall submit on the specified date and time through manual submission as indicated in paragraph 7 of the **IB**.

17. Opening and Preliminary Examination of Bids

¹ In the case of Framework Agreement, the undertaking shall refer to entering into contract with the Procuring Entity and furnishing of the performance security or the performance securing declaration within ten (10) calendar days from receipt of Notice to Execute Framework Agreement.

17.1. The BAC shall open the Bids in public at the time, on the date, and at the place specified in paragraph 9 of the **IB**. The Bidders' representatives who are present shall sign a register evidencing their attendance. In case videoconferencing, webcasting or other similar technologies will be used, attendance of participants shall likewise be recorded by the BAC Secretariat.

In case the Bids cannot be opened as scheduled due to justifiable reasons, the rescheduling requirements under Section 29 of the 2016 revised IRR of RA No. 9184 shall prevail.

17.2. The preliminary examination of bids shall be governed by Section 30 of the 2016 revised IRR of RA No. 9184.

18. Domestic Preference

18.1. The Procuring Entity will grant a margin of preference for the purpose of comparison of Bids in accordance with Section 43.1.2 of the 2016 revised IRR of RA No. 9184.

19. Detailed Evaluation and Comparison of Bids

- 19.1. The Procuring BAC shall immediately conduct a detailed evaluation of all Bids rated "*passed*," using non-discretionary pass/fail criteria. The BAC shall consider the conditions in the evaluation of Bids under Section 32.2 of the 2016 revised IRR of RA No. 9184.
- 19.2. If the Project allows partial bids, bidders may submit a proposal on any of the lots or items, and evaluation will be undertaken on a per lot or item basis, as the case maybe. In this case, the Bid Security as required by **ITB** Clause 15 shall be submitted for each lot or item separately.
- 19.3. The descriptions of the lots or items shall be indicated in Section VII (Technical Specifications), although the ABCs of these lots or items are indicated in the BDS for purposes of the NFCC computation pursuant to Section 23.4.2.6 of the 2016 revised IRR of RA No. 9184. The NFCC must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder.
- 19.4. The Project shall be awarded as follows:

One Project having several items that shall be awarded as one contract.

19.5. Except for bidders submitting a committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation, all Bids must include the NFCC computation pursuant to Section 23.4.1.4 of the 2016 revised IRR of RA No. 9184, which must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder. For bidders submitting the committed Line of Credit, it must be at least equal to ten percent (10%) of the ABCs for all the lots or items participated in by the prospective Bidder.

20. Post-Qualification

20.1. Within a non-extendible period of five (5) calendar days from receipt by the Bidder of the notice from the BAC that it submitted the Lowest Calculated Bid, the Bidder shall submit its latest income and business tax returns filed and paid through the BIR Electronic Filing and Payment System (eFPS) and other appropriate licenses and permits required by law and stated in the **BDS**.

21. Signing of the Contract

21.1. The documents required in Section 37.2 of the 2016 revised IRR of RA No. 9184 shall form part of the Contract. Additional Contract documents are indicated in the **BDS**.

Notes on the Bid Data Sheet

The Bid Data Sheet (BDS) consists of provisions that supplement, amend, or specify in detail, information, or requirements included in the ITB found in Section II, which are specific to each procurement.

This Section is intended to assist the Procuring Entity in providing the specific information in relation to corresponding clauses in the ITB and has to be prepared for each specific procurement.

The Procuring Entity should specify in the BDS information and requirements specific to the circumstances of the Procuring Entity, the processing of the procurement, and the bid evaluation criteria that will apply to the Bids. In preparing the BDS, the following aspects should be checked:

- a. Information that specifies and complements provisions of the ITB must be incorporated.
- b. Amendments and/or supplements, if any, to provisions of the ITB as necessitated by the circumstances of the specific procurement, must also be incorporated.

Bid Data	Sheet
-----------------	-------

	Diu Data Sileet
ITB	
Clause	
5.3	For this purpose, contracts similar to the Project shall be:
	a. A single contract similar to the items to be bid and must be at least fifty percent (50%) of the ABC.
	b. Completed within the last three (3) years prior to the deadline for the submission and receipt of bids substantially in a FORM prescribed by the QC-BAC-GOODS AND SERVICES, must be accompanied by a copy of Certificate of Acceptance by the end-user or Official Receipt (O.R) or Sales Invoice (S.I.) issued for the Contract.
7.1	Subcontracting is not allowed.
12	The price of the Goods shall be quoted DDP <i>within Quezon City</i> or the applicable International Commercial Terms (INCOTERMS) for this Project.
14.1	The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:
	 a. The amount of not less than <i>Php640,000.00</i> or equivalent to two percent (2%) of ABC if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or
	 b. The amount of not less than <i>Php1,600,000.00</i> or equivalent to five percent (5%) of ABC if bid security is in Surety Bond.
19.3	[In case the Project will be awarded by lot, list the grouping of lots by specifying the group title, items, and the quantity for every identified lot, and the corresponding ABC for each lot.]
	[In case the project will be awarded by item, list each item indicating its quantity and ABC.]
20.2	List of required licenses and permits relevant to the Project and the corresponding law requiring it.
	No additional requirements
21.2	Additional required documents relevant to the Project that are required by existing laws and/or the Procuring Entity.
	 Organizational Chart for the project Notarized Affidavit of Undertakings stating compliance to the following: a. The bidder must have demonstrated capability, experience and have implemented and managed citywide sensor equipment networks and meteorological information alerting to highly urbanized city (HUC) in the Philippines. These systems must be already existing and proven working by other highly urbanized cities (HUC) in the Philippines. b. The bidder will conduct the required training and documentation.
	 Statement of Warranty For Warning Station Equipment Set - Three (3) Years Maintenance, support, warranty, and data plan for 7 Warning Station Equipment including replacement of Battery For Control Station & Command Center Software - Three (3) Years software license, maintenance, support warranty, and technical support for All equipment and software
	(Please refer to the attached Terms of Reference)

Section IV. General Conditions of Contract

Notes on the General Conditions of Contract

The General Conditions of Contract (GCC) in this Section, read in conjunction with the Special Conditions of Contract in Section V and other documents listed therein, should be a complete document expressing all the rights and obligations of the parties.

Matters governing performance of the Supplier, payments under the contract, or matters affecting the risks, rights, and obligations of the parties under the contract are included in the GCC and Special Conditions of Contract.

Any complementary information, which may be needed, shall be introduced only through the Special Conditions of Contract.

1. Scope of Contract

This Contract shall include all such items, although not specifically mentioned, that can be reasonably inferred as being required for its completion as if such items were expressly mentioned herein. All the provisions of RA No. 9184 and its 2016 revised IRR, including the Generic Procurement Manual, and associated issuances, constitute the primary source for the terms and conditions of the Contract, and thus, applicable in contract implementation. Herein clauses shall serve as the secondary source for the terms and conditions of the Contract.

This is without prejudice to Sections 74.1 and 74.2 of the 2016 revised IRR of RA No. 9184 allowing the GPPB to amend the IRR, which shall be applied to all procurement activities, the advertisement, posting, or invitation of which were issued after the effectivity of the said amendment.

Additional requirements for the completion of this Contract shall be provided in the **Special Conditions of Contract (SCC).**

2. Advance Payment and Terms of Payment

- 2.1. Advance payment of the contract amount is provided under Annex "D" of the revised 2016 IRR of RA No. 9184.
- 2.2. The Procuring Entity is allowed to determine the terms of payment on the partial or staggered delivery of the Goods procured, provided such partial payment shall correspond to the value of the goods delivered and accepted in accordance with prevailing accounting and auditing rules and regulations. The terms of payment are indicated in the **SCC**.

3. Performance Security

Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than prior to the signing of the Contract by both parties, the successful Bidder shall furnish the performance security in any of the forms prescribed in Section 39 of the 2016 revised IRR of RA No. 9184.

4. Inspection and Tests

The Procuring Entity or its representative shall have the right to inspect and/or to test the Goods to confirm their conformity to the Project specifications at no extra cost to the Procuring Entity in accordance with the Generic Procurement Manual. In addition to tests in the **SCC**, **Section IV** (**Technical Specifications**) shall specify what inspections and/or tests the Procuring Entity requires, and where they are to be conducted. The Procuring Entity shall notify the Supplier in writing, in a timely manner, of the identity of any representatives retained for these purposes.

All reasonable facilities and assistance for the inspection and testing of Goods, including access to drawings and production data, shall be provided by the Supplier to the authorized inspectors at no charge to the Procuring Entity.

5. Warranty

6.1. In order to assure that manufacturing defects shall be corrected by the Supplier, a warranty shall be required from the Supplier as provided under Section 62.1 of the 2016 revised IRR of RA No. 9184.

6.2. The Procuring Entity shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall, repair or replace the defective Goods or parts thereof without cost to the Procuring Entity, pursuant to the Generic Procurement Manual.

6. Liability of the Supplier

The Supplier's liability under this Contract shall be as provided by the laws of the Republic of the Philippines.

If the Supplier is a joint venture, all partners to the joint venture shall be jointly and severally liable to the Procuring Entity.

Section V. Special Conditions of Contract

Notes on the Special Conditions of Contract

Similar to the BDS, the clauses in this Section are intended to assist the Procuring Entity in providing contract-specific information in relation to corresponding clauses in the GCC found in Section IV.

The Special Conditions of Contract (SCC) complement the GCC, specifying contractual requirements linked to the special circumstances of the Procuring Entity, the Procuring Entity's country, the sector, and the Goods purchased. In preparing this Section, the following aspects should be checked:

- a. Information that complements provisions of the GCC must be incorporated.
- b. Amendments and/or supplements to provisions of the GCC as necessitated by the circumstances of the specific purchase, must also be incorporated.

However, no special condition which defeats or negates the general intent and purpose of the provisions of the GCC should be incorporated herein.

GCC Clause 1 [List here any additional requirements for the completion of this Contract. The following requirements and the corresponding provisions may be deleted, amended, or retained depending on its applicability to this Contract:] **Delivery and Documents –** For purposes of the Contract, "EXW," "FOB," "FCA," "CIF," "CIP," "DDP" and other trade terms used to describe the obligations of the parties shall have the meanings assigned to them by the current edition of INCOTERMS published by the International Chamber of Commerce, Paris. The Delivery terms of this Contract shall be as follows: [For Goods supplied from abroad, state:] "The delivery terms applicable to the Contract are DDP delivered [indicate place of destination]. In accordance with **INCOTERMS.**" [For Goods supplied from within the Philippines, state:] "The delivery terms applicable to this Contract are delivered [indicate place of destination]. Risk and title will pass from the Supplier to the Procuring Entity upon receipt and final acceptance of the Goods at their final destination." Delivery of the Goods shall be made by the Supplier in accordance with the terms specified in Section VI (Schedule of Requirements). For purposes of this Clause the Procuring Entity's Representative at the Project Site is [indicate name(s)]. Incidental Services -The Supplier is required to provide all of the following services, including additional services, if any, specified in Section VI. Schedule of Requirements: Select appropriate requirements and delete the rest. performance or supervision of on-site assembly and/or start-up of a. the supplied Goods; furnishing of tools required for assembly and/or maintenance of the b. supplied Goods; furnishing of a detailed operations and maintenance manual for each c. appropriate unit of the supplied Goods; performance or supervision or maintenance and/or repair of the d. supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and training of the Procuring Entity's personnel, at the Supplier's plant e. and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied Goods. f. [Specify additional incidental service requirements, as needed.] The Contract price for the Goods shall include the prices charged by the Supplier for incidental services and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.

Special Conditions of Contract

Spare Parts –

The Supplier is required to provide all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the Supplier:

Select appropriate requirements and delete the rest.

- a. such spare parts as the Procuring Entity may elect to purchase from the Supplier, provided that this election shall not relieve the Supplier of any warranty obligations under this Contract; and
- b. in the event of termination of production of the spare parts:
 - i. advance notification to the Procuring Entity of the pending termination, in sufficient time to permit the Procuring Entity to procure needed requirements; and
 - ii. following such termination, furnishing at no cost to the Procuring Entity, the blueprints, drawings, and specifications of the spare parts, if requested.

The spare parts and other components required are listed in **Section VI** (**Schedule of Requirements**) and the cost thereof are included in the contract price.

The Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spare parts or components for the Goods for a period of [*indicate here the time period specified. If not used indicate a time period of three times the warranty period*].

Spare parts or components shall be supplied as promptly as possible, but in any case, within [*insert appropriate time period*] months of placing the order.

Packaging –

The Supplier shall provide such packaging of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in this Contract. The packaging shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit, and open storage. Packaging case size and weights shall take into consideration, where appropriate, the remoteness of the Goods' final destination and the absence of heavy handling facilities at all points in transit.

The packaging, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified below, and in any subsequent instructions ordered by the Procuring Entity.

The outer packaging must be clearly marked on at least four (4) sides as follows:

Name of the Procuring Entity Name of the Supplier

I F	Contract Description
	Contract Description Final Destination
	Gross weight
	Any special lifting instructions
	Any special handling instructions
	Any relevant HAZCHEM classifications
	Any recvant HAZeriew classifications
	A packaging list identifying the contents and quantities of the package is to be placed on an accessible point of the outer packaging if practical. If not practical the packaging list is to be placed inside the outer packaging but outside the secondary packaging.
	Transportation –
	Where the Supplier is required under Contract to deliver the Goods CIF, CIP, or DDP, transport of the Goods to the port of destination or such other named place of destination in the Philippines, as shall be specified in this Contract, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract Price.
	Where the Supplier is required under this Contract to transport the Goods to a specified place of destination within the Philippines, defined as the Project Site, transport to such place of destination in the Philippines, including insurance and storage, as shall be specified in this Contract, shall be arranged by the Supplier, and related costs shall be included in the contract price.
	Where the Supplier is required under Contract to deliver the Goods CIF, CIP or DDP, Goods are to be transported on carriers of Philippine registry. In the event that no carrier of Philippine registry is available, Goods may be shipped by a carrier which is not of Philippine registry provided that the Supplier obtains and presents to the Procuring Entity certification to this effect from the nearest Philippine registry are available but their schedule delays the Supplier in its performance of this Contract the period from when the Goods were first ready for shipment and the actual date of shipment the period of delay will be considered force majeure.
	The Procuring Entity accepts no liability for the damage of Goods during transit other than those prescribed by INCOTERMS for DDP deliveries. In the case of Goods supplied from within the Philippines or supplied by domestic Suppliers risk and title will not be deemed to have passed to the Procuring Entity until their receipt and final acceptance at the final destination.
	Intellectual Property Rights –
	The Supplier shall indemnify the Procuring Entity against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the Goods or any part thereof.
2.2	[If partial payment is allowed, state] "The terms of payment shall be as follows:
4	The inspections and tests that will be conducted are: <i>Product Presentation/Demonstration/Site Inspection, if applicable.</i>

Section VI. Schedule of Requirements

PROJECT NAME: SUPPLY, INSTALLATION, TESTING AND COMMISSIONING OF EMERGENCY OUTDOOR ALERTING SYSTEM FOR DISASTER RISK REDUCTION AND MANAGEMENT PROJECT NO. QCDRRMO-23-IT-1078

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

With minimum technical specifications: Image: Constant of the OUTDOOR Iot 1 Requirements and specifications of the OUTDOOR Iot 1 1 ALERTING SYSTEM FOR DISASTER RISK Iot 1 1 SEVEN (7) WARNING STATION EQUIPMENT SET A. Electronic Digital Siren with PA (Public Address) Iom Directional (Per Station), SB-2-41 12dB # 30m Omi Directional with 4 Horn Speaker, 12V/38Ah insight battery weatherproof cabinet with standalone power supply solar panel, solar charger, fully functional up to 1km range radius 2x Amplifier 400 Watts, Siren Pole Cap, IP66 steel enclosure. B. Primary Communication Interface per Siren, • Seven (7) pieces CMC-4 Serial HMI Control Console for CS Siren • Seven (7) pieces CMC-4 Serial HMI Control Console for CS Siren • Seven (7) pieces CMC-4 Serial HMI Control Console for CS Siren • Seven (7) pieces Radio Transceiver UHF 403-470 MHz for CS • Seven (7) pieces Radio Transceiver UHF 403-470 MHz for CS • Seven (7) pieces Antenna found UHF 403-470 MHz • Seven (7) pieces Antenna found UHF 403-470 MHz • Seven (7) pieces Antenna found UHF 403-470 MHz • Seven (7) pieces Se Antenna found UHF 403-470 MHz • Seven (7) pieces Se Antenna found UHF 403-470 MHz • Seven (7) pieces Se Antenna found UHF 403-470 MHz • Seven (7) pieces Pole mounting Junction Box for 4-6 horns 30cm x30cm x30	Item Number	Description	Unit of Issue	Quantity	Delivered, Weeks / Months
ALEXTING SYSTEM FOR DISASTER RISK REDUCTION AND MANAGEMENT 1. SEVEN (7) WARNING STATION EQUIPMENT SET A. Electronic Digital Siren with PA (Public Address) Omni Directional (Per Station), SB-24-1 12dB @ 30m Omni Directional with 4 Horn Speaker, 12V/38 Ah insight battery weatherproof cabinet with standalone power supply solar panel, solar charger, fully functional up to 1 km range radius 2x Amplifier 400 Watts, Siren Pole Cap, IP66 steel enclosure. B. Primary Communication Interface per Siren, • Seven (7) pieces CMC-4 Serial HMI Control Console for CS Siren Sirty (56) Sirty (56) C. Secondary Communication: UHF Radio, • Seven (7) pieces Radio Transceiver UHF 403-470 MHz for CS Sirty (60) Sisty (56) Seven (7) pieces Rick and Transceiver UHF 403-470 MHz for CS Seven (7) pieces FFSK Radio Modem • Seven (7) pieces Antenna Geder LMR/CFD200, feeder to radio N male to BNC framale Days Upon Issuance to Proceed D. Junction Box Set (cables and mounting accessories), • Seven (7) pieces PifskI Int, LMR/CFD200, feeder to radio N male to BNC male D. Junction Box Set (cables and mounting accessories), • Seven (7) pieces Pig tail Int, LMR/CFD200, feeder to radio N male to BNC male E Installation a seven (7) Sites, • Polemast Installation E. Installation at seven (7) sites, • Polemast Installation F. Three (3) Years Maintenance, support, warranty, and data plan for seven (7) Warning Station Equipment including replacement of battery. C. Includes Three (3) Days Training and User Manuals, • Two (2) Days End-User Equipment Maintenance • One (1) Days End-User Training • Digital User Manuals • All training will have eight (8) hours per		With minimum technical specifications:			
A. Electronic Digital Siren with PA (Public Address) Omni Directional (Per Station), SB-2-41 12dB # 30m Omni Directional with 4 Horn Speaker, 12V/38Ah insight battery weatherproof cabinet with standalone power supply solar panel, solar charger, fully functional up to 1km range radius 2x Amplifier 400 Watts, Siren Pole Cap. IP66 steel enclosure. B. Primary Communication Interface per Siren, • Seven (7) pieces CMC-4 Serial HMI Control Console for CS Siren • Seven (7) pieces Compact Microphone • Fourteen (14) pieces Battery, 12V/38 Ah Lead Acid, maintenance free C. Secondary Communication: UHF Radio, • Seven (7) pieces Antenna Omni UHF 403-470 MHz for CS • Seven (7) pieces Antenna feeder LMR/CFD200, feeder to radio N value to BNC female 17m • Seven (7) pieces pig tail 1m, LMR/CFD200, feeder to radio N male to BNC male D. Junction Box Set (cables and mounting accessories), • Seven (7) pieces Piets mounting Junction Box for 4-6 horms 30cm x 30cm x 15cm, 1P66 E. Installation at seven (7) sites, • Polemast Installation • Digital Electronic Siren Installation F. Three (3) Years Maintenance, support, warranty, and data plan for seven (7) Warning Station Equipment including replacement of battery. C. Includes Three (3) Days Training and User Manuals, • Two (2) Days End-User Fraining • Digital User Manuals • Two (2) Days End-User Fraining • Digital User Manuals • Nut ratining will have eight (8) hours per day with five (5) </td <td>1</td> <td>ALERTING SYSTEM FOR DISASTER RISK</td> <td>lot</td> <td>1</td> <td></td>	1	ALERTING SYSTEM FOR DISASTER RISK	lot	1	
Omni Directional (Per Station), SB-24-1 12dB @ 30m Omni Directional with 4 Horn Speaker, 12V/38Ah insight battery weatherproof cabinet with standalone power supply solar panel, solar charger, fully functional up to 1km range radius 2x Amplifier 400 Watts, Siren Pole Cap, IP66 steel enclosure. B. Primary Communication Interface per Siren, • Seven (7) pieces CMC-4 Serial HMI Control Console for CS Siren • Seven (7) pieces Compact Microphone • Fourteen (14) pieces Battery, 12V/38 Ah Lead Acid, maintenance free C. Secondary Communication: UHF Radio, • Seven (7) pieces Antenna Chard HMI Control Console for CS • Seven (7) pieces Antenna feeder LMR/CFD400, N Hz for CS even (7) pieces Antenna feeder LMR/CFD200, feeder to radio N male to BNC male D. Junction Box Set (cables and mounting accessories), • Seven (7) pieces Pole mounting Junction Box for 4-6 horms 30cm x 30cm x 15cm, IP66 E. Installation at seven (7) sites, • Polemast Installation • Digital Electronic Siren Installation F. Three (3) Years Maintenance, support, warranty, and data plan for seven (7) Warning Station Equipment including replacement of battery.		1. SEVEN (7) WARNING STATION EQUIPMENT SET			
 Seven (7) pieces CMC-4 Serial HMI Control Console for CS Siren Seven (7) pieces Compact Microphone Fourteen (14) pieces Battery, 12V/38 Ah Lead Acid, maintenance free C. Secondary Communication: UHF Radio, Seven (7) pieces Radio Transceiver UHF 403-470 MHz for CS Seven (7) pieces FFSK Radio Modem Seven (7) pieces FFSK Radio Modem Seven (7) pieces Antenna Gender LMR/CFD400, N male to BNC female 17m Seven (7) pieces pig tail 1m, LMR/CFD200, feeder to radio N male to BNC male D. Junction Box Set (cables and mounting accessories), Seven (7) pieces Pole mounting Junction Box for 4-6 horns 30cm x 30cm x 15cm, IP66 E. Installation at seven (7) sites, Polemast Installation Digital Electronic Siren Installation F. Three (3) Years Maintenance, support, warranty, and data plan for seven (7) Warning Station Equipment including replacement of battery. G. Includes Three (3) Days Training and User Manuals, Two (2) Days End-User Equipment Maintenance One (1) Day End-User Taqing Digital User Manuals All training will have eight (8) hours per day with five (5) 		Omni Directional (Per Station), SB-2-4-1 12dB @ 30m Omni Directional with 4 Horn Speaker, 12V/38Ah insight battery weatherproof cabinet with standalone power supply solar panel, solar charger, fully functional up to 1km range radius 2x Amplifier 400			
C. Secondary Communication: UHF Radio, Seven (7) pieces Radio Transceiver UHF 403-470 MHz for CS Seven (7) pieces FFSK Radio Modem Seven (7) pieces FFSK Radio Modem Seven (7) pieces Antenna Omni UHF 403-470 MHz Seven (7) pieces Antenna feeder LMR/CFD400, N male to BNC female 17m Seven (7) pieces pig tail 1m, LMR/CFD200, feeder to radio N male to BNC male D. Junction Box Set (cables and mounting accessories), Seven (7) pieces Pole mounting Junction Box for 4-6 horns 30cm x 30cm x 15cm, IP66 E. Installation at seven (7) sites, Polemast Installation Digital Electronic Siren Installation F. Three (3) Years Maintenance, support, warranty, and data plan for seven (7) Warning Station Equipment including replacement of battery. C. Includes Three (3) Days Training and User Manuals, Two (2) Days End-User Equipment Maintenance One (1) Day End-User Training Digital User Manuals All training will have eight (8) hours per day with five (5)		 Seven (7) pieces CMC-4 Serial HMI Control Console for CS Siren Seven (7) pieces Compact Microphone Fourteen (14) pieces Battery, 12V/38 Ah Lead Acid, 			
 Seven (7) pieces FFSK Radio Modem Seven (7) pieces Antenna Omni UHF 403-470 MHz Seven (7) pieces Antenna feeder LMR/CFD400, N male to BNC female 17m Seven (7) pieces pig tail 1m, LMR/CFD200, feeder to radio N male to BNC male D. Junction Box Set (cables and mounting accessories), Seven (7) pieces Pole mounting Junction Box for 4-6 horns 30cm x 30cm x 15cm, IP66 E. Installation at seven (7) sites, Polemast Installation Digital Electronic Siren Installation F. Three (3) Years Maintenance, support, warranty, and data plan for seven (7) Warning Station Equipment including replacement of battery. G. Includes Three (3) Days Training and User Manuals, Two (2) Days End-User Equipment Maintenance One (1) Day End-User Training Digital User Manuals All training will have eight (8) hours per day with five (5) 		C. Secondary Communication: UHF Radio,Seven (7) pieces Radio Transceiver UHF 403-470 MHz for			Calendar Days Upon Issuance
N male to BNC male D. Junction Box Set (cables and mounting accessories), • Seven (7) pieces Pole mounting Junction Box for 4-6 horns 30cm x 30cm x 15cm, IP66 E. Installation at seven (7) sites, • Polemast Installation • Digital Electronic Siren Installation F. Three (3) Years Maintenance, support, warranty, and data plan for seven (7) Warning Station Equipment including replacement of battery. G. Includes Three (3) Days Training and User Manuals, • Two (2) Days End-User Equipment Maintenance • One (1) Day End-User Training • Digital User Manuals • All training will have eight (8) hours per day with five (5)		 Seven (7) pieces Antenna Omni UHF 403-470 MHz Seven (7) pieces Antenna feeder LMR/CFD400, N male to BNC female 17m 			
 Seven (7) pieces Pole mounting Junction Box for 4-6 horns 30cm x 30cm x 15cm, IP66 E. Installation at seven (7) sites, Polemast Installation Digital Electronic Siren Installation F. Three (3) Years Maintenance, support, warranty, and data plan for seven (7) Warning Station Equipment including replacement of battery. G. Includes Three (3) Days Training and User Manuals, Two (2) Days End-User Equipment Maintenance One (1) Day End-User Training Digital User Manuals All training will have eight (8) hours per day with five (5) 					
 Polemast Installation Digital Electronic Siren Installation F. Three (3) Years Maintenance, support, warranty, and data plan for seven (7) Warning Station Equipment including replacement of battery. G. Includes Three (3) Days Training and User Manuals, Two (2) Days End-User Equipment Maintenance One (1) Day End-User Training Digital User Manuals All training will have eight (8) hours per day with five (5) 		• Seven (7) pieces Pole mounting Junction Box for 4-6 horns			
 support, warranty, and data plan for seven (7) Warning Station Equipment including replacement of battery. G. Includes Three (3) Days Training and User Manuals, Two (2) Days End-User Equipment Maintenance One (1) Day End-User Training Digital User Manuals All training will have eight (8) hours per day with five (5) 		Polemast Installation			
 Two (2) Days End-User Equipment Maintenance One (1) Day End-User Training Digital User Manuals All training will have eight (8) hours per day with five (5) 		support, warranty, and data plan for seven (7) Warning			
• All training will have eight (8) hours per day with five (5)		Two (2) Days End-User Equipment MaintenanceOne (1) Day End-User Training			
		• All training will have eight (8) hours per day with five (5)			

 Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Processor: Intel i5 10th Generation or better 	
 Windows Based Software Capable on running on Windows 10 and 11 operating system MySQL Database 1TB Storage 32CB RAM 16VCPUs Easy to use interface that will be installed a desktop PC mentioned in item B Three (3) years license and technical support B. Computer and Communications Equipment, One (1) piece Desktop PC Other (1) piece TC/IP VoIP and Data Card O Precipherals (Same Brand): Starge: T23 in the start of th	
 Windows Based Software Capable on running on Windows 10 and 11 operating system MySQL Database 1TB Storage 32CB RAM 16VCPUs Basy to use interface that will be installed a desktop PC mentioned in item B Three (3) years license and technical support B. Computer and Communications Equipment, One (1) piece Desktop PC Otage: 1TB HDD Kam: 16GB Power Supply: AC and Uninterrupted Power Supply Output Capacity: 1200VA/650W Output Capacity: 1200VA/650W Output Capacity: 1200VA/650W Output Capacity: 1200VA/650W Output con battery: 2200-240V, 50/60Hz +/-1Hz Input Power: 2200-240V, 50/60Hz +/-5Hz (auto sensing) Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 6020 C13 Dimension: 39m x 19m x 14cm Battery Type: Lead-Acid O Processor: Intel 5 10th Generation or better Monitor: Size: 23 is inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 169 Peripherals (Same Brand): Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Compatibility: Windows -Installation: Plug-and-Play Keyboard One (1) piece TC/IP VoIP and Data Card One (1) piece TC/IP VoIP and Data Card One (1) piece atlerent Media Converter; 12VDC One (1) piece Sternet Bachster 6cm x 50cm x 20cm, IP66 Basic AC One (1) piece Sternet Media Converter; 12VDC One (1) piece Sternet Media Converte	
 operating system MySQL Database o 1TB Storage o 32GB RAM o 16vCPUs Easy to use interface that will be installed a desktop PC mentioned in item B Three (3) years license and technical support B. Computer and Communications Equipment, Orne (1) piece Desktop PC o Operating System: Windows 11 Pro o Storage: 1TB HDD o Ram: 16CB o Dutput Capacity: 1200VA/650W Output (capacity: 1200VA/650W Surge: 723 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid o Processor: Intel i5 10th Generation or better o Monitor: Size: 23.8 inches Pamel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16-9 O Feripherals (Same Brand): Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece TC/IP VoIP and Data Card One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece SPT module 100M/Single mode/2-fiber// 1310/LC connected to be connected to Desktop PC 	
 operating system MySQL Database 0 1TB Storage 0 32GB RAM 0 160 CPUs Easy to use interface that will be installed a desktop PC mentioned in item B Three (3) years license and technical support B. Computer and Communications Equipment, Orne (1) piece Desktop PC o Operating System: Windows 11 Pro o Storage: 1TB HDD o Ram: 16CB o Dutput Capacity: 1200VA/650W Output (capacity: 1200VA/650W Surge: 77a J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39m x 19em x 14cm Battery Type: Lead-Acid o Processor: Intel 51 10th Generation or better o Monitor: Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16-9 Periphertals (Same Brand): Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card Orne (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Stery 12V/27Ah lead acid maintenance free One (1) piece Stery 12V/27Ah lead acid maintenance free One (1) piece Stery 12V/27Ah lead acid maintenance free One (1) piece Stery 12V/27Ah lead acid maintenance free One (1) piece Stery 12V/27Ah lead acid maintenance free One (1) piece Stery 12V/27Ah lead acid maintenance free One (1) piece Stery 12V/27Ah lead acid maintenance One (1) piece Stery 12V/27Ah lead acid maintenance One (1) piece Stery 12V/27Ah lead acid maintenance One (1) piece Stery	
 MySQL Database TIB Storage 32GB RAM 166/CPUs Easy to use interface that will be installed a desktop PC mentioned in item B Three (3) years license and technical support B. Computer and Communications Equipment, One (1) piece Desktop PC Operating System: Windows 11 Pro Storage: TB HDD Ram: 16GB Power Supply: AC and Uninterrupted Power Supply Output Capacity: 1200VA/650W Output Capacity: 1200VA/650W Output constant; 1200VA/650W Output constant; 200VA/650W Output Constant; 200VA/650W Output Constant; 1200VA/650W Output Power: 220V-240V, 50/60Hz +/-1Hz Input Power: 220V-240V, 50/60Hz +/-5Hz (auto sensing) Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 EC 6020 C13 Dimension: 39(m x 19cm x 14cm Battery Type: Lead-Acid Dimension: 1920 x 1080 Aspect Ratio: 16-9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Connectivity: USB -Connectivity: USB -Connectivity: USB -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -One (1) piece TC/IP VoIP and Data Card One (1) piece TC/IP VOIP	
 o 1TB Storage o 32GB RAM o 164 CPUs Easy to use interface that will be installed a desktop PC mentioned in item B Three (3) years license and technical support B. Computer and Communications Equipment, • One (1) piece Desktop PC o Operating System: Windows 11 Pro o Storage: 1TB HDD o Ram: 16CB o Output Capacity: 1200 VA/650W • Output Capacity: 1200 VA/650W • Output Capacity: 1200 VA/650W • Output (on battery): 220V-240V, 50/60Hz +/-1Hz • Input Power: 220V-240V, 50/60Hz +/-5Hz (auto sensing) • Surge: 273 J • Battery recharge time: 8 hours • No. of Outlets: 5-4 universal, 1EC 60320 C13 • Dimension: 39 cm x 19 cm x 14 cm • Battery Type: Lead-Acid • Panel Type: IPS • Resolution: 1920 x 1080 • Aspect Ratio: 16-9 • Peripherals Game Brand): • Mouse: • Connectivity: USB • Compatibility: Windows -Installation: Plug-and-Play • Keyboard • Orne (1) piece TC/IP VoIP and Data Card • One (1) piece TC/IP VoIP and Data Card • One (1) piece TC/IP VoIP and Data Card • One (1) piece To: 90 Universal Enclosure 60 cm x 50 cm x 20 cm, 1P66 Basic AC • One (1) piece To: 91 To: 92 To:	
 o 32GB RAM o 16vCPUs Fasy to use interface that will be installed a desktop PC mentioned in item B Three (3) years license and technical support B. Computer and Communications Equipment, One (1) piece Desktop PC o Operating System: Windows 11 Pro o Storage: 1TB HDD 0 Ram: 16CB o Power Supply: AC and Uninterrupted Power Supply Output Capacity: 1200VA/650W Operation: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Commetivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Compatibility: Windows -Installation: Plug-and-Play Keyboard One (1) piece TC/IP VoIP and Data Card One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece Stirg 20/27Ah lead acid maintenance free One (1) piece Stirg 2	
 Easy to use interface that will be installed a desktop PC mentioned in item B Three (3) years license and technical support B. Computer and Communications Equipment, Ome (1) piece Desktop PC Operating System: Windows 11 Pro Storage: ITB HDD Ram: 16CB Power Supply: AC and Uninterrupted Power Supply Output (apacity: 1200VA/650W Output Chapacity: 1200VA/650W Output Chapacity: 1200VA/650W Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Conne (
mentioned in item B • Three (3) years license and technical support B. Computer and Communications Equipment, • One (1) piece Desktop PC • Operating System: Windows 11 Pro • Storage: 1TB HDD • Ram: 16GB • Power Supply: AC and Uninterrupted Power Supply • Output Capacity: 1200VA/650W • Output Copacity: 1200VA/650W • Output to battery: 220V-240V, 50/60Hz +/-1Hz • Input Power: 220V-240V, 50/60Hz +/- 5Hz (auto sensing) • Surge: 273 J • Battery recharge time: 8 hours • No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 • Dimension: 39Cm x 19Cm x 14cm • Battery Type: Lead-Acid • Dimension: 39Cm x 19Cm x 14cm • Monitor: • Size: 23.8 inches • Panel Type: IPS • Resolution: 1920 x 1080 • Aspect Ratio: 16:9 • Peripherals (Same Brand): • Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play • Keyboard • One (1) piece TC/IP VoIP and Data Card • One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC • One (1) piece Bittery 12V/27Ah lead acid maintenance free • One (1) piece SPI module 100M/Single mode/2-fiber/ 1310/LC connector to be connected to Desktop PC	
mentioned in item B • Three (3) years license and technical support B. Computer and Communications Equipment, • One (1) piece Desktop PC • Operating System: Windows 11 Pro • Storage: 1TB HDD • Ram: 16GB • Power Supply: AC and Uninterrupted Power Supply • Output Capacity: 1200VA/650W • Output Copacity: 1200VA/650W • Output to battery: 220V-240V, 50/60Hz +/-1Hz • Input Power: 220V-240V, 50/60Hz +/- 5Hz (auto sensing) • Surge: 273 J • Battery recharge time: 8 hours • No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 • Dimension: 39Cm x 19Cm x 14cm • Battery Type: Lead-Acid • Dimension: 39Cm x 19Cm x 14cm • Monitor: • Size: 23.8 inches • Panel Type: IPS • Resolution: 1920 x 1080 • Aspect Ratio: 16:9 • Peripherals (Same Brand): • Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play • Keyboard • One (1) piece TC/IP VoIP and Data Card • One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC • One (1) piece Bittery 12V/27Ah lead acid maintenance free • One (1) piece SPI module 100M/Single mode/2-fiber/ 1310/LC connector to be connected to Desktop PC	
 B. Computer and Communications Equipment, One (1) piece Desktop PC Operating System: Windows 11 Pro Storage: 1TB HDD Ram: 16GB Power Supply: AC and Uninterrupted Power Supply Output Capacity: 1200VA/650W Output (on battery): 220V-240V, 50/60Hz +/-1Hz Input Power: 220V-240V, 50/60Hz +/-5Hz (auto sensing) Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Dattery Type: Lead-Acid Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16.9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece TC/IP VoIP and Data Card One (1) piece TC/IP VoIP and Data Card One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece StP module 100M/Single mode/2-fiber/1310/LC connector to be connected to Desktop PC 	
 One (1) piece Desktop PC Operating System: Windows 11 Pro Storage: 1TB HDD Ram: 16GB Power Supply: AC and Uninterrupted Power Supply Output Capacity: 1200VA/650W Output (on battery): 220V-240V, 50/60Hz +/-1Hz Input Power: 220V-240V, 50/60Hz +/-1Hz Input Power: 220V-240V, 50/60Hz +/-1Hz Input Power: 220V-240V, 50/60Hz +/-1Hz Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Datery Type: Lead-Acid Battery Type: Lead-Acid Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Conne (1) piece SEN power supply with battery	
 One (1) piece Desktop PC Operating System: Windows 11 Pro Storage: 1TB HDD Ram: 16GB Power Supply: AC and Uninterrupted Power Supply Output Capacity: 1200VA/650W Output (on battery): 220V-240V, 50/60Hz +/-1Hz Input Power: 220V-240V, 50/60Hz +/- 5Hz (auto sensing) Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Dattery Type: Lead-Acid Battery Type: Lead-Acid Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece V8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece SEP module 100M/Single mode/2-fiber/1310/LC connector to be connected to Desktop PC 	
 o Operating System: Windows 11 Pro o Storage: 1TB HDD o Ram: 16GB o Power Supply: AC and Uninterrupted Power Supply Output Capacity: 1200VA/650W Output (on battery): 220V-240V, 50/60Hz +/-1Hz Input Power: 220V-240V, 50/60Hz +/- 5Hz (auto sensing) Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Days Battery Type: Lead-Acid o Processor: Intel 15 10th Generation or better o Monitor: Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16-9 o Pertipherals (Same Brand): Mouse: -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Compatibility: Windows -Installation: Plug-and-Play Sceybard One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece Bttery 12V/27Ah lead acid maintenance free One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece SFP module 100M/Single mode/2-fiber/1310/LC connector to be connected to Desktop PC 	
 Storage: 1TB HDD Ram: 16GB Power Supply: AC and Uninterrupted Power Supply Output Capacity: 1200VA/650W Output Capacity: 1200VA/650W Output Power: 220V-240V, 50/60Hz +/-1Hz Input Power: 220V-240V, 50/60Hz +/- 5Hz (auto sensing) Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Compatibility: Windows -Installation: Plug-and-Play One (1) piece Ethernet Media Converter; 12VDC One (1) piece U-/IP VoIP and Data Card One (1) piece U-/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece Ward Success and Success and	
 Ram: 16GB Power Supply: AC and Uninterrupted Power Supply Output: Qapacity: 1200VA/650W Output (on battery): 220V-240V, 50/60Hz +/-1Hz Input Power: 220V-240V, 50/60Hz +/-5Hz (auto sensing) Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Processor: Intel i5 10th Generation or better Mouse: -Connectivity: USB -Connectivity: Windows -Installation: Plug-and-Play Keyboard -Connectivity: USB -Connectivity: Wise -Installation: Plug-and-Play Nedopard-Play One (1) piece TC/IP VoIP and Data Card One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece attery 12V/27Ah lead acid maintenance free One (1) piece Bthernet Media Converter; 12VDC 	
 o Power Supply: AC and Uninterrupted Power Supply Outlets: Five (5) Output Capacity: 1200VA/650W Output I con battery): 220V-240V, 50/60Hz +/-1Hz Input Power: 220V-240V, 50/60Hz +/-5Hz (auto sensing) Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Processor: Intel 15 10th Generation or better o Monitor: Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 o Peripherals (Same Brand): Mouse: -Connectivity: USB -Connectivity: USB -Connectivity: Windows -Installation: Plug-and-Play Keyboard -Connectivity: Windows -Installation: Plug-and-Play One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece SPIP module 100M/Single mode/2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 Outlets: Five (5) Output Capacity: 1200VA/650W Output (on battery): 220V-240V, 50/60Hz +/-1Hz Input Power: 220V-240V, 50/60Hz +/- 5Hz (auto sensing) Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Days Battery Type: Lead-Acid Processor: Intel i5 10th Generation or better Monitor: Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard Orne (1) piece TC/IP VoIP and Data Card One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece SFP module 100M/Single mode/2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 Output Capacity: 1200VA/650W Output (on battery): 220V-240V, 50/60Hz +/-1Hz Input Power: 220V-240V, 50/60Hz +/- 5Hz (auto sensing) Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Days Battery Type: Lead-Acid O Processor: Intel i5 10th Generation or better Monitor: Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: Connectivity: USB Compatibility: Windows Installation: Plug-and-Play Keyboard Compatibility: Windows Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece SEP module 100M/Single mode/2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 Output (on battery): 220V-240V, 50/60Hz +/-1Hz Input Power: 220V-240V, 50/60Hz +/- 5Hz (auto sensing) Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Dattery Type: Lead-Acid Surge: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece SFP module 100M/Single mode/2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 Input Power: 220V-240V, 50/60Hz +/- 5Hz (auto sensing) Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid O Processor: Intel i5 10th Generation or better Monitor: Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Connectivity: USB -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece athery 12V/27Ah lead acid maintenance free One (1) piece abling and assembly One (1) piece abling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 sensing) Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid o Processor: Intel i5 10th Generation or better o Monitor: Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Connectivity: USB -Connectivity: Windows -Installation: Plug-and-Play Keyboard -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece STP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 Surge: 273 J Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Days Battery Type: Lead-Acid O Processor: Intel i5 10th Generation or better Monitor: Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethermet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece attery 12V/27Ah lead acid maintenance free One (1) piece abiling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 Battery recharge time: 8 hours No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid O Processor: Intel i5 10th Generation or better Monitor: Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece cabling and assembly One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid O Processor: Intel 15 10th Generation or better Monitor: Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 A spect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece abttery 12V/27Ah lead acid maintenance free One (1) piece abttery 12V/27Ah lead acid maintenance free One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	thin
 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Days o Processor: Intel i5 10th Generation or better o Monitor: Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece tEchernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece abttery 12V/27Ah lead acid maintenance free One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	y (60)
 Battery Type: Lead-Acid Days Issu Processor: Intel i5 10th Generation or better Monitor: Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Connectivity: USB -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece abling and assembly One (1) piece abling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	endar
 o Processor: Intel i5 10th Generation or better o Monitor: Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 o Peripherals (Same Brand): Mouse: -Connectivity: USB -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece SFP module 100M//Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	Upon
 b Processor: Intel 15 forth Ceneration of Detter o Monitor: Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 o Peripherals (Same Brand): Mouse: Connectivity: USB Compatibility: Windows Installation: Plug-and-Play Keyboard Compatibility: Windows Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	ance
 o Monitor: Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 o Peripherals (Same Brand): Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece Ethernet Media Converter; 12VDC One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	otice
 Size: 23.8 inches Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	oceed
 Panel Type: IPS Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece abttery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 Resolution: 1920 x 1080 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 Aspect Ratio: 16:9 Peripherals (Same Brand): Mouse: Connectivity: USB Compatibility: Windows Installation: Plug-and-Play Keyboard Connectivity: USB Compatibility: Windows Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 o Peripherals (Same Brand): Mouse: Connectivity: USB Compatibility: Windows Installation: Plug-and-Play Keyboard Connectivity: USB Compatibility: Windows Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 Mouse: Connectivity: USB Compatibility: Windows Installation: Plug-and-Play Keyboard Connectivity: USB Compatibility: Windows Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 -Compatibility: Windows -Installation: Plug-and-Play Keyboard -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 Keyboard Connectivity: USB Compatibility: Windows Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 -Connectivity: USB -Compatibility: Windows -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 -Compatibility: Windows Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/1310/LC connector to be connected to Desktop PC 	
 -Installation: Plug-and-Play One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 One (1) piece TC/IP VoIP and Data Card One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 One (1) piece Ethernet Media Converter; 12VDC One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 x 20cm, IP66 Basic AC One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 One (1) piece 2V/8A DC Power supply with battery monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 monitor UPS One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
 One (1) piece Battery 12V/27Ah lead acid maintenance free One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
free • One (1) piece cabling and assembly • One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC	
 One (1) piece cabling and assembly One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC 	
One (1) piece SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector to be connected to Desktop PC	
1310/LC connector to be connected to Desktop PC	
-	
• One (1) piece Radio Transceiver UHF 403-470 MHz for CS	
One (1) piece Kadio Maliscerver Offr 405-470 MHz for C5 One (1) piece FFSK Radio Modem	
One (1) piece Antenna feeder LMR/CFD500, N male to N	
female 30m	
One (1) piece Pig-tail 1m, LMR/CFD200, feeder to radio N	
male to BNC male	

 One (1) piece CMC-8 HMI Warning System Control Console for CS Siren One (1) piece TW-100 Communication Audio Pane One (1) piece RS485 adapter to USB C. Control Station Software, Ability to control up to individual siren and siren as Scheduled or user requested diagnostic tests. Can activate Warning or Emergency Signals and prerecord Voice Messages. Provides alert notifications for Siren's Errors in Ho Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support Scheduled or user requested activation. D. Automated Trigger Middleware, Integration connectors to CS siren for manual and automated trigger. Integration connectors to existing iRISE-UP hazard 	I groups.
 API call management to call iRISE-UP data flexible 5min, 15min, 30min, 45min, 1 hour to 12 hours upot time for hazards that require API calling. Development of automated and programmed trigg and messaging depending on alert category and ha flexibility for adjustment. Siren location and status shall be integrated and pu to be displayed in the iRISE-UP GIS interface of bo applications and native mobile application (IOS/An QC DRRMO will provide access to the developmen environment. 	per lating Within Sixty (60) Calendar Days Upon Issuance of Notice to Proceed
 E. Historical Data Storage, Store and Manage alert data thru a simple historica management interface 	l data
 F. Three (3) years, 1 standard Internet Broadband Line for 24 x 7 use, One-time installation Bandwidth: Unlimited (No Data Cap) Speed: Up to 50Mbps G. Three (3) years software license, maintenance, so warranty, and technical support, 	ıpport,
 All equipment Software H. Includes Three (3) Days Training and User Man One (1) Day Train the Trainer 	uals
 One (1) Day System Admin Training One (1) Day End-User Training Digital User Manuals All training will have eight (8) hours per da five (5) attendees. 	y with
with attached Terms of Reference which will form integral part of this bidding document	1 an

Schedule of Requirements Page 3 of 4

I hereby certify to comply and deliver all the above requirements.

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____

Schedule of Requirements Page 4 of 4

Notes for Preparing the Technical Specifications

A set of precise and clear specifications is a prerequisite for Bidders to respond realistically and competitively to the requirements of the Procuring Entity without qualifying their Bids. In the context of Competitive Bidding, the specifications (*e.g.* production/delivery schedule, manpower requirements, and after-sales service/parts, descriptions of the lots or items) must be prepared to permit the widest possible competition and, at the same time, present a clear statement of the required standards of workmanship, materials, and performance of the goods and services to be procured. Only if this is done will the objectives of transparency, equity, efficiency, fairness, and economy in procurement be realized, responsiveness of bids be ensured, and the subsequent task of bid evaluation and post-qualification facilitated. The specifications should require that all items, materials and accessories to be included or incorporated in the goods be new, unused, and of the most recent or current models, and that they include or incorporate all recent improvements in design and materials unless otherwise provided in the Contract.

Samples of specifications from previous similar procurements are useful in this respect. The use of metric units is encouraged. Depending on the complexity of the goods and the repetitiveness of the type of procurement, it may be advantageous to standardize the General Technical Specifications and incorporate them in a separate subsection. The General Technical Specifications should cover all classes of workmanship, materials, and equipment commonly involved in manufacturing similar goods. Deletions or addenda should then adapt the General Technical Specifications to the particular procurement.

Care must be taken in drafting specifications to ensure that they are not restrictive. In the specification of standards for equipment, materials, and workmanship, recognized Philippine and international standards should be used as much as possible. Where other particular standards are used, whether national standards or other standards, the specifications should state that equipment, materials, and workmanship that meet other authoritative standards, and which ensure at least a substantially equal quality than the standards mentioned, will also be acceptable. The following clause may be inserted in the Special Conditions of Contract or the Technical Specifications.

Sample Clause: Equivalency of Standards and Codes

Wherever reference is made in the Technical Specifications to specific standards and codes to be met by the goods and materials to be furnished or tested, the provisions of the latest edition or revision of the relevant standards and codes shall apply, unless otherwise expressly stated in the Contract. Where such standards and codes are national or relate to a particular country or region, other authoritative standards that ensure substantial equivalence to the standards and codes specified will be acceptable.

Reference to brand name and catalogue number should be avoided as far as possible; where unavoidable they should always be followed by the words "*or at least equivalent*." References to brand names cannot be used when the funding source is the GOP.

Where appropriate, drawings, including site plans as required, may be furnished by the Procuring Entity with the Bidding Documents. Similarly, the Supplier may be requested to provide drawings or samples either with its Bid or for prior review by the Procuring Entity during contract execution.

Bidders are also required, as part of the technical specifications, to complete their statement of compliance demonstrating how the items comply with the specification.

Technical Specifications

PROJECT NAME: SUPPLY, INSTALLATION, TESTING AND COMMISSIONING OF EMERGENCY OUTDOOR ALERTING SYSTEM FOR DISASTER RISK REDUCTION AND MANAGEMENT PROJECT NO. QCDRRMO-23-IT-1078

Item	Specification	Statement of Compliance
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post- qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]
A.	With minimum technical specifications:	
1	Requirements and specifications of the OUTDOOR ALERTING SYSTEM FOR DISASTER RISK REDUCTION AND MANAGEMENT	
	1. SEVEN (7) WARNING STATION EQUIPMENT SET	
	A. Electronic Digital Siren with PA (Public Address) Omni Directional (Per Station), SB-2-4-1 12dB @ 30m Omni Directional with 4 Horn Speaker, 12V/38Ah insight battery weatherproof cabinet with standalone power supply solar panel, solar charger, fully functional up to 1km range radius 2x Amplifier 400 Watts, Siren Pole Cap, IP66 steel enclosure.	
	 B. Primary Communication Interface per Siren, Seven (7) pieces CMC-4 Serial HMI Control Console for CS Siren Seven (7) pieces Compact Microphone Fourteen (14) pieces Battery, 12V/38 Ah Lead Acid, maintenance free 	
	 C. Secondary Communication: UHF Radio, Seven (7) pieces Radio Transceiver UHF 403-470 MHz for CS Seven (7) pieces FFSK Radio Modem Seven (7) pieces Antenna Omni UHF 403-470 MHz Seven (7) pieces Antenna feeder LMR/CFD400, N male to BNC female 17m Seven (7) pieces pig tail 1m, LMR/CFD200, feeder to radio N male to BNC male 	

D. Junction Box Set (cables and mounting accessories),	
• Seven (7) pieces Pole mounting Junction Box for 4-6	
horns 30cm x 30cm x 15cm, IP66	
E. Installation at seven (7) sites,	
Polemast Installation	
Digital Electronic Siren Installation	
F. Three (3) Years Maintenance,	
support, warranty, and data plan for seven (7) Warning	
Station Equipment including replacement of battery.	
G. Includes Three (3) Days Training and User Manuals,	
• Two (2) Days End-User Equipment Maintenance	
• One (1) Day End-User Training	
• Digital User Manuals	
• All training will have eight (8) hours per day with five	
(5) attendees.	
2. CONTROL STATION AND COMMAND CENTER	
SOFTWARE	
A. Software Backend and Frontend,	
Windows Based Software	
• Capable on running on Windows 10 and 11	
operating system	
• MySQL Database	
o 1TB Storage	
o 32GB RAM	
o 16vCPUs	
• Easy to use interface that will be installed a desktop PC	
mentioned in item B	
• Three (3) years license and technical support	
R Computer and Communications Equipment	
B. Computer and Communications Equipment,One (1) piece Desktop PC	
o Operating System: Windows 11 Pro	
o Storage: 1TB HDD	
o Ram: 16GB	
o Power Supply: AC and Uninterrupted Power	
Supply	
 Outlets: Five (5) 	
 Output Capacity: 1200VA/650W 	
• Output (on battery): 220V-240V, 50/60Hz +/-1Hz	
 Input Power: 220V-240V, 50/60Hz +/- 5Hz (auto 	
sensing)	
 Surge: 273 J 	
Battery recharge time: 8 hours	
■ No of Outlater 5 Augustrangel 1 IEC (0220 C12	
 No. of Outlets: 5 - 4 universal, 1 IEC 60320 C13 Dim on aircra 20 are at 10 are at 14 are 	
 Dimension: 39cm x 19cm x 14cm 	
 Dimension: 39cm x 19cm x 14cm 	
Dimension: 39cm x 19cm x 14cmBattery Type: Lead-Acid	
 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Processor: Intel i5 10th Generation or better 	
 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Processor: Intel i5 10th Generation or better Monitor: Size: 23.8 inches 	
 Dimension: 39cm x 19cm x 14cm Battery Type: Lead-Acid Processor: Intel i5 10th Generation or better Monitor: Size: 23.8 inches 	

o Peripherals (Same Brand):	
e l'empréries (curre stand).	
 Mouse: 	
-Connectivity: USB	
-Compatibility: Windows	
-Installation: Plug-and-Play	
 Keyboard 	
-Connectivity: USB	
-Compatibility: Windows	
-Installation: Plug-and-Play	
• One (1) piece TC/IP VoIP and Data Card	
• One (1) piece Ethernet Media Converter; 12VDC	
• One (1) piece UE-60-50 Universal Enclosure 60cm x 50cm	
x 20cm, IP66 Basic AC	
• One (1) piece 2V/8A DC Power supply with battery	
monitor UPS	
• One (1) piece Battery 12V/27Ah lead acid maintenance	
free	
• One (1) piece cabling and assembly	
• One (1) piece SFP module 100M/Single mode/ 2-fiber/	
1310/LC connector to be connected to Desktop PC	
mentioned in item B.	
• One (1) piece Radio Transceiver UHF 403-470 MHz for CS	
• One (1) piece FFSK Radio Modem	
• One (1) piece Antenna feeder LMR/CFD500, N male to N	
female 30m	
• One (1) piece Pig-tail 1m, LMR/CFD200, feeder to radio	
N male to BNC male	
• One (1) piece CMC-8 HMI Warning System Control	
Console for CS Siren	
• One (1) piece TW-100 Communication Audio Panel	
• One (1) piece RS485 adapter to USB	
C. Control Station Software,	
• Ability to control up to individual siren and siren groups	
• Scheduled or user requested diagnostic tests.	
Can activate Warning or Emergency Signals and	
prerecord Voice Messages.	
• Provides alert notifications for Siren's Errors in Horns,	
• Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication.	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support Scheduled or user requested activation. 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support Scheduled or user requested activation. D. Automated Trigger Middleware, 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support Scheduled or user requested activation. D. Automated Trigger Middleware, Integration connectors to CS siren for manual and 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support Scheduled or user requested activation. D. Automated Trigger Middleware, Integration connectors to CS siren for manual and automated trigger. 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support Scheduled or user requested activation. D. Automated Trigger Middleware, Integration connectors to CS siren for manual and 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support Scheduled or user requested activation. D. Automated Trigger Middleware, Integration connectors to CS siren for manual and automated trigger. 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support Scheduled or user requested activation. D. Automated Trigger Middleware, Integration connectors to CS siren for manual and automated trigger. Integration connectors to existing iRISE-UP hazard API's 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support Scheduled or user requested activation. D. Automated Trigger Middleware, Integration connectors to CS siren for manual and automated trigger. Integration connectors to existing iRISE-UP hazard API's for automated trigger (up to 50 data parameters). 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support Scheduled or user requested activation. D. Automated Trigger Middleware, Integration connectors to CS siren for manual and automated trigger. Integration connectors to existing iRISE-UP hazard API's for automated trigger (up to 50 data parameters). API call management to call iRISE-UP data flexible per 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support Scheduled or user requested activation. D. Automated Trigger Middleware, Integration connectors to CS siren for manual and automated trigger. Integration connectors to existing iRISE-UP hazard API's for automated trigger (up to 50 data parameters). API call management to call iRISE-UP data flexible per 5min, 15min, 30min, 45min, 1 hour to 12 hours updating 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support Scheduled or user requested activation. D. Automated Trigger Middleware, Integration connectors to CS siren for manual and automated trigger. Integration connectors to existing iRISE-UP hazard API's for automated trigger (up to 50 data parameters). API call management to call iRISE-UP data flexible per 5min, 15min, 30min, 45min, 1 hour to 12 hours updating time for hazards that require API calling. Development of automated and programmed triggers 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support Scheduled or user requested activation. D. Automated Trigger Middleware, Integration connectors to CS siren for manual and automated trigger. Integration connectors to existing iRISE-UP hazard API's for automated trigger (up to 50 data parameters). API call management to call iRISE-UP data flexible per 5min, 15min, 30min, 45min, 1 hour to 12 hours updating time for hazards that require API calling. Development of automated and programmed triggers and messaging depending on alert category and have 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support Scheduled or user requested activation. D. Automated Trigger Middleware, Integration connectors to CS siren for manual and automated trigger. Integration connectors to existing iRISE-UP hazard API's for automated trigger (up to 50 data parameters). API call management to call iRISE-UP data flexible per 5min, 15min, 30min, 45min, 1 hour to 12 hours updating time for hazards that require API calling. Development of automated and programmed triggers and messaging depending on alert category and have the flexibility for adjustment. 	
 Provides alert notifications for Siren's Errors in Horns, Amplifiers, Batteries and Communication. Can perform live voice announcements for Public Addressing. Three (3) User Access Levels and i-button support Scheduled or user requested activation. D. Automated Trigger Middleware, Integration connectors to CS siren for manual and automated trigger. Integration connectors to existing iRISE-UP hazard API's for automated trigger (up to 50 data parameters). API call management to call iRISE-UP data flexible per 5min, 15min, 30min, 45min, 1 hour to 12 hours updating time for hazards that require API calling. Development of automated and programmed triggers and messaging depending on alert category and have 	

	(IOS/Android), QC DRRMO will provide access to the	
	development environment.	
	development environment.	
	E. Historical Data Storage,	
	• Store and Manage alert data thru a simple historical data	
	management interface	
	-	
	F. Three (3) years, 1 standard Internet Broadband	
	Line for 24 x 7 use,	
	One-time installation	
	 Bandwidth: Unlimited (No Data Cap) 	
	• Speed: Up to 50Mbps	
	C. Three (2) means a fitzen lisener maintenance	
	G. Three (3) years software license, maintenance,	
	support, warranty, and technical support,All equipment	
	 An equipment Software 	
	• Software	
	H. Includes Three (3) Days Training and User Manuals	
	• One (1) Day Train the Trainer	
	One (1) Day System Admin Training	
	• One (1) Day End-User Training	
	Digital User Manuals	
	• All training will have eight (8) hours per day with	
	five (5) attendees.	
	*** with attached Terms of Reference which will form an	
B.	<i>integral part of this bidding document</i> *** Compliance to the Schedule of Requirements (Section VI)	
D.	Compliance to the Schedule of Requirements (Section VI)	

I hereby certify to comply and deliver all the above requirements.

Name: _____

Legal Capacity: _____

Signature: ______

Duly authorized to sign the Bid for and behalf of:

Technical Specifications Page 4 of 4

PROJECT NAME: SUPPLY, INSTALLATION, TESTING AND COMMISSIONING OF EMERGENCY OUTDOOR ALERTING SYSTEM FOR DISASTER RISK REDUCTION AND MANAGEMENT PROJECT NO. QCDRRMO-23-IT-1078

	PARTICULAR	UNIT OF ISSUE	QUANTITY	UNIT COST	TOTAL
	SEVEN (7) WARNING STATION EQUIPMENT SET				
	Electronic Digital Siren with PA (Public Address) Omni Directional	set	7		
	Primary Communication Interface per Siren				
	CMC-4 Serial HMI Control Console for CS Siren	piece	7		
	Compact Microphone	piece	7		
	• Battery ,12V/38 Ah Lead Acid, maintenance free	piece	14		
	Secondary Communication: UHF Radio				
	• Radio Transceiver UHF 403-470 MHz for CS	piece	7		
	• FFSK Radio Modem	piece	7		
	• Antenna Omni UHF 403-470 MHz	piece	7		
	Antenna feeder LMR/CFD400, N male to BNC female 17m	piece	7		
	• pig tail 1m, LMR/CFD200, feeder to radio N male to BNC male	piece	7		
	Junction Box Set (cables and mounting accessories)				
	Pole mounting Junction Box for 4-6 horns 30cm x 30cm x 15cm, IP66	piece	7		
	Installation at seven (7) sites				
	Polemast Installation	lot	7		
	Digital Electronic Siren Installation	lot	7		
	Maintenance, support, warranty , and data plan for seven (7) Warning Station Equipment including replacement of Battery	lot	7		
	Training and User Manuals	lot	1		
2	CONTROL STATION AND COMMAND CENTER SOFTWARE				
	Software Backend and Frontend	lot	1		
	Computer and Communications Equipment				
	Desktop PC	set	1		
	TC/IP VoIP and Data Card	piece	1		
	Ethernet Media Converter; 12VDC	piece	1		

Cost Derivation Page 1 of 2

• UE-60-50 Universal Enclosure 60cm x 50cm x 20cm, IP66 Basic AC	piece	1		
2V/8A DC Power supply with battery monitor UPS	piece	1		
Battery 12V/27Ah lead acid maintenance free	piece	1		
• cabling and assembly	piece	1		
• SFP module 100M/Single mode/ 2-fiber/ 1310/LC connector	piece	1		
Radio Transceiver UHF 403-470 MHz for CS	piece	1		
FFSK Radio Modem	piece	1		
Antenna feeder LMR/CFD500, N male to N female 30m	piece	1		
• Pig-tail 1m, LMR/CFD200, feeder to radio N male to BNC male	piece	1		
CMC-8 HMI Warning System Control Console for CS Siren	piece	1		
TW-100 Communication Audio Panel One (1) piece RS485 adapter to USB	piece	1		
Control Station Software	lot	1		
Automated Trigger Middleware	lot	1		
Historical Data Storage	lot	1		
Internet Broadband line for 24 x 7 use	lot	1		
software license, maintenance, support, warranty, and technical support	lot	1		
Training and User Manuals	lot	1		
	TOTAL BID AMOUNT			

I hereby certify to comply and deliver all the above requirements.

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: ______

Cost Derivation Page 2 of 2

Section VIII. Checklist of Technical and Financial Documents

Notes on the Checklist of Technical and Financial Documents

The prescribed documents in the checklist are mandatory to be submitted in the Bid, but shall be subject to the following:

- a. GPPB Resolution No. 09-2020 on the efficient procurement measures during a State of Calamity or other similar issuances that shall allow the use of alternate documents in lieu of the mandated requirements; or
- b. Any subsequent GPPB issuances adjusting the documentary requirements after the effectivity of the adoption of the PBDs.

The BAC shall be checking the submitted documents of each Bidder against this checklist to ascertain if they are all present, using a non-discretionary "pass/fail" criterion pursuant to Section 30 of the 2016 revised IRR of RA No. 9184.

Checklist of Technical and Financial Documents I. TECHNICAL COMPONENT ENVELOPE

Class "A" Documents

Legal Documents

(a) Valid PhilGEPS Registration Certificate (Platinum Membership) (all pages) in accordance with Section 8.5.2 of the IRR;

Technical Documents

(b) Statement of the prospective bidder of all its ongoing government and private contracts, including contracts awarded but not yet started, if any, whether similar or not similar in nature and complexity to the contract to be bid (in a FORM prescribed by the QC-BAC-GOODS AND SERVICES); and

- (c) Statement of the bidder's Single Largest Completed Contract (SLCC) similar to the contract to be bid, except under conditions provided for in Sections 23.4.1.3 and 23.4.2.4 of the 2016 revised IRR of RA No. 9184, within the relevant period as provided in the Bidding Documents (in a FORM prescribed by the QC-BAC-GOODS AND SERVICES); and
- (d) Original copy of Bid Security. If in the form of a Surety Bond, submit also a certification issued by the Insurance Commission;
 or

Original copy of Notarized Bid Securing Declaration; and

(e) Conformity with Section VI. (Schedule of Requirements) and Section VII. (Technical Specifications), which may include production/delivery schedule, manpower requirements, and/or after-sales/parts, if applicable; <u>and</u>
 (f) Original duly signed Omnibus Sworn Statement (OSS);

and if applicable, Original Notarized Secretary's Certificate in case of a corporation, partnership, or cooperative; or Original Special Power of Attorney of all members of the joint venture giving full power and authority to its officer to sign the OSS and do acts to represent the Bidder.

Financial Documents

(g) The prospective bidder's computation of Net Financial Contracting Capacity (NFCC) (in a FORM prescribed by the QC-BAC-GOODS AND SERVICES);

<u>or</u>

A committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation.

Class "B" Documents

(h) If applicable, a duly signed joint venture agreement (JVA) in case the joint venture is already in existence;

<u>or</u>

duly notarized statements from all the potential joint venture partners stating that they will enter into and abide by the provisions of the JVA in the instance that the bid is successful.

Other documentary requirements under RA No. 9184 (as applicable)

- (i) [For foreign bidders claiming by reason of their country's extension of reciprocal rights to Filipinos] Certification from the relevant government office of their country stating that Filipinos are allowed to participate in government procurement activities for the same item or product.
- (j) Certification from the DTI if the Bidder claims preference as a Domestic Bidder or Domestic Entity.

II. FINANCIAL COMPONENT ENVELOPE

- (a) Original of duly signed and accomplished Financial Bid Form;
- (b) Original of duly signed and accomplished Price Schedule(s); and
- (c) Cost derivation.

III. REQUIRED DOCUMENTS in BDS SECTION 20.2 and 21.2

- Organizational Chart for the project
- Notarized Affidavit of Undertakings stating compliance to the following:
 - c. The bidder must have demonstrated capability, experience and have implemented and managed citywide sensor equipment networks and meteorological information alerting to highly urbanized city (HUC) in the Philippines. These systems must be already existing and proven working by other highly urbanized cities (HUC) in the Philippines.
 - d. The bidder will conduct the required training and documentation.
- Statement of Warranty
 - For Warning Station Equipment Set Three (3) Years Maintenance, support, warranty, and data plan for 7 Warning Station Equipment including replacement of Battery
 - For Control Station & Command Center Software Three (3) Years software license, maintenance, support warranty, and technical support for All equipment and software

(Please refer to the attached Terms of Reference)

Note:

- Please refer to
 [https://drive.google.com/file/d/1uiYurh5WrpBL5B_pqpzAb62yucAblR1p/view?usp=sharinq] for the following requirements:
 - a. Computation of NFCC;
 - b. List of All Ongoing Contracts/List of Contracts already awarded but not yet started;
 - c. Statement of Single Largest Completed Contract
- 2. Please refer to GPPB Resolution No. 16-2020 for the following requirements:
 - a. Bid Form;
 - b. Price Schedule (for Goods Offered from Abroad/ Within the Philippines)
 - c. Bid Securing Declaration; and
 - d. Omnibus Sworn Statement

