

REPUBLIC OF THE PHILIPPINES QUEZON CITY GOVERNMENT BIDS AND AWARDS COMMITTEE – GOODS AND SERVICES



PHILIPPINE BIDDING DOCUMENTS

(As Harmonized with Development Partners)

PROCUREMENT OF VARIOUS HOSPITAL EQUIPMENT

PROJECT NO. QCGH-24-HME-1557

LINE 1: ANESTHESIA MACHINE WITH PATIENT MONITOR AND BRAIN FUNCTION AND CEREBRAL OXIMETRY MONITORING

LINE 2: ANESTHESIA MACHINE WITH PATIENT MONITOR

LINE 3: SUPPLY AND DELIVERY OF X-RAY FLAT PANEL DIGITIZER WITH DIGITAL RADIOGRAPHY SYSTEM

Government of the Republic of the Philippines

Sixth Edition July 2020

Preface

These Philippine Bidding Documents (PBDs) for the procurement of Goods through Competitive Bidding have been prepared by the Government of the Philippines for use by any branch, constitutional commission or office, agency, department, bureau, office, or instrumentality of the Government of the Philippines, National Government Agencies, including Government-Owned and/or Controlled Corporations, Government Financing Institutions, State Universities and Colleges, and Local Government Unit. The procedures and practices presented in this document have been developed through broad experience, and are for mandatory use in projects that are financed in whole or in part by the Government of the Philippines or any foreign government/foreign or international financing institution in accordance with the provisions of the 2016 revised Implementing Rules and Regulations of Republic Act No. 9184.

The Bidding Documents shall clearly and adequately define, among others: (i) the objectives, scope, and expected outputs and/or results of the proposed contract or Framework Agreement, as the case may be; (ii) the eligibility requirements of Bidders; (iii) the expected contract or Framework Agreement duration, the estimated quantity in the case of procurement of goods, delivery schedule and/or time frame; and (iv) the obligations, duties, and/or functions of the winning bidder.

Care should be taken to check the relevance of the provisions of the PBDs against the requirements of the specific Goods to be procured. If duplication of a subject is inevitable in other sections of the document prepared by the Procuring Entity, care must be exercised to avoid contradictions between clauses dealing with the same matter.

Moreover, each section is prepared with notes intended only as information for the Procuring Entity or the person drafting the Bidding Documents. They shall not be included in the final documents. The following general directions should be observed when using the documents:

- a. All the documents listed in the Table of Contents are normally required for the procurement of Goods. However, they should be adapted as necessary to the circumstances of the particular Procurement Project.
- b. Specific details, such as the "name of the Procuring Entity" and "address for bid submission," should be furnished in the Instructions to Bidders, Bid Data Sheet, and Special Conditions of Contract. The final documents should contain neither blank spaces nor options.
- c. This Preface and the footnotes or notes in italics included in the Invitation to Bid, Bid Data Sheet, General Conditions of Contract, Special Conditions of Contract, Schedule of Requirements, and Specifications are not part of the text of the final document, although they contain instructions that the Procuring Entity should strictly follow.
- d. The cover should be modified as required to identify the Bidding Documents as to the Procurement Project, Project Identification Number, and Procuring Entity, in addition to the date of issue.

- e. Modifications for specific Procurement Project details should be provided in the Special Conditions of Contract as amendments to the Conditions of Contract. For easy completion, whenever reference has to be made to specific clauses in the Bid Data Sheet or Special Conditions of Contract, these terms shall be printed in bold typeface on Sections I (Instructions to Bidders) and III (General Conditions of Contract), respectively.
- f. For guidelines on the use of Bidding Forms and the procurement of Foreign-Assisted Projects, these will be covered by a separate issuance of the Government Procurement Policy Board.

Table of Contents

| Glossa | ry of Acronyms, Terms, and Abbreviations | 4 |
|---------|--|----|
| Section | I. Invitation to Bid | 7 |
| Section | II. Instructions to Bidders | 11 |
| 1. | Scope of Bid | 12 |
| 2. | Funding Information | 12 |
| 3. | Bidding Requirements | 12 |
| 4. | Corrupt, Fraudulent, Collusive, and Coercive Practices | 12 |
| 5. | Eligible Bidders | 13 |
| 6. | Origin of Goods | 13 |
| 7. | Subcontracts | 13 |
| 8. | Pre-Bid Conference | 13 |
| 9. | Clarification and Amendment of Bidding Documents | 14 |
| 10. | Documents comprising the Bid: Eligibility and Technical Components | 14 |
| 11. | Documents comprising the Bid: Financial Component | 14 |
| 12. | Bid Prices | 14 |
| 13. | Bid and Payment Currencies | 15 |
| 14. | Bid Security | 15 |
| 15. | Sealing and Marking of Bids | 15 |
| 16. | Deadline for Submission of Bids | 16 |
| 17. | Opening and Preliminary Examination of Bids | 16 |
| 18. | Domestic Preference | 16 |
| 19. | Detailed Evaluation and Comparison of Bids | 16 |
| 20. | Post-Qualification | 17 |
| 21. | Signing of the Contract | 17 |
| Section | III. Bid Data Sheet | 18 |
| Section | IV. General Conditions of Contract | 21 |
| 1. | Scope of Contract | 22 |
| 2. | Advance Payment and Terms of Payment | 22 |
| 3. | Performance Security | 22 |
| 4. | Inspection and Tests | 22 |
| 5. | Warranty | 22 |
| 6. | Liability of the Supplier | 23 |
| Section | V. Special Conditions of Contract | 24 |
| Section | VI. Schedule of Requirements | 28 |
| Section | VII. Technical Specifications | 35 |
| Section | VIII. Checklist of Technical and Financial Documents | 38 |

Glossary of Acronyms, Terms, and Abbreviations

ABC – Approved Budget for the Contract.

BAC - Bids and Awards Committee.

Bid – A signed offer or proposal to undertake a contract submitted by a bidder in response to and in consonance with the requirements of the bidding documents. Also referred to as *Proposal* and *Tender*. (2016 revised IRR, Section 5[c])

Bidder – Refers to a contractor, manufacturer, supplier, distributor and/or consultant who submits a bid in response to the requirements of the Bidding Documents. (2016 revised IRR, Section 5[d])

Bidding Documents – The documents issued by the Procuring Entity as the bases for bids, furnishing all information necessary for a prospective bidder to prepare a bid for the Goods, Infrastructure Projects, and/or Consulting Services required by the Procuring Entity. (2016 revised IRR, Section 5[e])

BIR - Bureau of Internal Revenue.

BSP – Bangko Sentral ng Pilipinas.

Consulting Services – Refer to services for Infrastructure Projects and other types of projects or activities of the GOP requiring adequate external technical and professional expertise that are beyond the capability and/or capacity of the GOP to undertake such as, but not limited to: (i) advisory and review services; (ii) pre-investment or feasibility studies; (iii) design; (iv) construction supervision; (v) management and related services; and (vi) other technical services or special studies. (2016 revised IRR, Section 5[i])

CDA - Cooperative Development Authority.

Contract – Refers to the agreement entered into between the Procuring Entity and the Supplier or Manufacturer or Distributor or Service Provider for procurement of Goods and Services; Contractor for Procurement of Infrastructure Projects; or Consultant or Consulting Firm for Procurement of Consulting Services; as the case may be, as recorded in the Contract Form signed by the parties, including all attachments and appendices thereto and all documents incorporated by reference therein.

CIF - Cost Insurance and Freight.

CIP - Carriage and Insurance Paid.

CPI - Consumer Price Index.

DDP – Refers to the quoted price of the Goods, which means "delivered duty paid."

DTI – Department of Trade and Industry.

EXW - Ex works.

FCA - "Free Carrier" shipping point.

FOB - "Free on Board" shipping point.

Foreign-funded Procurement or Foreign-Assisted Project—Refers to procurement whose funding source is from a foreign government, foreign or international financing institution as specified in the Treaty or International or Executive Agreement. (2016 revised IRR, Section 5[b]).

Framework Agreement – Refers to a written agreement between a procuring entity and a supplier or service provider that identifies the terms and conditions, under which specific purchases, otherwise known as "Call-Offs," are made for the duration of the agreement. It is in the nature of an option contract between the procuring entity and the bidder(s) granting the procuring entity the option to either place an order for any of the goods or services identified in the Framework Agreement List or not buy at all, within a minimum period of one (1) year to a maximum period of three (3) years. (GPPB Resolution No. 27-2019)

GFI – Government Financial Institution.

GOCC - Government-owned and/or -controlled corporation.

Goods – Refer to all items, supplies, materials and general support services, except Consulting Services and Infrastructure Projects, which may be needed in the transaction of public businesses or in the pursuit of any government undertaking, project or activity, whether in the nature of equipment, furniture, stationery, materials for construction, or personal property of any kind, including non-personal or contractual services such as the repair and maintenance of equipment and furniture, as well as trucking, hauling, janitorial, security, and related or analogous services, as well as procurement of materials and supplies provided by the Procuring Entity for such services. The term "related" or "analogous services" shall include, but is not limited to, lease or purchase of office space, media advertisements, health maintenance services, and other services essential to the operation of the Procuring Entity. (2016 revised IRR, Section 5[r])

GOP – Government of the Philippines.

GPPB – Government Procurement Policy Board.

INCOTERMS – International Commercial Terms.

Infrastructure Projects – Include the construction, improvement, rehabilitation, demolition, repair, restoration or maintenance of roads and bridges, railways, airports, seaports, communication facilities, civil works components of information technology projects, irrigation, flood control and drainage, water supply, sanitation, sewerage and solid waste management systems, shore protection, energy/power and electrification facilities, national buildings, school buildings, hospital buildings, and other related construction projects of the government. Also referred to as *civil works or works*. (2016 revised IRR, Section 5[u])

LGUs - Local Government Units.

NFCC – Net Financial Contracting Capacity.

NGA – National Government Agency.

PhilGEPS - Philippine Government Electronic Procurement System.

Procurement Project – refers to a specific or identified procurement covering goods, infrastructure project or consulting services. A Procurement Project shall be described, detailed, and scheduled in the Project Procurement Management Plan prepared by the agency

which shall be consolidated in the procuring entity's Annual Procurement Plan. (GPPB Circular No. 06-2019 dated 17 July 2019)

PSA – Philippine Statistics Authority.

SEC – Securities and Exchange Commission.

SLCC – Single Largest Completed Contract.

Supplier – refers to a citizen, or any corporate body or commercial company duly organized and registered under the laws where it is established, habitually established in business and engaged in the manufacture or sale of the merchandise or performance of the general services covered by his bid. (Item 3.8 of GPPB Resolution No. 13-2019, dated 23 May 2019). Supplier as used in these Bidding Documents may likewise refer to a distributor, manufacturer, contractor, or consultant.

UN – United Nations.

Section I. Invitation to Bid

Notes on the Invitation to Bid

The Invitation to Bid (IB) provides information that enables potential Bidders to decide whether to participate in the procurement at hand. The IB shall be posted in accordance with Section 21.2 of the 2016 revised IRR of RA No. 9184.

Apart from the essential items listed in the Bidding Documents, the IB should also indicate the following:

- a. The date of availability of the Bidding Documents, which shall be from the time the IB is first advertised/posted until the deadline for the submission and receipt of bids;
- b. The place where the Bidding Documents may be acquired or the website where it may be downloaded;
- c. The deadline for the submission and receipt of bids; and
- d. Any important bid evaluation criteria (e.g., the application of a margin of preference in bid evaluation).

The IB should be incorporated in the Bidding Documents. The information contained in the IB must conform to the Bidding Documents and in particular to the relevant information in the Bid Data Sheet.

Section II. Instructions to Bidders

Notes on the Instructions to Bidders

This Section on the Instruction to Bidders (ITB) provides the information necessary for bidders to prepare responsive bids, in accordance with the requirements of the Procuring Entity. It also provides information on bid submission, eligibility check, opening and evaluation of bids, post-qualification, and on the award of contract.

1. Scope of Bid

The Procuring Entity, Quezon City Local Government wishes to receive Bids for the PROCUREMENT OF VARIOUS HOSPITAL EQUIPMENT with identification number QCGH-24-HME-1557.

[Note: The Project Identification Number is assigned by the Procuring Entity based on its own coding scheme and is not the same as the PhilGEPS reference number, which is generated after the posting of the bid opportunity on the PhilGEPS website.]

The Procurement Project (referred to herein as "Project") is composed of *three (3) line items*, the details of which are described in Section VII (Technical Specifications).

2. Funding Information

- 2.1. The GOP through the source of funding as indicated below for 2024 in the amount of ELEVEN MILLION NINE HUNDRED EIGHTY-TWO THOUSAND PESOS AND 00/100 ONLY (Php11,982,000.00).
- 2.2. The source of funding is:
 - a) LGUs, the Annual or Supplemental Budget, as approved by the Sanggunian.

3. Bidding Requirements

The Bidding for the Project shall be governed by all the provisions of RA No. 9184 and its 2016 revised IRR, including its Generic Procurement Manuals and associated policies, rules and regulations as the primary source thereof, while the herein clauses shall serve as the secondary source thereof.

Any amendments made to the IRR and other GPPB issuances shall be applicable only to the ongoing posting, advertisement, or **IB** by the BAC through the issuance of a supplemental or bid bulletin.

The Bidder, by the act of submitting its Bid, shall be deemed to have verified and accepted the general requirements of this Project, including other factors that may affect the cost, duration and execution or implementation of the contract, project, or work and examine all instructions, forms, terms, and project requirements in the Bidding Documents.

4. Corrupt, Fraudulent, Collusive, and Coercive Practices

The Procuring Entity, as well as the Bidders and Suppliers, shall observe the highest standard of ethics during the procurement and execution of the contract. They or through an agent shall not engage in corrupt, fraudulent, collusive, coercive, and obstructive practices defined under Annex "I" of the 2016 revised IRR of RA No. 9184 or other integrity violations in competing for the Project.

A book of

5. Eligible Bidders

- 5.1. Only Bids of Bidders found to be legally, technically, and financially capable will be evaluated.
- 5.2. Foreign ownership exceeding those allowed under the rules may participate pursuant to:
 - i. When a Treaty or International or Executive Agreement as provided in Section 4 of the RA No. 9184 and its 2016 revised IRR allow foreign bidders to participate;
 - ii. Citizens, corporations, or associations of a country, included in the list issued by the GPPB, the laws or regulations of which grant reciprocal rights or privileges to citizens, corporations, or associations of the Philippines;
 - iii. When the Goods sought to be procured are not available from local suppliers; or
 - iv. When there is a need to prevent situations that defeat competition or restrain trade.
- 5.3. Pursuant to Section 23.4.1.3 of the 2016 revised IRR of RA No.9184, the Bidder shall have an SLCC that is at least one (1) contract similar to the Project the value of which, adjusted to current prices using the PSA's CPI, must be at least equivalent to:
 - a. For the procurement of Non-Expendable Supplies and Services: The Bidder must have completed a single contract that is similar to this Project, equivalent to at least fifty percent (50%) of the ABC.
- 5.4. The Bidders shall comply with the eligibility criteria under Section 23.4.1 of the 2016 IRR of RA No. 9184.

6. Origin of Goods

There is no restriction on the origin of goods other than those prohibited by a decision of the UN Security Council taken under Chapter VII of the Charter of the UN, subject to Domestic Preference requirements under ITB Clause 18.

7. Subcontracts

7.1. The Bidder may subcontract portions of the Project to the extent allowed by the Procuring Entity as stated herein, but in no case more than twenty percent (20%) of the Project.

The Procuring Entity has prescribed that: Subcontracting is not allowed.

8. Pre-Bid Conference

A 600 4 24

The Procuring Entity will hold a pre-bid conference for this Project on the specified date and time and either at its physical address and/or through videoconferencing as indicated in paragraph 6 of the IB.

9. Clarification and Amendment of Bidding Documents

Prospective bidders may request for clarification on and/or interpretation of any part of the Bidding Documents. Such requests must be in writing and received by the Procuring Entity, either at its given address or through electronic mail indicated in the IB, at least ten (10) calendar days before the deadline set for the submission and receipt of Bids.

10. Documents comprising the Bid: Eligibility and Technical Components

- 10.1. The first envelope shall contain the eligibility and technical documents of the Bid as specified in Section VIII (Checklist of Technical and Financial Documents).
- 10.2. The Bidder's SLCC as indicated in **ITB** Clause 5.3 should have been completed within *the last three (3) years* prior to the deadline for the submission and receipt of bids.
- 10.3. If the eligibility requirements or statements, the bids, and all other documents for submission to the BAC are in foreign language other than English, it must be accompanied by a translation in English, which shall be authenticated by the appropriate Philippine foreign service establishment, post, or the equivalent office having jurisdiction over the foreign bidder's affairs in the Philippines. Similar to the required authentication above, for Contracting Parties to the Apostille Convention, only the translated documents shall be authenticated through an Apostille pursuant to GPPB Resolution No. 13-2019 dated 23 May 2019. The English translation shall govern, for purposes of interpretation of the bid.

11. Documents comprising the Bid: Financial Component

- 11.1. The second bid envelope shall contain the financial documents for the Bid as specified in Section VIII (Checklist of Technical and Financial Documents).
- 11.2. If the Bidder claims preference as a Domestic Bidder or Domestic Entity, a certification issued by DTI shall be provided by the Bidder in accordance with Section 43.1.3 of the 2016 revised IRR of RA No. 9184.
- 11.3. Any bid exceeding the ABC indicated in paragraph 1 of the IB shall not be accepted.
- 11.4. For Foreign-funded Procurement, a ceiling may be applied to bid prices provided the conditions are met under Section 31.2 of the 2016 revised IRR of RA No. 9184.

12. Bid Prices

- 12.1. Prices indicated on the Price Schedule shall be entered separately in the following manner:
 - a. For Goods offered from within the Procuring Entity's country:

- i. The price of the Goods quoted EXW (ex-works, ex-factory, exwarehouse, ex-showroom, or off-the-shelf, as applicable);
- ii. The cost of all customs duties and sales and other taxes already paid or payable;
- iii. The cost of transportation, insurance, and other costs incidental to delivery of the Goods to their final destination; and
- iv. The price of other (incidental) services, if any, listed in e.

b. For Goods offered from abroad:

- i. Unless otherwise stated in the **BDS**, the price of the Goods shall be quoted delivered duty paid (DDP) with the place of destination in the Philippines as specified in the **BDS**. In quoting the price, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible source country.
- ii. The price of other (incidental) services, if any, as listed in Section VII (Technical Specifications).

13. Bid and Payment Currencies

- 13.1. For Goods that the Bidder will supply from outside the Philippines, the bid prices may be quoted in the local currency or tradeable currency accepted by the BSP at the discretion of the Bidder. However, for purposes of bid evaluation, Bids denominated in foreign currencies, shall be converted to Philippine currency based on the exchange rate as published in the BSP reference rate bulletin on the day of the bid opening.
- 13.2. Payment of the contract price shall be made in:
 - a. Philippine Pesos.

14. Bid Security

- 14.1. The Bidder shall submit a Bid Securing Declaration¹ or any form of Bid Security in the amount indicated in the **BDS**, which shall be not less than the percentage of the ABC in accordance with the schedule in the **BDS**.
- 14.2. The Bid and bid security in no case shall exceed One Hundred Twenty (120) calendar days from the date of opening of bids, unless duly extended by the bidder upon the request of the Head of the Procuring Entity (HoPE) of the Quezon City Local Government. Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.

15. Sealing and Marking of Bids

Each Bidder shall submit one copy of the first and second components of its Bid.

¹ In the case of Framework Agreement, the undertaking shall refer to entering into contract with the Procuring Entity and furnishing of the performance security or the performance securing declaration within ten (10) calendar days from receipt of Notice to Execute Framework Agreement.

The Procuring Entity may request additional hard copies and/or electronic copies of the Bid. However, failure of the Bidders to comply with the said request shall not be a ground for disqualification.

If the Procuring Entity allows the submission of bids through online submission or any other electronic means, the Bidder shall submit an electronic copy of its Bid, which must be digitally signed. An electronic copy that cannot be opened or is corrupted shall be considered non-responsive and, thus, automatically disqualified.

16. Deadline for Submission of Bids

16.1. The Bidders shall submit on the specified date and time through manual submission as indicated in paragraph 7 of the IB.

17. Opening and Preliminary Examination of Bids

17.1. The BAC shall open the Bids in public at the time, on the date, and at the place specified in paragraph 9 of the **IB**. The Bidders' representatives who are present shall sign a register evidencing their attendance. In case videoconferencing, webcasting or other similar technologies will be used, attendance of participants shall likewise be recorded by the BAC Secretariat.

In case the Bids cannot be opened as scheduled due to justifiable reasons, the rescheduling requirements under Section 29 of the 2016 revised IRR of RA No. 9184 shall prevail.

17.2. The preliminary examination of bids shall be governed by Section 30 of the 2016 revised IRR of RA No. 9184.

18. Domestic Preference

18.1. The Procuring Entity will grant a margin of preference for the purpose of comparison of Bids in accordance with Section 43.1.2 of the 2016 revised IRR of RA No. 9184.

19. Detailed Evaluation and Comparison of Bids

- 19.1. The Procuring BAC shall immediately conduct a detailed evaluation of all Bids rated "passed," using non-discretionary pass/fail criteria. The BAC shall consider the conditions in the evaluation of Bids under Section 32.2 of the 2016 revised IRR of RA No. 9184.
- 19.2. If the Project allows partial bids, bidders may submit a proposal on any of the lots or items, and evaluation will be undertaken on a per lot or item basis, as the case maybe. In this case, the Bid Security as required by ITB Clause 15 shall be submitted for each lot or item separately.
- 19.3. The descriptions of the lots or items shall be indicated in **Section VII (Technical Specifications)**, although the ABCs of these lots or items are indicated in the **BDS** for purposes of the NFCC computation pursuant to Section 23.4.2.6 of the 2016 revised IRR of RA No. 9184. The NFCC must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder.
- 19.4. The Project shall be awarded as follows:

One Project having several items that shall be awarded as one contract.

19.5. Except for bidders submitting a committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation, all Bids must include the NFCC computation pursuant to Section 23.4.1.4 of the 2016 revised IRR of RA No. 9184, which must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder. For bidders submitting the committed Line of Credit, it must be at least equal to ten percent (10%) of the ABCs for all the lots or items participated in by the prospective Bidder.

20. Post-Qualification

20.1. Within a non-extendible period of five (5) calendar days from receipt by the Bidder of the notice from the BAC that it submitted the Lowest Calculated Bid, the Bidder shall submit its latest income and business tax returns filed and paid through the BIR Electronic Filing and Payment System (eFPS) and other appropriate licenses and permits required by law and stated in the BDS.

21. Signing of the Contract

21.1. The documents required in Section 37.2 of the 2016 revised IRR of RA No. 9184 shall form part of the Contract. Additional Contract documents are indicated in the **BDS**.

Section III. Bid Data Sheet

Notes on the Bid Data Sheet

The Bid Data Sheet (BDS) consists of provisions that supplement, amend, or specify in detail, information, or requirements included in the ITB found in Section II, which are specific to each procurement.

This Section is intended to assist the Procuring Entity in providing the specific information in relation to corresponding clauses in the ITB and has to be prepared for each specific procurement.

The Procuring Entity should specify in the BDS information and requirements specific to the circumstances of the Procuring Entity, the processing of the procurement, and the bid evaluation criteria that will apply to the Bids. In preparing the BDS, the following aspects should be checked:

- a. Information that specifies and complements provisions of the ITB must be incorporated.
- b. Amendments and/or supplements, if any, to provisions of the ITB as necessitated by the circumstances of the specific procurement, must also be incorporated.

Bid Data Sheet

| ITB | " | | | |
|--------|--|--|--|--|
| Clause | | | | |
| 5.3 | For this purpose, contracts similar to the | Project shall be: | | |
| | a. A single contract similar to the items to be bid and must be at least fifty percent (50%) of the ABC. | | | |
| | b. Completed within the last three (3) years prior to the deadline for the submission and receipt of bids substantially in a FORM prescribed by the QC-BAC-GOODS AND SERVICES, must be accompanied by a copy of Certificate of Acceptance by the end-user or Official Receipt (O.R) or Sales Invoice (S.I.) issued for the Contract. | | | |
| 7.1 | Subcontracting is not allowed. | | | |
| 12 | The price of the Goods shall be quote International Commercial Terms (INCO | d DDP within Quezon City or the applicable FERMS) for this Project. | | |
| 14.1 | The bid security shall be in the form of following forms and amounts: | of a Bid Securing Declaration, or any of the | | |
| | LINE 1 | | | |
| | | 96,000.00 or equivalent to two percent (2%) | | |
| | of ABC if bid security is in cash draft/guarantee or irrevocable let | , cashier's/manager's check, bank | | |
| | | 240,000.00 or equivalent to five percent (5%) | | |
| | of ABC if bid security is in Sure | ty Bond. | | |
| | LINE, 2 | LINE 2 | | |
| | | a. The amount of not less than Php50,000.00 or equivalent to two percent (2%) | | |
| | of ABC if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or | | | |
| | b. The amount of not less than Php125,000.00 or equivalent to five percent (5%) of ABC if bid security is in Surety Bond. | | | |
| | LINE 3 | | | |
| | a. The amount of not less than Php93,640.00 or equivalent to two percent (2%) of ABC if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or b. The amount of not less than Php234,100.00 or equivalent to five percent (5%) | | | |
| | of ABC if bid security is in Sure | ty Bond. | | |
| } | | | | |
| 19.3 | A BRITANIES DATE OF | T POD THE CONTRA CO | | |
| | LINE 1 | T FOR THE CONTRACT P4,800,000.00 | | |
| | LINE 2 | P 2,500,000.00 | | |
| | LINE 3 > | P 4,682,000.00 | | |
| | TOTAL | P 11,982,000.00 | | |
| 20.2 | List of required licenses and permits relevant to the Project and the corresponding law requiring it. | | | |
| | | nse to Operate from DOH Accreditation as Manufacturer for Medical or Hospital | | |
| 21.2 | Additional required documents relevant to the Project that are required by existing laws and/or the Procuring Entity. | | | |
| | 1 | | | |

EOB FINE I

- Statement of Warranty: Minimum of One (I) Year on parts and service (with Project
- parts for the next five (5) years and that parts shall be available at the Certification from the Manufacturer guaranteeing the availability of all spare Name and Project No.)
- authorized Philippine service center/s for a period of (5) years after the
- Certification that the bidder is an authorized or exclusive distributor of the warranty period.
- maintenance personnel of the hospital (with Project Name and Project No.) Certification that the bidder can provide applications training for users and brand/model being offered.

FOR LINE 2

- Statement of Warranty: Minimum of One (1) Year on parts and service (with Project
- Certification on the availability of spare parts for five (5) years after the warranty Name and Project No.)
- Certification that the bidder is an authorized or exclusive distributor of the

EOB TINE 3

brand/model being offered.

- Project Vame and Project No.) Statement of Warranty: Minimum of Three (3) years on parts and service (with
- Certification of Preventive Maintenance Services (PMS) with semi-annual
- Certification that the offered imaging equipment and its Ai and non-Ai schedule of visit up to three (3) years free of charge
- accessories of the offered imaging equipment in the next from (3) years and will not be Certification from the manufacturer ensuring the availability of spare parts and software are of the same brand.
- discontinued in the next free (5) years.
- for more than ten (10) years, to attach proof of documents. Certification from the manufacturer that the brand has been in the Philippine market
- Certification from the manufacturer of their valid ISO 13485:2016 (QMS for Medical
- ISO 9001:2015 (Quality Management System). Device Manufacturing), ISO 14001:2015 (Environmental Management Systems) and
- year of delivery and installation must be clearly specified on the list. Philippines of the specific model being offered. Hospital/client name, contact details, Certification from the manufacturer declaring a list of at least 150 installations in the
- berelto gnied (votoeleb land tall eselevia) inemqiupe gnigami Authority to sell from the manufacturer / authorized distributorship of the offered
- spare part is required to be purchased. Availability of service engineers 24/7 with guaranteed uptime mithin 24 hours unless

Section IV. General Conditions of Contract

Notes on the General Conditions of Contract

The General Conditions of Contract (GCC) in this Section, read in conjunction with the Special Conditions of Contract in Section V and other documents listed therein, should be a complete document expressing all the rights and obligations of the parties.

Matters governing performance of the Supplier, payments under the contract, or matters affecting the risks, rights, and obligations of the parties under the contract are included in the GCC and Special Conditions of Contract.

Any complementary information, which may be needed, shall be introduced only through the Special Conditions of Contract.

1. Scope of Contract

This Contract shall include all such items, although not specifically mentioned, that can be reasonably inferred as being required for its completion as if such items were expressly mentioned herein. All the provisions of RA No. 9184 and its 2016 revised IRR, including the Generic Procurement Manual, and associated issuances, constitute the primary source for the terms and conditions of the Contract, and thus, applicable in contract implementation. Herein clauses shall serve as the secondary source for the terms and conditions of the Contract.

This is without prejudice to Sections 74.1 and 74.2 of the 2016 revised IRR of RA No. 9184 allowing the GPPB to amend the IRR, which shall be applied to all procurement activities, the advertisement, posting, or invitation of which were issued after the effectivity of the said amendment.

Additional requirements for the completion of this Contract shall be provided in the Special Conditions of Contract (SCC).

2. Advance Payment and Terms of Payment

- 2.1. Advance payment of the contract amount is provided under Annex "D" of the revised 2016 IRR of RA No. 9184.
- 2.2. The Procuring Entity is allowed to determine the terms of payment on the partial or staggered delivery of the Goods procured, provided such partial payment shall correspond to the value of the goods delivered and accepted in accordance with prevailing accounting and auditing rules and regulations. The terms of payment are indicated in the SCC.

3. Performance Security

Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than prior to the signing of the Contract by both parties, the successful Bidder shall furnish the performance security in any of the forms prescribed in Section 39 of the 2016 revised IRR of RA No. 9184.

4. Inspection and Tests

The Procuring Entity or its representative shall have the right to inspect and/or to test the Goods to confirm their conformity to the Project specifications at no extra cost to the Procuring Entity in accordance with the Generic Procurement Manual. In addition to tests in the SCC, Section IV (Technical Specifications) shall specify what inspections and/or tests the Procuring Entity requires, and where they are to be conducted. The Procuring Entity shall notify the Supplier in writing, in a timely manner, of the identity of any representatives retained for these purposes.

All reasonable facilities and assistance for the inspection and testing of Goods, including access to drawings and production data, shall be provided by the Supplier to the authorized inspectors at no charge to the Procuring Entity.

5. Warranty

6.1. In order to assure that manufacturing defects shall be corrected by the Supplier, a warranty shall be required from the Supplier as provided under Section 62.1 of the 2016 revised IRR of RA No. 9184.

6.2. The Procuring Entity shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall, repair or replace the defective Goods or parts thereof without cost to the Procuring Entity, pursuant to the Generic Procurement Manual.

6. Liability of the Supplier

The Supplier's liability under this Contract shall be as provided by the laws of the Republic of the Philippines.

If the Supplier is a joint venture, all partners to the joint venture shall be jointly and severally liable to the Procuring Entity.

Section V. Special Conditions of Contract

Notes on the Special Conditions of Contract

Similar to the BDS, the clauses in this Section are intended to assist the Procuring Entity in providing contract-specific information in relation to corresponding clauses in the GCC found in Section IV.

The Special Conditions of Contract (SCC) complement the GCC, specifying contractual requirements linked to the special circumstances of the Procuring Entity, the Procuring Entity's country, the sector, and the Goods purchased. In preparing this Section, the following aspects should be checked:

- a. Information that complements provisions of the GCC must be incorporated.
- b. Amendments and/or supplements to provisions of the GCC as necessitated by the circumstances of the specific purchase, must also be incorporated.

However, no special condition which defeats or negates the general intent and purpose of the provisions of the GCC should be incorporated herein.

Special Conditions of Contract

| GCC Clause | |
|---------------|---|
| 1 | [List here any additional requirements for the completion of this Contract. The following requirements and the corresponding provisions may be deleted, amended, or retained depending on its applicability to this Contract:] |
| | Delivery and Documents – |
| | For purposes of the Contract, "EXW," "FOB," "FCA," "CIF," "CIP," "DDP" and other trade terms used to describe the obligations of the parties shall have the meanings assigned to them by the current edition of INCOTERMS published by the International Chamber of Commerce, Paris. The Delivery terms of this Contract shall be as follows: |
| | [For Goods supplied from abroad, state:] "The delivery terms applicable to the Contract are DDP delivered [indicate place of destination]. In accordance with INCOTERMS." |
| | [For Goods supplied from within the Philippines, state:] "The delivery terms applicable to this Contract are delivered [indicate place of destination]. Risk and title will pass from the Supplier to the Procuring Entity upon receipt and final acceptance of the Goods at their final destination." |
| | Delivery of the Goods shall be made by the Supplier in accordance with the terms specified in Section VI (Schedule of Requirements). |
| | For purposes of this Clause the Procuring Entity's Representative at the Project Site is <i>[indicate name(s)]</i> . |
| | Incidental Services – |
| | The Supplier is required to provide all of the following services, including additional services, if any, specified in Section VI. Schedule of Requirements: Select appropriate requirements and delete the rest. |
| | a. performance or supervision of on-site assembly and/or start-up of the supplied Goods; |
| | b. furnishing of tools required for assembly and/or maintenance of the supplied Goods; |
| | c. furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied Goods; |
| | d. performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and |
| | e. training of the Procuring Entity's personnel, at the Supplier's plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied Goods. f. [Specify additional incidental service requirements, as needed.] |
| | The Contract price for the Goods shall include the prices charged by the Supplier for incidental services and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services. |

Spare Parts -

The Supplier is required to provide all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the Supplier:

Select appropriate requirements and delete the rest.

- a. such spare parts as the Procuring Entity may elect to purchase from the Supplier, provided that this election shall not relieve the Supplier of any warranty obligations under this Contract; and
- b. in the event of termination of production of the spare parts:
 - i. advance notification to the Procuring Entity of the pending termination, in sufficient time to permit the Procuring Entity to procure needed requirements; and
 - ii. following such termination, furnishing at no cost to the Procuring Entity, the blueprints, drawings, and specifications of the spare parts, if requested.

The spare parts and other components required are listed in **Section VI** (**Schedule of Requirements**) and the cost thereof are included in the contract price.

The Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spare parts or components for the Goods for a period of [indicate here the time period specified. If not used indicate a time period of three times the warranty period].

Spare parts or components shall be supplied as promptly as possible, but in any case, within [insert appropriate time period] months of placing the order.

Packaging -

The Supplier shall provide such packaging of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in this Contract. The packaging shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit, and open storage. Packaging case size and weights shall take into consideration, where appropriate, the remoteness of the Goods' final destination and the absence of heavy handling facilities at all points in transit.

The packaging, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified below, and in any subsequent instructions ordered by the Procuring Entity.

The outer packaging must be clearly marked on at least four (4) sides as follows:

Name of the Procuring Entity Name of the Supplier

Contract Description Final Destination Gross weight Any special lifting instructions Any special handling instructions Any relevant HAZCHEM classifications A packaging list identifying the contents and quantities of the package is to be placed on an accessible point of the outer packaging if practical. If not practical the packaging list is to be placed inside the outer packaging but outside the secondary packaging. Transportation -Where the Supplier is required under Contract to deliver the Goods CIF, CIP, or DDP, transport of the Goods to the port of destination or such other named place of destination in the Philippines, as shall be specified in this Contract, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract Price. Where the Supplier is required under this Contract to transport the Goods to a specified place of destination within the Philippines, defined as the Project Site, transport to such place of destination in the Philippines, including insurance and storage, as shall be specified in this Contract, shall be arranged by the Supplier, and related costs shall be included in the contract price. Where the Supplier is required under Contract to deliver the Goods CIF, CIP or DDP, Goods are to be transported on carriers of Philippine registry. In the event that no carrier of Philippine registry is available, Goods may be shipped by a carrier which is not of Philippine registry provided that the Supplier obtains and presents to the Procuring Entity certification to this effect from the nearest Philippine consulate to the port of dispatch. In the event that carriers of Philippine registry are available but their schedule delays the Supplier in its performance of this Contract the period from when the Goods were first ready for shipment and the actual date of shipment the period of delay will be considered force majeure. The Procuring Entity accepts no liability for the damage of Goods during transit other than those prescribed by INCOTERMS for DDP deliveries. In the case of Goods supplied from within the Philippines or supplied by domestic Suppliers risk and title will not be deemed to have passed to the Procuring Entity until their receipt and final acceptance at the final destination. Intellectual Property Rights -The Supplier shall indemnify the Procuring Entity against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the Goods or any part thereof. 2.2 [If partial payment is allowed, state] "The terms of payment shall be as follows: 4 The inspections and tests that will be conducted are: Product Presentation/Demonstration/Site Inspection, if applicable.

Section VI. Schedule of Requirements

PROJECT NAME: LINE 1: ANESTHESIA MACHINE WITH PATIENT MONITOR AND BRAIN FUNCTION AND CEREBRAL OXIMETRY MONITORING PROJECT NO. QCGH-24-HME-1557

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

| Item Number | Description | Unit of Issue | Quantity | Delivered, Weeks/ Months |
|----------------|--|------------------|----------|--|
| 1 | Anesthesia Machine with Patient Monitor and Brain Function and Cerebral Oximetry Monitoring At least 15-inch touch screen monitor with adjustable brightness Weight: not exceeding 150 kg (without vaporizers and | Unit | 1 | TYZOATEATO |
| | cylinders) Top Shelf: Weight limit: 25-35kg, Width: 255-355mm, Length: 500-550mm Work Surface: Height: 800-900mm, Area 1500-1650cm2 Pneumatic-driven, electronically controlled anesthetic ventilator | | | |
| | Ventilator Specifications: Modes of Ventilation: Manual/ Spontaneous Ventilation/Bypass, Volume Control Ventilation (VCV) with PLV function. Pressure Control Ventilation (PCV) with/without volume guarantee (VG), Synchronized Intermittent Mandatory Ventilation (SIMV-Volume Controlled and SIMV-Pressure Controlled), Pressure Support Ventilation (PS) with apnea backup, Synchronized Intermittent Mandatory Ventilation Volume Guarantee (SIMV-VG), Continuous Positive Airway Pressure/Pressure Support Ventilation (CPAP/PS) with 2 vaporizers selectatec with interlock (Sevoflurane and Isoflurane) Monitoring and Alarm Monitoring part must be able to all setting and alarm parameters (including Breath rate, I/E ratio, Tidal volume, Minute volume, PEEP, MEAN, PEAK, PLAT, and O2 concentration, EtCO2, N2O, Aesthesia gas concentration) Ventilator Performance: Peak gas flow: 100L/min plus fresh gas, Flow valve range: 1- 100 L/min Plug-and-play modular CO2 monitoring Trolley with central brake and cable pusher on each caster | | | Within Ninety (90) Calendar Days upon issuance of Notice to Proceed |
| | Battery backup: at least 60 min, Battery type: Build in Li-ion battery, 11.1 VDC, 4400 mAh Pneumatic Specifications Switching Auxiliary Common Gas Outlet (ACGO) Connector: ISO 22mm OD and 15 mm ID the outlet locates at the inspiratory limb. Gas Supply pipeline input range: 0.28~0.6mPa, Pipeline connections: NIST or DISS Cylinder input: PISS APL valve range 1~75 cmH20, with tactile knob indication at: >30 cm H20 Breathing Circuit Specifications | | | |

Sytem Pressure Gauge: Range: -20~100 cmH20, Accuracy: +/-2.5% full scale, Bag/Mechanical Ventilation Switch: Type: Bistable **Breathing Circuit Parameters** Compliance: <4mL/100Pa Automatically compensates for compression loss with in the breathing circuit on mechanical mode, Expiration resistance: < 6cm H20 @ 60 L/min, Inspiration resistance: <6cm H20 @ 60 L/min **Brain Function and Cerebral Oximetry Monitoring** Equipment with Non Invasive Pulse co-oximeter Specifications: 1. Base Unit a. Multi-modal patient monitoring platform b. Can measure and display the following parameters i. Brain function (i.e. processed EEG)" "ii. Regional Oximetry iii. Non-invasive pulse oximetry Display is at least 1280 x 800 pixels, and at least 10.1inch diagonal" "Multi-touch screen and customizable with multiple screen i. Can display tissue oxygen saturation (%rSO2) of specific area (i.e. forehead) ii. Can display the difference between current rSO2 and userdefined baseline iii. Can display the difference between sPO2 and rSO2" "iv. Can display four channels of bilateral EEG waveforms v. Can display a value of processed EEG parameter related to the effect of anesthetic agents vi. Can display a value of spectrogram representing activity in both sides of the brain. e. Must have ports or docks that are compatible to plug-andplay modules for the measurement of brain function monitoring, regional oximetry, and non-invasive pulse oximetry" at least four hours' battery capacity and maximum of four hours charging time 2. Modules: a. One (1) Module for brain function monitoring/Unit with patient cable b. One (1) Module for Cerebral Oximetry Monitoring c. One (1) Module for Non-Invasive Pulse Oximetry i. Can be used as a bedside, handheld, or transport monitor ii. Has high-definition multi-touch customizable display that can show oxygen saturation (sPO2), pulse rate, perfusion index, pleth variability index, and high resolution plethysmographic waveform

| I hereby certify to comply | and deliver all the | above requirements. |
|----------------------------|---------------------|---------------------|
| | | |

| Name: | |
|---|--|
| Legal Capacity: | |
| Signature: | |
| Duly authorized to sign the Bid for and behalf of | |

Q

Section VI. Schedule of Requirements

PROJECT NAME: LINE 2: ANESTHESIA MACHINE WITH PATIENT MONITOR PROJECT NO. QCGH-24-HME-1557

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

| Electronically Controlled and pneumatically driven with rising bellow, Maximum Machine Weight must not exceed 128kg. Workbench supporting max weight capacity of 20kg, Pop Pate supporting max weight capacity of 20kg, Working Temperature of 10-487 degrees C. = /-93% of Humidity, Input Voltage of 100-240V, Input Current of 3.8-8.5A. Input Frequency of 50/601z +/- 11kz, 7000mAn 11.1/DC. Automatic Breathing Circuit Compliance at -/-64ml/100pa and Automatically compensates for cipcuit compression loss within the breathing circuit in mechanical mode, Tidal Volume range of VCV [45-1500 ml and PCV 5-1500 ml, 2000 Rechargeable Lithium-lon Battery with four hours charging time and at least 2 fours continuous working operation time Gas Source (Oxygen and AIR) with Electronic Flowmeter Oxygen Flush switch gives 100% fast oxygen Pressure Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isofftrane) ACOO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Govern Air, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Gricuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSVPR, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro, Ventilation Principle is Chronometric, volumetric and barometric, Electronically controlled and Pneumatically Driven, Electronically controlled and Pneumatically Driven, Electronic Selective Air or Oxygen | Item Number | Description | Unit of Issue | Quantity | Delivered, Weeks/ Months |
|---|----------------|--|------------------|----------|--------------------------------|
| rising bel/bw, Maximum Machine Weight must not exceed 128kg, Workbench supporting max weight capacity of 20kg, Top Plate supporting max weight capacity of 20kg, Working Temperature of 10-467 degrees C. → 93% of Humidity, Input Voltage of 100-240V, Input Current of 3.58-85, Al prut Frequency of 50/60Hz +/- 1Hz, 7000mAh 11.1VDC, Automatic Breathing Circuit Compliance at → (4ml/100pa and Automatically compensates for cipauit compression loss within the breathing circuit in mechanical mode, Tidal Volume range of VCV 1≥-1500 ml and PCV 5-1500 ml, 2600 Rechargeable Lithium-Ion Battery with four hours charging time and at least 2 flours continuous working operation time Gas Source (Oxygen and AIR) with Electronic Flowmeter Oxygen Flush switch gives 100% fast oxygen Pressure Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isoflurane) ACØO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Gauge, Manual Bag Port, Bellowa, Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-VC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro, Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | 1 | Anesthesia Machine with Patient Monitor | Unit | _1 | |
| rising bel/bw, Maximum Machine Weight must not exceed 128kg, Workbench supporting max weight capacity of 20kg, Top Plate supporting max weight capacity of 20kg, Working Temperature of 10-467 degrees C. → 93% of Humidity, Input Voltage of 100-240V, Input Current of 3.58-85, Al prut Frequency of 50/60Hz +/- 1Hz, 7000mAh 11.1VDC, Automatic Breathing Circuit Compliance at → (4ml/100pa and Automatically compensates for cipauit compression loss within the breathing circuit in mechanical mode, Tidal Volume range of VCV 1≥-1500 ml and PCV 5-1500 ml, 2600 Rechargeable Lithium-Ion Battery with four hours charging time and at least 2 flours continuous working operation time Gas Source (Oxygen and AIR) with Electronic Flowmeter Oxygen Flush switch gives 100% fast oxygen Pressure Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isoflurane) ACØO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Gauge, Manual Bag Port, Bellowa, Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-VC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro, Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| 128kg, Workbench supporting max weight capacity of 20kg, Top Plate supporting max weight capacity of 20kg, Working Temperature of 10-40f degrees C, =6-39% of Humidity, Input Voltage of 100-240V, Input Current of 3.5-8.5.A., Input Frequency of 50/60Hz +/- Hlz, 7000mAh 11.1VDC, Automatic Breathing Circuit Compliance at =1-64ml/100pa and Automatically compensates for ciguit compression loss within the breathing circuit in mechanical punde, Tidal Volume range of VCV 15-1500 ml and PCV 5-1500 ml, 2600 Rechargeable Lithium-Ion Battery with four hours charging time and at least 2 hours continuous working operation time Gas Source (Oxygen and AIR) with Electronic Flowmeter Oxygen Flush switch gives 100% fast oxygen Pressure Regulation With Total Flowmeter With Back-up Flow Control System Button With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevofiurane and Isoffurane) ACCO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O (with Minimum opening pressure of 0.3/0.5 cmH2O (dryhumid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-VC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPC, Petro, Ventilation Principle is Chronometric, volumetric and barometric, Electronics Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | · | | | | |
| Top Plate supporting max weight capacity of 20kg, Working Temperature of 10-40*degrees C, =/-93% of Humidity, Input Voltage of 100-240V, Input Current of 3.5-8.5A. Input Frequency of 50/60Hz +/- HTz, 7000mAh 11.1VDC, Automatic Breathing Circuit Compliance at =/4ml/100pa and Automatically compensates for circuit compression loss within the breathing circuit fompliance at =/4ml/100pa and Automatically compensates for circuit compression loss within the breathing circuit in mechanical mode, Tidal Volume range of VCV 15-1500 ml, 2000 Rechargeable Lithium-Ion Battery with four hours charging time and at least 2 flours continuous working operation time Gas Source (Oxygen and AIR) with Electronic Flowmeter Oxygen Flush switch gives 100% fast oxygen Pressure Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isoflurane) Accoff (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmpter Autoclavable Breathing System must have the following: Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Gircuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro, Ventilation Principle is Chronometric, volumetric and bărometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen | | | | | |
| Temperature of 10-4d'degrees C, =/-93% of Humidity, Input Voltage of 100-240V, Input Current of 3-8-8.5A, Input Frequency of 50/601z +/- Htz, 7000mAh 11.1VDC, Automatic Breathing Circuit Compliance at =/ <dmi (auxiliary="" (dry="" (oxygen="" (peep)="" (sevoflurane="" 0.30.5="" 1-75="" 100%="" 100pa="" 15mm="" 2="" 22mm="" 2600="" 5-1500="" [3-1500="" absorbent,="" accommodates="" acgo="" air="" air)="" airway="" and="" apl="" arm,="" assembly,="" at="" autoclavable="" automatic="" automatically="" auxiliary="" back-up="" bag="" barometric,="" battery="" bellows="" breathing="" by-pass="" canister="" charging="" check="" chronometric,="" circuit="" cmh2o="" coaxial="" common="" compensates="" compression="" connector="" continuous="" control="" controlled="" cpap,="" cq2="" drain="" driven,="" electronic="" electronically="" end="" expiration="" expiratory="" fast="" fitting="" flow="" flowmeter="" flush="" following:="" for="" four="" function="" gas="" gauge,="" gives="" have="" hold="" hook,="" hours="" humid),="" in="" inside="" inspiration="" inspiratory="" integrated,<="" interlocking="" is="" isofturane)="" knob="" leak="" least="" lever,="" lithium-ion="" loss="" manual="" mechanical="" minimum="" ml="" ml,="" mode,="" modes:="" must="" of="" opening="" operation="" or="" outlet)="" outside="" oxygen="" pcv="" plug,="" pneumatically="" port="" port,="" positive="" pressure="" principle="" prvc,="" psv="" psv-pro.="" range="" rechargeable="" regulation="" release="" selective="" simv-pc,="" simv-prvc,="" sodalime="" source="" supply="" switch="" switch,="" system="" taper="" td="" test="" the="" tidal="" time="" two="" type="" using="" valve="" valve,="" vaporizers="" vc,="" vcv="" ventilation="" ventilation,="" ventilator="" volume="" volumetric="" vpc,="" with="" within="" working=""><td></td><td></td><td></td><td></td><td></td></dmi> | | | | | |
| Voltage of 100-240V, Input Current of 3.5-8.5A, Input Frequency of 50/60Hz +/- HHz, 7000mAh 11.1VDC, Automatic Breathing Circuit Compliance at =/ Amil/100pa and Automatically compensates for circuit compression loss within the breathing circuit in mechanical mode, Tidal Volume range of VCV 15-1500 ml, 2600 Rechargeable Lithium-lon Battery with four hours charging time and at least 2 hours continuous working operation time Gas Source (Oxygen and AIR) with Electronic Flowmeter Oxygen Flush switch gives 100% fast oxygen Pressure Regulation With Total Flowmeter With Back-up Flow Control System Button With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isoflurane) ACAO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Air Flowmeter Autoclavable Breathing System must have the following: Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| Frequency of 50/60 Iz +/- IHz, 7000mAh 11.1VPC, Automatic Breathing Circuit Compliance at =/ <am (auxiliary="" (dry="" (oxygen="" (peep)="" (sevoflurane="" 0.30.5="" 1-75="" 100%="" 100pa="" 12-1500="" 15mm="" 2="" 2000="" 22mm="" 5-1500="" absorbent,="" acco="" accommodates="" air="" air)="" airway="" and="" apl="" arm,="" assembly,="" at="" autoclavable="" automatic="" automatically="" auxiliary="" back-up="" bag="" barometric,="" battery="" bellows="" breathing="" button="" by-pass="" canister="" carnister="" charging="" check="" chronometric,="" circuit="" cmh2c="" cmh2o="" co2="" coaxial="" common="" compensates="" compression="" connector="" continuous="" control="" controlled="" cpap,="" drain="" driven,="" electronic="" electronically="" end="" expiration="" expiratory="" expisation="" fast="" fitting="" flow="" flowmeter="" flush="" following;="" for="" four="" fours="" function="" gas="" gauge,="" gives="" have="" hold="" hook,="" hours="" humid),="" in="" inside="" inspiration="" inspiratory="" integrated,<="" interlocking="" is="" isoffurane)="" knob="" leak="" least="" lever,="" lithium-ion="" loss="" manual="" mechanical="" minimum="" ml="" ml,="" mode,="" modes:="" must="" of="" opening="" operation="" or="" outlet)="" outside="" oxygen="" pcv="" plug,="" pneumatically="" port="" port,="" positive="" pressure="" principle="" prvc,="" psv="" psvpro.="" range="" rechargeable="" regulation="" release="" selective="" simv-pc,="" simv-prvc,="" sodalime="" source="" supply="" switch="" switch,="" system="" taper="" td="" test="" the="" tidal="" time="" total="" two="" type="" using="" valve="" valve,="" vaporizers="" vc,="" vcv="" ventilation="" ventilation,="" ventilator="" volume="" volumetric="" vpc,="" with="" within="" working=""><td></td><td></td><td></td><td></td><td></td></am> | | | | | |
| Automatic Breathing Circuit Compliance at =/<4ml/100pa and Automatically compensates for circuit compression loss within the breathing circuit in mechanical mode, Tidal Volume range of VCV [x-1500 ml and PCV 5-1500 ml, 2800 Rechargeable Lithium-Ion Battery with four hours charging time and at least 2 fours continuous working operation time Gas Source (Oxygen and AIR) with Electronic Flowmeter Oxygen Flush switch gives 100% fast oxygen Pressure Regulator With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isoflurane) ACCO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CQ2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Methanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| and Automatically compensates for circuit compression loss within the breathing circuit in mechanical mode, Tidal Volume range of VCV [3-150 ml and PCV 5-150 ml, 2000 Rechargeable Lithium-Ion Battery with four hours charging time and at least 2 fours continuous working operation time Gas Source (Oxygen and AIR) with Electronic Flowmeter Oxygen Flush switch gives 100% fast oxygen Pressure Regulator With Total Flowmeter With Back-up Flow Control System Button With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isoflurane) ACGO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following: Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Provingle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| within the breathing circuit in mechanical mode, Tidal Volume range of VCV Iz-1500 ml and PCV 5-1500 ml, 2000 Rechargeable Lithium-Ion Battery with four hours charging time and at least 2 fours continuous working operation time Gas Source (Oxygen and AIR) with Electronic Flowmeter Oxygen Flush switch gives 100% fast oxygen Pressure Regulation With Total Flowmeter With Back-up Flow Control System Button With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isoflurane) ACCÓ (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Auxiliary Oxygen Supply Flowmeter Auxiliary Oxygen Supply Flowmeter Auxiliary Oxygen Supply Flowmeter Auxiliary Air Flowmeter Auxiliary Air Flowmeter Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Auxiliary Air Flowmeter Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Gauge, Manual Bag Port, Eellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Provingle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | · · · · · · · · · · · · · · · · · · · | | | |
| Volume range of VCV 1,5-1500 ml and PCV 5-1500 ml, 2,800 Rechargeable Lithium-Ion Battery with four hours charging time and at least 2 hours continuous working operation time Gas Source (Oxygen and AIR) with Electronic Flowmeter Oxygen Flush switch gives 100% fast oxygen Pressure Regulation With Total Flowmeter With Back-up Flow Control System Button With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isoflurane) ACCO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CAPA, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro, Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | l | | | |
| Rechargeable Lithium-Ion Battery with four hours charging time and at least 2 Kours continuous working operation time Gas Source (Oxygen and AIR) with Electronic Flowmeter Oxygen Flush switch gives 100% fast oxygen Pressure Regulator With Total Flowmeter With Back-up Flow Control System Button With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isoffurane) ACAO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Air Flowmeter Auxiliary Air Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | : | | | | |
| time and at least 2 hours continuous working operation time Gas Source (Oxygen and AIR) with Electronic Flowmeter Oxygen Flush switch gives 100% fast oxygen Pressure Regulation With Total Flowmeter With Back-up Flow Control System Button With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isoflurane) ACO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O (with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| Gas Source (Oxygen and AIR) with Electronic Flowmeter Oxygen Flush switch gives 100% fast oxygen Pressure Regulator With Total Flowmeter With Back-up Flow Control System Button With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isofturane) ACCO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| Oxygen Flush switch gives 100% fast oxygen Pressure Regulator With Total Flowmeter With Back-up Flow Control System Button With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isoflurane) ACCO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| Oxygen Flush switch gives 100% fast oxygen Pressure Regulator With Total Flowmeter With Back-up Flow Control System Button With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isoflurane) ACCO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | with Electronic Flowmstor | | | |
| Pressure Regulator With Total Flowmeter With Back-up Flow Control System Button With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isofturane) ACCO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Carnister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| With Total Flowmeter With Back-up Flow Control System Button With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isofturane) ACCO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVP70. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| With Back-up Flow Control System Button With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isofturane) ACCO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isofkurane) ACCO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isofturane) ACCO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | _ · · · · · · · · · · · · · · · · · · · | | | _ |
| Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isoflurane) ACCO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | _ | | | |
| (Sevoflurane and Isoflurane) ACEO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | 1 7 | | | |
| ACCO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | - · | | | |
| ACGO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | l = | | | |
| Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| Expiration Port, Canister for CO2 sodalime absorbent, Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | T | | | |
| Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | 1 | | | |
| Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| By-Pass Function Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | • · · · · · · · · · · · · · · · · · · | | | |
| PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | Vantilator Modes: VCV/VC PCV/VPC SIMV_VC SIMV_PC | | | |
| SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | l / | | | |
| volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | | | | |
| Pneumatically Driven, Electronic Selective Air or Oxygen Positive End Expiratory Pressure (PEEP) Type Integrated, | | 1 | | | |
| | | _ | | | |
| | | Positive End Evnisatory Processes (DEED) Toma Interest | | | |
| | 1 | electronically controlled, Range 0~70 cmH2O | | | |

30

At least 12.1" 800 x 600 Resolution Touchscreen Ventilator and Monitoring Main Display Screen

At least two Monitoring Module Slots with Anesthetic Gas (Ag) module with Bi Spectral Index (BJS) module with ETCO2 module (capnográphy)

Module must be compatible with the Patient Monitor AGSS (Anesthesia Gas Scavenging System) must have the following; Waste gas exhaust noozle connector, AGSS waste gases outlet, Outer cone connector for hose of transfer system, Pressure Compensation Port, Main Body of AGSS System, Red Color Float, and Flow regulation Knob, Suction flow rate 50-80LPM, Stainless steel mesh filter with pore size of 60-100um

4 pcs of Caster Wheel with Central Brake System With at least two Storage Drawer with lock and key With Amrest/Handle for easy maneuvering to and from Operating Room

At least four Auxiliary power output socket with individual socket breaker with output voltage of 100-240V, output frequency of 50/60Hz.

Must have comprehensive ventilation mode such as the following; Manual (Induction), VCV, PCV, and PRCV (Maintaining Anethesia), SIMV-VC, SIMV-PC, and SIMV-PRVC (Anesthesia Recovery), PSV, and PSVPro (Practice of Spontaneous Breathing)

Must have at least 10 seconds shutdown delay Must have a Ventilator Operation Risk Reminder/Prompt Must have a central lock braking system

Patient Monitor:

Main Display Monitor of not less than 12" TFT Touchscreen and Navigation Knob with 3 built in module slots for various monitoring parameters, must have a built in thermal printer, must not exceed 5kg, 360' visible (Dual Alarm) Indicator Light

Must have fixed hidden handle, must have Al-Mg alloy stent design, Easy to maintain, must have these six conventional monitoring parameters of 5-Lead ECG (optional 3/12-lead) and provides 26 arrhythmia analysis, Heart Rate, Respiration, Dual-Temperature that can measure skin and oral/anal temperature simultaneously, SpO2 provides isports and anti-low perfusion, AcuTec NIBP High Accuracy technology for Hypertension. Power Supply 100-240V, 50/60Hz +/-1Hz, must use 4400mAh Lithium-Ion rechargeable battery with charging time of 6 hours and 3-6 hours continuous battery back up

Various User Interface such as Standard Interface, Big Font, Trendm OxyCRG, List, Bed to Bed and 7-Lead ECG/12-Lead Full Screen Cascade ECG

Date storage Alarm Event Recall: 200 groups, Wave Recall: 2 hours, NIBP Recall: 2000 groups, Trend Graph: 120hours, Trend Table: 120hours, ARR Event: 100 groups

Recorder Type: Built-in; Thermal array, Channel: 3 channel waveforms, Speed: 25mm/s,50mm/s, Record width: 50mm, External printer: Yes

Respiration Method: RA-LL Impedance Method, RR measurement range: (Adult: 0-120rpm, Pediatric/Neonate: 0 - 150bpm), Accuracy: 7-150rpm ±2rpm or ±2% (whichever



| isgreater), 0-6rpm: unspecified, RESP Apnea Alarm: Adult: | | |
|---|--|---|
| 10s-60s Ped/Neo: 10s~20s, Alarm: Audible and visual alarm; | | |
| alarm events reviewable, Sweep Speed: 6.25,12.5, 25 mm/s, | | |
| Gain Selection: X0.25, X0.5, X1, X2, x4 | | |
| SpO2 Measurement & alarm range: 0~100%, Resolution: 1%, | | |
| Accuracy: ±2% (70~100%, Ped/Adu, non-motion), ±3% (70- | | |
| 100%, Neo, non-motion); 0-69% unspecified, | | |
| PR Measurement Range: 25250bpm, Accuracy: ±1bpm, | | |
| Alarm range: 20~350bpm, | | |
| Temperature (Dual Channel) Measurement & alarm range: | | |
| 0~50°C, XEMP sensor: Standard configuration-skin TEMP | | |
| sensor, Resolution: 0.1°C, Accuracy: ±0.1°C (exclusive of | | |
| error of sensor), Channel type: T1, T2, TD (Temperature | | |
| Difference) | | |
| Dual IBP Module | | |
| *** | | |
| | | _ |

I hereby certify to comply and deliver all the above requirements.

| Name: | · · · · · · · · · · · · · · · · · · · |
|--|---|
| Legal Capacity: | |
| Signature: | |
| Duly authorized to sign the Bid for and behalf of: | |

Schedule of Requirements Page 3 of 3 LINE 2: QCGH-24-HME-1557

y my

Section VI. Schedule of Requirements

PROJECT NAME: LINE 3: SUPPLY AND DELIVERY OF X-RAY FLAT PANEL DIGITIZER WITH DIGITAL RADIOGRAPHY SYSTEM PROJECT NO. QCGH-24-HME-1557-L3

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

| Item Number | Description | Unit of Issue | Quantity | Delivered, Weeks/ Months |
|----------------|--|------------------|----------|---|
| 1 | Supply and Delivery of 14 x 17 inches (35 x 43 cm) X-ray Flat Panel Digitizer with DR Bystem a. Wireless capability CSI Scintillator b. With at least 150 um pixel pitch and 16 bits acquisition depth c. Conforms to ingress protection standard of at least 1P44. d. Aluminum housing material with carbon fiber and aluminum plate as sensor protection material weighing < 3.5 kg. e. Must have a mini-PACS system that can store up to 15,000 images f. Capable of at least 150 kg load capacity. g. Time from exposure to preview (wireless): at least < 2 sec h. Detector must be capable of functioning with both artificial intelligence (Ai) software and non-artificial intelligence (Ai) software. i. Capable of capturing up to 200 images to the detector while kilometers away from the main console. Warranty a. Product Warranty shall include 3 years for parts and labor. b. Preventive maintenance services with semi-annual schedule of visit shall be provided up to 3 years free of charge. Accessories: a. One (1) set of charger and batteries (lithium-ion) b. One (1) set of work station (computer, tablet or laptop), Windows 19, equipped with most advanced software that has advanced reading and measurement from horizontal, lippman-cobb angle, goniometry and coxometry), Mini-PACS, web viewer licenses, image-rejection analysis feature, and custom measurement tools for chiropractic and orthopedic practices. The installed software must be upgradable with options to purchase additional feature like mini-PACS connection to Ortho View and remote image viewing. c. Twenty-five (25) can/160 of medical disinfecting towelettes for cleaning of imaging equipment. Must be durable, non-woven towels with 2-minute efficacy against multi-drug resistant bacteria and enveloped viruses. Should be soaked in cavicide and effective against SARS-CoV-2 on hard non-porous surfaces and kills TB in 3 mins., and MRSA, HIV-1, and HCV in 2 minutes. EPA List N Compliant. Soaked in cavicide. To provide copy of distributorship from manufacturer or importer. | Unit | 1 | Within Ninety (90) Calendar Days upon issuance of Notice to Proceed |
| | | | | |

Schedule of Requirements Page 1 of 2 LINE 3: QCGH-24-HME-1557

Terms and Conditions: a. With Three (3) Year Warranty for Parts and Services of the offered imaging equipment (wireless flat panel detector) b. Performs the required preventive maintenance of the machine during warranty period. c. Delivery and installation lead time of Ninety (90) calendar days from the date of receipt of Notice to Proceed. d. Certification from the bidder stating that the offered imaging equipment and its Ai and non-Ai software are of the same brand. e. Certification from the manufacturer ensuring the availability of spare parts and accessories of the offered imaging equipment in the next five (5) years and will not be discontinued in the next five (5) years. f. Certification from the manufacturer that the brand has been in the Philippine market for more than ten (10) years, to attach proof of documents. g. Certification from the manufacturer of their valid ISO 13485:2016 (QMS for Medical Device Manufacturing), ISO 14001:2015 (Environmental Management Systems) and ISO 9001:2015 (Quality Management System). h. Certification from the manufacturer declaring a list of at least 150 installations in the Philippines of the specific model being offered. Hospital/client name, contact details, year of delivery and installation must be clearly specified on the list. i. Authority to sell from the manufacturer / authorized distributorship of the offered imaging equipment (wireless flat panel detector) being offered j. Availability of service engineers 24/7 with guaranteed uptime

I hereby certify to comply and deliver all the above requirements.

within 24 hours unless spare part is required to be purchased.

| Name: | |
|--|--|
| Legal Capacity: | |
| Signature: | |
| Duly authorized to sign the Bid for and behalf of: | |

Schedule of Requirements Page 2 of 2 LINE 3; QCGH-24-HME-1557

+ how y }

Section VII. Technical Specifications

Notes for Preparing the Technical Specifications

A set of precise and clear specifications is a prerequisite for Bidders to respond realistically and competitively to the requirements of the Procuring Entity without qualifying their Bids. In the context of Competitive Bidding, the specifications (e.g. production/delivery schedule, manpower requirements, and after-sales service/parts, descriptions of the lots or items) must be prepared to permit the widest possible competition and, at the same time, present a clear statement of the required standards of workmanship, materials, and performance of the goods and services to be procured. Only if this is done will the objectives of transparency, equity, efficiency, fairness, and economy in procurement be realized, responsiveness of bids be ensured, and the subsequent task of bid evaluation and post-qualification facilitated. The specifications should require that all items, materials and accessories to be included or incorporated in the goods be new, unused, and of the most recent or current models, and that they include or incorporate all recent improvements in design and materials unless otherwise provided in the Contract.

Samples of specifications from previous similar procurements are useful in this respect. The use of metric units is encouraged. Depending on the complexity of the goods and the repetitiveness of the type of procurement, it may be advantageous to standardize the General Technical Specifications and incorporate them in a separate subsection. The General Technical Specifications should cover all classes of workmanship, materials, and equipment commonly involved in manufacturing similar goods. Deletions or addenda should then adapt the General Technical Specifications to the particular procurement.

Care must be taken in drafting specifications to ensure that they are not restrictive. In the specification of standards for equipment, materials, and workmanship, recognized Philippine and international standards should be used as much as possible. Where other particular standards are used, whether national standards or other standards, the specifications should state that equipment, materials, and workmanship that meet other authoritative standards, and which ensure at least a substantially equal quality than the standards mentioned, will also be acceptable. The following clause may be inserted in the Special Conditions of Contract or the Technical Specifications.

Sample Clause: Equivalency of Standards and Codes

Wherever reference is made in the Technical Specifications to specific standards and codes to be met by the goods and materials to be furnished or tested, the provisions of the latest edition or revision of the relevant standards and codes shall apply, unless otherwise expressly stated in the Contract. Where such standards and codes are national or relate to a particular country or region, other authoritative standards that ensure substantial equivalence to the standards and codes specified will be acceptable.

Reference to brand name and catalogue number should be avoided as far as possible; where unavoidable they should always be followed by the words "or at least equivalent." References to brand names cannot be used when the funding source is the GOP.

Where appropriate, drawings, including site plans as required, may be furnished by the Procuring Entity with the Bidding Documents. Similarly, the Supplier may be requested to provide drawings or samples either with its Bid or for prior review by the Procuring Entity during contract execution.

Bidders are also required, as part of the technical specifications, to complete their statement of compliance demonstrating how the items comply with the specification.

Technical Specifications

PROJECT NAME: LINE 1: ANESTHESIA MACHINE WITH PATIENT MONITOR AND BRAIN FUNCTION AND CEREBRAL OXIMETRY MONITORING PROJECT NO. QCGH-24-HME-1557

| Item | Specification | Statement of Compliance |
|------|--|--|
| Item | Specification | [Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable |
| | | for prosecution subject to the applicable laws and |
| | With minimum Technical Specifications | issuances.] |
| A.1 | Anesthesia Machine with Patient Monitor and Brain | |
| | Function and Cerebral Oximetry Monitoring | |
| | At least 15-inch touch screen monitor with adjustable | |
| | brightness | |
| | Weight: not exceeding 150 kg (without vaporizers and | |
| | cylinders) | |
| | Top Shelf: Weight limit: 25-35kg, Width: 255-355mm, | |
| | Length: 500-550mm | |
| | Work Surface: Height: 800-900mm, Area 1500-1650cm2 | |
| | Pneumatic-driven, electronically controlled anesthetic ventilator | |
| | Ventilator Specifications: | |
| | Modes of Ventilation: Manual/Spontaneous | |
| | Ventilation/Bypass, Volume Control Ventilation (VCV) | |
| | with PLV function. Pressure Control Ventilation (PCV) | |
| | with/without volume guarantee (VG), Synchronized | |
| | Intermittent Mandatory Ventilation (SIMV-Volume | |
| | Controlled and SIMV-Pressure Controlled), Pressure | |
| | Support Ventilation (PS) with apnea backup, | |
| | Synchronized Intermittent Mandatory Ventilation | |
| | Volume Guarantee (SIMV-VG), Continuous Positive Airway Pressure/Pressure Support Ventilation | |
| | (CPAP/PS) | |
| | with 2 vaporizers selectatec with interlock (Sevoflurane | |
| | and Isoflurane) | |
| | Monitoring and Alarm | |
| | Monitoring part must be able to all setting and alarm | |
| | parameters (including Breath rate, I/E ratio, Tidal | |
| | volume, Minute volume, PEEP, MEAN, PEAK, PLAT, | |
| | and O2 concentration, EtCO2, N2O, Aesthesia gas | |
| | concentration) | |

Ventilator Performance:

Peak gas flow: 100L/min plus fresh gas, Flow valve

range: 1-100 L/min

Plug-and-play modular CO2 monitoring

Trolley with central brake and cable pusher on each caster Battery backup: at least 60 min, Battery type: Build in Li-

ion battery, 11.1 VDC, 4400 mAh

Pneumatic Specifications

Switching Auxiliary Common Gas Outlet (ACGO)

Connector: ISO 22mm OD and 15 mm ID the outlet locates at the inspiratory limb. Gas Supply pipeline input range: 0.28~0.6mPa, Pipeline connections: NIST or DISS

Cylinder input: PISS

APL valve range 1~75 cmH20, with tactile knob indication at: >30 cm H20

Breathing Circuit Specifications

Sytem Pressure Gauge: Range: -20~100 cmH20, Accuracy: +/-2.5% full scale, Bag/Mechanical Ventilation Switch:

Type: Bi-stable

Breathing Circuit Parameters

Compliance: <4mL/100Pa Automatically compensates for compression loss with in the breathing circuit on mechanical mode, Expiration resistance: < 6cm H20 @ 60 L/min, Inspiration resistance: <6cm H20 @ 60 L/min

Brain Function and Cerebral Oximetry Monitoring Equipment with Non Invasive Pulse co-oximeter Specifications:

1. Base Unit

- a. Multi-modal patient monitoring platform
- b. Can measure and display the following parameters
- i. Brain function (i.e. processed EEG)"
- "ii. Regional Oximetry
- iii. Non-invasive pulse oximetry
- c. Display is at least 1280 x 800 pixels, and at least 10.1-inch diagonal"
- "Multi-touch screen and customizable with multiple screen views
- i. Can display tissue oxygen saturation (%rSO2) of specific area (i.e. forehead)
- ii. Can display the difference between current rSO2 and user-defined baseline
- iii. Can display the difference between sPO2 and rSO2"
- "iv. Can display four channels of bilateral EEG waveforms
- v. Can display a value of processed EEG parameter related to the effect of anesthetic agents
- vi. Can display a value of spectrogram representing activity in both sides of the brain.
- e. Must have ports or docks that are compatible to plugand-play modules for the measurement of brain function monitoring, regional oximetry, and non-invasive pulse oximetry"
- at least four hours' battery capacity and maximum of four hours charging time

2. Modules:

- a. One (1) Module for brain function monitoring/Unit with patient cable
- b. One (1) Module for Cerebral Oximetry Monitoring

| | c. One (1) Module for Non-Invasive Pulse Oximetry | | |
|----|--|---|--|
| | i. Can be used as a bedside, handheld, or transport | | |
| | monitor | | |
| | ii. Has high-definition multi-touch customizable display | 1 | |
| | that can show oxygen saturation (sPO2), pulse rate, | | |
| | perfusion index, pleth variability index, and high | | |
| | resolution plethysmographic waveform | | |
| В. | Compliance to the Schedule of Requirements | | |
| | (Section VI) | | |

I hereby certify to comply and deliver all the above requirements.


| Name: | |
|--|--|
| Legal Capacity: | |
| Signature: | |
| Duly authorized to sign the Bid for and behalf of: | |

Technical Specifications Page 3 of 3 LINE 1: QCGH-24-HME-1557

Technical Specifications

PROJECT NAME: LINE 2: ANESTHESIA MACHINE WITH PATIENT MONITOR PROJECT NO. QCGH-24-HME-1557

| Item | Specification | Statement of Compliance |
|------|--|--|
| | | [Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and crossreferenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.] |
| | With minimum Technical Specifications | |
| A.1 | Electronically Controlled and pneumatically driven with rising bellow, Maximum Machine Weight must not exceed 128kg, Workbench supporting max weight capacity of 20kg, Top Plate supporting max weight capacity of 20kg, Working Temperature of 10-40 degrees C, =/<93% of Humidity, Input Voltage of 100-240V, Input Current of 3.5-8.5 A, Input Frequency of 50/60Hz +/- 1Hz, 7000mAh 11.1VDC, Automatic Breathing Circuit Compliance at =/<4ml/100pa and Automatically compensates for circuit compression loss within the breathing circuit in mechanical mode, Tidal Volume range of VCV 15-1500 ml and PCV 5-1500 ml, 2000 Rechargeable Lithium-Ion Battery with four hours charging time and at least 2 hours continuous working operation time Gas Source (Oxygen and AIR) | |
| | with Electronic Flowmeter Oxygen Flush switch gives 100% fast oxygen Pressure Regulator With Total Flowmeter With Back-up Flow Control System Button With Flow Regulation Knob of the back-up flow control system Accommodates Two Vaporizers with Interlocking System (Sevoflurane and Isoflurane) | |
| | ACGO (Auxiliary Common Gas Outlet) using connector type Taper coaxial fitting of 22mm outside and 15mm inside Auxiliary Air Flowmeter Auxiliary Oxygen Supply Flowmeter Autoclavable Breathing System must have the following; Expiration Port, Canister for CO2 sodalime absorbent, | |



Canister hold and release lever, Manual Drain Valve, Leak Test Plug, Breathing Circuit Hook, Expiration Check Valve, APL Valve range 1-75 cmH2O with Minimum opening pressure of 0.3/0.5 cmH2O (dry/humid), Airway Pressure Gauge, Manual Bag Port, Bellows Assembly, Manual Bag Port Arm, Manual and Mechanical Ventilation Switch, Inspiratory Check Valve and Inspiration Port, and Circuit By-Pass Function

Ventilator Modes: VCV/VC, PCV/VPC, SIMV-VC, SIMV-PC, PRVC, PSV/CPAP, Manual and Automatic Ventilation, SIMV-PRVC, PSVPro. Ventilation Principle is Chronometric, volumetric and barometric, Electronically Controlled and Pneumatically Driven, Electronic Selective Air or Oxygen

Positive End Expiratory Pressure (PEEP) Type Integrated, electronically controlled, Range 0~70 cmH2O At least 12.1" 800 x 600 Resolution Touchscreen Ventilator and Monitoring Main Display Screen

At least two Monitoring Module Slots with Anesthetic Gas (Ag) module with Bi Spectral Index (BIS) module with ETCO2 module (capnography) Module must be compatible with the Patient Monitor AGSS (Anesthesia Gas Scavenging System) must have the following; Waste gas exhaust noozle connector, AGSS waste gases outlet, Outer cone connector for hose of transfer system, Pressure Compensation Port, Main Body of AGSS System, Red Color Float, and Flow regulation Knob, Suction flow rate 50-80LPM, Stainless steel mesh filter with pore size of 60-100um 4 pcs of Caster Wheel with Central Brake System With at least two Storage Drawer with lock and key With Amrest/Handle for easy maneuvering to and from Operating Room

At least four Auxiliary power output socket with individual socket breaker with output voltage of 100-240V, output frequency of 50/60Hz.

Must have comprehensive ventilation mode such as the following; Manual (Induction), VCV, PCV, and PRCV (Maintaining Anethesia), SIMV-VC, SIMV-PC, and SIMV-PRVC (Anesthesia Recovery), PSV, and PSVPro (Practice of Spontaneous Breathing)

Must have at least 10 seconds shutdown delay Must have a Ventilator Operation Risk Reminder/Prompt Must have a central lock braking system

Patient Monitor:

Main Display Monitor of not less than 12" TFT Touchscreen and Navigation Knob with 3 built in module slots for various monitoring parameters, must have a built in thermal printer, must not exceed 5kg, 360' visible (Dual Alarm) Indicator Light

Must have fixed hidden handle, must have Al-Mg alloy stent design, Easy to maintain, must have these six conventional monitoring parameters of 5-Lead ECG (optional 3/12-lead) and provides 26 arrhythmia analysis, Heart Rate, Respiration, Dual-Temperature that can

measure skin and oral/anal temperature simultaneously, SpO2 provides i-sports and anti-low perfusion, AcuTec NIBP High Accuracy technology for Hypertension. Power Supply 100-240V, 50/60Hz +/-1Hz, must use 4400mAh Lithium-Ion rechargeable battery with charging time of 6 hours and 3-6 hours continuous battery back up Various User Interface such as Standard Interface, Big Font, Trendm OxyCRG, List, Bed to Bed and 7-Lead ECG/12-Lead Full Screen Cascade ECG Date storage Alarm Event Recall: 200 groups, Wave Recall: 2 hours, NIBP Recall: 2000 groups, Trend Graph: 120hours, Trend Table: 120hours, ARR Event: 100 groups Recorder Type: Built-in; Thermal array, Channel: 3 channel waveforms, Speed: 25mm/s,50mm/s, Record width: 50mm, External printer: Yes Respiration Method: RA-LL Impedance Method, RR measurement range: (Adult: 0-120rpm, Pediatric/Neonate: 0 -150bpm), Accuracy: 7-150rpm ±2rpm or ±2% (whichever isgreater), 0-6rpm: unspecified, RESP Apnea Alarm: Adult: 10s-60s Ped/Neo: 10s~20s, Alarm: Audible and visual alarm; alarm events reviewable, Sweep Speed: 6.25,12.5, 25 mm/s, Gain Selection: X0.25, X0.5, X1, X2, x4 **SpO2 Measurement & alarm** range: 0~100%, Resolution: 1%, Accuracy: ±2% (70~100%, Ped/Adu, non-motion), ±3% (70-100%, Neo, non-motion); 0-69% unspecified, PR Measurement Range: 25--250bpm, Accuracy: ±1bpm, Alarm range: 20~350bpm, Temperature (Dual Channel) Measurement & alarm range: 0~50°C, TEMP sensor: Standard configuration-skin TEMP sensor, Resolution: 0.1°C, Accuracy: ±0.1°C (exclusive of error of sensor), Channel type: T1, T2, TD (Temperature Difference) **Dual IBP Module** R. Compliance to the Schedule of Requirements (Section VI)

I hereby certify to comply and deliver all the above requirements.

| Name: | |
|--|--|
| Legal Capacity: | |
| Signature: | |
| Duly authorized to sign the Bid for and behalf of: | |

Technical Specifications Page 3 of 3 LINE 2: QCGH-24-HME-1557

cm }

Technical Specifications

PROJECT NAME: LINE 3: SUPPLY AND DELIVERY OF X-RAY FLAT PANEL DIGITIZER WITH DIGITAL RADIOGRAPHY SYSTEM PROJECT NO. QCGH-24-HME-1557

| Item | Specification | Statement of Compliance |
|------|---|---|
| | | [Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered. Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and crossreferenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, postqualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.] |
| | With minimum Technical Specifications | |
| A.1 | Supply and Delivery of 14 x 17 inches (35 x 43 cm) X-ray Flat Panel Digitizer with DR System | |
| | a. Wireless capability CSI Scintillator b. With at least 150 um pixel pitch and 16 bits acquisition depth | |
| | c. Conforms to ingress protection standard of at least IP44. | |
| | d. Aluminum housing material with carbon fiber and aluminum | |
| | plate as sensor protection material weighing < 3.5 kg. | |
| | e. Must have a mini-PACS system that can store up to 15,000 images | |
| | f. Capable of at least 150 kg load capacity. | |
| | g. Time from exposure to preview (wireless): at least < 2 sec | |
| | h. Detector must be capable of functioning with both artificial | |
| | intelligence (Ai) software and non-artificial intelligence (Ai) software. | |
| | i. Capable of capturing up to 200 images to the detector while kilometers away from the main console. | |
| | Warranty a. Product Warranty shall include 3 years for parts and labor. | |
| | b. Preventive maintenance services with semi-annual schedule of | |
| | visit shall be provided up to 3 years free of charge. | |
| | Accessories: a. One (1) set of charger and batteries (lithium-ion) | |
| | b. One (1) set of work station (computer, tablet or laptop), | |
| | Windows 10, equipped with most advanced software that has | |
| | advanced reading and measurement tools (absolute rotation | |
| | angle, extended cobb angle, measurement from horizontal, | |
| | lippman-cobb angle, goniometry and coxometry), Mini-PACS, | |
| | web viewer licenses, image-rejection analysis feature, and | |
| | custom measurement tools for chiropractic and orthopedic | |
| | practices. The installed software must be upgradable with | |



options to purchase additional feature like mini-PACS connection to Ortho View and remote image viewing. c. Twenty-five (25) can/160 of medical disinfecting towelettes for cleaning of imaging equipment. Must be durable, non-woven towels with 2-minute efficacy against multi-drug resistant bacteria and enveloped viruses. Should be soaked in cavicide and effective against SARS-CoV-2 on hard non-porous surfaces and kills TB in 3 mins., and MRSA, HIV-1, and HCV in 2 minutes. EPA List N Compliant. Soaked in cavicide. To provide copy of distributorship from manufacturer or importer. Terms and Conditions: a. With Three (3) Year Warranty for Parts and Services of the offered imaging equipment (wireless flat panel detector) b. Performs the required preventive maintenance of the machine

- during warranty period.
- c. Delivery and installation lead time of Ninety (90) calendar days from the date of receipt of Notice to Proceed.
- d. Certification from the bidder stating that the offered imaging equipment and its Ai and non-Ai software are of the same brand.
- e. Certification from the manufacturer ensuring the availability of spare parts and accessories of the offered imaging equipment in the next five (5) years and will not be discontinued in the next five (5) years.
- f. Certification from the manufacturer that the brand has been in the Philippine market for more than ten (10) years, to attach proof of documents.
- g. Certification from the manufacturer of their valid ISO 13485:2016 (QMS for Medical Device Manufacturing), ISO 14001:2015 (Environmental Management Systems) and ISO 9001:2015 (Quality Management System).
- h. Certification from the manufacturer declaring a list of at least 150 installations in the Philippines of the specific model being offered. Hospital/client name, contact details, year of delivery and installation must be clearly specified on the list.
- i. Authority to sell from the manufacturer / authorized distributorship of the offered imaging equipment (wireless flat panel detector) being offered
- j. Availability of service engineers 24/7 with guaranteed uptime within 24 hours unless spare part is required to be purchased.

B. Compliance to the Schedule of Requirements (Section VI)

I hereby certify to comply and deliver all the above requirements.

| Name: | |
|--|--|
| Legal Capacity: | |
| Signature: | |
| Duly authorized to sign the Bid for and behalf of: | |

Technical Specifications Page 2 of 2 LINE 3: QCGH-24-HME-1557

Section VIII. Checklist of Technical and Financial Documents

Notes on the Checklist of Technical and Financial Documents

The prescribed documents in the checklist are mandatory to be submitted in the Bid, but shall be subject to the following:

- a. GPPB Resolution No. 09-2020 on the efficient procurement measures during a State of Calamity or other similar issuances that shall allow the use of alternate documents in lieu of the mandated requirements; or
- b. Any subsequent GPPB issuances adjusting the documentary requirements after the effectivity of the adoption of the PBDs.

The BAC shall be checking the submitted documents of each Bidder against this checklist to ascertain if they are all present, using a non-discretionary "pass/fail" criterion pursuant to Section 30 of the 2016 revised IRR of RA No. 9184.

Checklist of Technical and Financial Documents

I. TECHNICAL COMPONENT ENVELOPE

Class "A" Documents

| Lega | <u>ıl Doc</u> | <u>uments</u> |
|-------------|---------------|--|
| | (a) | Valid PhilGEPS Registration Certificate (Platinum Membership) (all pages) in accordance with Section 8.5.2 of the IRR; |
| <u>Tech</u> | <u>inical</u> | <u>Documents</u> |
| | (b) | Statement of the prospective bidder of all its ongoing government and private contracts, including contracts awarded but not yet started, if any, whether similar or not similar in nature and complexity to the contract to be bid (in a FORM prescribed by the QC-BAC-GOODS AND SERVICES); and |
| | (c) | Statement of the bidder's Single Largest Completed Contract (SLCC) similar to the contract to be bid, except under conditions provided for in Sections 23.4.1.3 and 23.4.2.4 of the 2016 revised IRR of RA No. 9184, within the relevant period as provided in the Bidding Documents (in a FORM prescribed by the QC-BAC-GOODS AND SERVICES); and |
| | (d) | Original copy of Bid Security. If in the form of a Surety Bond, submit also a certification issued by the Insurance Commission; or Original copy of Notarized Bid Securing Declaration; and |
| | (e) | Conformity with Section VI. (Schedule of Requirements) and Section VII. (Technical Specifications), which may include production/delivery schedule, manpower requirements, and/or after-sales/parts, if applicable; <u>and</u> |
| | (f) | Original duly signed Omnibus Sworn Statement (OSS); and if applicable, Original Notarized Secretary's Certificate in case of a corporation, partnership, or cooperative; or Original Special Power of Attorney of all members of the joint venture giving full power and authority to its officer to sign the OSS and do acts to represent the Bidder. |
| Fina | ncial . | Documents |
| | (g) | The prospective bidder's computation of Net Financial Contracting Capacity (NFCC) (in a FORM prescribed by the QC-BAC-GOODS AND SERVICES); |
| | | A committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation. |
| | (h) | Class "B" Documents If applicable, a duly signed joint venture agreement (JVA) in case the joint venture is already in existence; or |
| | | duly notarized statements from all the potential joint venture partners stating that they will enter into and abide by the provisions of the JVA in the instance that the bid is successful. |
| <u>Othe</u> | r doci | umentary requirements under RA No. 9184 (as applicable) |
| | (i) | [For foreign bidders claiming by reason of their country's extension of reciprocal rights to Filipinos] Certification from the relevant government office of their country stating that Filipinos are allowed to participate in government procurement activities for the same item or product. |
| | (j) | Certification from the DTI if the Bidder claims preference as a Domestic Bidder or Domestic Entity. |

II. FINANCIAL COMPONENT ENVELOPE

□ (a) Original of duly signed and accomplished Financial Bid Form; and
 □ (b) Original of duly signed and accomplished Price Schedule(s).

III.REQUIRED DOCUMENTS in BDS SECTION 20.2 and 21.2

 Copy of valid, current License to Operate from DOH Accreditation as Supplier, Distributor or Manufacturer for Medical or Hospital Equipment or Devices

FOR LINE 1

- Statement of Warranty: Minimum of One (1) Year on parts and service (with Project Name and Project No.)
- Certification from the Manufacturer guaranteeing the availability of all spare parts for the next five (5) years and that parts shall be available at the authorized Philippine service center/s for a period of (5) years after the warranty period.
- Certification that the bidder is an authorized or exclusive distributor of the brand/model being offered.
- Certification that the bidder can provide applications training for users and maintenance personnel of the hospital (with Project Name and Project No.)
- Certification of Current and Valid Certificate of Manufacturer's compliance with ISO and/or CE Certificate

FOR LINE 2

- Statement of Warranty: Minimum of One (1) Year on parts and service (with Project Name and Project No.)
- Certification on the availability of spare parts for five (5) years after the warranty period.
- Certification that the bidder is an authorized or exclusive distributor of the brand/model being offered.

FOR LINE 3

- Statement of Warranty: Minimum of Three (3) years on parts and service (with Project Name and Project No.)
- Certification of Preventive Maintenance Services (PMS) with semi-annual schedule of visit up to three (3) years free of charge
- Certification that the offered imaging equipment and its Ai and non-Ai software are of the same brand.
- Certification from the manufacturer ensuring the availability of spare parts and accessories of the offered imaging equipment in the next five (5) years and will not be discontinued in the next five (5) years.
- Certification from the manufacturer that the brand has been in the Philippine market for more than ten (10) years, to attach proof of documents.
- Certification from the manufacturer of their valid ISO 13485:2016 (QMS for Medical Device Manufacturing), ISO 14001:2015 (Environmental Management Systems) and ISO 9001:2015 (Quality Management System).
- Certification from the manufacturer declaring a list of at least 150 installations in the Philippines of the specific model being offered. Hospital/client name, contact details, year of delivery and installation must be clearly specified on the list.
- Authority to sell from the manufacturer / authorized distributorship of the offered imaging equipment (wireless flat panel detector) being offered
- Availability of service engineers 24/7 with guaranteed uptime within 24 hours unless spare part is required to be purchased.

of hu ,

Note:

- 1. Please refer to
 - [https://drive.google.com/file/d/1uiYurh5WrpBL5B_pqpzAb62yucAblR1p/view?usp=sh_aring] for the following requirements:
 - a. Computation of NFCC;
 - b. List of All Ongoing Contracts/List of Contracts already awarded but not yet started;
 - c. Statement of Single Largest Completed Contract
- 2. Please refer to GPPB Resolution No. 16-2020 for the following requirements:
 - a. Bid Form;
 - b. Price Schedule (for Goods Offered from Abroad/ Within the Philippines)
 - c. Bid Securing Declaration; and
 - d. Omnibus Sworn Statement

