

REPUBLIC OF THE PHILIPPINES QUEZON CITY GOVERNMENT BIDS AND AWARDS COMMITTEE – GOODS AND SERVICES



PHILIPPINE BIDDING DOCUMENTS

(As Harmonized with Development Partners)

PROCUREMENT OF ULTRASOUND MACHINE

PROJECT NO. RMBGH-25-HME-0742

Government of the Republic of the Philippines

Sixth Edition July 2020

Preface

These Philippine Bidding Documents (PBDs) for the procurement of Goods through Competitive Bidding have been prepared by the Government of the Philippines for use by any branch, constitutional commission or office, agency, department, bureau, office, or instrumentality of the Government of the Philippines, National Government Agencies, including Government-Owned and/or Controlled Corporations, Government Financing Institutions, State Universities and Colleges, and Local Government Unit. The procedures and practices presented in this document have been developed through broad experience, and are for mandatory use in projects that are financed in whole or in part by the Government of the Philippines or any foreign government/foreign or international financing institution in accordance with the provisions of the 2016 revised Implementing Rules and Regulations of Republic Act No. 9184.

The Bidding Documents shall clearly and adequately define, among others: (i) the objectives, scope, and expected outputs and/or results of the proposed contract or Framework Agreement, as the case may be; (ii) the eligibility requirements of Bidders; (iii) the expected contract or Framework Agreement duration, the estimated quantity in the case of procurement of goods, delivery schedule and/or time frame; and (iv) the obligations, duties, and/or functions of the winning bidder.

Care should be taken to check the relevance of the provisions of the PBDs against the requirements of the specific Goods to be procured. If duplication of a subject is inevitable in other sections of the document prepared by the Procuring Entity, care must be exercised to avoid contradictions between clauses dealing with the same matter.

Moreover, each section is prepared with notes intended only as information for the Procuring Entity or the person drafting the Bidding Documents. They shall not be included in the final documents. The following general directions should be observed when using the documents:

- a. All the documents listed in the Table of Contents are normally required for the procurement of Goods. However, they should be adapted as necessary to the circumstances of the particular Procurement Project.
- b. Specific details, such as the "name of the Procuring Entity" and "address for bid submission," should be furnished in the Instructions to Bidders, Bid Data Sheet, and Special Conditions of Contract. The final documents should contain neither blank spaces nor options.
- c. This Preface and the footnotes or notes in italics included in the Invitation to Bid, Bid Data Sheet, General Conditions of Contract, Special Conditions of Contract, Schedule of Requirements, and Specifications are not part of the text of the final document, although they contain instructions that the Procuring Entity should strictly follow.
- d. The cover should be modified as required to identify the Bidding Documents as to the Procurement Project, Project Identification Number, and Procuring Entity, in addition to the date of issue.

- e. Modifications for specific Procurement Project details should be provided in the Special Conditions of Contract as amendments to the Conditions of Contract. For easy completion, whenever reference has to be made to specific clauses in the Bid Data Sheet or Special Conditions of Contract, these terms shall be printed in bold typeface on Sections I (Instructions to Bidders) and III (General Conditions of Contract), respectively.
- f. For guidelines on the use of Bidding Forms and the procurement of Foreign-Assisted Projects, these will be covered by a separate issuance of the Government Procurement Policy Board.

Table of Contents

Glossa	ry of Acronyms, Terms, and Abbreviations	4
Section	n I. Invitation to Bid	7
Section	n II. Instructions to Bidders	11
1.	Scope of Bid	12
2.	Funding Information	12
3.	Bidding Requirements	12
4.	Corrupt, Fraudulent, Collusive, and Coercive Practices	12
5.	Eligible Bidders	12
6.	Origin of Goods	13
7.	Subcontracts	13
8.	Pre-Bid Conference	13
9.	Clarification and Amendment of Bidding Documents	14
10.	Documents comprising the Bid: Eligibility and Technical Components	14
11.	Documents comprising the Bid: Financial Component	14
12.	Bid Prices	14
13.	Bid and Payment Currencies	15
14.	Bid Security	15
15.	Sealing and Marking of Bids	15
16.	Deadline for Submission of Bids	16
17.	Opening and Preliminary Examination of Bids	16
18.	Domestic Preference	16
1 9 .	Detailed Evaluation and Comparison of Bids	16
20.	Post-Qualification	17
21.	Signing of the Contract	17
Section	n III. Bid Data Sheet	18
Section	n IV. General Conditions of Contract	20
1.	Scope of Contract	21
2.	Advance Payment and Terms of Payment	21
3.	Performance Security	21
4.	Inspection and Tests	21
5.	Warranty	21
6.	Liability of the Supplier	22
Section	n V. Special Conditions of Contract	23
Section	n VI. Schedule of Requirements	27
Section	n VII. Technical Specifications	28
Sectio	n VIII. Checklist of Technical and Financial Documents	31

Glossary of Acronyms, Terms, and Abbreviations

ABC – Approved Budget for the Contract.

BAC - Bids and Awards Committee.

Bid – A signed offer or proposal to undertake a contract submitted by a bidder in response to and in consonance with the requirements of the bidding documents. Also referred to as *Proposal* and *Tender*. (2016 revised IRR, Section 5[c])

Bidder – Refers to a contractor, manufacturer, supplier, distributor and/or consultant who submits a bid in response to the requirements of the Bidding Documents. (2016 revised IRR, Section 5[d])

Bidding Documents – The documents issued by the Procuring Entity as the bases for bids, furnishing all information necessary for a prospective bidder to prepare a bid for the Goods, Infrastructure Projects, and/or Consulting Services required by the Procuring Entity. (2016 revised IRR, Section 5[e])

BIR – Bureau of Internal Revenue.

BSP – Bangko Sentral ng Pilipinas.

Consulting Services – Refer to services for Infrastructure Projects and other types of projects or activities of the GOP requiring adequate external technical and professional expertise that are beyond the capability and/or capacity of the GOP to undertake such as, but not limited to: (i) advisory and review services; (ii) pre-investment or feasibility studies; (iii) design; (iv) construction supervision; (v) management and related services; and (vi) other technical services or special studies. (2016 revised IRR, Section 5[i])

CDA - Cooperative Development Authority.

Contract – Refers to the agreement entered into between the Procuring Entity and the Supplier or Manufacturer or Distributor or Service Provider for procurement of Goods and Services; Contractor for Procurement of Infrastructure Projects; or Consultant or Consulting Firm for Procurement of Consulting Services; as the case may be, as recorded in the Contract Form signed by the parties, including all attachments and appendices thereto and all documents incorporated by reference therein.

CIF – Cost Insurance and Freight.

CIP – Carriage and Insurance Paid.

CPI – Consumer Price Index.

DDP - Refers to the quoted price of the Goods, which means "delivered duty paid."

DTI – Department of Trade and Industry.

EXW – Ex works.

FCA – "Free Carrier" shipping point.

FOB – "Free on Board" shipping point.

Foreign-funded Procurement or Foreign-Assisted Project– Refers to procurement whose funding source is from a foreign government, foreign or international financing institution as specified in the Treaty or International or Executive Agreement. (2016 revised IRR, Section 5[b]).

Framework Agreement – Refers to a written agreement between a procuring entity and a supplier or service provider that identifies the terms and conditions, under which specific purchases, otherwise known as "Call-Offs," are made for the duration of the agreement. It is in the nature of an option contract between the procuring entity and the bidder(s) granting the procuring entity the option to either place an order for any of the goods or services identified in the Framework Agreement List or not buy at all, within a minimum period of one (1) year to a maximum period of three (3) years. (GPPB Resolution No. 27-2019)

GFI – Government Financial Institution.

GOCC – Government-owned and/or –controlled corporation.

Goods – Refer to all items, supplies, materials and general support services, except Consulting Services and Infrastructure Projects, which may be needed in the transaction of public businesses or in the pursuit of any government undertaking, project or activity, whether in the nature of equipment, furniture, stationery, materials for construction, or personal property of any kind, including non-personal or contractual services such as the repair and maintenance of equipment and furniture, as well as trucking, hauling, janitorial, security, and related or analogous services, as well as procurement of materials and supplies provided by the Procuring Entity for such services. The term "related" or "analogous services" shall include, but is not limited to, lease or purchase of office space, media advertisements, health maintenance services, and other services essential to the operation of the Procuring Entity. (2016 revised IRR, Section 5[r])

GOP – Government of the Philippines.

GPPB – Government Procurement Policy Board.

INCOTERMS – International Commercial Terms.

Infrastructure Projects – Include the construction, improvement, rehabilitation, demolition, repair, restoration or maintenance of roads and bridges, railways, airports, seaports, communication facilities, civil works components of information technology projects, irrigation, flood control and drainage, water supply, sanitation, sewerage and solid waste management systems, shore protection, energy/power and electrification facilities, national buildings, school buildings, hospital buildings, and other related construction projects of the government. Also referred to as *civil works or works*. (2016 revised IRR, Section 5[u])

LGUs – Local Government Units.

NFCC – Net Financial Contracting Capacity.

NGA – National Government Agency.

PhilGEPS - Philippine Government Electronic Procurement System.

Procurement Project – refers to a specific or identified procurement covering goods, infrastructure project or consulting services. A Procurement Project shall be described, detailed, and scheduled in the Project Procurement Management Plan prepared by the agency

which shall be consolidated in the procuring entity's Annual Procurement Plan. (GPPB Circular No. 06-2019 dated 17 July 2019)

PSA – Philippine Statistics Authority.

SEC – Securities and Exchange Commission.

SLCC – Single Largest Completed Contract.

Supplier – refers to a citizen, or any corporate body or commercial company duly organized and registered under the laws where it is established, habitually established in business and engaged in the manufacture or sale of the merchandise or performance of the general services covered by his bid. (Item 3.8 of GPPB Resolution No. 13-2019, dated 23 May 2019). Supplier as used in these Bidding Documents may likewise refer to a distributor, manufacturer, contractor, or consultant.

UN – United Nations.

Notes on the Invitation to Bid

The Invitation to Bid (IB) provides information that enables potential Bidders to decide whether to participate in the procurement at hand. The IB shall be posted in accordance with Section 21.2 of the 2016 revised IRR of RA No. 9184.

Apart from the essential items listed in the Bidding Documents, the IB should also indicate the following:

- a. The date of availability of the Bidding Documents, which shall be from the time the IB is first advertised/posted until the deadline for the submission and receipt of bids;
- b. The place where the Bidding Documents may be acquired or the website where it may be downloaded;
- c. The deadline for the submission and receipt of bids; and
- d. Any important bid evaluation criteria (*e.g.*, the application of a margin of preference in bid evaluation).

The IB should be incorporated in the Bidding Documents. The information contained in the IB must conform to the Bidding Documents and in particular to the relevant information in the Bid Data Sheet.

Notes on the Instructions to Bidders

This Section on the Instruction to Bidders (ITB) provides the information necessary for bidders to prepare responsive bids, in accordance with the requirements of the Procuring Entity. It also provides information on bid submission, eligibility check, opening and evaluation of bids, post-qualification, and on the award of contract.

1. Scope of Bid

The Procuring Entity, Quezon City Local Government wishes to receive Bids for the **PROCUREMENT OF ULTRASOUND MACHINE** with identification number **RMBGH-25-HME-0742.**

[Note: The Project Identification Number is assigned by the Procuring Entity based on its own coding scheme and is not the same as the PhilGEPS reference number, which is generated after the posting of the bid opportunity on the PhilGEPS website.]

The Procurement Project (referred to herein as "Project") is composed of *one (1) item*, the details of which are described in Section VII (Technical Specifications).

2. Funding Information

- 2.1. The GOP through the source of funding as indicated below for 2025 in the amount of FOUR MILLION FIVE HUNDRED THOUSAND PESOS AND '00/100 ONLY (Php 4,500,000.00). '
- 2.2. The source of funding is:
 - a. LGUs, the Annual or Supplemental Budget, as approved by the Sanggunian.

3. Bidding Requirements

The Bidding for the Project shall be governed by all the provisions of RA No. 9184 and its 2016 revised IRR, including its Generic Procurement Manuals and associated policies, rules and regulations as the primary source thereof, while the herein clauses shall serve as the secondary source thereof.

Any amendments made to the IRR and other GPPB issuances shall be applicable only to the ongoing posting, advertisement, or **IB** by the BAC through the issuance of a supplemental or bid bulletin.

The Bidder, by the act of submitting its Bid, shall be deemed to have verified and accepted the general requirements of this Project, including other factors that may affect the cost, duration and execution or implementation of the contract, project, or work and examine all instructions, forms, terms, and project requirements in the Bidding Documents.

4. Corrupt, Fraudulent, Collusive, and Coercive Practices

The Procuring Entity, as well as the Bidders and Suppliers, shall observe the highest standard of ethics during the procurement and execution of the contract. They or through an agent shall not engage in corrupt, fraudulent, collusive, coercive, and obstructive practices defined under Annex "I" of the 2016 revised IRR of RA No. 9184 or other integrity violations in competing for the Project.

5. Eligible Bidders

- 5.1. Only Bids of Bidders found to be legally, technically, and financially capable will be evaluated.
- 5.2. Foreign ownership exceeding those allowed under the rules may participate pursuant to:

- i. When a Treaty or International or Executive Agreement as provided in Section 4 of the RA No. 9184 and its 2016 revised IRR allow foreign bidders to participate;
- ii. Citizens, corporations, or associations of a country, included in the list issued by the GPPB, the laws or regulations of which grant reciprocal rights or privileges to citizens, corporations, or associations of the Philippines;
- iii. When the Goods sought to be procured are not available from local suppliers; or
- iv. When there is a need to prevent situations that defeat competition or restrain trade.
- 5.3. Pursuant to Section 23.4.1.3 of the 2016 revised IRR of RA No.9184, the Bidder shall have an SLCC that is at least one (1) contract similar to the Project the value of which, adjusted to current prices using the PSA's CPI, must be at least equivalent to:
 - a. For the procurement of Non-Expendable Supplies and Services: The Bidder must have completed a single contract that is similar to this Project, equivalent to at least fifty percent (50%) of the ABC.
- 5.4. The Bidders shall comply with the eligibility criteria under Section 23.4.1 of the 2016 IRR of RA No. 9184.

6. Origin of Goods

There is no restriction on the origin of goods other than those prohibited by a decision of the UN Security Council taken under Chapter VII of the Charter of the UN, subject to Domestic Preference requirements under ITB Clause 18.

7. Subcontracts

7.1. The Bidder may subcontract portions of the Project to the extent allowed by the Procuring Entity as stated herein, but in no case more than twenty percent (20%) of the Project.

The Procuring Entity has prescribed that: Subcontracting is not allowed.

8. Pre-Bid Conference

The Procuring Entity will hold a pre-bid conference for this Project on the specified date and time and either at its physical address and/or through videoconferencing as indicated in paragraph 6 of the **IB**.

9. Clarification and Amendment of Bidding Documents

Prospective bidders may request for clarification on and/or interpretation of any part of the Bidding Documents. Such requests must be in writing and received by the Procuring Entity, either at its given address or through electronic mail indicated in the **IB**, at least ten (10) calendar days before the deadline set for the submission and receipt of Bids.

10. Documents comprising the Bid: Eligibility and Technical Components

- 10.1. The first envelope shall contain the eligibility and technical documents of the Bid as specified in Section VIII (Checklist of Technical and Financial Documents).
- 10.2. The Bidder's SLCC as indicated in **ITB** Clause 5.3 should have been completed within *the last three (3) years* prior to the deadline for the submission and receipt of bids.
- 10.3. If the eligibility requirements or statements, the bids, and all other documents for submission to the BAC are in foreign language other than English, it must be accompanied by a translation in English, which shall be authenticated by the appropriate Philippine foreign service establishment, post, or the equivalent office having jurisdiction over the foreign bidder's affairs in the Philippines. Similar to the required authentication above, for Contracting Parties to the Apostille Convention, only the translated documents shall be authenticated through an Apostille pursuant to GPPB Resolution No. 13-2019 dated 23 May 2019. The English translation shall govern, for purposes of interpretation of the bid.

11. Documents comprising the Bid: Financial Component

- 11.1. The second bid envelope shall contain the financial documents for the Bid as specified in Section VIII (Checklist of Technical and Financial Documents).
- 11.2. If the Bidder claims preference as a Domestic Bidder or Domestic Entity, a certification issued by DTI shall be provided by the Bidder in accordance with Section 43.1.3 of the 2016 revised IRR of RA No. 9184.
- 11.3. Any bid exceeding the ABC indicated in paragraph 1 of the IB shall not be accepted.
- 11.4. For Foreign-funded Procurement, a ceiling may be applied to bid prices provided the conditions are met under Section 31.2 of the 2016 revised IRR of RA No. 9184.

12. Bid Prices

- 12.1. Prices indicated on the Price Schedule shall be entered separately in the following manner:
 - a. For Goods offered from within the Procuring Entity's country:
 - i. The price of the Goods quoted EXW (ex-works, ex-factory, exwarehouse, ex-showroom, or off-the-shelf, as applicable);

- ii. The cost of all customs duties and sales and other taxes already paid or payable;
- iii. The cost of transportation, insurance, and other costs incidental to delivery of the Goods to their final destination; and
- iv. The price of other (incidental) services, if any, listed in e.
- b. For Goods offered from abroad:
 - i. Unless otherwise stated in the **BDS**, the price of the Goods shall be quoted delivered duty paid (DDP) with the place of destination in the Philippines as specified in the **BDS**. In quoting the price, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible source country.
 - ii. The price of other (incidental) services, if any, as listed in Section VII (Technical Specifications).

13. Bid and Payment Currencies

- 13.1. For Goods that the Bidder will supply from outside the Philippines, the bid prices may be quoted in the local currency or tradeable currency accepted by the BSP at the discretion of the Bidder. However, for purposes of bid evaluation, Bids denominated in foreign currencies, shall be converted to Philippine currency based on the exchange rate as published in the BSP reference rate bulletin on the day of the bid opening.
- 13.2. Payment of the contract price shall be made in:
 - a. Philippine Pesos.

14. Bid Security

- 14.1. The Bidder shall submit a Bid Securing Declaration¹ or any form of Bid Security in the amount indicated in the **BDS**, which shall be not less than the percentage of the ABC in accordance with the schedule in the **BDS**.
- 14.2. The Bid and bid security in no case shall exceed One Hundred Twenty (120) calendar days from the date of opening of bids, unless duly extended by the bidder upon the request of the Head of the Procuring Entity (HoPE) of the Quezon City Local Government. Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.

15. Sealing and Marking of Bids

Each Bidder shall submit one copy of the first and second components of its Bid.

The Procuring Entity may request additional hard copies and/or electronic copies of the Bid. However, failure of the Bidders to comply with the said request shall not be a ground for disqualification.

¹ In the case of Framework Agreement, the undertaking shall refer to entering into contract with the Procuring Entity and furnishing of the performance security or the performance securing declaration within ten (10) calendar days from receipt of Notice to Execute Framework Agreement.

If the Procuring Entity allows the submission of bids through online submission or any other electronic means, the Bidder shall submit an electronic copy of its Bid, which must be digitally signed. An electronic copy that cannot be opened or is corrupted shall be considered non-responsive and, thus, automatically disqualified.

16. Deadline for Submission of Bids

16.1. The Bidders shall submit on the specified date and time through manual submission as indicated in paragraph 7 of the IB.

17. Opening and Preliminary Examination of Bids

17.1. The BAC shall open the Bids in public at the time, on the date, and at the place specified in paragraph 9 of the **IB**. The Bidders' representatives who are present shall sign a register evidencing their attendance. In case videoconferencing, webcasting or other similar technologies will be used, attendance of participants shall likewise be recorded by the BAC Secretariat.

In case the Bids cannot be opened as scheduled due to justifiable reasons, the rescheduling requirements under Section 29 of the 2016 revised IRR of RA No. 9184 shall prevail.

17.2. The preliminary examination of bids shall be governed by Section 30 of the 2016 revised IRR of RA No. 9184.

18. Domestic Preference

18.1. The Procuring Entity will grant a margin of preference for the purpose of comparison of Bids in accordance with Section 43.1.2 of the 2016 revised IRR of RA No. 9184.

19. Detailed Evaluation and Comparison of Bids

- 19.1. The Procuring BAC shall immediately conduct a detailed evaluation of all Bids rated "*passed*," using non-discretionary pass/fail criteria. The BAC shall consider the conditions in the evaluation of Bids under Section 32.2 of the 2016 revised IRR of RA No. 9184.
- 19.2. If the Project allows partial bids, bidders may submit a proposal on any of the lots or items, and evaluation will be undertaken on a per lot or item basis, as the case maybe. In this case, the Bid Security as required by **ITB** Clause 15 shall be submitted for each lot or item separately.
- 19.3. The descriptions of the lots or items shall be indicated in Section VII (Technical Specifications), although the ABCs of these lots or items are indicated in the BDS for purposes of the NFCC computation pursuant to Section 23.4.2.6 of the 2016 revised IRR of RA No. 9184. The NFCC must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder.
- 19.4. The Project shall be awarded as follows:

One Project having several items that shall be awarded as one contract.

19.5. Except for bidders submitting a committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation, all Bids must include the

NFCC computation pursuant to Section 23.4.1.4 of the 2016 revised IRR of RA No. 9184, which must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder. For bidders submitting the committed Line of Credit, it must be at least equal to ten percent (10%) of the ABCs for all the lots or items participated in by the prospective Bidder.

20. Post-Qualification

20.1. Within a non-extendible period of five (5) calendar days from receipt by the Bidder of the notice from the BAC that it submitted the Lowest Calculated Bid, the Bidder shall submit its latest income and business tax returns filed and paid through the BIR Electronic Filing and Payment System (eFPS) and other appropriate licenses and permits required by law and stated in the **BDS**.

21. Signing of the Contract

21.1. The documents required in Section 37.2 of the 2016 revised IRR of RA No. 9184 shall form part of the Contract. Additional Contract documents are indicated in the **BDS**.

Notes on the Bid Data Sheet

The Bid Data Sheet (BDS) consists of provisions that supplement, amend, or specify in detail, information, or requirements included in the ITB found in Section II, which are specific to each procurement.

This Section is intended to assist the Procuring Entity in providing the specific information in relation to corresponding clauses in the ITB and has to be prepared for each specific procurement.

The Procuring Entity should specify in the BDS information and requirements specific to the circumstances of the Procuring Entity, the processing of the procurement, and the bid evaluation criteria that will apply to the Bids. In preparing the BDS, the following aspects should be checked:

- a. Information that specifies and complements provisions of the ITB must be incorporated.
- b. Amendments and/or supplements, if any, to provisions of the ITB as necessitated by the circumstances of the specific procurement, must also be incorporated.

Bid Data Sheet

For this purpose, contracts similar to the Project shall be:						
a. A single contract similar to the item/s to be bid and must be at least fift y percent (50%) of the ABC.						
 b. Completed within the last three (3) years prior to the deadline for the submission and receipt of bids substantially in a FORM prescribed by the QC-BAC-GOODS AND SERVICES, must be accompanied by a cop of Certificate of Acceptance by the end-user or Official Receipt (O.I. or Sales Invoice (S.I.) issued for the Contract. 						
Subcontracting is not allowed.						
The price of the Goods shall be quoted DDP within Quezon City or the applicable International Commercial Terms (INCOTERMS) for this Project.						
The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:						
 a. The amount of not less than <i>Php 90,000.00</i> or equivalent to two percent (2%) of ABC if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or 						
 b. The amount of not less than <i>Php 225,000.00</i> or equivalent to five percent (5%) of ABC if bid security is in Surety Bond. 						
[In case the Project will be awarded by lot, list the grouping of lots by specifying the group title, items, and the quantity for every identified lot, and the corresponding ABC for each lot.]						
[In case the project will be awarded by item, list each item indicating its quantity and ABC.]						
List of required licenses and permits relevant to the Project and the corresponding law requiring it.						
 No additional requirement 						
Additional required documents relevant to the Project that are required by existing laws and/or the Procuring Entity.						
• Copy of valid, current License to Operate from DOH Accreditation as Supplier, Distributor or Manufacturer for Medical or Hospital Equipment or Devices.						
• Statement of Warranty with Project Number and Project Title - Minimum of One (1) year for parts and service including Preventive Maintenance Schedule (PMS) and Calibration of Machine						

/

Notes on the General Conditions of Contract

The General Conditions of Contract (GCC) in this Section, read in conjunction with the Special Conditions of Contract in Section V and other documents listed therein, should be a complete document expressing all the rights and obligations of the parties.

Matters governing performance of the Supplier, payments under the contract, or matters affecting the risks, rights, and obligations of the parties under the contract are included in the GCC and Special Conditions of Contract.

Any complementary information, which may be needed, shall be introduced only through the Special Conditions of Contract.

1. Scope of Contract

This Contract shall include all such items, although not specifically mentioned, that can be reasonably inferred as being required for its completion as if such items were expressly mentioned herein. All the provisions of RA No. 9184 and its 2016 revised IRR, including the Generic Procurement Manual, and associated issuances, constitute the primary source for the terms and conditions of the Contract, and thus, applicable in contract implementation. Herein clauses shall serve as the secondary source for the terms and conditions of the Contract.

This is without prejudice to Sections 74.1 and 74.2 of the 2016 revised IRR of RA No. 9184 allowing the GPPB to amend the IRR, which shall be applied to all procurement activities, the advertisement, posting, or invitation of which were issued after the effectivity of the said amendment.

Additional requirements for the completion of this Contract shall be provided in the **Special Conditions of Contract (SCC)**.

2. Advance Payment and Terms of Payment

- 2.1. Advance payment of the contract amount is provided under Annex "D" of the revised 2016 IRR of RA No. 9184.
- 2.2. The Procuring Entity is allowed to determine the terms of payment on the partial or staggered delivery of the Goods procured, provided such partial payment shall correspond to the value of the goods delivered and accepted in accordance with prevailing accounting and auditing rules and regulations. The terms of payment are indicated in the SCC.

3. Performance Security

Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than prior to the signing of the Contract by both parties, the successful Bidder shall furnish the performance security in any of the forms prescribed in Section 39 of the 2016 revised IRR of RA No. 9184.

4. Inspection and Tests

The Procuring Entity or its representative shall have the right to inspect and/or to test the Goods to confirm their conformity to the Project specifications at no extra cost to the Procuring Entity in accordance with the Generic Procurement Manual. In addition to tests in the SCC, Section IV (Technical Specifications) shall specify what inspections and/or tests the Procuring Entity requires, and where they are to be conducted. The Procuring Entity shall notify the Supplier in writing, in a timely manner, of the identity of any representatives retained for these purposes.

All reasonable facilities and assistance for the inspection and testing of Goods, including access to drawings and production data, shall be provided by the Supplier to the authorized inspectors at no charge to the Procuring Entity.

5. Warranty

6.1. In order to assure that manufacturing defects shall be corrected by the Supplier, a warranty shall be required from the Supplier as provided under Section 62.1 of the 2016 revised IRR of RA No. 9184. 6.2. The Procuring Entity shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall, repair or replace the defective Goods or parts thereof without cost to the Procuring Entity, pursuant to the Generic Procurement Manual.

6. Liability of the Supplier

The Supplier's liability under this Contract shall be as provided by the laws of the Republic of the Philippines.

If the Supplier is a joint venture, all partners to the joint venture shall be jointly and severally liable to the Procuring Entity.

Section V. Special Conditions of Contract

Notes on the Special Conditions of Contract

Similar to the BDS, the clauses in this Section are intended to assist the Procuring Entity in providing contract-specific information in relation to corresponding clauses in the GCC found in Section IV.

The Special Conditions of Contract (SCC) complement the GCC, specifying contractual requirements linked to the special circumstances of the Procuring Entity, the Procuring Entity's country, the sector, and the Goods purchased. In preparing this Section, the following aspects should be checked:

- a. Information that complements provisions of the GCC must be incorporated.
- b. Amendments and/or supplements to provisions of the GCC as necessitated by the circumstances of the specific purchase, must also be incorporated.

However, no special condition which defeats or negates the general intent and purpose of the provisions of the GCC should be incorporated herein.

Special Conditions of Contract

GCC Clause	
1	[List here any additional requirements for the completion of this Contract. The following requirements and the corresponding provisions may be deleted, amended, or retained depending on its applicability to this Contract:]
	Delivery and Documents –
	For purposes of the Contract, "EXW," "FOB," "FCA," "CIF," "CIP," "DDP" and other trade terms used to describe the obligations of the parties shall have the meanings assigned to them by the current edition of INCOTERMS published by the International Chamber of Commerce, Paris. The Delivery terms of this Contract shall be as follows:
	[For Goods supplied from abroad, state:] "The delivery terms applicable to the Contract are DDP delivered [indicate place of destination]. In accordance with INCOTERMS."
	[For Goods supplied from within the Philippines, state:] "The delivery terms applicable to this Contract are delivered [indicate place of destination]. Risk and title will pass from the Supplier to the Procuring Entity upon receipt and final acceptance of the Goods at their final destination."
	Delivery of the Goods shall be made by the Supplier in accordance with the terms specified in Section VI (Schedule of Requirements).
	For purposes of this Clause the Procuring Entity's Representative at the Project Site is [indicate name(s)].
	Incidental Services –
	The Supplier is required to provide all of the following services, including additional services, if any, specified in Section VI. Schedule of Requirements: Select appropriate requirements and delete the rest.
	a. performance or supervision of on-site assembly and/or start-up of the supplied Goods;
	 b. furnishing of tools required for assembly and/or maintenance of the supplied Goods; c. furnishing of a detailed operations and maintenance manual for each
	 appropriate unit of the supplied Goods; d. performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and
	e. training of the Procuring Entity's personnel, at the Supplier's plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied Goods.
	f. [Specify additional incidental service requirements, as needed.]
	The Contract price for the Goods shall include the prices charged by the Supplier for incidental services and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.

Spare Parts
The Supplier is required to provide all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the Supplier:
Select appropriate requirements and delete the rest.
a. such spare parts as the Procuring Entity may elect to purchase from the Supplier, provided that this election shall not relieve the Supplier of any warranty obligations under this Contract; and
b. in the event of termination of production of the spare parts:
i. advance notification to the Procuring Entity of the pending termination, in sufficient time to permit the Procuring Entity to procure needed requirements; and
ii. following such termination, furnishing at no cost to the Procuring Entity, the blueprints, drawings, and specifications of the spare parts, if requested.
The spare parts and other components required are listed in Section VI (Schedule of Requirements) and the cost thereof are included in the contract price.
The Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spare parts or components for the Goods for a period of [indicate here the time period specified. If not used indicate a time period of three times the warranty period].
Spare parts or components shall be supplied as promptly as possible, but in any case, within [<i>insert appropriate time period</i>] months of placing the order.
Packaging –
The Supplier shall provide such packaging of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in this Contract. The packaging shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit, and open storage. Packaging case size and weights shall take into consideration, where appropriate, the remoteness of the Goods' final destination and the absence of heavy handling facilities at all points in transit.
The packaging, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified below, and in any subsequent instructions ordered by the Procuring Entity.
The outer packaging must be clearly marked on at least four (4) sides as follows:
Name of the Procuring Entity Name of the Supplier

	Contract Description
	Final Destination
	Gross weight
	Any special lifting instructions
;	Any special handling instructions
	Any relevant HAZCHEM classifications
	A packaging list identifying the contents and quantities of the package is to be placed on an accessible point of the outer packaging if practical. If not practical the packaging list is to be placed inside the outer packaging but outside the secondary packaging.
	Transportation –
	Where the Supplier is required under Contract to deliver the Goods CIF, CIP, or DDP, transport of the Goods to the port of destination or such other named place of destination in the Philippines, as shall be specified in this Contract, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract Price.
	Where the Supplier is required under this Contract to transport the Goods to a specified place of destination within the Philippines, defined as the Project Site, transport to such place of destination in the Philippines, including insurance and storage, as shall be specified in this Contract, shall be arranged by the Supplier, and related costs shall be included in the contract price.
	Where the Supplier is required under Contract to deliver the Goods CIF, CIP or DDP, Goods are to be transported on carriers of Philippine registry. In the event that no carrier of Philippine registry is available, Goods may be shipped by a carrier which is not of Philippine registry provided that the Supplier obtains and presents to the Procuring Entity certification to this effect from the nearest Philippine registry are available but their schedule delays the Supplier in its performance of this Contract the period from when the Goods were first ready for shipment and the actual date of shipment the period of delay will be considered force majeure.
	The Procuring Entity accepts no liability for the damage of Goods during transit other than those prescribed by INCOTERMS for DDP deliveries. In the case of Goods supplied from within the Philippines or supplied by domestic Suppliers risk and title will not be deemed to have passed to the Procuring Entity until their receipt and final acceptance at the final destination.
	Intellectual Property Rights –
	The Supplier shall indemnify the Procuring Entity against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the Goods or any part thereof.
2.2	[If partial payment is allowed, state] "The terms of payment shall be as follows:
4	The inspections and tests that will be conducted are: Product Presentation/Demonstration/Site Inspection, if applicable.

Section VI. Schedule of Requirements

PROJECT NAME: PROCUREMENT OF ULTRASOUND MACHINE PROJECT NO. RMBGH-25-HME-0742

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item Number	Description	Unit of Issue	Quantity	Delivered, Weeks / Months
1.	ULTRASOUND MACHINE	Unit	1 '	
	PARAMETERS:			
	Dimensions and Weight			
	Height with Monitor - at least 1565 mm			
	Width - at least 585 mm			
	Depth - at least 670 mm			
	Weight (no peripherals): not more than 60 kg			
	Power Supply			
	Voltage: 100 - 240 VAC			
	Frequency: 50/60 Hz			
	Power consumption maximum of not more than			
	350VA with peripherals			
	Console Design			
	at least 3 active probe ports			i
	Integrated SSD: at least 512 GB			
	Integrated Speakers			
	Easily removable air filters			
	Probe holders, removable for cleaning and washing			
	Gel Holder, removable for cleaning and washing			
	Gel warmer			Within Sixty (60)
	Probe cable management slots			Calendar Days
	Front and rear handles			upon issuance of
	Wheels: Locking mechanism that provides rolling			Notice to Proceed
-	lock and caster swivel back			
	User Interface Operator Keyboard			
	Height Adjustable			
	Full alphanumeric keypad covered with washable			
	protection film			
	at least 8 TGC pods			
ļ	at least 13.3 inches touch panel support multi-			
	gesture recognition and user configuration lay-out			
	Monitor			
	at least 21.5 inches high-resolution			
	LED Monitor: 1920 x 1080 resolution			
	Fixed monitor arm			
	Full-down and lock mechanism for transportation			
	Brightness and contrast adjustment			
	Tilt angle: at least +25/-90 degrees			
	Rotate angle: at least -90, +90 degrees			
-	Applications			

\$0

Abdomen, Pediatric, Obstetrical, Transcranial, Cynecological, Cardiac, Small Parts, Thoracic/Pleural, Musculoskeletal, Transvaginal Vascular/Peripheral Vascular, Urological, Transvaginal Vascular/Peripheral Vascular, Urological, Transvaginal Vascular/Peripheral Vascular, Urological, Transvaginal Coder Phase Inversion Harmonic Imaging M-Mode Color M-Mode Color M-Mode Color M-Mode Color M-Mode Power Doppler Imaging PDI Directional PDI PW Doppler with high PRF Cine Memory Image Memory 384 MB of Cine Memory Selectable Cine Sequence for Cine review Proceptive Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Quad Image Cine Display Quad Least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamforming Continous dynamic r						
Gynecological, Cardiac, Small Parts, Thoracic/Pleural, Musculos/eletal, Transvaginal Vascular/Peripheral Vascular, Urological, Transrectal, Interventional Guidance Operating Modes B-Mode Coded Phase Inversion Harmonic Imaging M-Mode Color M-Mode Color M-Mode Color M-Mode Power Doppler Imaging PDI Directional PDI PW Doppler with high PRF Cine Memory Image Memory 364 MB of Cine Memory 364 MB of Cine Memory 364 MB of Cine Memory 364 MB of Cine Memory 364 Guide Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Quad Image Cine Display Cine review loop Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 2 pre-programmable categories Digital agile beamformer architecture at least 2 23,907 system scanning channels Displayed imaging depth: at least 1-33 cm Transmission focus: 1-8 focal points selectable Quad beamforming Continous dynamic renegie at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 17 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 <i>lfs</i> (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.		Abdomen, Pediatric,				
Small Parts, Thoracic/Ploual, Musculaskeletal, Transvaginal Vascular/Peripheral Vascular, Urological, Transrectal, Interventional Guidance Operating Modes B-Mode Coded Phase Inversion Harmonic Imaging M-Mode Color M-Mode Colar M-Mode Colar M-Mode Scollin Instructure </td <td></td> <td>Obstetrical, Transcranial,</td> <td></td> <td></td> <td></td> <td></td>		Obstetrical, Transcranial,				
Small Parts, Thoracic//Peupl. Musculaskeletal, Transvaginal Vascular/Peripheral Vascular, Urological, Transrectal, Interventional Guidance Operating Modes B-Mode Coded Phase Inversion Harmonic Imaging M-Mode Color Flow Mode Power Doppler Imaging PDI Directional PDI Pirectional PDI PW Doppler with high PRF Cine Memory Image Memory 384 MB of Cine Memory 384 MB of Cine Memory 384 MB of Cine Memory Selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Quad Image Cine Display Cine review loop Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 22,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive		Gynecological, Cardiac,				
Musculoskeletal, Transvaginal Vascular/Peripheral Vascular, Urological, Transrectal, Interventional Guidance Operating Modes B-Mode Coder Phase Inversion Harmonic Imaging M-Mode Color M-Mode Color M-Mode Color M-Mode Color M-Mode Color M-Mode Power Doppler Imaging PDI Directional PDI PW Doppler with high PRF Cine Memory Image Memory 384 MB of Cine Memory 384 MB of Cine Memory 384 MB of Cine Memory Selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/Calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Quad Image Cine Display Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 22,907 system scanning channels Displayed imaging depth: at least 1- 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Contin						
Vascular/Peripheral Vascular, Urological, Transrectal, Interventional Guidance Operating Modes B-Mode Coded Phase Inversion Harmonic Imaging M-Mode Color M-Mode Color Flow Mode Power Doppler Imaging PD1 Directional PD1 PW Doppler with high PRF Cine Memory Image Memory 384 MB of Cine Memory 384 MB of Cine Memory 384 MB of Cine Memory 384 MB of Cine Memory 5 selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Quad Image Cine Display Cine gauge and Cine image number display Cine review loop Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,007 system scanning channels Displayed imaging depth: at least 1- 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 226 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 17 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 (% (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.						
Urological, Transrectal, Interventional Guidance Operating Modes B-Mode Coded Phase Inversion Harmonic Imaging M-Mode Color Flow Mode Power Doppler Imaging PD1 Directional PD1 PW Doppler with high PRF Cine Memory Image Memory 384 MB of Cine Image Memory 384 MB of Cine Memory 384 MB of Cine Image Memory 384 MB of Cine Memory 384 MB of Cine Image Memory Dual Image Cine Display Cine gauge and Cine Image number display Cine gauge and Cine Image number display Cine review loop Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 223,907 system scanning channels Displayed Imaging depti: at least 1- 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continuum Gover a least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.		0				
Operating Modes B-Mode Code Phase Inversion Harmonic Imaging M-Mode Color M-Mode Color Flow Mode Power Doppler Imaging PDI Directional PDI PW Doppler with high PRF Cine Memory Image Memory 384 MB of Cine Memory 384 MB of Cine Memory Selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 23,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mh		-				
B-Mode Coded Phase Inversion Harmonic Imaging M-Mode Color Flow Mode Color Flow Mode Power Doppler imaging PD1 Directional PD1 PW Doppler with high PRF Cine Memory Image Memory 384 MB of Cine Memory 384 MB of Cine Memory Selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Cine review loop Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic reage: at least 226 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.		÷				
Coded Phase Inversion Harmonic Imaging M-Mode Color M-Mode Color Flow Mode Power Doppler Imaging PDI Directional PDI PW Doppler with high PRF Cine Memory Image Memory 384 MB of Cine Memory 384 MB of Cine Memory Selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Quad Image Cine Display Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 23,907 system scanning channels Displayed imaging depth: at least 1-33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.		1 0				
 M-Mode Color How Mode Color Flow Mode Power Doppler Imaging PDI Directional PDI PW Doppler with high PRF Cine Memory 384 MB of Cine Memory 384 MB of Cine Memory Selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Quad Image Cine Display Quad Image Cine Display Cine review loop Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformar architecture at least 23.907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic race: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder. 						
Color M-Mode Color Flow Mode Power Doppler Imaging PDI Directional PDI PW Doppler with high PRF Cine Memory Image Memory 384 MB of Cine Memory Selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Quad Image Cine Display Quad Image Cine Display Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 23,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency/wideband technology Max frame rate: at least 1790 I/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features						
Color Flow Mode Power Doppler Imaging PD1 Directional PD1 PW Doppler with high PRF Cine Memory Image Memory 384 MB of Cine Memory 384 MB of Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 2269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 i/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees <td< td=""><td></td><td></td><td></td><td></td><td></td><td></td></td<>						
Power Doppler Imaging PDI Directional PDI PW Doppler with high PRF Cine Memory Image Memory 384 MB of Cine Memory 384 MB of Cine Memory Selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Cine review loop Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 9 pre-programmable categories Digital agile beamformer architecture at least 9 ystem scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic raceive focus/aperture Shades of gray: at least 256 Systematic dynamic raceive focus/aperture Shades of gray: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard						
Directional PDI PW Doppler with high PRF Cine Memory Image Memory 384 MB of Cine Memory Selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Quad Image Cine Display Cine gauge and Cine image number display Cine review loop Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 226 Systematic dynamic range: at least 226 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.		Color Flow Mode				
PW Doppler with high PRF Cine Memory Image Memory 384 MB of Cine Memory Selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Quad Image Cine Display Cine gauge and Cine image number display Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 23,907 system scanning channels Displayed imaging depth: at least 1-33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 27		Power Doppler Imaging PDI				
Cine Memory Image Memory 384 MB of Cine Memory Selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Quad Image Cine Display Cine gauge and Cine image number display Cine review loop Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.		Directional PDI				
384 MB of Cine Memory Selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Cine gauge and Cine image number display Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto		PW Doppler with high PRF				
384 MB of Cine Memory Selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Cine gauge and Cine image number display Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto						
384 MB of Cine Memory Selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Cine gauge and Cine image number display Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto		Cine Memory Image Memory				
Selectable Cine Sequence for Cine review Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Quad Image Cine Display Cine gauge and Cine image number display Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi		, , ,				
Prospective Cine Mark Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Quad Image Cine Display Cine gauge and Cine image number display Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission foccus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of	.	-				
Measurements/calculations and annotations on Cine Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Quad Image Cine Display Cine gauge and Cine image number display Cine review loop Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1- 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic ca		-				
Playback Scrolling timeline memory Dual Image Cine Display Quad Image Cine Display Cine gauge and Cine image number display Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 23,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.		A 1				
Scrolling timeline memoryDual Image Cine DisplayQuad Image Cine DisplayCine gauge and Cine image number displayCine review loopCine review loopCine review speed: at least 11 stepsSystem Scanning Parametersat least 9 pre-programmable categoriesDigital agile beamformer architectureat least 223,907 system scanning channelsDisplayed imaging depth: at least 1 - 33 cmTransmission focus: 1 - 8 focal points selectableQuad beamformingContinous dynamic receive focus/apertureShades of gray: at least 256Systematic dynamic range: at least 269 dBAdjustable Field of View: up to 168 degrees(depending on probe)Frequency range: 1.7 to 18Mhz (probe dependent)Multifrequency/wideband technologyMax frame rate: at least 1790 f/s (depends on probesand modes)Image rotation: 0, 90, 180, 270 degreesSystem Standard FeaturesAuto Detection, Auto Measurement and AutoAnnotation (Optional)Semi-automatic calculation of the volume ofbladder.						
Dual Image Cine Display Quad Image Cine Display Cine gauge and Cine image number display Cine review loop Cine review speed: at least 11 stepsSystem Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.		•				
Quad Image Cine Display Cine gauge and Cine image number display Cine review loop Cine review speed: at least 11 stepsSystem Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamformingContinous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.						
Cine gauge and Cine image number display Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.		e				
Cine review loop Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.						
Cine review speed: at least 11 steps System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.						
System Scanning Parameters at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.		-				
at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.		Cine review speed: at least 11 steps				
at least 9 pre-programmable categories Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.						
Digital agile beamformer architecture at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamformingContinous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.		,				
at least 223,907 system scanning channels Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.						
Displayed imaging depth: at least 1 - 33 cm Transmission focus: 1 - 8 focal points selectable Quad beamforming Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.		•				
Transmission focus: 1 - 8 focal points selectableQuad beamformingContinous dynamic receive focus/apertureShades of gray: at least 256Systematic dynamic range: at least 269 dBAdjustable Field of View: up to 168 degrees(depending on probe)Frequency range: 1.7 to 18Mhz (probe dependent)Multifrequency/wideband technologyMax frame rate: at least 1790 f/s (depends on probesand modes)Image rotation: 0, 90, 180, 270 degreesSystem Standard FeaturesAuto Detection, Auto Measurement and AutoAnnotation (Optional)Semi-automatic calculation of the volume ofbladder.						
Quad beamformingContinous dynamic receive focus/apertureShades of gray: at least 256Systematic dynamic range: at least 269 dBAdjustable Field of View: up to 168 degrees(depending on probe)Frequency range: 1.7 to 18Mhz (probe dependent)Multifrequency/wideband technologyMax frame rate: at least 1790 f/s (depends on probesand modes)Image rotation: 0, 90, 180, 270 degreesSystem Standard FeaturesAuto Detection, Auto Measurement and AutoAnnotation (Optional)Semi-automatic calculation of the volume ofbladder.						
 Continous dynamic receive focus/aperture Shades of gray: at least 256 Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder. 		-				
Shades of gray: at least 256Systematic dynamic range: at least 269 dBAdjustable Field of View: up to 168 degrees(depending on probe)Frequency range: 1.7 to 18Mhz (probe dependent)Multifrequency/wideband technologyMax frame rate: at least 1790 f/s (depends on probesand modes)Image rotation: 0, 90, 180, 270 degreesSystem Standard FeaturesAuto Detection, Auto Measurement and AutoAnnotation (Optional)Semi-automatic calculation of the volume ofbladder.		Quad beamforming				
Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.	-	Continous dynamic receive focus/aperture		1		
Systematic dynamic range: at least 269 dB Adjustable Field of View: up to 168 degrees (depending on probe) Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes) Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.		Shades of gray: at least 256				
Adjustable Field of View: up to 168 degrees (depending on probe)Frequency range: 1.7 to 18Mhz (probe dependent) Multifrequency/wideband technology Max frame rate: at least 1790 f/s (depends on probes and modes)Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.						
(depending on probe)Frequency range: 1.7 to 18Mhz (probe dependent)Multifrequency/wideband technologyMax frame rate: at least 1790 f/s (depends on probesand modes)Image rotation: 0, 90, 180, 270 degreesSystem Standard FeaturesAuto Detection, Auto Measurement and AutoAnnotation (Optional)Semi-automatic calculation of the volume ofbladder.						
Frequency range: 1.7 to 18Mhz (probe dependent)Multifrequency/wideband technologyMax frame rate: at least 1790 f/s (depends on probesand modes)Image rotation: 0, 90, 180, 270 degreesSystem Standard FeaturesAuto Detection, Auto Measurement and AutoAnnotation (Optional)Semi-automatic calculation of the volume ofbladder.		,			1	
Multifrequency/wideband technologyMax frame rate: at least 1790 f/s (depends on probes and modes)Image rotation: 0, 90, 180, 270 degreesSystem Standard FeaturesAuto Detection, Auto Measurement and Auto Annotation (Optional)Semi-automatic calculation of the volume of bladder.						
Max frame rate: at least 1790 f/s (depends on probes and modes)Image rotation: 0, 90, 180, 270 degrees System Standard FeaturesAuto Detection, Auto Measurement and Auto Annotation (Optional)Semi-automatic calculation of the volume of bladder.						
and modes)Image rotation: 0, 90, 180, 270 degreesSystem Standard FeaturesAuto Detection, Auto Measurement and AutoAnnotation (Optional)Semi-automatic calculation of the volume ofbladder.						
Image rotation: 0, 90, 180, 270 degrees System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.						
System Standard Features Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.						
Auto Detection, Auto Measurement and Auto Annotation (Optional) Semi-automatic calculation of the volume of bladder.		•				
Annotation (Optional) Semi-automatic calculation of the volume of bladder.						
Semi-automatic calculation of the volume of bladder.				1		
bladder.						
Provides 3,5,7,9 frames of spatial compounding						
		Provides 3,5,7,9 frames of spatial compounding	İ	<u> </u>		
			[4-

Schedule of Requirements Page 2 of 5

,

1	Optimize the brightness, contrast and uniformity of				
	B-Mode images when scanning different tissues.				
	High Definition Speckle Reduction Imaging				
	Coded Phase Inversion Harmonic Imaging				
	Provides a convex field of view				
	Extended field of view imaging				
	Patient Information database				
	Image archive on integrated SSD				
	Raw Data Analysis				
	Voice Comments (Optional)				
	Real-time automatic Doppler Calculations				
	OB Calculations				
	SonoBiometry				
	Fetal trending				
	Multi gestational calculations				
	Gynelogical calculations				
	Breast Productivity (Optional)				
	Renal Calculations				
	Urological Calculations				
	Cardiac Calculations				
	On board reporting package				
	MPEGvue				
	Network storage				
	Remote Capability				
	Stress Echo (Optional)				
	TVM / TVI				
1	Anatomical and Curved M-Mode (AMM and				
	Curved AMM) (Optional)				
	Auto EF (Optional)				
	Auto IMT				
	Elastography (Optional)				
	Idle mode				
	Lateral Gain Compensation				
	Technology that allow operator to rotate light source				
	by tracking ball under 3D mode. Operator could				
	adjust the light from different direction to highlight				
	the cared area for a better clarity on targeted issue				
	Scan Assistant that provides an automated exam				
	script that moves you through an exam step by step.				
	Abstracted from basic user manual, it lists out FAQs				
	from customers and instructs customer how to solve				
	problems by themselves timely.				
	Static 3D / Real-time 4D				
	Advance 3D (Easy 3D)				
	Sectional Planes				
	B-Flow (B-Flow Color)		!		
	Allows you to obtain precise needle imaging in the				ľ
	dashed box. (Optional)				
1	Simplify the workflow that need to label or annotate				
	ultrasound images for documentation purposes		ļ		
	similar to manual annotation of ultrasound images.		ĺ		
	Contextual reference tool with clinical guidance for				
	scan plane acquisition and references for anatomical		1		
,	structures. It can be displayed on-demand by the				
	succession of the be uppingen on neurinin of the	L	· · · · · · · · · · · · · · · · · · ·	L	

Schedule of Requirements Page 3 of 5

,

user. Clinical reference images and animations to depict information related to each step.	
1r	
Thyroid measurement with including measurement	
and relevant description (includes TI-RADS ACR)	
Li Rads	
Intended to more accurately perform serial scans on	
a patient, and compare the images of a previous	
ultrasound exam with the current exam. (Optional)	
Transducer Types	
1. Convex Probe	
Application: Abdomen, OB, Gyn, Vascular,	
Urology, Thoracic, Pediatric, MSK	
FOV: at least 58 degrees	
Footprint: at least 66.2 x 18.3 mm	
Frequency: at least 2.0 and 5.0 MHz	
2. Linear Probe	
Application: Vascular, Small parts, Pediatrics,	
Musculoskeletal (conventional and superficial)	
Footprint: at least 47 x 11.4 mm	
Frequency: at least 4.0 and 13 MHz	
3. Endocavity Probe	
Application: OB, Gyn, Urology, Transvaginal,	
Transrectal	
FOV: at least 128 degrees	
Footprint: at least 16.9 x 21.2 mm	
Frequency: at least 4.2 and 10 MHz	
4. Micro Convex Probe	
Application: Pediatric, MSK Conventional, Cardiac	
Pediatric, Transcranial, Interventional Guidance	
FOV: at least 131 degrees	
Footprint: at least 22.0 x 12.0 mm	
Frequency: at least 4.2 and 10 MHz	
Inputs and Outputs	
RCA	
S-video output	
VGA output	
HDMI output	
Ethernet (RJ45)	
USB (at least 4 ports for use)	
Peripheral/Add-Ons	
Medical grade high resolution Integrated printers:	
B&W thermal printer for image documentation	
Integrated DVD/CD burning capability for storage	
of DICOM images or export in JPEG and Avi for PC	
Compatibility	
USB ECG kits	
Footswitch	
With AVR and UPS dedicated to each machine	
, and a set of the second to be and the second seco	
Warranty: One (1) year parts and service	

30

¥9

I hereby certify to comply and deliver all the above requirements.

Name:	 	
Legal Capacity:		
Signature:	 	
Duly authorized to sign the Bid for and behalf of:	 	

.

YN

Notes for Preparing the Technical Specifications

A set of precise and clear specifications is a prerequisite for Bidders to respond realistically and competitively to the requirements of the Procuring Entity without qualifying their Bids. In the context of Competitive Bidding, the specifications (*e.g.* production/delivery schedule, manpower requirements, and after-sales service/parts, descriptions of the lots or items) must be prepared to permit the widest possible competition and, at the same time, present a clear statement of the required standards of workmanship, materials, and performance of the goods and services to be procured. Only if this is done will the objectives of transparency, equity, efficiency, fairness, and economy in procurement be realized, responsiveness of bids be ensured, and the subsequent task of bid evaluation and post-qualification facilitated. The specifications should require that all items, materials and accessories to be included or incorporated in the goods be new, unused, and of the most recent or current models, and that they include or incorporate all recent improvements in design and materials unless otherwise provided in the Contract.

Samples of specifications from previous similar procurements are useful in this respect. The use of metric units is encouraged. Depending on the complexity of the goods and the repetitiveness of the type of procurement, it may be advantageous to standardize the General Technical Specifications and incorporate them in a separate subsection. The General Technical Specifications should cover all classes of workmanship, materials, and equipment commonly involved in manufacturing similar goods. Deletions or addenda should then adapt the General Technical Specifications to the particular procurement.

Care must be taken in drafting specifications to ensure that they are not restrictive. In the specification of standards for equipment, materials, and workmanship, recognized Philippine and international standards should be used as much as possible. Where other particular standards are used, whether national standards or other standards, the specifications should state that equipment, materials, and workmanship that meet other authoritative standards, and which ensure at least a substantially equal quality than the standards mentioned, will also be acceptable. The following clause may be inserted in the Special Conditions of Contract or the Technical Specifications.

Sample Clause: Equivalency of Standards and Codes

Wherever reference is made in the Technical Specifications to specific standards and codes to be met by the goods and materials to be furnished or tested, the provisions of the latest edition or revision of the relevant standards and codes shall apply, unless otherwise expressly stated in the Contract. Where such standards and codes are national or relate to a particular country or region, other authoritative standards that ensure substantial equivalence to the standards and codes specified will be acceptable.

Reference to brand name and catalogue number should be avoided as far as possible; where unavoidable they should always be followed by the words "or at least equivalent." References to brand names cannot be used when the funding source is the GOP.

Where appropriate, drawings, including site plans as required, may be furnished by the Procuring Entity with the Bidding Documents. Similarly, the Supplier may be requested to provide drawings or samples either with its Bid or for prior review by the Procuring Entity during contract execution.

Bidders are also required, as part of the technical specifications, to complete their statement of compliance demonstrating how the items comply with the specification.

Technical Specifications

PROJECT NAME: PROCUREMENT OF ULTRASOUND MACHINE * PROJECT NO. RMBGH-25-HME-0742 *

Item	Specification	Statement of Compliance
		[Bidders must state here either "Comply" or
		"Not Comply" against each of the individual
		parameters of each Specification stating the corresponding performance parameter of the
		equipment offered. Statements of "Comply" or
		"Not Comply" must be supported by evidence in
		a Bidders Bid and cross-referenced to that
		evidence. Evidence shall be in the form of
		manufacturer's un-amended sales literature,
		unconditional statements of specification and
		compliance issued by the manufacturer, samples,
		independent test data etc., as appropriate. A statement that is not supported by evidence or is
		subsequently found to be contradicted by the
		evidence presented will render the Bid under
		evaluation liable for rejection. A statement either
		in the Bidder's statement of compliance or the
		supporting evidence that is found to be false
		either during Bid evaluation, post-qualification
		or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier
		liable for prosecution subject to the applicable
		laws and issuances.]
A.1	ULTRASOUND MACHINE	······································
	PARAMETERS:	
	Dimensions and Weight	
	Height with Monitor - at least 1565 mm	
	Width - at least 585 mm	
	Depth - at least 670 mm	
	Weight (no peripherals): not more than 60 kg	
	Power Supply	
	Voltage: 100 - 240 VAC	
	Frequency: 50/60 Hz	
	Power consumption maximum of not more than	
	350VA with peripherals	
	Console Design	
	at least 3 active probe ports	
	Integrated SSD: at least 512 GB	
	Integrated Speakers	
	Easily removable air filters	
	Probe holders, removable for cleaning and washing	
	Gel Holder, removable for cleaning and washing	
	Gel warmer	
	Probe cable management slots	
	Front and rear handles	
	Wheels: Locking mechanism that provides rolling lock	
	and caster swivel back	
	User Interface Operator Keyboard	
	Height Adjustable	
	L Melont Adilistanie	

v0

	Full alphanumeric keypad covered with washable	
	protection film	
	at least 8 TGC pods	
	at least 13.3 inches touch panel support multi-gesture	
	recognition and user configuration lay-out	
	Monitor	
	at least 21.5 inches high-resolution	
	LED Monitor: 1920 x 1080 resolution	
	Fixed monitor arm	
	Full-down and lock mechanism for transportation	
	Brightness and contrast adjustment	
	Tilt angle: at least +25/-90 degrees	
:	Rotate angle: at least -90, +90 degrees	
	Applications	
	Abdomen, Pediatric,	
	Obstetrical, Transcranial,	
	Gynecological, Cardiac,	
	Small Parts, Thoracic/Pleural,	
	Musculoskeletal, Transvaginal Vasqular/Peripheral Vasqular	
	Vascular/Peripheral Vascular,	
	Urological, Transrectal, Interventional Guidance	
	Operating Modes	
	B-Mode	
	Coded Phase Inversion Harmonic Imaging	
	M-Mode	
	Color M-Mode	
	Color Flow Mode	
	Power Doppler Imaging PDI	
	Directional PDI	
	PW Doppler with high PRF	
	Cine Memory Image Memory	
	384 MB of Cine Memory	
	Selectable Cine Sequence for Cine review	
	Prospective Cine Mark	
	Measurements/calculations and annotations on Cine	
	Playback	
	Scrolling timeline memory	
	Dual Image Cine Display	
	Quad Image Cine Display	
	Cine gauge and Cine image number display	
	Cine review loop	
	Cine review speed: at least 11 steps	
	System Scanning Parameters	
	at least 9 pre-programmable categories	
	Digital agile beamformer architecture	
	at least 223,907 system scanning channels	
	Displayed imaging depth: at least 1 - 33 cm	
	Transmission focus: 1 - 8 focal points selectable	
	Quad beamforming	
	Continous dynamic receive focus/aperture	

.

Shades of gray: at least 256	
Systematic dynamic range: at least 269 dB	
Adjustable Field of View: up to 168 degrees (depending	
on probe)	
Frequency range: 1.7 to 18Mhz (probe dependent)	
Multifrequency/wideband technology	
Max frame rate: at least 1790 f/s (depends on probes	
and modes)	
Image rotation: 0, 90, 180, 270 degrees	
System Standard Features	
Auto Detection, Auto Measurement and Auto	
Annotation (Optional)	
Semi-automatic calculation of the volume of bladder.	
Provides 3,5,7,9 frames of spatial compounding	
Optimize the brightness, contrast and uniformity of B-	
Mode images when scanning different tissues.	
High Definition Speckle Reduction Imaging	
Coded Phase Inversion Harmonic Imaging	
Provides a convex field of view	
Extended field of view imaging	
Patient Information database	
Image archive on integrated SSD	
Raw Data Analysis	
Voice Comments (Optional)	
Real-time automatic Doppler Calculations	
OB Calculations	
SonoBiometry	
Fetal trending	
Multi gestational calculations	
Gynelogical calculations	
Breast Productivity (Optional)	
Renal Calculations	
Urological Calculations	
Cardiac Calculations	
On board reporting package	
MPEGvue	
Network storage	
Remote Capability	
Stress Echo (Optional)	
TVM / TVI	
Anatomical and Curved M-Mode (AMM and Curved	
AMM) (Optional)	
Auto EF (Optional)	
Auto IMT	
Elastography (Optional)	
Idle mode	
Lateral Gain Compensation	
Technology that allow operator to rotate light source	
by tracking ball under 3D mode. Operator could adjust	
the light from different direction to highlight the cared	
area for a better clarity on targeted issue	
Scan Assistant that provides an automated exam script	
that moves you through an exam step by step.	

Technical Specifications Page 3 of 5

Abstracted from basic user manual, it lists out FAQs	
from customers and instructs customer how to solve	
problems by themselves timely.	
Static 3D / Real-time 4D	
Advance 3D (Easy 3D)	
Sectional Planes	
B-Flow (B-Flow Color)	
Allows you to obtain precise needle imaging in the	
dashed box. (Optional)	
Simplify the workflow that need to label or annotate	
· · ·	
ultrasound images for documentation purposes similar	
to manual annotation of ultrasound images.	
Contextual reference tool with clinical guidance for	
scan plane acquisition and references for anatomical	
structures. It can be displayed on-demand by the user.	
Clinical reference images and animations to depict	
information related to each step.	
Thyroid measurement with including measurement	
and relevant description (includes TI-RADS ACR)	
Li Rads	
Intended to more accurately perform serial scans on a	
patient, and compare the images of a previous	
ultrasound exam with the current exam. (Optional)	
Transducer Types	
1. Convex Probe	
Application: Abdomen, OB, Gyn, Vascular,	
Urology, Thoracic, Pediatric, MSK	
FOV: at least 58 degrees	
Footprint: at least 66.2 x 18.3 mm	
Frequency: at least 2.0 and 5.0 MHz	
2. Linear Probe	
Application: Vascular, Small parts, Pediatrics,	
Musculoskeletal (conventional and superficial)	
Footprint: at least 47 x 11.4 mm	
Frequency: at least 4.0 and 13 MHz	
3. Endocavity Probe	
Application: OB, Gyn, Urology, Transvaginal,	
Transrectal	
FOV: at least 128 degrees	
Footprint: at least 16.9 x 21.2 mm	
Frequency: at least 4.2 and 10 MHz	
4. Micro Convex Probe	
Application: Pediatric, MSK Conventional, Cardiac	
Pediatric, Transcranial, Interventional Guidance	
FOV: at least 131 degrees	
Footprint: at least 22.0 x 12.0 mm	
Frequency: at least 4.2 and 10 MHz	
Inputs and Outputs	
RCA	
S-video output	
VGA output	
HDMI output	
Ethernet (RJ45)	

.

	USB (at least 4 ports for use)	
	Peripheral/Add-Ons Medical grade high resolution Integrated printers:	
	B&W thermal printer for image documentation	
	Integrated DVD/CD burning capability for storage of	
	DICOM images or export in JPEG and Avi for PC	
	Compatibility	
	USB ECG kits	
	Footswitch	
	With AVR and UPS dedicated to each machine	
	Warranty: One (1) year parts and service	
B.	Compliance to the Schedule of Requirements	
	(Section VI)	

I hereby certify to comply and deliver all the above requirements.

Name: _____

Legal Capacity: _____

Signature: _____

Duly authorized to sign the Bid for and behalf of: _____

Technical Specifications Page 5 of 5

Section VIII. Checklist of Technical and Financial Documents

Notes on the Checklist of Technical and Financial Documents

The prescribed documents in the checklist are mandatory to be submitted in the Bid, but shall be subject to the following:

- a. GPPB Resolution No. 09-2020 on the efficient procurement measures during a State of Calamity or other similar issuances that shall allow the use of alternate documents in lieu of the mandated requirements; or
- b. Any subsequent GPPB issuances adjusting the documentary requirements after the effectivity of the adoption of the PBDs.

The BAC shall be checking the submitted documents of each Bidder against this checklist to ascertain if they are all present, using a non-discretionary "pass/fail" criterion pursuant to Section 30 of the 2016 revised IRR of RA No. 9184.

Checklist of Technical and Financial Documents I. TECHNICAL COMPONENT ENVELOPE

Class "A" Documents

Legal Documents

(a) Valid PhilGEPS Registration Certificate (Platinum Membership) (all pages) in accordance with Section 8.5.2 of the IRR;

Technical Documents

- (b) Statement of the prospective bidder of all its ongoing government and private contracts, including contracts awarded but not yet started, if any, whether similar or not similar in nature and complexity to the contract to be bid (in a FORM prescribed by the QC-BAC-GOODS AND SERVICES); and
- (c) Statement of the bidder's Single Largest Completed Contract (SLCC) similar to the contract to be bid, except under conditions provided for in Sections 23.4.1.3 and 23.4.2.4 of the 2016 revised IRR of RA No. 9184, within the relevant period as provided in the Bidding Documents (in a FORM prescribed by the QC-BAC-GOODS AND SERVICES); and
- (d) Original copy of Bid Security. If in the form of a Surety Bond, submit also a certification issued by the Insurance Commission;
 Or

Original copy of Notarized Bid Securing Declaration; and

- (e) Conformity with Section VI. (Schedule of Requirements) and Section VII. (Technical Specifications), which may include production/delivery schedule, manpower requirements, and/or after-sales/parts, if applicable; **and**
- (f) Original duly signed Omnibus Sworn Statement (OSS);
 and if applicable, Original Notarized Secretary's Certificate in case of a corporation, partnership, or cooperative; or Original Special Power of Attorney of all members of the joint venture giving full power and authority to its officer to sign the OSS and do acts to represent the Bidder.

Financial Documents

(g) The prospective bidder's computation of Net Financial Contracting Capacity (NFCC) (in a FORM prescribed by the QC-BAC-GOODS AND SERVICES);

<u>or</u>

A committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation.

Class "B" Documents

(h) If applicable, a duly signed joint venture agreement (JVA) in case the joint venture is already in existence;

<u>or</u>

duly notarized statements from all the potential joint venture partners stating that they will enter into and abide by the provisions of the JVA in the instance that the bid is successful.

Other documentary requirements under RA No. 9184 (as applicable)

- (i) [For foreign bidders claiming by reason of their country's extension of reciprocal rights to Filipinos] Certification from the relevant government office of their country stating that Filipinos are allowed to participate in government procurement activities for the same item or product.
- (j) Certification from the DTI if the Bidder claims preference as a Domestic Bidder or Domestic Entity.

II. FINANCIAL COMPONENT ENVELOPE

- (a) Original of duly signed and accomplished Financial Bid Form; and
- (b) Original of duly signed and accomplished Price Schedule(s).

7

III. REQUIRED DOCUMENTS in BDS SECTION 20.2 and 21.2

• Copy of valid, current License to Operate from DOH Accreditation as Supplier, Distributor or Manufacturer for Medical or Hospital Equipment or Devices.

1

\$

• Statement of Warranty with Project Number and Project Title - Minimum of One (1) year for parts and service including Preventive Maintenance Schedule (PMS) and Calibration of Machine

Note:

- 1. Please refer to [https://drive.google.com/file/d/1uiYurh5WrpBL5B_pqpzAb62yucAblR1p/view?usp=sh aring] for the following requirements:
 - a. Computation of NFCC;
 - b. List of All Ongoing Contracts/List of Contracts already awarded but not yet started;
 - c. Statement of Single Largest Completed Contract
- 2. Please refer to GPPB Resolution No. 16-2020 for the following requirements:
 - a. Bid Form;
 - b. Price Schedule (for Goods Offered from Abroad/ Within the Philippines)
 - c. Bid Securing Declaration; and
 - d. Omnibus Sworn Statement

