

REPUBLIC OF THE PHILIPPINES QUEZON CITY GOVERNMENT BIDS AND AWARDS COMMITTEE – GOODS AND SERVICES



PHILIPPINE BIDDING DOCUMENTS

(As Harmonized with Development Partners)

PROCUREMENT OF INFANT INCUBATOR, RADIANT WARMER AND OTHERS

PROJECT NO. RMBGH-25-HME-0766

Government of the Republic of the Philippines

Sixth Edition July 2020

Preface

These Philippine Bidding Documents (PBDs) for the procurement of Goods through Competitive Bidding have been prepared by the Government of the Philippines for use by any branch, constitutional commission or office, agency, department, bureau, office, or instrumentality of the Government of the Philippines, National Government Agencies, including Government-Owned and/or Controlled Corporations, Government Financing Institutions, State Universities and Colleges, and Local Government Unit. The procedures and practices presented in this document have been developed through broad experience, and are for mandatory use in projects that are financed in whole or in part by the Government of the Philippines or any foreign government/foreign or international financing institution in accordance with the provisions of the 2016 revised Implementing Rules and Regulations of Republic Act No. 9184.

The Bidding Documents shall clearly and adequately define, among others: (i) the objectives, scope, and expected outputs and/or results of the proposed contract or Framework Agreement, as the case may be; (ii) the eligibility requirements of Bidders; (iii) the expected contract or Framework Agreement duration, the estimated quantity in the case of procurement of goods, delivery schedule and/or time frame; and (iv) the obligations, duties, and/or functions of the winning bidder.

Care should be taken to check the relevance of the provisions of the PBDs against the requirements of the specific Goods to be procured. If duplication of a subject is inevitable in other sections of the document prepared by the Procuring Entity, care must be exercised to avoid contradictions between clauses dealing with the same matter.

Moreover, each section is prepared with notes intended only as information for the Procuring Entity or the person drafting the Bidding Documents. They shall not be included in the final documents. The following general directions should be observed when using the documents:

- a. All the documents listed in the Table of Contents are normally required for the procurement of Goods. However, they should be adapted as necessary to the circumstances of the particular Procurement Project.
- b. Specific details, such as the "name of the Procuring Entity" and "address for bid submission," should be furnished in the Instructions to Bidders, Bid Data Sheet, and Special Conditions of Contract. The final documents should contain neither blank spaces nor options.
- c. This Preface and the footnotes or notes in italics included in the Invitation to Bid, Bid Data Sheet, General Conditions of Contract, Special Conditions of Contract, Schedule of Requirements, and Specifications are not part of the text of the final document, although they contain instructions that the Procuring Entity should strictly follow.
- d. The cover should be modified as required to identify the Bidding Documents as to the Procurement Project, Project Identification Number, and Procuring Entity, in addition to the date of issue.

- e. Modifications for specific Procurement Project details should be provided in the Special Conditions of Contract as amendments to the Conditions of Contract. For easy completion, whenever reference has to be made to specific clauses in the Bid Data Sheet or Special Conditions of Contract, these terms shall be printed in bold typeface on Sections I (Instructions to Bidders) and III (General Conditions of Contract), respectively.
- f. For guidelines on the use of Bidding Forms and the procurement of Foreign-Assisted Projects, these will be covered by a separate issuance of the Government Procurement Policy Board.

Table of Contents

Gloss	ary of Acronyms, Terms, and Abbreviations	4
Sectio	on I. Invitation to Bid	7
Sectio	on II. Instructions to Bidders	11
1.	Scope of Bid	12
2.	Funding Information	12
3.	Bidding Requirements	12
4.	Corrupt, Fraudulent, Collusive, and Coercive Practices	12
5.	Eligible Bidders	12
6.	Origin of Goods	13
7.	Subcontracts	13
8.	Pre-Bid Conference	13
9.	Clarification and Amendment of Bidding Documents	13
10.	Documents comprising the Bid: Eligibility and Technical Components 144	
11.	Documents comprising the Bid: Financial Component	14
12.	Bid Prices	14
13.	Bid and Payment Currencies	
14.	Bid Security	
15.	Sealing and Marking of Bids	
16.	Deadline for Submission of Bids	
17.	Opening and Preliminary Examination of Bids	
18.	Domestic Preference	
19.	Detailed Evaluation and Comparison of Bids	
20.	Post-Qualification	
21.	Signing of the Contract	
Sectio	on III. Bid Data Sheet	18
Section	on IV. General Conditions of Contract	20
1.	Scope of Contract	21
2.	Advance Payment and Terms of Payment	21
3.	Performance Security	21
4.	Inspection and Tests	
5.	Warranty	21
6.	Liability of the Supplier	
Section	on V. Special Conditions of Contract	23
	on VI. Schedule of Requirements	
	on VII. Technical Specifications	
	on VIII. Checklist of Technical and Financial Documents	

Glossary of Acronyms, Terms, and Abbreviations

ABC - Approved Budget for the Contract.

BAC - Bids and Awards Committee.

Bid – A signed offer or proposal to undertake a contract submitted by a bidder in response to and in consonance with the requirements of the bidding documents. Also referred to as *Proposal* and *Tender*. (2016 revised IRR, Section 5[c])

Bidder – Refers to a contractor, manufacturer, supplier, distributor and/or consultant who submits a bid in response to the requirements of the Bidding Documents. (2016 revised IRR, Section 5[d])

Bidding Documents – The documents issued by the Procuring Entity as the bases for bids, furnishing all information necessary for a prospective bidder to prepare a bid for the Goods, Infrastructure Projects, and/or Consulting Services required by the Procuring Entity. (2016 revised IRR, Section 5[e])

BIR - Bureau of Internal Revenue.

BSP - Bangko Sentral ng Pilipinas.

Consulting Services – Refer to services for Infrastructure Projects and other types of projects or activities of the GOP requiring adequate external technical and professional expertise that are beyond the capability and/or capacity of the GOP to undertake such as, but not limited to: (i) advisory and review services; (ii) pre-investment or feasibility studies; (iii) design; (iv) construction supervision; (v) management and related services; and (vi) other technical services or special studies. (2016 revised IRR, Section 5[i])

CDA - Cooperative Development Authority.

Contract – Refers to the agreement entered into between the Procuring Entity and the Supplier or Manufacturer or Distributor or Service Provider for procurement of Goods and Services; Contractor for Procurement of Infrastructure Projects; or Consultant or Consulting Firm for Procurement of Consulting Services; as the case may be, as recorded in the Contract Form signed by the parties, including all attachments and appendices thereto and all documents incorporated by reference therein.

CIF - Cost Insurance and Freight.

CIP - Carriage and Insurance Paid.

CPI – Consumer Price Index.

DDP - Refers to the quoted price of the Goods, which means "delivered duty paid."

DTI – Department of Trade and Industry.

EXW – Ex works.

FCA - "Free Carrier" shipping point.

FOB - "Free on Board" shipping point.

Foreign-funded Procurement or Foreign-Assisted Project—Refers to procurement whose funding source is from a foreign government, foreign or international financing institution as specified in the Treaty or International or Executive Agreement. (2016 revised IRR, Section 5[b]).

Framework Agreement – Refers to a written agreement between a procuring entity and a supplier or service provider that identifies the terms and conditions, under which specific purchases, otherwise known as "Call-Offs," are made for the duration of the agreement. It is in the nature of an option contract between the procuring entity and the bidder(s) granting the procuring entity the option to either place an order for any of the goods or services identified in the Framework Agreement List or not buy at all, within a minimum period of one (1) year to a maximum period of three (3) years. (GPPB Resolution No. 27-2019)

GFI - Government Financial Institution.

GOCC – Government-owned and/or –controlled corporation.

Goods – Refer to all items, supplies, materials and general support services, except Consulting Services and Infrastructure Projects, which may be needed in the transaction of public businesses or in the pursuit of any government undertaking, project or activity, whether in the nature of equipment, furniture, stationery, materials for construction, or personal property of any kind, including non-personal or contractual services such as the repair and maintenance of equipment and furniture, as well as trucking, hauling, janitorial, security, and related or analogous services, as well as procurement of materials and supplies provided by the Procuring Entity for such services. The term "related" or "analogous services" shall include, but is not limited to, lease or purchase of office space, media advertisements, health maintenance services, and other services essential to the operation of the Procuring Entity. (2016 revised IRR, Section 5[r])

GOP – Government of the Philippines.

GPPB - Government Procurement Policy Board.

INCOTERMS – International Commercial Terms.

Infrastructure Projects – Include the construction, improvement, rehabilitation, demolition, repair, restoration or maintenance of roads and bridges, railways, airports, seaports, communication facilities, civil works components of information technology projects, irrigation, flood control and drainage, water supply, sanitation, sewerage and solid waste management systems, shore protection, energy/power and electrification facilities, national buildings, school buildings, hospital buildings, and other related construction projects of the government. Also referred to as *civil works or works*. (2016 revised IRR, Section 5[u])

LGUs - Local Government Units.

NFCC - Net Financial Contracting Capacity.

NGA - National Government Agency.

PhilGEPS - Philippine Government Electronic Procurement System.

Procurement Project — refers to a specific or identified procurement covering goods, infrastructure project or consulting services. A Procurement Project shall be described, detailed, and scheduled in the Project Procurement Management Plan prepared by the agency

which shall be consolidated in the procuring entity's Annual Procurement Plan. (GPPB Circular No. 06-2019 dated 17 July 2019)

PSA – Philippine Statistics Authority.

SEC - Securities and Exchange Commission.

SLCC – Single Largest Completed Contract.

Supplier – refers to a citizen, or any corporate body or commercial company duly organized and registered under the laws where it is established, habitually established in business and engaged in the manufacture or sale of the merchandise or performance of the general services covered by his bid. (Item 3.8 of GPPB Resolution No. 13-2019, dated 23 May 2019). Supplier as used in these Bidding Documents may likewise refer to a distributor, manufacturer, contractor, or consultant.

UN – United Nations.

Section I. Invitation to Bid

Notes on the Invitation to Bid

The Invitation to Bid (IB) provides information that enables potential Bidders to decide whether to participate in the procurement at hand. The IB shall be posted in accordance with Section 21.2 of the 2016 revised IRR of RA No. 9184.

Apart from the essential items listed in the Bidding Documents, the IB should also indicate the following:

- a. The date of availability of the Bidding Documents, which shall be from the time the IB is first advertised/posted until the deadline for the submission and receipt of bids;
- b. The place where the Bidding Documents may be acquired or the website where it may be downloaded;
- c. The deadline for the submission and receipt of bids; and
- d. Any important bid evaluation criteria (e.g., the application of a margin of preference in bid evaluation).

The IB should be incorporated in the Bidding Documents. The information contained in the IB must conform to the Bidding Documents and in particular to the relevant information in the Bid Data Sheet.

Section II. Instructions to Bidders

Notes on the Instructions to Bidders

This Section on the Instruction to Bidders (ITB) provides the information necessary for bidders to prepare responsive bids, in accordance with the requirements of the Procuring Entity. It also provides information on bid submission, eligibility check, opening and evaluation of bids, post-qualification, and on the award of contract.

1. Scope of Bid

The Procuring Entity, Quezon City Local Government wishes to receive Bids for the PROCUREMENT OF INFANT INCUBATOR, RADIANT WARMER AND OTHERS with identification number RMBGH-25-HME-0766.

[Note: The Project Identification Number is assigned by the Procuring Entity based on its own coding scheme and is not the same as the PhilGEPS reference number, which is generated after the posting of the bid opportunity on the PhilGEPS website.]

The Procurement Project (referred to herein as "Project") is composed of ten (10) items, the details of which are described in Section VII (Technical Specifications).

2. Funding Information

- 2.1. The GOP through the source of funding as indicated below for 2025 in the amount of EIGHTEEN MILLION FOUR THOUSAND PESOS AND 00/100 ONLY (Php18,004,000.00).
- 2.2. The source of funding is:
 - a. LGUs, the Annual or Supplemental Budget, as approved by the Sanggunian.

3. Bidding Requirements

The Bidding for the Project shall be governed by all the provisions of RA No. 9184 and its 2016 revised IRR, including its Generic Procurement Manuals and associated policies, rules and regulations as the primary source thereof, while the herein clauses shall serve as the secondary source thereof.

Any amendments made to the IRR and other GPPB issuances shall be applicable only to the ongoing posting, advertisement, or **IB** by the BAC through the issuance of a supplemental or bid bulletin.

The Bidder, by the act of submitting its Bid, shall be deemed to have verified and accepted the general requirements of this Project, including other factors that may affect the cost, duration and execution or implementation of the contract, project, or work and examine all instructions, forms, terms, and project requirements in the Bidding Documents.

4. Corrupt, Fraudulent, Collusive, and Coercive Practices

The Procuring Entity, as well as the Bidders and Suppliers, shall observe the highest standard of ethics during the procurement and execution of the contract. They or through an agent shall not engage in corrupt, fraudulent, collusive, coercive, and obstructive practices defined under Annex "I" of the 2016 revised IRR of RA No. 9184 or other integrity violations in competing for the Project.

5. Eligible Bidders

- 5.1. Only Bids of Bidders found to be legally, technically, and financially capable will be evaluated.
- 5.2. Foreign ownership exceeding those allowed under the rules may participate pursuant to:

- i. When a Treaty or International or Executive Agreement as provided in Section 4 of the RA No. 9184 and its 2016 revised IRR allow foreign bidders to participate;
- ii. Citizens, corporations, or associations of a country, included in the list issued by the GPPB, the laws or regulations of which grant reciprocal rights or privileges to citizens, corporations, or associations of the Philippines;
- iii. When the Goods sought to be procured are not available from local suppliers; or
- iv. When there is a need to prevent situations that defeat competition or restrain trade.
- 5.3. Pursuant to Section 23.4.1.3 of the 2016 revised IRR of RA No.9184, the Bidder shall have an SLCC that is at least one (1) contract similar to the Project the value of which, adjusted to current prices using the PSA's CPI, must be at least equivalent to:
 - a. For the procurement Non-Expendable Supplies and Services: The Bidder must have completed a single contract that is similar to this Project, equivalent to at least fifty percent (50%) of the ABC.
- 5.4. The Bidders shall comply with the eligibility criteria under Section 23.4.1 of the 2016 IRR of RA No. 9184.

6. Origin of Goods

There is no restriction on the origin of goods other than those prohibited by a decision of the UN Security Council taken under Chapter VII of the Charter of the UN, subject to Domestic Preference requirements under ITB Clause 18.

7. Subcontracts

7.1. The Bidder may subcontract portions of the Project to the extent allowed by the Procuring Entity as stated herein, but in no case more than twenty percent (20%) of the Project.

The Procuring Entity has prescribed that: Subcontracting is not allowed.

8. Pre-Bid Conference

The Procuring Entity will hold a pre-bid conference for this Project on the specified date and time and either at its physical address and/or through videoconferencing as indicated in paragraph 6 of the **IB**.

9. Clarification and Amendment of Bidding Documents

Prospective bidders may request for clarification on and/or interpretation of any part of the Bidding Documents. Such requests must be in writing and received by the Procuring Entity, either at its given address or through electronic mail indicated in the IB, at least ten (10) calendar days before the deadline set for the submission and receipt of Bids.

10. Documents comprising the Bid: Eligibility and Technical Components

- 10.1. The first envelope shall contain the eligibility and technical documents of the Bid as specified in Section VIII (Checklist of Technical and Financial Documents).
- 10.2. The Bidder's SLCC as indicated in **ITB** Clause 5.3 should have been completed within *the last three (3) years* prior to the deadline for the submission and receipt of bids.
- 10.3. If the eligibility requirements or statements, the bids, and all other documents for submission to the BAC are in foreign language other than English, it must be accompanied by a translation in English, which shall be authenticated by the appropriate Philippine foreign service establishment, post, or the equivalent office having jurisdiction over the foreign bidder's affairs in the Philippines. Similar to the required authentication above, for Contracting Parties to the Apostille Convention, only the translated documents shall be authenticated through an Apostille pursuant to GPPB Resolution No. 13-2019 dated 23 May 2019. The English translation shall govern, for purposes of interpretation of the bid.

11. Documents comprising the Bid: Financial Component

- 11.1. The second bid envelope shall contain the financial documents for the Bid as specified in Section VIII (Checklist of Technical and Financial Documents).
- 11.2. If the Bidder claims preference as a Domestic Bidder or Domestic Entity, a certification issued by DTI shall be provided by the Bidder in accordance with Section 43.1.3 of the 2016 revised IRR of RA No. 9184.
- 11.3. Any bid exceeding the ABC indicated in paragraph 1 of the IB shall not be accepted.
- 11.4. For Foreign-funded Procurement, a ceiling may be applied to bid prices provided the conditions are met under Section 31.2 of the 2016 revised IRR of RA No. 9184.

12. Bid Prices

- 12.1. Prices indicated on the Price Schedule shall be entered separately in the following manner:
 - a. For Goods offered from within the Procuring Entity's country:
 - i. The price of the Goods quoted EXW (ex-works, ex-factory, exwarehouse, ex-showroom, or off-the-shelf, as applicable);
 - ii. The cost of all customs duties and sales and other taxes already paid or payable;
 - iii. The cost of transportation, insurance, and other costs incidental to delivery of the Goods to their final destination; and
 - iv. The price of other (incidental) services, if any, listed in e.

- b. For Goods offered from abroad:
 - i. Unless otherwise stated in the BDS, the price of the Goods shall be quoted delivered duty paid (DDP) with the place of destination in the Philippines as specified in the BDS. In quoting the price, the Bidder shall be free to use transportation through carriers registered in any eligible country. Similarly, the Bidder may obtain insurance services from any eligible source country.
 - ii. The price of other (incidental) services, if any, as listed in Section VII (Technical Specifications).

13. Bid and Payment Currencies

- 13.1. For Goods that the Bidder will supply from outside the Philippines, the bid prices may be quoted in the local currency or tradeable currency accepted by the BSP at the discretion of the Bidder. However, for purposes of bid evaluation, Bids denominated in foreign currencies, shall be converted to Philippine currency based on the exchange rate as published in the BSP reference rate bulletin on the day of the bid opening.
- 13.2. Payment of the contract price shall be made in:
 - a. Philippine Pesos.

14. Bid Security

- 14.1. The Bidder shall submit a Bid Securing Declaration¹ or any form of Bid Security in the amount indicated in the **BDS**, which shall be not less than the percentage of the ABC in accordance with the schedule in the **BDS**.
- 14.2. The Bid and bid security in no case shall exceed One Hundred Twenty (120) calendar days from the date of opening of bids, unless duly extended by the bidder upon the request of the Head of the Procuring Entity (HoPE) of the Quezon City Local Government. Any Bid not accompanied by an acceptable bid security shall be rejected by the Procuring Entity as non-responsive.

15. Sealing and Marking of Bids

Each Bidder shall submit one copy of the first and second components of its Bid.

The Procuring Entity may request additional hard copies and/or electronic copies of the Bid. However, failure of the Bidders to comply with the said request shall not be a ground for disqualification.

If the Procuring Entity allows the submission of bids through online submission or any other electronic means, the Bidder shall submit an electronic copy of its Bid, which must be digitally signed. An electronic copy that cannot be opened or is corrupted shall be considered non-responsive and, thus, automatically disqualified.

16. Deadline for Submission of Bids

15

¹ In the case of Framework Agreement, the undertaking shall refer to entering into contract with the Procuring Entity and furnishing of the performance security or the performance securing declaration within ten (10) calendar days from receipt of Notice to Execute Framework Agreement.

16.1. The Bidders shall submit on the specified date and time through manual submission as indicated in paragraph 7 of the IB.

17. Opening and Preliminary Examination of Bids

17.1. The BAC shall open the Bids in public at the time, on the date, and at the place specified in paragraph 9 of the **IB**. The Bidders' representatives who are present shall sign a register evidencing their attendance. In case videoconferencing, webcasting or other similar technologies will be used, attendance of participants shall likewise be recorded by the BAC Secretariat.

In case the Bids cannot be opened as scheduled due to justifiable reasons, the rescheduling requirements under Section 29 of the 2016 revised IRR of RA No. 9184 shall prevail.

17.2. The preliminary examination of bids shall be governed by Section 30 of the 2016 revised IRR of RA No. 9184.

18. Domestic Preference

18.1. The Procuring Entity will grant a margin of preference for the purpose of comparison of Bids in accordance with Section 43.1.2 of the 2016 revised IRR of RA No. 9184.

19. Detailed Evaluation and Comparison of Bids

- 19.1. The Procuring BAC shall immediately conduct a detailed evaluation of all Bids rated "passed," using non-discretionary pass/fail criteria. The BAC shall consider the conditions in the evaluation of Bids under Section 32.2 of the 2016 revised IRR of RA No. 9184.
- 19.2. If the Project allows partial bids, bidders may submit a proposal on any of the lots or items, and evaluation will be undertaken on a per lot or item basis, as the case maybe. In this case, the Bid Security as required by ITB Clause 15 shall be submitted for each lot or item separately.
- 19.3. The descriptions of the lots or items shall be indicated in **Section VII (Technical Specifications)**, although the ABCs of these lots or items are indicated in the **BDS** for purposes of the NFCC computation pursuant to Section 23.4.2.6 of the 2016 revised IRR of RA No. 9184. The NFCC must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder.
- 19.4. The Project shall be awarded as follows:

One Project having several items that shall be awarded as one contract.

19.5. Except for bidders submitting a committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation, all Bids must include the NFCC computation pursuant to Section 23.4.1.4 of the 2016 revised IRR of RA No. 9184, which must be sufficient for the total of the ABCs for all the lots or items participated in by the prospective Bidder. For bidders submitting the committed Line of Credit, it must be at least equal to ten percent (10%) of the ABCs for all the lots or items participated in by the prospective Bidder.

20. Post-Qualification

20.1. Within a non-extendible period of five (5) calendar days from receipt by the Bidder of the notice from the BAC that it submitted the Lowest Calculated Bid, the Bidder shall submit its latest income and business tax returns filed and paid through the BIR Electronic Filing and Payment System (eFPS) and other appropriate licenses and permits required by law and stated in the BDS.

21. Signing of the Contract

21.1. The documents required in Section 37.2 of the 2016 revised IRR of RA No. 9184 shall form part of the Contract. Additional Contract documents are indicated in the **BDS**.

Section III. Bid Data Sheet

Notes on the Bid Data Sheet

The Bid Data Sheet (BDS) consists of provisions that supplement, amend, or specify in detail, information, or requirements included in the ITB found in Section II, which are specific to each procurement.

This Section is intended to assist the Procuring Entity in providing the specific information in relation to corresponding clauses in the ITB and has to be prepared for each specific procurement.

The Procuring Entity should specify in the BDS information and requirements specific to the circumstances of the Procuring Entity, the processing of the procurement, and the bid evaluation criteria that will apply to the Bids. In preparing the BDS, the following aspects should be checked:

- a. Information that specifies and complements provisions of the ITB must be incorporated.
- b. Amendments and/or supplements, if any, to provisions of the ITB as necessitated by the circumstances of the specific procurement, must also be incorporated.

Bid Data Sheet

ITB						
Clause 5.3	For this purpose, contracts similar to the Project shall be:					
	 a. A single contract similar to the item/s to be bid and must be at least Fifty percent (50%) of the ABC. b. Completed within the last three (3) years prior to the deadline for the submission and receipt of bids substantially in a FORM prescribed by the QC-BAC-GOODS AND SERVICES, must be accompanied by a copy of Certificate of Acceptance by the end-user or Official Receipt (O.R) 					
7.1	or Sales Invoice (S.I.) issued for the Contract. Subcontracting is not allowed.					
12	The price of the Goods shall be quoted DDP within Quezon City or the applicable International Commercial Terms (INCOTERMS) for this Project.					
14.1	The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts: a. The amount of not less than <i>Php 360,080.00</i> or equivalent to two percent (2%) of ABC if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or b. The amount of not less than <i>Php 900,200.00</i> or equivalent to five percent (5%) of ABC if bid security is in Surety Bond.					
19.3	BREAKDOWN OF APPROVED BUDGET FOR THE CONTRACT (ABC)					
	Item nos. 1-2 7,700,000.00					
	Item nos. 3-10 10,304,000.00 TOTAL P 18,004,000.00					
20.2	List of required licenses and permits relevant to the Project and the corresponding law requiring it.					
	No additional requirements					
21.2	Additional required documents relevant to the Project that are required by existing laws and/or the Procuring Entity. Copy of valid, current License to Operate from DOH Accreditation as Supplier, Distributor or Manufacturer for Medical or Hospital Equipment or Devices. Statement of Warranty with Project Number and Project Title - Minimum of One (1) Year.					

Section IV. General Conditions of Contract

Notes on the General Conditions of Contract

The General Conditions of Contract (GCC) in this Section, read in conjunction with the Special Conditions of Contract in Section V and other documents listed therein, should be a complete document expressing all the rights and obligations of the parties.

Matters governing performance of the Supplier, payments under the contract, or matters affecting the risks, rights, and obligations of the parties under the contract are included in the GCC and Special Conditions of Contract.

Any complementary information, which may be needed, shall be introduced only through the Special Conditions of Contract.

1. Scope of Contract

This Contract shall include all such items, although not specifically mentioned, that can be reasonably inferred as being required for its completion as if such items were expressly mentioned herein. All the provisions of RA No. 9184 and its 2016 revised IRR, including the Generic Procurement Manual, and associated issuances, constitute the primary source for the terms and conditions of the Contract, and thus, applicable in contract implementation. Herein clauses shall serve as the secondary source for the terms and conditions of the Contract.

This is without prejudice to Sections 74.1 and 74.2 of the 2016 revised IRR of RA No. 9184 allowing the GPPB to amend the IRR, which shall be applied to all procurement activities, the advertisement, posting, or invitation of which were issued after the effectivity of the said amendment.

Additional requirements for the completion of this Contract shall be provided in the Special Conditions of Contract (SCC).

2. Advance Payment and Terms of Payment

- 2.1. Advance payment of the contract amount is provided under Annex "D" of the revised 2016 IRR of RA No. 9184.
- 2.2. The Procuring Entity is allowed to determine the terms of payment on the partial or staggered delivery of the Goods procured, provided such partial payment shall correspond to the value of the goods delivered and accepted in accordance with prevailing accounting and auditing rules and regulations. The terms of payment are indicated in the SCC.

3. Performance Security

Within ten (10) calendar days from receipt of the Notice of Award by the Bidder from the Procuring Entity but in no case later than prior to the signing of the Contract by both parties, the successful Bidder shall furnish the performance security in any of the forms prescribed in Section 39 of the 2016 revised IRR of RA No. 9184.

4. Inspection and Tests

The Procuring Entity or its representative shall have the right to inspect and/or to test the Goods to confirm their conformity to the Project specifications at no extra cost to the Procuring Entity in accordance with the Generic Procurement Manual. In addition to tests in the SCC, Section IV (Technical Specifications) shall specify what inspections and/or tests the Procuring Entity requires, and where they are to be conducted. The Procuring Entity shall notify the Supplier in writing, in a timely manner, of the identity of any representatives retained for these purposes.

All reasonable facilities and assistance for the inspection and testing of Goods, including access to drawings and production data, shall be provided by the Supplier to the authorized inspectors at no charge to the Procuring Entity.

5. Warranty

6.1. In order to assure that manufacturing defects shall be corrected by the Supplier, a warranty shall be required from the Supplier as provided under Section 62.1 of the 2016 revised IRR of RA No. 9184.

6.2. The Procuring Entity shall promptly notify the Supplier in writing of any claims arising under this warranty. Upon receipt of such notice, the Supplier shall, repair or replace the defective Goods or parts thereof without cost to the Procuring Entity, pursuant to the Generic Procurement Manual.

6. Liability of the Supplier

The Supplier's liability under this Contract shall be as provided by the laws of the Republic of the Philippines.

If the Supplier is a joint venture, all partners to the joint venture shall be jointly and severally liable to the Procuring Entity.

Section V. Special Conditions of Contract

Notes on the Special Conditions of Contract

Similar to the BDS, the clauses in this Section are intended to assist the Procuring Entity in providing contract-specific information in relation to corresponding clauses in the GCC found in Section IV.

The Special Conditions of Contract (SCC) complement the GCC, specifying contractual requirements linked to the special circumstances of the Procuring Entity, the Procuring Entity's country, the sector, and the Goods purchased. In preparing this Section, the following aspects should be checked:

- a. Information that complements provisions of the GCC must be incorporated.
- b. Amendments and/or supplements to provisions of the GCC as necessitated by the circumstances of the specific purchase, must also be incorporated.

However, no special condition which defeats or negates the general intent and purpose of the provisions of the GCC should be incorporated herein.

Special Conditions of Contract

	Special Conditions of Contract
GCC Clause	
1	[List here any additional requirements for the completion of this Contract. The following requirements and the corresponding provisions may be deleted, amended, or retained depending on its applicability to this Contract:]
	Delivery and Documents –
	For purposes of the Contract, "EXW," "FOB," "FCA," "CIF," "CIP," "DDP" and other trade terms used to describe the obligations of the parties shall have the meanings assigned to them by the current edition of INCOTERMS published by the International Chamber of Commerce, Paris. The Delivery terms of this Contract shall be as follows:
	[For Goods supplied from abroad, state:] "The delivery terms applicable to the Contract are DDP delivered [indicate place of destination]. In accordance with INCOTERMS."
	[For Goods supplied from within the Philippines, state:] "The delivery terms applicable to this Contract are delivered [indicate place of destination]. Risk and title will pass from the Supplier to the Procuring Entity upon receipt and final acceptance of the Goods at their final destination."
	Delivery of the Goods shall be made by the Supplier in accordance with the terms specified in Section VI (Schedule of Requirements).
	For purposes of this Clause the Procuring Entity's Representative at the Project Site is [indicate name(s)].
	Incidental Services –
	The Supplier is required to provide all of the following services, including additional services, if any, specified in Section VI. Schedule of Requirements: Select appropriate requirements and delete the rest.
	 a. performance or supervision of on-site assembly and/or start-up of the supplied Goods; b. furnishing of tools required for assembly and/or maintenance of the supplied Goods; c. furnishing of a detailed operations and maintenance manual for each
	appropriate unit of the supplied Goods; d. performance or supervision or maintenance and/or repair of the supplied Goods, for a period of time agreed by the parties, provided that this service shall not relieve the Supplier of any warranty obligations under this Contract; and
	 e. training of the Procuring Entity's personnel, at the Supplier's plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied Goods. f. [Specify additional incidental service requirements, as needed.]
	The Contract price for the Goods shall include the prices charged by the Supplier for incidental services and shall not exceed the prevailing rates charged to other parties by the Supplier for similar services.

Spare Parts -

The Supplier is required to provide all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the Supplier:

Select appropriate requirements and delete the rest.

- a. such spare parts as the Procuring Entity may elect to purchase from the Supplier, provided that this election shall not relieve the Supplier of any warranty obligations under this Contract; and
- b. in the event of termination of production of the spare parts:
 - i. advance notification to the Procuring Entity of the pending termination, in sufficient time to permit the Procuring Entity to procure needed requirements; and
 - ii. following such termination, furnishing at no cost to the Procuring Entity, the blueprints, drawings, and specifications of the spare parts, if requested.

The spare parts and other components required are listed in Section VI (Schedule of Requirements) and the cost thereof are included in the contract price.

The Supplier shall carry sufficient inventories to assure ex-stock supply of consumable spare parts or components for the Goods for a period of [indicate here the time period specified. If not used indicate a time period of three times the warranty period].

Spare parts or components shall be supplied as promptly as possible, but in any case, within [insert appropriate time period] months of placing the order.

Packaging -

The Supplier shall provide such packaging of the Goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in this Contract. The packaging shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit, and open storage. Packaging case size and weights shall take into consideration, where appropriate, the remoteness of the Goods' final destination and the absence of heavy handling facilities at all points in transit.

The packaging, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the Contract, including additional requirements, if any, specified below, and in any subsequent instructions ordered by the Procuring Entity.

The outer packaging must be clearly marked on at least four (4) sides as follows:

Name of the Procuring Entity Name of the Supplier

Contract Description Final Destination Gross weight Any special lifting instructions Any special handling instructions Any relevant HAZCHEM classifications A packaging list identifying the contents and quantities of the package is to be placed on an accessible point of the outer packaging if practical. If not practical the packaging list is to be placed inside the outer packaging but outside the secondary packaging. Transportation – Where the Supplier is required under Contract to deliver the Goods CIF, CIP, or DDP, transport of the Goods to the port of destination or such other named place of destination in the Philippines, as shall be specified in this Contract, shall be arranged and paid for by the Supplier, and the cost thereof shall be included in the Contract Price. Where the Supplier is required under this Contract to transport the Goods to a specified place of destination within the Philippines, defined as the Project Site, transport to such place of destination in the Philippines, including insurance and storage, as shall be specified in this Contract, shall be arranged by the Supplier, and related costs shall be included in the contract price. Where the Supplier is required under Contract to deliver the Goods CIF, CIP or DDP, Goods are to be transported on carriers of Philippine registry. In the event that no carrier of Philippine registry is available, Goods may be shipped by a carrier which is not of Philippine registry provided that the Supplier obtains and presents to the Procuring Entity certification to this effect from the nearest Philippine consulate to the port of dispatch. In the event that carriers of Philippine registry are available but their schedule delays the Supplier in its performance of this Contract the period from when the Goods were first ready for shipment and the actual date of shipment the period of delay will be considered force majeure. The Procuring Entity accepts no liability for the damage of Goods during transit other than those prescribed by INCOTERMS for DDP deliveries. In the case of Goods supplied from within the Philippines or supplied by domestic Suppliers risk and title will not be deemed to have passed to the Procuring Entity until their receipt and final acceptance at the final destination. Intellectual Property Rights -The Supplier shall indemnify the Procuring Entity against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the Goods or any part thereof. [If partial payment is allowed, state] "The terms of payment shall be as 2.2 follows:" that Product will be conducted are: tests

4

The

inspections

and

Presentation/Demonstration/Site Inspection, if applicable.

Section VI. Schedule of Requirements

PROJECT NAME: PROCUREMENT OF INFANT INCUBATOR, RADIANT WARMER AND OTHERS

PROJECT NO. RMBGH-25-HME-0766

The delivery schedule expressed as weeks/months stipulates hereafter a delivery date which is the date of delivery to the project site.

Item Number	Description	Unit of Issue	Quantity	Delivered, Weeks/Months
1	INFANT INCUBATOR	unit	2	
	Physical Characteristics Size: at least 1130mm×570mm×1350mm Weight: at least 83kg Mattress size: at least 408mm*635mm*20mm Side panel: at least 761mm*325mm Access door: at least 212mm*175mm Screen Size: at least 8 inches TFT touch screen with adjustable viewing and height Resolution: at least 800 × 600 Water Tank: Capacity: at least ≥1L Tray and bracket: Optional; Caster wheel four brakes Filter efficiency 95% with 0.3um particle, 99.9% with 0.5um particle		2	
	Operation Environment Working Temp: at least 20~30°C Humidity Relative humidity: 10% to 95% (noncondensing) Air flow: at least ≤0.3m/s; Noise in the hood: not more than 45db Power Supply: 100-240V~, 50/60Hz±1Hz Battery Type: Rechargeable Lithium-ion battery Battery Capacity: at least 2200mAh Battery Recharging Time: At least 2 hours in normal conditions Battery backup: at least 2 hours for continuous working Indicator: One alarm indicator, Power indicator, Battery indicator, Alarm sound, Operating key sound Interface: Parameter cable interface, AC power input, Three auxiliary power output, USB port			Within Sixty (60) Calendar Days upon issuance of the Notice to Proceed
	Date storage Alarm Review: 200 groups; Power-off storage: Yes Alarm: User-adjustable High and Low 3-level; Limits; Prioritized audible and visual alarm			
	Temperature Control Setting range: 35.0~37.5°C (Infant mode), 25.0~39.0°C (Air Mode) Measurement: range: 25.0~45.0°C Resolution: 0.1°C; Accuracy: Air temp sensor: ±0.3°C Skin temp sensor: ±0.1°C (37~39°C) Alarm range: 34~38.5°C (Infant mode); 23~41°C (Air Mode)			

Schedule of Requirements Page 1 of 7

2	Heating time: ≤35mins (Temperature inside the thermostat cover rises from 25°C to 36°C) Humidification time: <20mins (In indoor environment conditions of 25°C,50% relative humidity, reach to 33°C, 85% relative humidity) Operation without adding water time: ≥12 hours (full water tank, environment: 25°C indoor temp 50% relative humidity, humidity is set to 65%) Humidity Control Relative humidity range: 30%–95% Resolution: 1% Accuracy: ±5% (Set humidity less than 85%); ±10% (Set humidity over than 85%) Alarm: Audible and visual alarm; alarm events reviewable FEATURES Air flow rate inside thermostat cover: ≤0.2m/s Work mode: Air/ Infant Bed tilt range: -13 ~ +13° Input voltage 100-120V~or 200 - 240V~ Input current 10A or 6.6A Input frequency 50/60 Hz Weight Scale Measurement Range: 300g~8000g Accuracy: ±10g Trend View: 7days RADIANT WARMER Graphic 5 inches monochrome LCD screen Display of heating intensity, date & time, set and actual values, trends. State-of-the-art microprocessor controller Intelligent operating system, cont. self-test diagnostic during operation. Test button, self-test, silence, reset button Water resistant, foil-covered, soft touch control panel with lock. Visual and audible alarms and backup sensors Fix, large size lying surface, anti-decubitus mattress, x-ray compatible. Drawer for x-ray cassette tray under the lying surface (35x65 cm). Four foldable transparent side walls with name plate (17,5 cm). Tube ports & measure line on both side walls, IV pole on the column. Heavy duty mobile stand with four anti-static castors, two with brakes. Dimmable halogen examination light in the non-turn able heating head.	unit	1	Within Sixty (60) Calendar Days upon issuance of the Notice to Proceed
	compatible. Drawer for x-ray cassette tray under the lying surface (35x65 cm). Four foldable transparent side walls with name plate (17,5 cm). Tube ports & measure line on both side walls, IV pole on the column. Heavy duty mobile stand with four anti-static castors, two with brakes. Dimmable halogen examination light in the non-turn able heating head. Protector bumpers on each side and handle on front for movement.			
	Radiant ceramic heaters with parabolic reflectors and protecting grid. Automatic warm-up, pre-warm, manual mode Surface and skin temperature servo control, Apgar timer Mechanical continuous positive and reverse Trendelenburg lying surface (infant bed). Monitor shelves on the left & right side of the column Lower cabinet with two drawers on the stand (30x30x9cm)			

r				
	Two gas cylinder holders on the stand			
	Additional LED examination light			
	Heating head with horizontal rotation by 30° and 90° both			
	direction			
3	PHOTOTHERAPY MACHINE	unit	2	
3	THOTOTHERALT MACHINE	WIIIL		
	Phototherapy Light with Skin Temperature Display			
	1) LED therapy light head			
	Wave length: at least 440~460 nm			
	Radiant intensity (blue super light): 35~110µ W/cm²/nm			
	Blue LEDs irradiant area: at least 54 x 32 cm			
	Lamp type: Super blue LED (16 pcs)			
	2) Control system: Back light LCD screen			
	Lamp LED life time: at least 50,000 hours			
	3) Height adjustable stand: at least 1,068cm-1,754 cm			
	4) Exam LED lamp: White LED 4 pcs			
	5) Standard accessory			
	Mobile Cart with 4 wheels (Lock on 2 wheels)			
	Eye protection google			
	User manual; Power supply; Automatic voltage regulator			
	<u>Characteristics</u>			
	Noise Level <50 dBA			
	Screen: at least 2 x 16 LCD display			
	Lamp Type: 16 pcs Blue Super LEDlamps	i	İ	Within Sixty
•	Examination Lamps: 4 pcs White LED Lamps			(60) Calendar
		unit	3	Days upon
4	FETAL MONITOR	unit	3	issuance of
				the Notice to
ļ	Standard parameters: fetal heart rate, uterine contraction			Proceed
i	pressure, fetal movement, ECG, pulse, blood pressure, blood			110000
	oxygen, respiration, body temperature.			
	Ultrasonic operating frequency should not be greater than			
	1MHz.	1		
	Ultrasonic beam sound intensity: Iob<10mW/cm2			
	The measurement range of fetal heart rate is not less than at			
	least 30bpm~250bpm.	ļ		
	Fetal heart rate measurement accuracy should be ≤±1bpm			
	The measurement range of TOCO is 0-100.			
	Standard 3/5 lead ECG, optional 12 lead ECG measurement		1	
	Heart rate measurement range: 20-300bpm for adults, 20-		ĺ	
	350bpm for children/newborns, accuracy: ±1% or ±1bpm,	•		
	whichever is greater.			
	NIBP measurement range: 30-300mmHg, accuracy ±3 mmHg		i	
	SPO2 measurement range: 0-100%, measurement accuracy:			
	within the range of 70%-100%, adult/child ±2%, neonate ±3%.			
	Optional Masimo blood oxygen monitoring, with perfusion			
	index PI display, measurement range 0.02%-20%			
	Not less than 12 inches TFT LCD touch screen			
	The display screen supports 0-90° adjustable, which can be			
	observed from multiple angles.			
1	Adopt damping shaft and magnetic suction design, support			
	step-less adjustment of screen angle, and will not shake when			
	carrying.			
	Support sliding screen operation, quickly switch the display			
	interface.			
	There should be two alarm lights, and the physiological and			
	technical alarm lights should be displayed separately.			
L	technical alarm fights should be displayed separately.	L		
		Ι.	C-L11	eauirements Page 3 of 7

Schedule of Requirements Page 3 of 7 RMBGH-25-HME-0766

	Equipped with dual USB ports as standard, which can			
	connect external printer and U disk at the same time.			
	Uniform sensor interface design (fetal heart rate, uterine			
	contractions, fetal movement).			
	The probe is automatically recognized, and can be plugged			
	and unplugged at will.			
	Built-in 150mm or 152mm thermal printer, and support USB			
	external printer.			
	The medical record report can be saved in a picture format,			
	which is convenient for remote diagnosis and teaching.			
	It can be equipped with a probe holder			
	Support intelligent power management can trigger lock			
	screen, standby, shutdown, energy-saving, and			
	environmental protection.			
	It can store no less than 2000 medical record reports			
	Standard built-in large-capacity lithium battery, which can			
	work continuously for more than 4 hours.			
	Built-in communication interface, support wired/wireless			
	connection to the central monitoring system.			
	<u>Inclusions</u>			
	1. Recording Paper (150mm×90mm-150P, Z-folded)- 1 piece			Within Sixty
	2. 12-Pin 5Lead Clip Connector ECG Cable/Lead replaceable			(60) Calendar
	Anti-Defibrillation/IEC- 1 piece			Days upon
	3. Adult NIBP Cuff-1pc			issuance of
	4. Adult NIBP Tube (white) 2M- 1 piece			the Notice to
	5. Adult Reusable Analog SpO2 sensor- 1 piece			Proceed
	6. Adult Disposable Adhesive Electrodes (12 pieces/pack)-1 piece	ı		
	7. FHR1 Probe – 1 piece			
	8. TOCO Probe- 1 piece			
	9. Event Marker/water proof- 1piece			
	10. Belt- 1 piece			
	11. Power Cord- 1 piece			
	12. User Manual – 1 piece			
5	FETAL DOPPLER	unit	3	
	Waterproof Fetal Doppler			
	• Fetal Heart Rate Display with Fixed waterproof 2MHz			
	probe features:			
	Compact and Light-weight design for ease of portability Light and Light-weight design for EHR in RIC numbers.			
	 Large high contrast LCD display of FHR in BIG numbers Combines Power On/Off/Volume control for ease of use 			
	High sensitivity probes to maximize performance in a wide we risk a falinizations and procedures.			
	 variety of clinical applications and procedures Supports 2MHz waterproof probes suitable for use in water 			
	birth	 		
	Easy clean probes for improved infection control			
	• AA batteries for convenience and long-term cost			
	effectiveness			
	CICCHYCICS			
	Technical Specification:			
	Probe Frequency: at least 2MHz]		
	Attached Waterproof Probes			
	Noise Reduction: Active			
	Integrated Loudspeaker	1		
	Headphone Output			
				equirements Page 4 of 7

Schedule of Requirements Page 4 of 7 RMBGH-25-HME-0766

	Battery Type Supplied as Standard: 2 x AA Alkaline batteries			
	Battery: Life (no. of 1min exams): 500			
	Weight: 304gms (10.7oz)			
	Dimensions (Main unit): Height 140mm (5.5 inches), Width			
	75mm (3.0 inches), Depth 33mm (1.18 inches)			
6	HEAVY DUTY SUCTION MACHINE	unit	2	
	* Surgical & OBS suction function switched by selective knob.			
	* Two collection bottles			
	* Storage platform for instrument			
	* Aluminum bottle rack			
	* Reusable and disposable bottles are able to install			
	* Autoclavable and reusable PC collection bottle is factory			
	equipped.			
	* Depositing box			
	* Lower center of gravity			
	* Fan is designed to prevent heat			
	* Foot switch		ļ	
	* Dual-fuse safety design may shut off electricity when			
	overloaded			
	* Moving stable and smooth			
	* Caster with locking brake			
	Specifications:			
	Capacity: at least 3000 c.c + 3000 c.c	İ		
	Safety device: Floating control			
	N.W / G.W: at least 24kg/ 28kg			
	Machine size: at least L45xW40xH90 (cm)			
	Packing size: at least L53xW49xH96 (cm)			
	Pump Specifications:			Within Sixty
	Voltage: at least 120V / 240V			(60) Calendar
	Current: at least 3.5A / 1.7A			Days upon
	Capacitor: at least 40uf / 13uf			issuance of
	R.P.M: at least 1720 / 1450 rpm			the Notice to
	Output: at least 375 W			Proceed
	Max flow rate: at least 120 L/min			
	Vacuum range: at least >650mmHg			
	Weight: at least 8.3 kg			
	Standard Accessory:			
	Suction tube (S) 180cm 2 pieces			
	Suction tube (L) 180cm 1 piece			
	Suction catheter: 1 piece			
	Foot switch: 1 piece			
	Filter (L): 2 pieces			
	Adaptor (D): 1 piece			
7	PORTABLE SUCTION MACHINE	unit	10	
	Whole ABS plastic injection made outer shell. Designed with			
	smart appearance, light weight and easy to carry.			
	Low noise and powerful suction function which provides a			
	better quality care for users.			
	Regulator is designed to adjust the pressure which could			
	help suction function become more efficient.			
	The operation panel is made by PC material which can			
	prevent from scratch and discoloration.		L	

Schedule of Requirements Page 5 of 7 RMBGH-25-HME-0766

suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min.	8	Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. ≥10-inches TFT touch screen. 13.It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R-R analysis function. 15. Can store and review the latest 10 minutes of ECG waveforms. 16.The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy search by name and ID number. 18. Support external keyboard, mouse, barcode scanner.	unit	10	Within Sixty (60) Calendar Days upon issuance of the Notice to Proceed
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500.c.x 1 pc Vacuum Range: at least 5.60 mmHg Max. Plow: at least 6.61 mmHg Max. Plow: at least 6.63 kg Gross Weight: at least 6.3 kg Gross Weight: at least 6.3 kg Gross Weight: at least 7.6 kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (5): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: 515uVppp 4. Frequency response: 0.05Hz-150Hz 5.The time constant: 255 6. Input loop current: \$50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ±105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. 210-inches TET touch screen. 13. It has the function of intelligent ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R-R analysis function. 15. Can store and review the latest 10 minutes of ECG waveforms. 16. The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy search by name and 1D number. 18. Support external keyboard, mouse, barcode scanner. 19. With standby wake-up function, timed shutdown function. 20. An optional USB printer can be connected and support A4 print paper. 21. Support a bullet in Wi-Fi module.		function. 20. An optional USB printer can be connected and support A4 print paper. 21. Support a built-in Wi-Fi module.			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least 560 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 1.62 x W23 x H36 (cm) Net Weight: at least 1.63 x W33 x H36 (cm) Net Weight: at least 6.3 kg Gross Weight: at least 1.67 x Kg Sandard Accessories: Suction tube (180 cm): 1 piece Suction Catheer 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing, 2. ECG channel: 12 channels 3. Noise level: s15aVp-p 4. Frequency response: 0.05Hz-150Hz 5.1The time constant: 255 6. Input loop current: 590nA 7. Polarization resistance voltage: £650mV (provided with inspection report) 8. CMRR: 21054B 9. The heart rate measurement range should be 30~ 300bpm, and the measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. 216-inches TTC touch screen. 13. It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R R analysis function. 15. Can store and review the latest 10 minutes of ECC waveforms. 16. The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy scarch by name and 110 number. 18. Support external keyboard, mouse, barcode scanner. 19. With standard USB printer can be connected and support Ad print paper. 21. Support abuil-in Wi-Fi modulo. 22. Passed CFDA. CE certification.		function. 20. An optional USB printer can be connected and support A4 print paper. 21. Support a built-in Wi-Fi module. 22. Passed CFDA, CE certification.			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500.c.x 1 pc Vacuum Range: at least 550 mmHg Max. Plow: at least 545 liter/min. Safety Device: Level float control Machine Size: at least 142 x W16 x H40 (cm) Packing Size: at least 1.42 x W16 x H40 (cm) Packing Size: at least 1.63 x Kg Gross Weight: at least 6.3 x Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECC MACHINE 12 LEAD 1. FCG I ead: 12-leads synchronous acquisition, display, and printing. 2. ECC channel: 12 channels 3. Noise level: s15uVp-p 4. Frequency response: 0.05Hz-150Hz 5:The time constant: 255 6. Input loop current: \$50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ±105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 25mm/s, 10mm/s 12. ±10-inches TFT touch screen. 13. It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 fevels 14. With R-R analysis function. 15. Can store and review the latest 10 minutes of ECC waveforms. 16. The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy search by name and 1D number. 18. Support external keyboard, mouse, barcode scanner. 19. With standby wake-up function, timed shutdown function. 20. An optional USB printer can be connected and support A4 print paper. 21. Support a built-in Wi-Fi module.	9	function. 20. An optional USB printer can be connected and support A4 print paper. 21. Support a built-in Wi-Fi module. 22. Passed CFDA, CE certification.	unit	10	
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pe Vacuum Range: at least 560 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 1.42 x W16 x 1440 (cm) Packing Size: at least 1.63 x W23 x 136 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Sandard Accessories. Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: \$15aVp-p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: 265 6. Input loop current: \$50nA 7. Polarization resistance voltage: \$650mV (provided with inspection report) 8. CMRR: 21054B 9. The heart rate measurement range should be \$10pm or \$1.56 10. In heart rate measurement accuracy should be \$10pm or \$1.56 11. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: \$5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. 216-inchest TT touch screen. 13. It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R.R. analysis function. 15. Can store and review the latest 10 minutes of ECG waveforms. 16. The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy scarch by name and 1D number. 18. Support external keyboard, mouse, barcode scanner. 19. With standay wake-up function, timed shutdown function. 20. An optional USB printer can be connected and support Ad print paper. 21. Support a buill-in Wi-Fi module. 22. Passed CFDA. CE certification.		function. 20. An optional USB printer can be connected and support A4 print paper. 21. Support a built-in Wi-Fi module. 22. Passed CFDA, CE certification.		40	
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500.c.x 1 pc Vacuum Range: at least 550 mmHg Max. Plow: at least 545 liter/min. Safety Device: Level float control Machine Size: at least 142 x W16 x H40 (cm) Packing Size: at least 1.42 x W16 x H40 (cm) Packing Size: at least 1.63 x Kg Gross Weight: at least 6.3 x Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECC MACHINE 12 LEAD 1. FCG I ead: 12-leads synchronous acquisition, display, and printing. 2. ECC channel: 12 channels 3. Noise level: s15uVp-p 4. Frequency response: 0.05Hz-150Hz 5:The time constant: 255 6. Input loop current: \$50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ±105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 25mm/s, 10mm/s 12. ±10-inches TFT touch screen. 13. It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 fevels 14. With R-R analysis function. 15. Can store and review the latest 10 minutes of ECC waveforms. 16. The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy search by name and 1D number. 18. Support external keyboard, mouse, barcode scanner. 19. With standby wake-up function, timed shutdown function. 20. An optional USB printer can be connected and support A4 print paper. 21. Support a built-in Wi-Fi module.		function. 20. An optional USB printer can be connected and support A4 print paper. 21. Support a built-in Wi-Fi module.	i i		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 5500.c.x 1 pc Vacuum Range: at least 5650 mmHg Max. Plow: at least 5650 mmHg Max. Plow: at least 1628 VIB (Fmin. Safety Device: Level float control Machine Size: at least 1.42 x W16 x H40 (cm) Packing Size: at least 1.50 x W23 x H36 (cm) Net Weight: at least 1.50 x W23 x H36 (cm) Net Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. FCG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: <151Vp p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: 255 6. Input loop current: <50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: 2105dB 9. The heart rate measurement range should be 21bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: Smm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. 210-inches TTT touch screen. 13. It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R.R. analysis function. 15. Can store and review the latest 10 minutes of ECG waveforms. 16. The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy search by name and ID number. 18. Support external keyboard, mouse, barcode scanner. 19. With standby wake-up function, timed shutdown function. 20. An optional USB printer can be connected and support AM print paper.		function. 20. An optional USB printer can be connected and support A4 print paper.			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500.c.c.x 1 pc Vacuum Range: at least 5 650 mmHg Max. Plow: at least 650 termin. Safety Device: Level float control Machine Size: at least 142 x W16 x H40 (cm) Packing Size: at least 152 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 6.3 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (5): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing; 2. ECG channel: 12 channels 3. Noise level: s15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s. 25mm/s. 50mm/s. 12. ≥10-inches TFT touch screen. 13.1t has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R-R analysis function. 15. Can store and review the latest 10 minutes of ECG waveforms. 16.The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy search by name and 1D number. 18. Support external keyboard, mouse, barcode scanner. 19. With standby wake-up function, timed shutdown function. 20. An optional USB printer can be connected and support		function. 20. An optional USB printer can be connected and support			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500.c.c.x 1 pc Vacuum Range: at least 5 650 mmHg Max. Plow: at least 650 termin. Safety Device: Level float control Machine Size: at least 142 x W16 x H40 (cm) Packing Size: at least 152 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 6.3 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (5): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing; 2. ECG channel: 12 channels 3. Noise level: s15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s. 25mm/s. 50mm/s. 12. ≥10-inches TFT touch screen. 13.1t has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R-R analysis function. 15. Can store and review the latest 10 minutes of ECG waveforms. 16.The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy search by name and 1D number. 18. Support external keyboard, mouse, barcode scanner. 19. With standby wake-up function, timed shutdown function. 20. An optional USB printer can be connected and support		function. 20. An optional USB printer can be connected and support			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500.c.c.x 1 pc Vacuum Range: at least 5 650 mmHg Max. Plow: at least 650 termin. Safety Device: Level float control Machine Size: at least 142 x W16 x H40 (cm) Packing Size: at least 152 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 6.3 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (5): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing; 2. ECG channel: 12 channels 3. Noise level: s15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s. 25mm/s. 50mm/s. 12. ≥10-inches TFT touch screen. 13.1t has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R-R analysis function. 15. Can store and review the latest 10 minutes of ECG waveforms. 16.The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy search by name and 1D number. 18. Support external keyboard, mouse, barcode scanner. 19. With standby wake-up function, timed shutdown function. 20. An optional USB printer can be connected and support		function. 20. An optional USB printer can be connected and support			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500.c.x 1 pc Vacuum Range: at least 5 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 124 x W16 x 1140 (cm) Packing Size: at least 125 x W23 x 1136 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Cabheter: 1 piece Filter (S): 2 pieces BG MACHINE 12 LEAD 1. ECC Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.06Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: Smm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. 210-in-thes TFT touch screen. 13. It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R.R. analysis function. 15. Can store and review the latest 10 minutes of ECG waveforms. 16. The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy search by name and ID number. 18. Support external keyboard, mouse, barcode scanner. 19. With standby wake-up function, timed shutdown function.		function.			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least 55 liter/min. Safety Device: Level float control Machine Size: at least 142 x W16 x H40 (cm) Packing Size: at least 142 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Cross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filler (S): 2 pieces BEG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: £15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant £55 6. Input loop current: £50nA 7. Polarization resistance voltage: £650mV (provided with inspection report) 8. CM(R: £105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. 210-inches TFT touch screen. 13.It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R.R. analysis function. 15. Can store and review the latest 10 minutes of ECG waveforms. 16. The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy search by name and ID number. 18. Support external keyboard, mouse, barcode scanner. 19. With standby wake-up function, timed shutdown function.		function.			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 5600 mmHg Max. Flow: at least 51500c.c x 1 pc Vacuum Range: at least 5 650 mmHg Max. Flow: at least 51500c.c x 1 pc Vacuum Range: at least 5 650 mmHg Max. Flow: at least 51500c.c x 1000 mmHg Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L42 x W16 x H40 (cm) Packing Size: at least L42 x W16 x H40 (cm) Packing Size: at least L70 kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (3): 2 pieces ECG MACHINE 12 LEAD 1. ECC Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp.p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: ≥55 6. Input loop current: ≤50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. ≥10-mches TFT touch screen. 13. It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light detection, which can automatically adjust the order of the can be screen brightness according to the intensity of the ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light detection		· -			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 5600 mmHg Max. Flow: at least 51500c.c x 1 pc Vacuum Range: at least 5 650 mmHg Max. Flow: at least 51500c.c x 1 pc Vacuum Range: at least 5 650 mmHg Max. Flow: at least 51500c.c x 1000 mmHg Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L42 x W16 x H40 (cm) Packing Size: at least L42 x W16 x H40 (cm) Packing Size: at least L70 kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (3): 2 pieces ECG MACHINE 12 LEAD 1. ECC Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp.p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: ≥55 6. Input loop current: ≤50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. ≥10-mches TFT touch screen. 13. It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light detection, which can automatically adjust the order of the can be screen brightness according to the intensity of the ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light detection		· -	1		1
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least 5 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x 1140 (cm) Packing Size: at least L42 x W16 x 1140 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: 4:5uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: 255 6. Input loop current: 5:50nA 7. Polarization resistance voltage: 4:650mV (provided with inspection report) 8. CMRR: 2:105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. 210-inches TFT touch screen. 13.It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness and review the latest 10 minutes of ECC waveforms. 16. The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy search by name and 110 number. 18. Support external keyboard, mouse, barcode scanner.	1		1		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least 5 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x 1140 (cm) Packing Size: at least L42 x W16 x 1140 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: 4:5uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: 255 6. Input loop current: 5:50nA 7. Polarization resistance voltage: 4:650mV (provided with inspection report) 8. CMRR: 2:105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. 210-inches TFT touch screen. 13.It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness and review the latest 10 minutes of ECC waveforms. 16. The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy search by name and 110 number. 18. Support external keyboard, mouse, barcode scanner.		19. With standby wake-up function, timed shutdown	1		1
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L42 x W16 x H40 (cm) Packing Size: at least L42 x W16 x H40 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uV-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 1.0. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 5mm/s. 12. ≥10-inches TFT touch screen. 13.1t has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light detection, which can automatically adjust the screen brightness cacording to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R-R analysis function. 15. Can store and review the latest 10 minutes of ECG waveforms. 16.The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy search by name and ID number.	1		1		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 150 iter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L42 x W16 x H40 (cm) Packing Size: at least L42 x W16 x H40 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: 415uVp-p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: ≥55 6. Input loop current: ≥50xA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. ≥10-inches TFT touch screen. 13.It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R-R analysis function. 15. Can store and review the latest 10 minutes of ECG waveforms. 16. The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy	1	18 Support external keyboard, mouse, harcode scanner	1		Į i
suction function automatically which could prevent the pollution against the pump. Specificationse Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 150 iter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L42 x W16 x H40 (cm) Packing Size: at least L42 x W16 x H40 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: 415uVp-p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: ≥55 6. Input loop current: ≥50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. ≥10-inches TFT touch screen. 13.It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R-R analysis function. 15. Can store and review the latest 10 minutes of ECG waveforms. 16. The ECG machine can store 10,000 cases. 17. The medical record search function supports fuzzy		scarch by name and ID number.			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least ≥ 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 1.50 x W23 x H36 (cm) Packing Size: at least 1.50 x W23 x H36 (cm) Net Weight: at least 5.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.0Hz-150Hz 5.The time constant: 255 6. Input loop current: ≤50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30 ~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. ≥10-in-ches TIT touch screen. 13.It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With RR. analysis function. 15. Can store and review the latest 10 minutes of ECG waveforms. 16. The ECG machine can store 10,000 cases.	1	_	1 I]
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least ≥ 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 1.50 x W23 x H36 (cm) Packing Size: at least 1.50 x W23 x H36 (cm) Net Weight: at least 5.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.0Hz-150Hz 5.The time constant: 255 6. Input loop current: ≤50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30 ~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. ≥10-in-ches TIT touch screen. 13.It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With RR. analysis function. 15. Can store and review the latest 10 minutes of ECG waveforms. 16. The ECG machine can store 10,000 cases.		17. The medical record search function supports fuzzy	į l]
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 1.42 x W16 x H40 (cm) Packing Size: at least 1.50 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: ≥55 6. Input loop current: ≤50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CARR: ≥105dB 9. The heart rate measurement range should be 30 ~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. ≥10-inches TFT touch screen. 13.It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambi	1		1		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 1.42 x W16 x H40 (cm) Packing Size: at least 1.50 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: ≥55 6. Input loop current: ≤50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CARR: ≥105dB 9. The heart rate measurement range should be 30 ~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. ≥10-inches TFT touch screen. 13.It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambi		16. The ECG machine can store 10.000 cases.			<u> </u>
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least 5650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 150 x W23 x H36 (cm) Net Weight: at least 5.0 x W23 x H36 (cm) Net Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces BECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: s15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12.≥10-inches TFT touch screen. 13.It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the screen brightness according to the intensity of the ambient light and the sc	1		1		Į l
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c×1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 55 liter/min. Safety Device: Level float control Machine Size: at least 150 x W23 x H36 (cm) Packing Size: at least 150 x W23 x H36 (cm) Net Weight: at least 7.0 kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12.≥10-inches TIT touch screen. 13.It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R-R analysis function.		wayoforms	1		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c×1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 55 liter/min. Safety Device: Level float control Machine Size: at least 150 x W23 x H36 (cm) Packing Size: at least 150 x W23 x H36 (cm) Net Weight: at least 7.0 kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12.≥10-inches TIT touch screen. 13.It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R-R analysis function.	1	13. Call Store and Teview the fatest to fillitudes of LCC	1		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c×1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 55 liter/min. Safety Device: Level float control Machine Size: at least 150 x W23 x H36 (cm) Packing Size: at least 150 x W23 x H36 (cm) Net Weight: at least 7.0 kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12.≥10-inches TIT touch screen. 13.It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels 14. With R-R analysis function.		15 Can store and review the latest 10 minutes of ECC	1		l l
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vactum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 10 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: 415uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: 250 6. Input loop current: 550nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: 2105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. ≥10-inches TFT touch screen. 13. It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light date toton, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels			į l		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 5.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 10 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: :15uVp-p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: ≥55 6. Input loop current: >550nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: >105dB 9. The heart rate measurement range should be 30~ 300pm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. ≥10-inches TFT touch screen. 13. It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the screen brightness can be adjusted to 10 levels		14 With R-R analysis function			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 42 x W16 x H40 (cm) Packing Size: at least 42 x W16 x H40 (cm) Packing Size: at least 42 x W16 x H40 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing, 2. ECG channel: 12 channels 3. Noise level: :15uVp-p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: 25S 6. Input loop current: :50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: 21054B 9. The heart rate measurement range should be 30~ 300pm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. 210-inches TFT touch screen. 13. It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the					
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 42 x W16 x H40 (cm) Packing Size: at least 42 x W16 x H40 (cm) Packing Size: at least 42 x W16 x H40 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing, 2. ECG channel: 12 channels 3. Noise level: :15uVp-p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: 25S 6. Input loop current: :50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: 21054B 9. The heart rate measurement range should be 30~ 300pm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. 210-inches TFT touch screen. 13. It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness according to the intensity of the ambient light and the		screen brightness can be adjusted to 10 levels			j l
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Nct Weight: at least 5.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filler (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing, 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≤55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. 210-inches TFT touch screen. 13. It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness			1		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Nct Weight: at least 5.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filler (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing, 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≤55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. 210-inches TFT touch screen. 13. It has the function of intelligent ambient light detection, which can automatically adjust the screen brightness		according to the intensity of the ambient light and the	1		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. ≥10-inches TFT touch screen. 13. It has the function of intelligent ambient light detection,	1		1]
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. ≥10-inches TFT touch screen. 13. It has the function of intelligent ambient light detection,		which can automatically adjust the screen brightness	1		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L50 x W23 x H36 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. ≥10-inches TFT touch screen.			1		j
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L50 x W23 x H36 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s. 12. ≥10-inches TFT touch screen.		13 It has the function of intelligent ambient light detection.	Į į		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L50 x W23 x H36 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 kg Gross Weight: at least 7.0 kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: s15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥50 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30∼ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s, 25mm/s, 50mm/s.			1		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 1.20 x W13 x H140 (cm) Packing Size: at least 1.50 x W23 x H36 (cm) Net Weight: at least 6.3 kg Gross Weight: at least 5.0 kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: ≥55 6. Input loop current: ≤50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s,		12 Mainchas THT touch screen	1		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 1.2 x W16 x H40 (cm) Packing Size: at least 1.50 x W23 x H36 (cm) Net Weight: at least 6.3 kg Gross Weight: at least 5.0 kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: ≥55 6. Input loop current: ≤50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm. 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s,	1	Zomm/s, oumm/s.	[
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500.c. x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 1.42 x W16 x H40 (cm) Packing Size: at least 1.42 x W16 x H40 (cm) Packing Size: at least 1.50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: ≥55 6. Input loop current: ≤50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and rhythm.	1	•	[
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500.c.x1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 5.3 Kg Gross Weight: at least 5.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: ≥55 6. Input loop current: ≤50nA 7. Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and thythm.]	11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s, 12.5mm/s,	[
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c×1 pc Vacuum Range: at least 550 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and	1		[
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500.c. x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 1.42 x W16 x H40 (cm) Packing Size: at least 1.42 x W16 x H40 (cm) Net Weight: at least 1.50 x W23 x H36 (cm) Net Weight: at least 5.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30∼ 300bpm, and the measurement accuracy should be ±1bpm or ±1%. 10. It has three display modes: automatic, manual, and	1	rhythm.	į l		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%.			ļ l		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm or ±1%.		10. It has three display modes: automatic, manual, and			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 55 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥58 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm			Į l		1
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 55 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥58 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~ 300bpm, and the measurement accuracy should be ±1bpm		or ±1%.	1		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~					
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB 9. The heart rate measurement range should be 30~	1	300bpm, and the measurement accuracy should be ±1bpm	1		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing, 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥5S 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB			Į l		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥5S 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report) 8. CMRR: ≥105dB	!	9. The heart rate measurement range should be 30~	j l		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 1.42 x W16 x H40 (cm) Packing Size: at least 1.50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 10 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥58 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report)			1		to rroceea
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 10 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5. The time constant: ≥55 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with inspection report)		8. CMRR: ≥105dB	1		to Proceed
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥58 6. Input loop current: ≤50nA 7.Polarization resistance voltage: ±650mV (provided with			į l		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 10 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥5S 6. Input loop current: ≤50nA The literature is the state of Econ V (apprinted with the constant to the constant		inspection report)	[[
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 10 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥5S 6. Input loop current: ≤50nA The literature is the state of Econ V (apprinted with the constant to the constant		_			issuance of
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 10 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥58 Calendar					
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 5.0 kg Gross Weight: at least 7.0 kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz 5.The time constant: ≥5S Within Sixty (60) Calendar	1	6. Input loop current: ≤50nA			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least 1.42 x W16 x H40 (cm) Packing Size: at least 1.50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Frequency response: 0.05Hz-150Hz Within Sixty (60)					Calendar
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 10 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p 4. Evenueus transporter 0.05Hz 150Hz		5. The time constant: ≥5S]		• ` '
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels 3. Noise level: ≤15uVp-p Within		, · · · · · · · · · · · · · · · · · · ·	ļ l		Sixty (60)
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 10 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels		1			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD 1. ECG Lead: 12-leads synchronous acquisition, display, and printing. 2. ECG channel: 12 channels		3. Noise level: ≤15uVp-p	l i		TA7241- 1
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 10 1. ECG Lead: 12-leads synchronous acquisition, display, and printing.					
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 10	\	2. ECG channel: 12 channels			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 10		1			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 10					
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces 8 ECG MACHINE 12 LEAD unit 10		1. ECG Lead: 12-leads synchronous acquisition, display, and			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces	1	4 0001 - 4 101-4 4 4 11-4 3			1
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces					
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece Filter (S): 2 pieces	0	ECG MACHINE 12 LEAD	will	10	1
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece	Q		unit	10	
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece Suction Catheter: 1 piece		Filter (S): 2 pieces			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories: Suction tube (180 cm): 1 piece		•	Ţ		
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg Standard Accessories:		The state of the s			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg	İ	Suction tube (180 cm): 1 piece		ļ	
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg					
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg		0. 1 1 4			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg					
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm) Net Weight: at least 6.3 Kg		Gross Weight: at least 7.0 Kg			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm)					
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control Machine Size: at least L42 x W16 x H40 (cm)					
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control		Packing Size: at least L50 x W23 x H36 (cm)			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control		• •	1	1	
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min.		*			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min.		Safety Device: Level float control			
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg	ĺ				
suction function automatically which could prevent the pollution against the pump. Specifications: Capacity: at least 1500c.c x 1 pc					
suction function automatically which could prevent the pollution against the pump. Specifications:	ĺ	Vacuum Range: at least > 650 mmHg			
suction function automatically which could prevent the pollution against the pump. Specifications:					
suction function automatically which could prevent the pollution against the pump.	İ	=			1
suction function automatically which could prevent the		Specifications			
suction function automatically which could prevent the	į				1
suction function automatically which could prevent the		pollution against the pump.			
While the bottle is full, the level float control will stop the		·			
7 17 17 17 17 17 17 17 17 17 17 17 17 17	ł	While the bottle is full, the level float control will stop the			

	Max Pressure: at least 2.5 bar 36 psi 250 kPa Operating Pressure: 1.1 bar 16 psi 110 kPa Max Air Flow: at least 14 V/min Operating Air Flow: at least 5.1 V/min Neb-rate with at least 4 ml solution: at least 0,40 ml/min Mass Median Aerodynamic Diameter: at least 2,44um (MMAD) Respirable Fraction (< 5pm): at least > 80% Duty Cycle: Non-stop operation; Weight: 1,65 kg; Size (Main Unit): on 15 x 22 x 12 Noise Level EN 13544-1 Standard): Approx. 55 Dba			Within Sixty (60) Calendar Days upon
10	PULSE OXIMETER TABLETOP (Adult, Pedia/Neonate)	unit	3	issuance of
	1) Display Color at least 6 inches TFT LCD			the Notice
	Pulse rate, SpO2 percentage, SpO2 limit, Pulse rate limits,			to Proceed
	Waveform, Alarm, Trend.			
	2) Perfusion rates from at least 0.05~20%			
	3) Adult/Pediatric/Neonate Mode			
	4) Rechargeable battery: Li-Ion Internal Battery			
	5) Power: Free volts input: at least 100-240Vac, 50/60Hz	:		
	6) Pivoting Heater Module			
	7) Standard Accessory			
	*Reusable SpO2 finger probe for Adult			
	*Reusable "Y" type sensor for Pediatric/Neonate			Į Į
	*Power Cord			
	*User Manual			
	8) Free item			
	*Hanging Handle			
	*Automatic Voltage Regulator			1

I hereby certify to comply and deliver all the above requirements.

Name:	
Legal Capacity:	
Signature:	
Duly authorized to sign the Bid for and behalf of:	

Schedule of Requirements Page 7 of 7 RMBGH-25-HME-0766

Section VII. Technical Specifications

Notes for Preparing the Technical Specifications

A set of precise and clear specifications is a prerequisite for Bidders to respond realistically and competitively to the requirements of the Procuring Entity without qualifying their Bids. In the context of Competitive Bidding, the specifications (e.g. production/delivery schedule, manpower requirements, and after-sales service/parts, descriptions of the lots or items) must be prepared to permit the widest possible competition and, at the same time, present a clear statement of the required standards of workmanship, materials, and performance of the goods and services to be procured. Only if this is done will the objectives of transparency, equity, efficiency, fairness, and economy in procurement be realized, responsiveness of bids be ensured, and the subsequent task of bid evaluation and post-qualification facilitated. The specifications should require that all items, materials and accessories to be included or incorporated in the goods be new, unused, and of the most recent or current models, and that they include or incorporate all recent improvements in design and materials unless otherwise provided in the Contract.

Samples of specifications from previous similar procurements are useful in this respect. The use of metric units is encouraged. Depending on the complexity of the goods and the repetitiveness of the type of procurement, it may be advantageous to standardize the General Technical Specifications and incorporate them in a separate subsection. The General Technical Specifications should cover all classes of workmanship, materials, and equipment commonly involved in manufacturing similar goods. Deletions or addenda should then adapt the General Technical Specifications to the particular procurement.

Care must be taken in drafting specifications to ensure that they are not restrictive. In the specification of standards for equipment, materials, and workmanship, recognized Philippine and international standards should be used as much as possible. Where other particular standards are used, whether national standards or other standards, the specifications should state that equipment, materials, and workmanship that meet other authoritative standards, and which ensure at least a substantially equal quality than the standards mentioned, will also be acceptable. The following clause may be inserted in the Special Conditions of Contract or the Technical Specifications.

Sample Clause: Equivalency of Standards and Codes

Wherever reference is made in the Technical Specifications to specific standards and codes to be met by the goods and materials to be furnished or tested, the provisions of the latest edition or revision of the relevant standards and codes shall apply, unless otherwise expressly stated in the Contract. Where such standards and codes are national or relate to a particular country or region, other authoritative standards that ensure substantial equivalence to the standards and codes specified will be acceptable.

Reference to brand name and catalogue number should be avoided as far as possible; where unavoidable they should always be followed by the words "or at least equivalent." References to brand names cannot be used when the funding source is the GOP.

Where appropriate, drawings, including site plans as required, may be furnished by the Procuring Entity with the Bidding Documents. Similarly, the Supplier may be requested to provide drawings or samples either with its Bid or for prior review by the Procuring Entity during contract execution.

Bidders are also required, as part of the technical specifications, to complete their statement of compliance demonstrating how the items comply with the specification.

Technical Specifications PROJECT NAME: PROCUREMENT OF INFANT INCUBATOR, RADIANT WARMER AND OTHERS PROJECT NO. RMBGH-25-HME-0766

em	Specification	Statement of Compliance
		[Bidders must state here either "Comply" or "Not Comply" against each of the individual parameters of each Specification stating the corresponding performance parameter of the equipment offered Statements of "Comply" or "Not Comply" must be supported by evidence in a Bidders Bid and cross-referenced to that evidence. Evidence shall be in the form of manufacturer's un-amended sales literature, unconditional statements of specification and compliance issued by the manufacturer, samples, independent test data etc., as appropriate. A statement that is not supported by evidence or is subsequently found to be contradicted by the evidence presented will render the Bid under evaluation liable for rejection. A statement either in the Bidder's statement of compliance or the supporting evidence that is found to be false either during Bid evaluation, post-qualification or the execution of the Contract may be regarded as fraudulent and render the Bidder or supplier liable for prosecution subject to the applicable laws and issuances.]
	With the following minimum technical specifications	
 1	INFANT INCUBATOR	
	Physical Characteristics	
	Size: at least 1130mm×570mm×1350mm	
	Weight: at least 83kg	
	Mattress size: at least 408mm*635mm*20mm	
	Side panel: at least 761mm*325mm	
	Access door: at least 212mm*175mm	
	Screen Size: at least 8 inches TFT touch screen with	
	adjustable viewing and height	
	Resolution: at least 800 × 600	
	Water Tank: Capacity: at least ≥1L	
	Tray and bracket: Optional; Caster wheel four brakes	
	Filter efficiency 95% with 0.3um particle, 99.9% with 0.5um	
	particle	
	Hood: Dual layer wall design	
	Operation Environment	
	Working Temp: at least 20~30°C	
	Humidity Relative humidity: 10% to 95% (noncondensing)	
	Air flow: at least ≤0.3m/s; Noise in the hood: not more than	
	45db	
	Power Supply: 100-240V~, 50/60Hz±1Hz	
	Battery Type: Rechargeable Lithium-ion battery	
	Battery Capacity: at least 2200mAh	
	Battery Recharging Time: At least 2 hours in normal conditions	
	Battery backup: at least 2 hours for continuous working	
	Indicator: One alarm indicator, Power indicator, Battery	
	indicator, Alarm sound, Operating key sound	
	Interface: Parameter cable interface, AC power input,	
	Three auxiliary power output, USB port	

Date storage

Alarm Review: 200 groups; Power-off storage: Yes Alarm: User-adjustable High and Low 3-level; Limits;

Prioritized audible and visual alarm

Temperature Control

Setting range: 35.0~37.5°C (Infant mode), 25.0~39.0°C (Air

Mode)

Measurement: range: 25.0~45.0°C

Resolution: 0.1°C; Accuracy: Air temp sensor: ±0.3°C

Skin temp sensor: ±0.1°C (37~39°C)

Alarm range: 34~38.5°C (Infant mode); 23~41°C (Air Mode) Heating time: ≤35mins (Temperature inside the thermostat

cover rises from 25°C to 36°C)

Humidification time: <20mins (In indoor environment conditions of 25°C,50% relative humidity, reach to 33°C,

85% relative humidity)

Operation without adding water time: ≥12 hours (full water tank, environment: 25°C indoor temp 50% relative

humidity, humidity is set to 65%)

Humidity Control

Relative humidity range: 30%~95%

Resolution: 1%

Accuracy: ±5% (Set humidity less than 85%); ±10% (Set

humidity over than 85%)

Alarm: Audible and visual alarm; alarm events reviewable

FEATURES

Air flow rate inside thermostat cover: ≤0.2m/s

Work mode: Air/ Infant Bed tilt range: -13 ~ +13°

Input voltage 100-120V~or 200 - 240V~

Input current 10A or 6.6A Input frequency 50/60 Hz

Weight Scale

Measurement Range: 300g~8000g

Accuracy: ±10g Trend View: 7days

2 RADIANT WARMER

Graphic 5 inches monochrome LCD screen

Display of heating intensity, date & time, set and actual values, trends.

State-of-the-art microprocessor controller

Intelligent operating system, cont. self-test diagnostic during operation.

Test button, self-test, silence, reset button

Water resistant, foil-covered, soft touch control panel with

Visual and audible alarms and backup sensors

Fix, large size lying surface, anti-decubitus mattress, x-ray compatible.

Drawer for x-ray cassette tray under the lying surface (35x65 cm). Four foldable transparent side walls with name plate (17,5 cm). Tube ports & measure line on both side walls, IV pole on the column.

Heavy duty mobile stand with four anti-static castors, two with brakes.

Technical Specifications Page 2 of 8 RMBGH-25-HME-0766 Dimmable halogen examination light in the non-turn able heating head.

Protector bumpers on each side and handle on front for movement.

Radiant ceramic heaters with parabolic reflectors and protecting grid.

Automatic warm-up, pre-warm, manual mode

Surface and skin temperature servo control, Apgar timer

Mechanical continuous positive and reverse Trendelenburg lying surface (infant bed).

Monitor shelves on the left & right side of the column

Lower cabinet with two drawers on the stand (30x30x9cm)

Two gas cylinder holders on the stand

Additional LED examination light

Heating head with horizontal rotation by 30° and 90° both direction

3 PHOTOTHERAPY MACHINE

Phototherapy Light with Skin Temperature Display

1) LED therapy light head

Wave length: at least 440~460 nm

Radiant intensity (blue super light): 35~110µ W/cm²/nm

Blue LEDs irradiant area: at least 54 x 32 cm

Lamp type: Super blue LED (16 pcs)

2) Control system: Back light LCD screen

Lamp LED life time: at least 50,000 hours

- 3) Height adjustable stand: at least 1,068cm-1,754 cm
- 4) Exam LED lamp: White LED 4 pcs
- 5) Standard accessory

Mobile Cart with 4 wheels (Lock on 2 wheels)

Eye protection google

User manual; Power supply; Automatic voltage regulator

Characteristics

Noise Level <50 dBA

Screen: at least 2 x 16 LCD display

Lamp Type: 16 pcs Blue Super LED --lamps

Examination Lamps: 4 pcs White LED Lamps

4 FETAL MONITOR

Standard parameters: fetal heart rate, uterine contraction pressure, fetal movement, ECG, pulse, blood pressure, blood oxygen, respiration, body temperature.

Ultrasonic operating frequency should not be greater than 1MHz.

Ultrasonic beam sound intensity: Iob<10mW/cm2

The measurement range of fetal heart rate is not less than at least 30bpm~250bpm.

Fetal heart rate measurement accuracy should be ≤±1bpm The measurement range of TOCO is 0-100.

Standard 3/5 lead ECG, optional 12 lead ECG measurement Heart rate measurement range: 20-300bpm for adults, 20-350bpm for children/newborns, accuracy: ±1% or ±1bpm, whichever is greater.

Technical Specifications Page 3 of 8 RMBGH-25-HME-0766 NIBP measurement range: 30-300mmHg, accuracy ± 3 mmHg

SPO2 measurement range: 0-100%, measurement accuracy: within the range of 70%-100%, adult/child $\pm 2\%$, neonate $\pm 3\%$.

Optional Masimo blood oxygen monitoring, with perfusion index PI display, measurement range 0.02%-20%

Not less than 12 inches TFT LCD touch screen

The display screen supports 0-90° adjustable, which can be observed from multiple angles.

Adopt damping shaft and magnetic suction design, support step-less adjustment of screen angle, and will not shake when carrying.

Support sliding screen operation, quickly switch the display interface.

There should be two alarm lights, and the physiological and technical alarm lights should be displayed separately.

Equipped with dual USB ports as standard, which can connect external printer and U disk at the same time.

Uniform sensor interface design (fetal heart rate, uterine contractions, fetal movement).

The probe is automatically recognized, and can be plugged and unplugged at will.

Built-in 150mm or 152mm thermal printer, and support USB external printer.

The medical record report can be saved in a picture format, which is convenient for remote diagnosis and teaching.

It can be equipped with a probe holder

Support intelligent power management can trigger lock screen, standby, shutdown, energy-saving, and environmental protection.

It can store no less than 2000 medical record reports Standard built-in large-capacity lithium battery, which can work continuously for more than 4 hours.

Built-in communication interface, support wired/wireless connection to the central monitoring system.

<u>Inclusions</u>

- 1. Recording Paper (150mm×90mm-150P, Z-folded)- 1 piece
- 2. 12-Pin 5Lead Clip Connector ECG Cable/Lead replaceable Anti-Defibrillation/IEC- 1 piece
- 3. Adult NIBP Cuff-1pc
- 4. Adult NIBP Tube (white) 2M-1 piece
- 5. Adult Reusable Analog SpO2 sensor- 1 piece
- 6. Adult Disposable Adhesive Electrodes (12 pieces/pack)-1 piece
- 7. FHR1 Probe 1 piece
- 8. TOCO Probe- 1 piece
- 9. Event Marker/water proof- 1piece
- 10. Belt- 1 piece
- 11. Power Cord-1 piece
- 12. User Manual 1 piece

5 FETAL DOPPLER

Waterproof Fetal Doppler

 Fetal Heart Rate Display with Fixed waterproof 2MHz probe features:

- Compact and Light-weight design for ease of portability
- Large high contrast LCD display of FHR in BIG numbers
- Combines Power On/Off/Volume control for ease of use
- High sensitivity probes to maximize performance in a wide variety of clinical applications and procedures
- Supports 2MHz waterproof probes suitable for use in water birth
- Easy clean probes for improved infection control
- AA batteries for convenience and long-term cost effectiveness

Technical Specification:

Probe Frequency: at least 2MHz

Attached Waterproof Probes

Noise Reduction: Active

Integrated Loudspeaker

Headphone Output

Battery Type Supplied as Standard: 2 x AA Alkaline

batteries Battery: Life (no. of 1min exams): 500

Weight: 304gms (10.7oz)

Dimensions (Main unit): Height 140mm (5.5 inches), Width

75mm (3.0 inches), Depth 33mm (1.18 inches)

HEAVY DUTY SUCTION MACHINE 6

- * Surgical & OBS suction function switched by selective knob.
- * Two collection bottles
- * Storage platform for instrument
- * Aluminum bottle rack
- * Reusable and disposable bottles are able to install
- * Autoclavable and reusable PC collection bottle is factory equipped.
- * Depositing box
- * Lower center of gravity
- * Fan is designed to prevent heat
- * Foot switch
- * Dual-fuse safety design may shut off electricity when overloaded
- * Moving stable and smooth
- * Caster with locking brake

Specifications:

Capacity:

at least 3000 c.c + 3000 c.c

Safety device: Floating control

N.W / G.W:

at least 24kg/28kg

Machine size:

at least L45xW40xH90 (cm)

Packing size:

at least L53xW49xH96 (cm)

Pump Specifications:

Voltage:

at least 120V / 240V

Current:

at least 3.5A / 1.7A

Capacitor:

at least 40uf / 13uf

R.P.M: Output: at least 1720 / 1450 rpm

at least 375 W

Max flow rate: at least 120 L/min

Weight:

Vacuum range: at least >650mmHg at least 8.3 kg

39

Technical Specifications Page 5 of 8 RMBGH-25-HME-0766

Standard Accessory:

Suction tube (S)... 180cm 2 pieces Suction tube (L)... 180cm 1 piece

Suction catheter: 1 piece Foot switch: 1 piece Filter (L): 2 pieces Adaptor (D): 1 piece

7 PORTABLE SUCTION MACHINE

Whole ABS plastic injection made outer shell. Designed with smart appearance, light weight and easy to carry. Low noise and powerful suction function which provides a better quality care for users.

Regulator is designed to adjust the pressure which could help suction function become more efficient.

The operation panel is made by PC material which can prevent from scratch and discoloration.

While the bottle is full, the level float control will stop the suction function automatically which could prevent the pollution against the pump.

Specifications:

Capacity: at least 1500c.c x 1 pc Vacuum Range: at least > 650 mmHg Max. Flow: at least 35 liter/min. Safety Device: Level float control

Machine Size: at least L42 x W16 x H40 (cm) Packing Size: at least L50 x W23 x H36 (cm)

Net Weight: at least 6.3 Kg Gross Weight: at least 7.0 Kg

Standard Accessories:

Suction tube (180 cm): 1 piece Suction Catheter: 1 piece

Filter (S): 2 pieces

8 ECG MACHINE 12 LEAD

- 1. ECG Lead: 12-leads synchronous acquisition, display, and printing.
- 2. ECG channel: 12 channels3. Noise level: ≤15uVp-p
- 4. Frequency response: 0.05Hz-150Hz
- 5.The time constant: ≥5S6. Input loop current: ≤50nA
- 7.Polarization resistance voltage: ±650mV (provided with inspection report)
- 8. CMRR: ≥105dB
- 9. The heart rate measurement range should be 30 \sim 300bpm, and the measurement accuracy should be ± 1 bpm or $\pm 1\%$.
- 10. It has three display modes: automatic, manual, and rhythm.
- 11. Recording speed: 5mm/s, 6.25mm/s, 10mm/s,
- 12.5mm/s, 25mm/s, 50mm/s.
- 12. ≥10-inches TFT touch screen.

Technical Specifications Page 6 of 8 RMBGH-25-HME-0766

	40 T. 1	,
	13.It has the function of intelligent ambient light detection,	
	which can automatically adjust the screen brightness	
	according to the intensity of the ambient light and the	
	screen brightness can be adjusted to 10 levels	·
	14. With R-R analysis function.	
	15. Can store and review the latest 10 minutes of ECG	
	waveforms.	
	16.The ECG machine can store 10,000 cases.	
	17. The medical record search function supports fuzzy	
	search by name and ID number.	
	18. Support external keyboard, mouse, barcode scanner.	
	19. With standby wake-up function, timed shutdown	
	function.	
	20. An optional USB printer can be connected and support	
	Λ4 print paper.	
	21. Support a built-in Wi-Fi module.	
	22. Passed CFDA, CE certification.	
9	PORTABLE NEBULIZER	
	Technical Features:	
	Compressor: Oil less and maintenance-free piston pump	
	Power Feeding: 230V/50Hz or 220V / 50-60Hz	
	Power consumption: 170 VA	
	Max Pressure: at least 2.5 bar 36 psi 250 kPa	
	Operating Pressure: 1.1 bar 16 psi 110 kPa	
	Max Air Flow: at least 14 V/min	
	Operating Air Flow: at least 5.1 V/min	
	Neb-rate with at least 4 ml solution: at least 0,40 ml/min	
	Mass Median Aerodynamic Diameter: at least 2,44um (MMAD)	
	Respirable Fraction (< 5pm): at least > 80%	·
	Duty Cycle: Non-stop operation; Weight: 1,65 kg; Size	
	(Main Unit): on 15 x 22 x 12	
	Noise Level EN 13544-1 Standard): Approx. 55 Dba	
	,	
10	PULSE OXIMETER TABLETOP (Adult, Pedia/Neonate)	
	1) Display Color at least 6 inches TFT LCD	
	Pulse rate, SpO2 percentage, SpO2 limit, Pulse rate limits,	
	Waveform, Alarm, Trend.	
	2) Perfusion rates from at least 0.05~20%	
	3) Adult/Pediatric/Neonate Mode	
	4) Rechargeable battery: Li-Ion Internal Battery	
	5) Power: Free volts input: at least 100-240Vac, 50/60Hz	
	6) Pivoting Heater Module	
	7) Standard Accessory	
	*Reusable SpO2 finger probe for Adult	
	*Reusable "Y" type sensor for Pediatric/Neonate	
	*Power Cord	
	*User Manual	
	8) Free item	
	*Hanging Handle	
D	*Automatic Voltage Regulator	
В.	Compliance to the Schedule of Requirements	
	(Section VI)	

Technical Specifications Page 7 of 8 RMBGH-25-HME-0766

I hereby certify to comply and deliver all the above requirements.

Name:	
Legal Capacity:	
Signature:	
Duly authorized to sign the Bid for and behalf of:	

Technical Specifications Page 8 of 8 RMBGH-25-HME-0766

Section VIII. Checklist of Technical and Financial Documents

Notes on the Checklist of Technical and Financial Documents

The prescribed documents in the checklist are mandatory to be submitted in the Bid, but shall be subject to the following:

- a. GPPB Resolution No. 09-2020 on the efficient procurement measures during a State of Calamity or other similar issuances that shall allow the use of alternate documents in lieu of the mandated requirements; or
- b. Any subsequent GPPB issuances adjusting the documentary requirements after the effectivity of the adoption of the PBDs.

The BAC shall be checking the submitted documents of each Bidder against this checklist to ascertain if they are all present, using a non-discretionary "pass/fail" criterion pursuant to Section 30 of the 2016 revised IRR of RA No. 9184.

Checklist of Technical and Financial Documents

I. TECHNICAL COMPONENT ENVELOPE

Class "A" Documents

Legal Do	cuments
(a)	Valid PhilGEPS Registration Certificate (Platinum Membership) (all pages) in accordance with Section 8.5.2 of the IRR;
<u>Technica</u>	<u>l Documents</u>
(b) S	statement of the prospective bidder of all its ongoing government and private contracts, including contracts awarded but not yet started, if any, whether similar or not similar in nature and complexity to the contract to be bid (in a
(c)	FORM prescribed by the QC-BAC-GOODS AND SERVICES); and Statement of the bidder's Single Largest Completed Contract (SLCC) similar to the contract to be bid, except under conditions provided for in Sections 23.4.1.3 and 23.4.2.4 of the 2016 revised IRR of RA No. 9184, within the
(d)	relevant period as provided in the Bidding Documents (in a FORM prescribed by the QC-BAC-GOODS AND SERVICES); and Original copy of Bid Security. If in the form of a Surety Bond, submit also a certification issued by the Insurance Commission;
(e)	or Original copy of Notarized Bid Securing Declaration; and Conformity with Section VI. (Schedule of Requirements) and Section VII. (Technical Specifications), which may include production/delivery schedule,
(f)	manpower requirements, and/or after-sales/parts, if applicable; <u>and</u> Original duly signed Omnibus Sworn Statement (OSS); and if applicable, Original Notarized Secretary's Certificate in case of a corporation, partnership, or cooperative; or Original Special Power of Attorney of all members of the joint venture giving full power and authority to its officer to sign the OSS and do acts to represent the Bidder.
Financia	The prospective bidder's computation of Net Financial Contracting Capacity (NFCC) (in a FORM prescribed by the QC-BAC-GOODS AND SERVICES); or
	A committed Line of Credit from a Universal or Commercial Bank in lieu of its NFCC computation.
	Class "B" Documents
(h)	If applicable, a duly signed joint venture agreement (JVA) in case the joint venture is already in existence;
	duly notarized statements from all the potential joint venture partners stating that they will enter into and abide by the provisions of the JVA in the instance that the bid is successful.
Other do	cumentary requirements under RA No. 9184 (as applicable)
(i)	[For foreign bidders claiming by reason of their country's extension of reciprocal rights to Filipinos] Certification from the relevant government office of their country stating that Filipinos are allowed to participate in
☐ (j)	government procurement activities for the same item or product. Certification from the DTI if the Bidder claims preference as a Domestic Bidder or Domestic Entity.

11.	TINANCIAL C	OTATE OTHER T	

II FINANCIAI COMPONENT ENVELOPE

(a)	Original of duly signed and accomplished Financial Bid Form; and
(b)	Original of duly signed and accomplished Price Schedule(s).

III. REQUIRED DOCUMENTS in BDS SECTION 20.2 and 21.2

- Copy of valid, current License to Operate from DOH Accreditation as Supplier, Distributor or Manufacturer for Medical or Hospital Equipment or Devices.
- Statement of Warranty with Project Number and Project Title Minimum of One (1) Year.

Note:

1. Please refer to

[https://drive.google.com/file/d/1uiYurh5WrpBL5B pqpzAb62yucAblR1p/view?usp=sharing] for the following requirements:

- a. Computation of NFCC;
- b. List of All Ongoing Contracts/List of Contracts already awarded but not yet started:
- c. Statement of Single Largest Completed Contract
- 2. Please refer to GPPB Resolution No. 16-2020 for the following requirements:
 - a. Bid Form;
 - b. Price Schedule (for Goods Offered from Abroad/ Within the Philippines)
 - c. Bid Securing Declaration; and
 - d. Omnibus Sworn Statement

Republic of the Philippines

Government Procurement Policy Board